

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN092921\
 Data File : BN016571.D
 Acq On : 29 Sep 2021 10:06
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4

Quant Time: Sep 29 11:39:41 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN092821.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 29 10:56:17 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.652	152	30472	0.400	ng	0.00
7) Naphthalene-d8	10.419	136	130024	0.400	ng	# 0.00
13) Acenaphthene-d10	14.281	164	65745	0.400	ng	# 0.00
19) Phenanthrene-d10	17.030	188	126367	0.400	ng	# 0.00
29) Chrysene-d12	21.249	240	114089	0.400	ng	# 0.00
35) Perylene-d12	23.508	264	106179	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.282	112	26460	0.349	ng	0.00
5) Phenol-d6	6.831	99	26110	0.343	ng	0.00
8) Nitrobenzene-d5	8.797	82	22356	0.324	ng	0.00
11) 2-Methylnaphthalene-d10	12.016	152	71676	0.388	ng	0.00
14) 2,4,6-Tribromophenol	15.779	330	8592	0.299	ng	0.00
15) 2-Fluorobiphenyl	12.898	172	103286	0.382	ng	0.00
27) Fluoranthene-d10	19.078	212	121760	0.357	ng	0.00
31) Terphenyl-d14	19.693	244	100046	0.407	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.263	88	4203	0.320	ng	# 58
3) n-Nitrosodimethylamine	3.613	42	8485	0.351	ng	# 54
6) bis(2-Chloroethyl)ether	7.089	93	31040	0.387	ng	90
9) Naphthalene	10.470	128	134388	0.383	ng	100
10) Hexachlorobutadiene	10.762	225	25973	0.375	ng	# 99
12) 2-Methylnaphthalene	12.092	142	85303	0.391	ng	99
16) Acenaphthylene	14.002	152	95207	0.327	ng	100
17) Acenaphthene	14.345	154	74160	0.375	ng	93
18) Fluorene	15.335	166	88774	0.368	ng	100
20) 4,6-Dinitro-2-methylph...	15.423	198	2031	0.415	ng	91
21) 4-Bromophenyl-phenylether	16.239	248	27456	0.367	ng	# 71
22) Hexachlorobenzene	16.336	284	34459	0.386	ng	95
23) Atrazine	16.519	200	14945	0.334	ng	95
24) Pentachlorophenol	16.689	266	8038	0.392	ng	99
25) Phenanthrene	17.078	178	148287	0.390	ng	100
26) Anthracene	17.164	178	114503	0.347	ng	100
28) Fluoranthene	19.108	202	156522	0.377	ng	98
30) Pyrene	19.475	202	153117	0.394	ng	100
32) Benzo(a)anthracene	21.227	228	124580	0.352	ng	98
33) Chrysene	21.281	228	152355	0.422	ng	98
34) Bis(2-ethylhexyl)phtha...	21.185	149	33920	0.322	ng	100
36) Indeno(1,2,3-cd)pyrene	25.802	276	128450	0.379	ng	98
37) Benzo(b)fluoranthene	22.827	252	137905	0.389	ng	99
38) Benzo(k)fluoranthene	22.872	252	137715	0.399	ng	99
39) Benzo(a)pyrene	23.409	252	116433	0.358	ng	99
40) Dibenzo(a,h)anthracene	25.822	278	99256	0.422	ng	98
41) Benzo(g,h,i)perylene	26.497	276	121431	0.394	ng	97


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN092921\
Data File : BN016571.D
Acq On : 29 Sep 2021 10:06
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Quant Time: Sep 29 11:39:41 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN092821.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Sep 29 10:56:17 2021
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.652 min Scan# 496
Delta R.T. -0.000 min
Lab File: BN016571.D
Acq: 29 Sep 2021 10:06




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:152 Resp: 30472
Ion Ratio Lower Upper
152 100
150 156.6 125.1 187.7
115 58.4 45.8 68.8



#2
1,4-Dioxane
Concen: 0.320 ng
RT: 3.263 min Scan# 20
Delta R.T. 0.000 min
Lab File: BN016571.D
Acq: 29 Sep 2021 10:06

Tgt Ion: 88 Resp: 4203
Ion Ratio Lower Upper
88 100
43 30.3 75.8 113.8#
58 81.9 81.6 122.4









#3
n-Nitrosodimethylamine
Concen: 0.351 ng
RT: 3.613 min Scan# 58
Delta R.T. 0.000 min
Lab File: BN016571.D
Acq: 29 Sep 2021 10:06




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion: 42 Resp: 8485
Ion Ratio Lower Upper
42 100
74 141.7 77.5 116.3#
44 8.7 0.0 0.0#



#4
2-Fluorophenol
Concen: 0.349 ng
RT: 5.282 min Scan# 239
Delta R.T. 0.000 min
Lab File: BN016571.D
Acq: 29 Sep 2021 10:06

Tgt Ion: 112 Resp: 26460
Ion Ratio Lower Upper
112 100
64 54.4 48.4 72.6
63 27.8 25.4 38.2









#5
Phenol-d6
Concen: 0.343 ng
RT: 6.831 min Scan# 407
Delta R.T. 0.000 min
Lab File: BN016571.D
Acq: 29 Sep 2021 10:06




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion: 99 Resp: 26110
Ion Ratio Lower Upper
99 100
42 17.3 19.6 29.4#
71 29.6 24.2 36.2



#6
bis(2-Chloroethyl)ether
Concen: 0.387 ng
RT: 7.089 min Scan# 435
Delta R.T. 0.000 min
Lab File: BN016571.D
Acq: 29 Sep 2021 10:06

Tgt Ion: 93 Resp: 31040
Ion Ratio Lower Upper
93 100
63 75.2 70.0 105.0
95 32.9 26.3 39.5









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.419 min Scan# 734
Delta R.T. 0.000 min
Lab File: BN016571.D
Acq: 29 Sep 2021 10:06




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:	136	Resp:	130024
Ion	Ratio	Lower	Upper
136	100		
137	10.9	8.8	13.2
54	6.5	5.8	8.6
68	5.4	3.6	5.4#



#8
Nitrobenzene-d5
Concen: 0.324 ng
RT: 8.797 min Scan# 606
Delta R.T. 0.000 min
Lab File: BN016571.D
Acq: 29 Sep 2021 10:06

Tgt Ion:	82	Resp:	22356
Ion	Ratio	Lower	Upper
82	100		
128	43.5	27.4	41.2#
54	46.6	58.5	87.7#









#9
Naphthalene
Concen: 0.383 ng
RT: 10.470 min Scan# 738
Delta R.T. 0.000 min
Lab File: BN016571.D
Acq: 29 Sep 2021 10:06

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Tgt Ion:128 Resp: 134388




Ion	Ratio	Lower	Upper
128	100		
129	10.9	8.9	13.3
127	12.7	10.1	15.1




#10
Hexachlorobutadiene
Concen: 0.375 ng
RT: 10.762 min Scan# 761
Delta R.T. 0.000 min
Lab File: BN016571.D
Acq: 29 Sep 2021 10:06

Tgt Ion:225 Resp: 25973

Ion	Ratio	Lower	Upper
225	100		
223	0.0	0.0	0.0
227	64.1	50.4	75.6









#11
2-Methylnaphthalene-d10
Concen: 0.388 ng
RT: 12.016 min Scan# 934
Delta R.T. 0.000 min
Lab File: BN016571.D
Acq: 29 Sep 2021 10:06




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:152 Resp: 71676
Ion Ratio Lower Upper
152 100
151 20.4 16.6 25.0



#12
2-Methylnaphthalene
Concen: 0.391 ng
RT: 12.092 min Scan# 951
Delta R.T. 0.000 min
Lab File: BN016571.D
Acq: 29 Sep 2021 10:06

Tgt Ion:142 Resp: 85303
Ion Ratio Lower Upper
142 100
141 86.9 69.9 104.9
115 32.5 27.0 40.4









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.281 min Scan# 1184
Delta R.T. 0.000 min
Lab File: BN016571.D
Acq: 29 Sep 2021 10:06




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:164 Resp: 65745
Ion Ratio Lower Upper
164 100
162 99.8 79.4 119.0
160 46.3 46.6 70.0#



#14
2,4,6-Tribromophenol
Concen: 0.299 ng
RT: 15.779 min Scan# 1302
Delta R.T. 0.000 min
Lab File: BN016571.D
Acq: 29 Sep 2021 10:06

Tgt Ion:330 Resp: 8592
Ion Ratio Lower Upper
330 100
332 0.0 0.0 0.0
141 33.6 34.4 51.6#









#15
2-Fluorobiphenyl
Concen: 0.382 ng
RT: 12.898 min Scan# 1075
Delta R.T. 0.000 min
Lab File: BN016571.D
Acq: 29 Sep 2021 10:06




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:172 Resp: 103286
Ion Ratio Lower Upper
172 100
171 34.9 27.4 41.2
170 24.1 18.9 28.3



#16
Acenaphthylene
Concen: 0.327 ng
RT: 14.002 min Scan# 1162
Delta R.T. 0.000 min
Lab File: BN016571.D
Acq: 29 Sep 2021 10:06

Tgt Ion:152 Resp: 95207
Ion Ratio Lower Upper
152 100
151 19.3 15.4 23.0
153 13.0 10.4 15.6









#17
Acenaphthene
Concen: 0.375 ng
RT: 14.345 min Scan# 1189
Delta R.T. 0.000 min
Lab File: BN016571.D
Acq: 29 Sep 2021 10:06




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:154 Resp: 74160
Ion Ratio Lower Upper
154 100
153 112.3 91.0 136.4
152 56.5 55.8 83.6



#18
Fluorene
Concen: 0.368 ng
RT: 15.335 min Scan# 1267
Delta R.T. 0.000 min
Lab File: BN016571.D
Acq: 29 Sep 2021 10:06

Tgt Ion:166 Resp: 88774
Ion Ratio Lower Upper
166 100
165 97.2 77.6 116.4
167 13.5 10.7 16.1









#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 17.030 min Scan# 1404
Delta R.T. 0.000 min
Lab File: BN016571.D
Acq: 29 Sep 2021 10:06




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:188 Resp: 126367
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 12.9 6.9 10.3#



#20
4,6-Dinitro-2-methylphenol
Concen: 0.415 ng
RT: 15.423 min Scan# 1274
Delta R.T. 0.000 min
Lab File: BN016571.D
Acq: 29 Sep 2021 10:06

Tgt Ion:198 Resp: 2031
Ion Ratio Lower Upper
198 100
182 17.8 17.5 26.3
77 0.0 0.0 0.0









#21
4-Bromophenyl-phenylether
Concen: 0.367 ng
RT: 16.239 min Scan# 1339
Delta R.T. 0.000 min
Lab File: BN016571.D
Acq: 29 Sep 2021 10:06




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:248 Resp: 27456
Ion Ratio Lower Upper
248 100
250 100.0 71.8 107.6
141 43.6 69.7 104.5#



#22
Hexachlorobenzene
Concen: 0.386 ng
RT: 16.336 min Scan# 1347
Delta R.T. 0.000 min
Lab File: BN016571.D
Acq: 29 Sep 2021 10:06

Tgt Ion:284 Resp: 34459
Ion Ratio Lower Upper
284 100
142 34.9 30.7 46.1
249 29.6 25.8 38.8









#23
Atrazine
Concen: 0.334 ng
RT: 16.519 min Scan# 1362
Delta R.T. 0.000 min
Lab File: BN016571.D
Acq: 29 Sep 2021 10:06




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:200 Resp: 14945
Ion Ratio Lower Upper
200 100
173 22.9 22.6 34.0
215 49.6 38.2 57.4



#24
Pentachlorophenol
Concen: 0.392 ng
RT: 16.689 min Scan# 1376
Delta R.T. 0.000 min
Lab File: BN016571.D
Acq: 29 Sep 2021 10:06

Tgt Ion:266 Resp: 8038
Ion Ratio Lower Upper
266 100
264 63.0 50.2 75.2
268 63.9 51.8 77.8









#25
Phenanthrene
Concen: 0.390 ng
RT: 17.078 min Scan# 1408
Delta R.T. 0.000 min
Lab File: BN016571.D
Acq: 29 Sep 2021 10:06




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:178 Resp: 148287
Ion Ratio Lower Upper
178 100
176 19.0 15.0 22.6
179 15.3 12.3 18.5



#26
Anthracene
Concen: 0.347 ng
RT: 17.164 min Scan# 1415
Delta R.T. 0.000 min
Lab File: BN016571.D
Acq: 29 Sep 2021 10:06

Tgt Ion:178 Resp: 114503
Ion Ratio Lower Upper
178 100
176 18.3 14.6 21.8
179 15.4 12.2 18.4









#27
Fluoranthene-d10
Concen: 0.357 ng
RT: 19.078 min Scan# 1699
Delta R.T. 0.000 min
Lab File: BN016571.D
Acq: 29 Sep 2021 10:06




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:212 Resp: 121760
Ion Ratio Lower Upper
212 100
106 13.5 8.8 13.2#
104 7.4 5.0 7.4



#28
Fluoranthene
Concen: 0.377 ng
RT: 19.108 min Scan# 1705
Delta R.T. 0.000 min
Lab File: BN016571.D
Acq: 29 Sep 2021 10:06

Tgt Ion:202 Resp: 156522
Ion Ratio Lower Upper
202 100
101 11.3 7.5 11.3
203 17.1 13.9 20.9









#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.249 min Scan# 2003
Delta R.T. 0.000 min
Lab File: BN016571.D
Acq: 29 Sep 2021 10:06




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:240 Resp: 114089
Ion Ratio Lower Upper
240 100
120 7.9 9.0 13.4#
236 25.9 21.7 32.5



#30
Pyrene
Concen: 0.394 ng
RT: 19.475 min Scan# 1779
Delta R.T. 0.000 min
Lab File: BN016571.D
Acq: 29 Sep 2021 10:06

Tgt Ion:202 Resp: 153117
Ion Ratio Lower Upper
202 100
200 20.6 16.3 24.5
203 17.5 13.9 20.9









#31
Terphenyl-d14
Concen: 0.407 ng
RT: 19.693 min Scan# 1823
Delta R.T. 0.000 min
Lab File: BN016571.D
Acq: 29 Sep 2021 10:06




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:244 Resp: 100046
Ion Ratio Lower Upper
244 100
212 7.4 6.1 9.1
122 11.7 8.4 12.6



#32
Benzo(a)anthracene
Concen: 0.352 ng
RT: 21.227 min Scan# 2001
Delta R.T. 0.000 min
Lab File: BN016571.D
Acq: 29 Sep 2021 10:06

Tgt Ion:228 Resp: 124580
Ion Ratio Lower Upper
228 100
226 27.7 21.2 31.8
229 19.1 15.6 23.4









#33
Chrysene
Concen: 0.422 ng
RT: 21.281 min Scan# 2006
Delta R.T. 0.000 min
Lab File: BN016571.D
Acq: 29 Sep 2021 10:06




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:228 Resp: 152355
Ion Ratio Lower Upper
228 100
226 30.0 23.1 34.7
229 19.4 15.8 23.8



#34
Bis(2-ethylhexyl)phthalate
Concen: 0.322 ng
RT: 21.185 min Scan# 1997
Delta R.T. 0.000 min
Lab File: BN016571.D
Acq: 29 Sep 2021 10:06

Tgt Ion:149 Resp: 33920
Ion Ratio Lower Upper
149 100
167 26.8 21.4 32.2
279 3.6 3.1 4.7









#35
Perylene-d12
Concen: 0.400 ng
RT: 23.508 min Scan# 2434
Delta R.T. 0.000 min
Lab File: BN016571.D
Acq: 29 Sep 2021 10:06




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:264 Resp: 106179
Ion Ratio Lower Upper
264 100
260 24.9 19.9 29.9
265 24.0 20.7 31.1



#36
Indeno(1,2,3-cd)pyrene
Concen: 0.379 ng
RT: 25.802 min Scan# 3104
Delta R.T. 0.000 min
Lab File: BN016571.D
Acq: 29 Sep 2021 10:06

Tgt Ion:276 Resp: 128450
Ion Ratio Lower Upper
276 100
138 24.9 18.2 27.2
277 24.9 19.8 29.8









#37
Benzo(b)fluoranthene
Concen: 0.389 ng
RT: 22.827 min Scan# 2235
Delta R.T. 0.000 min
Lab File: BN016571.D
Acq: 29 Sep 2021 10:06




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:252 Resp: 137905
Ion Ratio Lower Upper
252 100
253 22.0 17.8 26.6
125 10.1 7.3 10.9



#38
Benzo(k)fluoranthene
Concen: 0.399 ng
RT: 22.872 min Scan# 2248
Delta R.T. 0.000 min
Lab File: BN016571.D
Acq: 29 Sep 2021 10:06

Tgt Ion:252 Resp: 137715
Ion Ratio Lower Upper
252 100
253 22.2 17.8 26.6
125 10.8 7.4 11.0









#39
Benzo(a)pyrene
Concen: 0.358 ng
RT: 23.409 min Scan# 2405
Delta R.T. 0.000 min
Lab File: BN016571.D
Acq: 29 Sep 2021 10:06




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:252 Resp: 116433
Ion Ratio Lower Upper
252 100
253 22.3 17.9 26.9
125 11.5 8.2 12.4



#40
Dibenzo(a,h)anthracene
Concen: 0.422 ng
RT: 25.822 min Scan# 3110
Delta R.T. 0.000 min
Lab File: BN016571.D
Acq: 29 Sep 2021 10:06

Tgt Ion:278 Resp: 99256
Ion Ratio Lower Upper
278 100
139 17.2 12.1 18.1
279 24.1 19.2 28.8





#41
Benzo(g,h,i)perylene
Concen: 0.394 ng
RT: 26.497 min Scan# 3307
Delta R.T. 0.000 min
Lab File: BN016571.D
Acq: 29 Sep 2021 10:06

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Tgt Ion:276 Resp: 121431
Ion Ratio Lower Upper
276 100
277 24.0 18.9 28.3
138 21.8 15.7 23.5

