

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN092922\
 Data File : BN021957.D
 Acq On : 29 Sep 2022 16:33
 Operator : CG/JU
 Sample : N4846-02
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 TT158S1-20220923

Quant Time: Sep 30 02:45:33 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN092922.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Sep 30 02:27:15 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.033	152	4140	0.400	ng	0.00
7) Naphthalene-d8	10.855	136	12808	0.400	ng	0.00
13) Acenaphthene-d10	14.686	164	6930	0.400	ng	0.00
19) Phenanthrene-d10	17.437	188	13940	0.400	ng	0.00
29) Chrysene-d12	21.636	240	9847	0.400	ng	0.00
35) Perylene-d12	24.134	264	6904	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.563	112	1172	0.106	ng	0.00
5) Phenol-d6	7.188	99	890	0.064	ng	0.00
8) Nitrobenzene-d5	9.211	82	2481	0.232	ng	0.00
11) 2-Methylnaphthalene-d10	12.448	152	7397	0.268	ng	0.00
14) 2,4,6-Tribromophenol	16.184	330	740	0.249	ng	0.00
15) 2-Fluorobiphenyl	13.317	172	7911	0.258	ng	0.00
27) Fluoranthene-d10	19.471	212	11208	0.297	ng	0.00
31) Terphenyl-d14	20.060	244	7969	0.366	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.374	88	4205	0.716	ng	Qvalue # 95
34) Bis(2-ethylhexyl)phtha...	21.529	149	2366	0.103	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN092922\
Data File : BN021957.D
Acq On : 29 Sep 2022 16:33
Operator : CG/JU
Sample : N4846-02
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
TT158S1-20220923

Quant Time: Sep 30 02:45:33 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN092922.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Sep 30 02:27:15 2022
Response via : Initial Calibration

