

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN092922\
 Data File : BN021948.D
 Acq On : 28 Sep 2022 19:36
 Operator : CG/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 ICVBN092922

Quant Time: Sep 30 02:28:18 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN092922.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Sep 30 02:27:15 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	122	0.00
2	1,4-Dioxane	0.400	0.373	6.8	117	0.00
3	n-Nitrosodimethylamine	0.400	0.388	3.0	117	0.00
4 S	2-Fluorophenol	0.400	0.387	3.3	120	0.00
5 S	Phenol-d6	0.400	0.381	4.8	121	0.00
6	bis(2-Chloroethyl)ether	0.400	0.392	2.0	120	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	123	0.00
8 S	Nitrobenzene-d5	0.400	0.379	5.3	120	0.00
9	Naphthalene	0.400	0.400	0.0	123	0.00
10	Hexachlorobutadiene	0.400	0.406	-1.5	124	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.381	4.8	119	0.00
12	2-Methylnaphthalene	0.400	0.343	14.2	118	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	117	0.00
14 S	2,4,6-Tribromophenol	0.400	0.344	14.0	108	0.00
15 S	2-Fluorobiphenyl	0.400	0.406	-1.5	119	0.00
16	Acenaphthylene	0.400	0.370	7.5	118	0.00
17	Acenaphthene	0.400	0.396	1.0	117	0.00
18	Fluorene	0.400	0.379	5.3	111	0.01
19 I	Phenanthrene-d10	0.400	0.400	0.0	113	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.274	31.5#	92	0.00
21	4-Bromophenyl-phenylether	0.400	0.391	2.3	115	0.00
22	Hexachlorobenzene	0.400	0.405	-1.3	116	0.00
23	Atrazine	0.400	0.359	10.3	110	0.00
24	Pentachlorophenol	0.400	0.323	19.3	106	0.00
25	Phenanthrene	0.400	0.395	1.3	113	0.00
26	Anthracene	0.400	0.371	7.3	113	0.00
27 SURR	Fluoranthene-d10	0.400	0.379	5.3	110	0.00
28	Fluoranthene	0.400	0.379	5.3	110	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	105	0.00
30	Pyrene	0.400	0.392	2.0	104	0.00
31 S	Terphenyl-d14	0.400	0.422	-5.5	110	0.00
32	Benzo(a)anthracene	0.400	0.387	3.3	106	0.00
33	Chrysene	0.400	0.400	0.0	102	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.356	11.0	104	0.00
35 I	Perylene-d12	0.400	0.400	0.0	110	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.391	2.3	107	0.00
37	Benzo(b)fluoranthene	0.400	0.391	2.3	106	0.00
38	Benzo(k)fluoranthene	0.400	0.391	2.3	107	0.00
39 C	Benzo(a)pyrene	0.400	0.389	2.8	110	0.00
40	Dibenzo(a,h)anthracene	0.400	0.397	0.8	106	0.00
41	Benzo(g,h,i)perylene	0.400	0.377	5.8	101	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
ClientSampleId :
ICVBN092922