




Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN092925\
Data File : BN037930.D
Acq On : 29 Sep 2025 10:15
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Sep 29 11:50:13 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Sep 29 10:34:59 2025
Response via : Initial Calibration



TIC: BN037930.D\data.ms

(70) Pentachlorophenol (C)




16.875min (-0.011) 35044.31 ng

response 2284032

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	62.20	68.49
264.00	61.60	65.46
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN092925\
Data File : BN037930.D
Acq On : 29 Sep 2025 10:15
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Sep 29 11:50:13 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Sep 29 10:34:59 2025
Response via : Initial Calibration



TIC: BN037930.D\data.ms

(77) Benzidine**19.475min (+ 0.001) 0.00 ng****response 11301983**

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	15.50	15.77
183.00	13.20	13.39
0.00	0.00	0.00