

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN093020\
 Data File : BN012029.D
 Acq On : 30 Sep 2020 15:36
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Sep 30 16:46:41 2020
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN092320.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 30 16:28:17 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	77	0.00
2	1,4-Dioxane	0.696	0.616	11.5	75	0.00
3	n-Nitrosodimethylamine	1.001	1.103	-10.2	99	0.00
4 S	2-Fluorophenol	1.401	1.420	-1.4	86	0.00
5 S	Phenol-d6	1.767	1.755	0.7	86	0.00
6	bis(2-Chloroethyl)ether	1.511	1.412	6.6	81	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	76	0.00
8 S	Nitrobenzene-d5	0.413	0.412	0.2	86	0.00
9	Naphthalene	1.146	1.124	1.9	82	0.00
10	Hexachlorobutadiene	0.187	0.193	-3.2	84	0.00
11 SURR	2-Methylnaphthalene-d10	0.633	0.592	6.5	79	0.00
12	2-Methylnaphthalene	0.729	0.681	6.6	79	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	75	0.00
14 S	2,4,6-Tribromophenol	0.194	0.207	-6.7	91	0.00
15 S	2-Fluorobiphenyl	1.513	1.471	2.8	79	0.00
16	Acenaphthylene	1.928	1.810	6.1	79	0.00
17	Acenaphthene	1.320	1.215	8.0	77	0.00
18	Fluorene	1.612	1.504	6.7	77	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	71	0.00
20	4,6-Dinitro-2-methylphenol	0.076	0.078	-2.6	86	0.00
21	4-Bromophenyl-phenylether	0.212	0.209	1.4	78	-0.01
22	Hexachlorobenzene	0.247	0.240	2.8	76	0.00
23	Atrazine	0.203	0.211	-3.9	86	0.00
24	Pentachlorophenol	0.121	0.120	0.8	82	0.00
25	Phenanthrene	1.227	1.146	6.6	75	0.00
26	Anthracene	1.075	1.006	6.4	76	0.00
27 SURR	Fluoranthene-d10	1.174	1.117	4.9	76	0.00
28	Fluoranthene	1.371	1.274	7.1	75	0.00
29 I	Chrysene-d12	1.000	1.000	0.0	74	0.00
30	Pyrene	1.491	1.402	6.0	75	0.00
31 S	Terphenyl-d14	0.914	0.888	2.8	78	0.00
32	Benzo(a)anthracene	1.285	1.188	7.5	76	0.01
33	Chrysene	1.383	1.295	6.4	77	0.00
34	Bis(2-ethylhexyl)phthalate	0.907	1.037	-14.3	97	0.01
35	Indeno(1,2,3-cd)pyrene	1.570	1.652	-5.2	86	0.01
36 I	Perylene-d12	1.000	1.000	0.0	81	0.00
37	Benzo(b)fluoranthene	1.368	1.203	12.1	81	0.00
38	Benzo(k)fluoranthene	1.487	1.296	12.8	75	0.00
39 C	Benzo(a)pyrene	1.261	1.154	8.5	81	0.00
40	Dibenzo(a,h)anthracene	1.305	1.244	4.7	86	0.00
41	Benzo(g,h,i)perylene	1.410	1.337	5.2	84	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

LabSampleId :

SSTDCCC0.4