

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN093020\
 Data File : BN012031.D
 Acq On : 30 Sep 2020 18:01
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Sep 30 18:29:13 2020
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN092320.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 30 16:28:17 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	71	0.00
2	1,4-Dioxane	0.696	0.635	8.8	72	0.00
3	n-Nitrosodimethylamine	1.001	1.058	-5.7	88	0.00
4 S	2-Fluorophenol	1.401	1.387	1.0	79	0.00
5 S	Phenol-d6	1.767	1.752	0.8	79	0.00
6	bis(2-Chloroethyl)ether	1.511	1.422	5.9	76	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	70	0.00
8 S	Nitrobenzene-d5	0.413	0.417	-1.0	79	0.00
9	Naphthalene	1.146	1.085	5.3	73	0.00
10	Hexachlorobutadiene	0.187	0.195	-4.3	78	0.00
11 SURR	2-Methylnaphthalene-d10	0.633	0.590	6.8	72	0.00
12	2-Methylnaphthalene	0.729	0.671	8.0	71	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	68	0.00
14 S	2,4,6-Tribromophenol	0.194	0.195	-0.5	78	0.00
15 S	2-Fluorobiphenyl	1.513	1.464	3.2	72	0.00
16	Acenaphthylene	1.928	1.788	7.3	72	0.00
17	Acenaphthene	1.320	1.213	8.1	71	0.00
18	Fluorene	1.612	1.445	10.4	68	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	63	0.00
20	4,6-Dinitro-2-methylphenol	0.076	0.078	-2.6	76	0.00
21	4-Bromophenyl-phenylether	0.212	0.203	4.2	67	0.00
22	Hexachlorobenzene	0.247	0.246	0.4	69	0.00
23	Atrazine	0.203	0.205	-1.0	74	0.00
24	Pentachlorophenol	0.121	0.113	6.6	68	0.00
25	Phenanthrene	1.227	1.144	6.8	66	0.00
26	Anthracene	1.075	0.991	7.8	67	0.00
27 SURR	Fluoranthene-d10	1.174	1.100	6.3	67	0.00
28	Fluoranthene	1.371	1.228	10.4	64	0.00
29 I	Chrysene-d12	1.000	1.000	0.0	65	0.00
30	Pyrene	1.491	1.381	7.4	65	0.00
31 S	Terphenyl-d14	0.914	0.874	4.4	67	0.00
32	Benzo(a)anthracene	1.285	1.194	7.1	67	0.00
33	Chrysene	1.383	1.274	7.9	66	0.00
34	Bis(2-ethylhexyl)phthalate	0.907	0.971	-7.1	79	0.01
35	Indeno(1,2,3-cd)pyrene	1.570	1.723	-9.7	78	0.00
36 I	Perylene-d12	1.000	1.000	0.0	71	0.00
37	Benzo(b)fluoranthene	1.368	1.211	11.5	71	0.00
38	Benzo(k)fluoranthene	1.487	1.342	9.8	69	0.00
39 C	Benzo(a)pyrene	1.261	1.172	7.1	73	0.00
40	Dibenzo(a,h)anthracene	1.305	1.244	4.7	76	0.00
41	Benzo(g,h,i)perylene	1.410	1.377	2.3	76	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

LabSampleId :

SSTDCCC0.4