

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN093020\
 Data File : BN012031.D
 Acq On : 30 Sep 2020 18:01
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4EC

Quant Time: Sep 30 18:29:13 2020
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN092320.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 30 16:28:17 2020
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.667	152	2288	0.40	ng	0.00
7) Naphthalene-d8	10.440	136	9039	0.40	ng	# 0.00
13) Acenaphthene-d10	14.294	164	4478	0.40	ng	0.00
19) Phenanthrene-d10	17.041	188	8423	0.40	ng	# 0.00
29) Chrysene-d12	21.248	240	7804	0.40	ng	# 0.00
36) Perylene-d12	23.457	264	8443	0.40	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.275	112	3174	0.40	ng	0.00
5) Phenol-d6	6.837	99	4008	0.40	ng	0.00
8) Nitrobenzene-d5	8.805	82	3773	0.40	ng	0.00
11) 2-Methylnaphthalene-d10	12.033	152	5336	0.37	ng	0.00
14) 2,4,6-Tribromophenol	15.791	330	872	0.40	ng	0.00
15) 2-Fluorobiphenyl	12.923	172	6558	0.39	ng	0.00
27) Fluoranthene-d10	19.082	212	9266	0.37	ng	0.00
31) Terphenyl-d14	19.688	244	6823	0.38	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.246	88	1452	0.36	ng	95
3) n-Nitrosodimethylamine	3.552	42	2421	0.42	ng	96
6) bis(2-Chloroethyl)ether	7.103	93	3254	0.38	ng	94
9) Naphthalene	10.491	128	9803	0.38	ng	99
10) Hexachlorobutadiene	10.783	225	1765	0.42	ng	# 96
12) 2-Methylnaphthalene	12.109	142	6063	0.37	ng	97
16) Acenaphthylene	14.014	152	8008	0.37	ng	99
17) Acenaphthene	14.357	154	5433	0.37	ng	91
18) Fluorene	15.347	166	6470	0.36	ng	99
20) 4,6-Dinitro-2-methylph...	15.435	198	657	0.41	ng	# 27
21) 4-Bromophenyl-phenylether	16.250	248	1711	0.38	ng	# 91
22) Hexachlorobenzene	16.360	284	2075	0.40	ng	# 84
23) Atrazine	16.518	200	1725	0.40	ng	96
24) Pentachlorophenol	16.701	266	951	0.37	ng	96
25) Phenanthrene	17.090	178	9632	0.37	ng	99
26) Anthracene	17.175	178	8347	0.37	ng	99
28) Fluoranthene	19.112	202	10345	0.36	ng	# 96
30) Pyrene	19.479	202	10778	0.37	ng	100
32) Benzo(a)anthracene	21.227	228	9318	0.37	ng	97
33) Chrysene	21.280	228	9940	0.37	ng	98
34) Bis(2-ethylhexyl)phtha...	21.174	149	7577	0.43	ng	# 91
35) Indeno(1,2,3-cd)pyrene	25.664	276	13445	0.44	ng	# 95
37) Benzo(b)fluoranthene	22.792	252	10227	0.35	ng	# 80
38) Benzo(k)fluoranthene	22.840	252	11332	0.36	ng	# 82
39) Benzo(a)pyrene	23.357	252	9891	0.37	ng	# 69
40) Dibenzo(a,h)anthracene	25.681	278	10507	0.38	ng	# 83
41) Benzo(g,h,i)perylene	26.339	276	11626	0.39	ng	# 91


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN093020\
Data File : BN012031.D
Acq On : 30 Sep 2020 18:01
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC

Quant Time: Sep 30 18:29:13 2020
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN092320.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Sep 30 16:28:17 2020
Response via : Initial Calibration









#1
 1,4-Dichlorobenzene-d4
 Concen: 0.40 ng
 RT: 7.667 min Scan# 582
 Delta R.T. -0.000 min
 Lab File: BN012031.D
 Acq: 30 Sep 2020 18:01




Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4EC


Tgt Ion:152 Resp: 2288
 Ion Ratio Lower Upper
 152 100
 150 156.0 118.3 177.5
 115 76.2 51.3 76.9



#2
 1,4-Dioxane
 Concen: 0.36 ng
 RT: 3.246 min Scan# 33
 Delta R.T. 0.000 min
 Lab File: BN012031.D
 Acq: 30 Sep 2020 18:01

Tgt Ion: 88 Resp: 1452
 Ion Ratio Lower Upper
 88 100
 58 81.0 63.3 94.9
 43 48.6 33.0 49.4









#3
n-Nitrosodimethylamine
Concen: 0.42 ng
RT: 3.552 min Scan# 71
Delta R.T. -0.008 min
Lab File: BN012031.D
Acq: 30 Sep 2020 18:01

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC

Tgt Ion: 42 Resp: 2421




Ion	Ratio	Lower	Upper
42	100		
74	84.3	64.4	96.6
44	6.8	6.7	10.1




#4
2-Fluorophenol
Concen: 0.40 ng
RT: 5.275 min Scan# 285
Delta R.T. 0.000 min
Lab File: BN012031.D
Acq: 30 Sep 2020 18:01

Tgt Ion: 112 Resp: 3174

Ion	Ratio	Lower	Upper
112	100		
64	69.3	49.0	73.6
63	45.6	31.8	47.6









#5
Phenol-d6
Concen: 0.40 ng
RT: 6.837 min Scan# 479
Delta R.T. 0.000 min
Lab File: BN012031.D
Acq: 30 Sep 2020 18:01




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:	99	Resp:	4008
Ion	Ratio	Lower	Upper
99	100		
42	31.7	23.4	35.0
71	48.9	38.1	57.1



#6
bis(2-Chloroethyl)ether
Concen: 0.38 ng
RT: 7.103 min Scan# 512
Delta R.T. 0.000 min
Lab File: BN012031.D
Acq: 30 Sep 2020 18:01

Tgt Ion:	93	Resp:	3254
Ion	Ratio	Lower	Upper
93	100		
63	83.1	61.1	91.7
95	30.6	25.5	38.3









#7
Naphthalene-d8
Concen: 0.40 ng
RT: 10.440 min Scan# 830
Delta R.T. 0.000 min
Lab File: BN012031.D
Acq: 30 Sep 2020 18:01




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:	136	Resp:	9039
Ion Ratio	Lower	Upper	
136	100		
137	12.0	9.5	14.3
54	7.8	5.6	8.4
68	5.0	3.2	4.8#



#8
Nitrobenzene-d5
Concen: 0.40 ng
RT: 8.805 min Scan# 701
Delta R.T. 0.000 min
Lab File: BN012031.D
Acq: 30 Sep 2020 18:01

Tgt Ion:	82	Resp:	3773
Ion Ratio	Lower	Upper	
82	100		
128	31.6	34.2	51.2#
54	66.4	43.1	64.7#









#9
Naphthalene
Concen: 0.38 ng
RT: 10.491 min Scan# 834
Delta R.T. 0.000 min
Lab File: BN012031.D
Acq: 30 Sep 2020 18:01




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:128 Resp: 9803
Ion Ratio Lower Upper
128 100
129 11.9 9.7 14.5
127 13.2 11.0 16.4



#10
Hexachlorobutadiene
Concen: 0.42 ng
RT: 10.783 min Scan# 857
Delta R.T. 0.000 min
Lab File: BN012031.D
Acq: 30 Sep 2020 18:01

Tgt Ion:225 Resp: 1765
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 62.0 51.9 77.9









#11
2-Methylnaphthalene-d10
Concen: 0.37 ng
RT: 12.033 min Scan# 1031
Delta R.T. 0.000 min
Lab File: BN012031.D
Acq: 30 Sep 2020 18:01




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:152 Resp: 5336
Ion Ratio Lower Upper
152 100
151 19.4 16.7 25.1



#12
2-Methylnaphthalene
Concen: 0.37 ng
RT: 12.109 min Scan# 1048
Delta R.T. 0.000 min
Lab File: BN012031.D
Acq: 30 Sep 2020 18:01

Tgt Ion:142 Resp: 6063
Ion Ratio Lower Upper
142 100
141 87.4 70.0 105.0
115 50.4 35.8 53.6





#13

Acenaphthene-d10

Concen: 0.40 ng

RT: 14.294 min Scan# 1278

Delta R.T. 0.000 min

Lab File: BN012031.D

Acq: 30 Sep 2020 18:01





Tgt Ion:164 Resp: 4478

Ion Ratio Lower Upper

164 100

162 102.9 83.6 125.4

160 49.3 41.2 61.8



#14

2,4,6-Tribromophenol

Concen: 0.40 ng

RT: 15.791 min Scan# 1396

Delta R.T. 0.000 min

Lab File: BN012031.D

Acq: 30 Sep 2020 18:01




Tgt Ion:330 Resp: 872


Ion Ratio Lower Upper

330 100

332 0.0 0.0 0.0

141 57.2 33.7 50.5#









#15
2-Fluorobiphenyl
Concen: 0.39 ng
RT: 12.923 min Scan# 1170
Delta R.T. 0.000 min
Lab File: BN012031.D
Acq: 30 Sep 2020 18:01




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:	172	Resp:	6558
Ion Ratio	Lower	Upper	
172	100		
171	35.0	27.4	41.2
170	23.7	17.8	26.6



#16
Acenaphthylene
Concen: 0.37 ng
RT: 14.014 min Scan# 1256
Delta R.T. 0.000 min
Lab File: BN012031.D
Acq: 30 Sep 2020 18:01

Tgt Ion:	152	Resp:	8008
Ion Ratio	Lower	Upper	
152	100		
151	19.7	15.9	23.9
153	12.7	10.5	15.7









#17
Acenaphthene
Concen: 0.37 ng
RT: 14.357 min Scan# 1283
Delta R.T. 0.000 min
Lab File: BN012031.D
Acq: 30 Sep 2020 18:01




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:154 Resp: 5433
Ion Ratio Lower Upper
154 100
153 115.4 83.4 125.2
152 56.8 42.8 64.2



#18
Fluorene
Concen: 0.36 ng
RT: 15.347 min Scan# 1361
Delta R.T. 0.000 min
Lab File: BN012031.D
Acq: 30 Sep 2020 18:01

Tgt Ion:166 Resp: 6470
Ion Ratio Lower Upper
166 100
165 97.9 78.8 118.2
167 13.6 11.2 16.8





#19

Phenanthrene-d10

Concen: 0.40 ng

RT: 17.041 min Scan# 1498

Delta R.T. 0.000 min

Lab File: BN012031.D

Acq: 30 Sep 2020 18:01





Tgt Ion:188 Resp: 8423

Ion Ratio Lower Upper

188 100

94 0.0 0.0 0.0

80 12.8 7.7 11.5#



#20

4,6-Dinitro-2-methylphenol

Concen: 0.41 ng

RT: 15.435 min Scan# 1368

Delta R.T. 0.000 min

Lab File: BN012031.D

Acq: 30 Sep 2020 18:01




Tgt Ion:198 Resp: 657


Ion Ratio Lower Upper

198 100

51 129.2 45.0 67.4#

105 58.2 30.4 45.6#









#21
4-Bromophenyl-phenylether
Concen: 0.38 ng
RT: 16.250 min Scan# 1433
Delta R.T. 0.000 min
Lab File: BN012031.D
Acq: 30 Sep 2020 18:01




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:248 Resp: 1711
Ion Ratio Lower Upper
248 100
250 107.0 82.7 124.1
141 57.3 33.7 50.5#



#22
Hexachlorobenzene
Concen: 0.40 ng
RT: 16.360 min Scan# 1442
Delta R.T. 0.000 min
Lab File: BN012031.D
Acq: 30 Sep 2020 18:01

Tgt Ion:284 Resp: 2075
Ion Ratio Lower Upper
284 100
142 57.2 32.4 48.6#
249 31.8 26.8 40.2









#23
Atrazine
Concen: 0.40 ng
RT: 16.518 min Scan# 1455
Delta R.T. 0.000 min
Lab File: BN012031.D
Acq: 30 Sep 2020 18:01




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:200 Resp: 1725
Ion Ratio Lower Upper
200 100
173 25.2 18.7 28.1
215 47.2 40.2 60.4



#24
Pentachlorophenol
Concen: 0.37 ng
RT: 16.701 min Scan# 1470
Delta R.T. 0.000 min
Lab File: BN012031.D
Acq: 30 Sep 2020 18:01

Tgt Ion:266 Resp: 951
Ion Ratio Lower Upper
266 100
264 60.6 51.2 76.8
268 68.7 52.4 78.6









#25
Phenanthrene
Concen: 0.37 ng
RT: 17.090 min Scan# 1502
Delta R.T. 0.000 min
Lab File: BN012031.D
Acq: 30 Sep 2020 18:01




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:178 Resp: 9632
Ion Ratio Lower Upper
178 100
176 19.6 15.1 22.7
179 16.0 12.3 18.5



#26
Anthracene
Concen: 0.37 ng
RT: 17.175 min Scan# 1509
Delta R.T. 0.000 min
Lab File: BN012031.D
Acq: 30 Sep 2020 18:01

Tgt Ion:178 Resp: 8347
Ion Ratio Lower Upper
178 100
176 18.6 14.6 21.8
179 15.3 12.4 18.6









#27
Fluoranthene-d10
Concen: 0.37 ng
RT: 19.082 min Scan# 1793
Delta R.T. 0.000 min
Lab File: BN012031.D
Acq: 30 Sep 2020 18:01




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:212 Resp: 9266
Ion Ratio Lower Upper
212 100
106 11.6 6.4 9.6#
104 7.4 3.9 5.9#



#28
Fluoranthene
Concen: 0.36 ng
RT: 19.112 min Scan# 1799
Delta R.T. 0.000 min
Lab File: BN012031.D
Acq: 30 Sep 2020 18:01

Tgt Ion:202 Resp: 10345
Ion Ratio Lower Upper
202 100
101 10.9 5.4 8.2#
203 16.9 13.8 20.6









#29
Chrysene-d12
Concen: 0.40 ng
RT: 21.248 min Scan# 2096
Delta R.T. 0.000 min
Lab File: BN012031.D
Acq: 30 Sep 2020 18:01




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:240 Resp: 7804
Ion Ratio Lower Upper
240 100
120 10.6 3.5 5.3#
236 26.4 20.5 30.7



#30
Pyrene
Concen: 0.37 ng
RT: 19.479 min Scan# 1873
Delta R.T. 0.005 min
Lab File: BN012031.D
Acq: 30 Sep 2020 18:01

Tgt Ion:202 Resp: 10778
Ion Ratio Lower Upper
202 100
200 20.5 16.5 24.7
203 17.6 13.9 20.9









#31
 Terphenyl-d14
 Concen: 0.38 ng
 RT: 19.688 min Scan# 1915
 Delta R.T. 0.000 min
 Lab File: BN012031.D
 Acq: 30 Sep 2020 18:01




Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4EC


Tgt Ion:244 Resp: 6823
 Ion Ratio Lower Upper
 244 100
 212 9.8 7.3 10.9
 122 14.0 7.8 11.6#



#32
 Benzo(a)anthracene
 Concen: 0.37 ng
 RT: 21.227 min Scan# 2094
 Delta R.T. 0.000 min
 Lab File: BN012031.D
 Acq: 30 Sep 2020 18:01

Tgt Ion:228 Resp: 9318
 Ion Ratio Lower Upper
 228 100
 226 28.4 21.0 31.4
 229 19.7 16.0 24.0





#33

Chrysene

Concen: 0.37 ng

RT: 21.280 min Scan# 2099

Delta R.T. 0.000 min

Lab File: BN012031.D

Acq: 30 Sep 2020 18:01

Tgt Ion:228 Resp: 9940

Ion Ratio Lower Upper

228 100

226 30.8 23.3 34.9





229 20.2 15.8 23.8

Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4EC



#34

Bis(2-ethylhexyl)phthalate

Concen: 0.43 ng

RT: 21.174 min Scan# 2089

Delta R.T. 0.011 min

Lab File: BN012031.D

Acq: 30 Sep 2020 18:01




Tgt Ion:149 Resp: 7577


Ion Ratio Lower Upper

149 100

167 24.4 23.6 35.4

279 3.9 4.0 6.0#









#35
Indeno(1,2,3-cd)pyrene
Concen: 0.44 ng
RT: 25.664 min Scan# 3157
Delta R.T. 0.004 min
Lab File: BN012031.D
Acq: 30 Sep 2020 18:01




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:276 Resp: 13445
Ion Ratio Lower Upper
276 100
138 21.0 13.0 19.4#
277 24.5 20.0 30.0



#36
Perylene-d12
Concen: 0.40 ng
RT: 23.457 min Scan# 2512
Delta R.T. 0.000 min
Lab File: BN012031.D
Acq: 30 Sep 2020 18:01

Tgt Ion:264 Resp: 8443
Ion Ratio Lower Upper
264 100
260 25.2 19.6 29.4
265 73.4 38.6 58.0#









#37
Benzo(b)fluoranthene
Concen: 0.35 ng
RT: 22.792 min Scan# 2318
Delta R.T. -0.003 min
Lab File: BN012031.D
Acq: 30 Sep 2020 18:01




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:252 Resp: 10227
Ion Ratio Lower Upper
252 100
253 31.6 19.1 28.7#
125 19.8 6.2 9.2#



#38
Benzo(k)fluoranthene
Concen: 0.36 ng
RT: 22.840 min Scan# 2332
Delta R.T. 0.004 min
Lab File: BN012031.D
Acq: 30 Sep 2020 18:01

Tgt Ion:252 Resp: 11332
Ion Ratio Lower Upper
252 100
253 32.2 19.3 28.9#
125 15.8 6.0 9.0#









#39
Benzo(a)pyrene
Concen: 0.37 ng
RT: 23.357 min Scan# 2483
Delta R.T. -0.003 min
Lab File: BN012031.D
Acq: 30 Sep 2020 18:01




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:252 Resp: 9891
Ion Ratio Lower Upper
252 100
253 36.0 19.8 29.6#
125 28.6 7.3 10.9#



#40
Dibenzo(a,h)anthracene
Concen: 0.38 ng
RT: 25.681 min Scan# 3162
Delta R.T. 0.004 min
Lab File: BN012031.D
Acq: 30 Sep 2020 18:01

Tgt Ion:278 Resp: 10507
Ion Ratio Lower Upper
278 100
139 18.1 8.6 13.0#
279 32.6 19.8 29.8#





#41
Benzo(g,h,i)perylene
Concen: 0.39 ng
RT: 26.339 min Scan# 3354
Delta R.T. 0.004 min
Lab File: BN012031.D
Acq: 30 Sep 2020 18:01

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC

Tgt Ion:	276	Resp:	11626
Ion	Ratio	Lower	Upper
276	100		
277	25.1	19.4	29.0
138	22.5	11.3	16.9#

