

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN093022\
 Data File : BN021969.D
 Acq On : 30 Sep 2022 16:05
 Operator : CG/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 ICVBN093022

Quant Time: Sep 30 18:25:53 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN093022.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Sep 30 16:20:22 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	100	0.00
2	1,4-Dioxane	0.400	0.413	-3.2	100	0.00
3	n-Nitrosodimethylamine	0.400	0.396	1.0	98	0.00
4 S	2-Fluorophenol	0.400	0.386	3.5	97	0.00
5 S	Phenol-d6	0.400	0.381	4.8	98	0.00
6	bis(2-Chloroethyl)ether	0.400	0.396	1.0	99	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	98	0.00
8 S	Nitrobenzene-d5	0.400	0.387	3.3	98	0.00
9	Naphthalene	0.400	0.400	0.0	99	0.00
10	Hexachlorobutadiene	0.400	0.403	-0.8	100	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.413	-3.2	99	0.00
12	2-Methylnaphthalene	0.400	0.341	14.8	96	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	97	0.00
14 S	2,4,6-Tribromophenol	0.400	0.343	14.2	89	0.01
15 S	2-Fluorobiphenyl	0.400	0.405	-1.3	97	0.00
16	Acenaphthylene	0.400	0.364	9.0	95	0.00
17	Acenaphthene	0.400	0.394	1.5	97	0.00
18	Fluorene	0.400	0.374	6.5	91	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	95	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.340	15.0	91	0.00
21	4-Bromophenyl-phenylether	0.400	0.385	3.8	95	0.00
22	Hexachlorobenzene	0.400	0.405	-1.3	97	0.01
23	Atrazine	0.400	0.366	8.5	94	0.00
24	Pentachlorophenol	0.400	0.334	16.5	91	0.00
25	Phenanthrene	0.400	0.390	2.5	95	0.00
26	Anthracene	0.400	0.368	8.0	94	0.01
27 SURR	Fluoranthene-d10	0.400	0.385	3.8	94	0.00
28	Fluoranthene	0.400	0.387	3.3	95	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	94	0.00
30	Pyrene	0.400	0.394	1.5	94	0.00
31 S	Terphenyl-d14	0.400	0.422	-5.5	97	0.00
32	Benzo(a)anthracene	0.400	0.376	6.0	92	0.00
33	Chrysene	0.400	0.413	-3.2	96	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.360	10.0	91	0.00
35 I	Perylene-d12	0.400	0.400	0.0	96	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.401	-0.3	99	0.01
37	Benzo(b)fluoranthene	0.400	0.395	1.3	96	0.00
38	Benzo(k)fluoranthene	0.400	0.391	2.3	96	0.00
39 C	Benzo(a)pyrene	0.400	0.383	4.3	96	0.00
40	Dibenzo(a,h)anthracene	0.400	0.402	-0.5	99	0.01
41	Benzo(g,h,i)perylene	0.400	0.408	-2.0	99	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

ClientSampleId :

ICVBN093022