

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN100124\
 Data File : BN034295.D
 Acq On : 01 Oct 2024 09:59
 Operator : RC/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTD020323

Quant Time: Oct 01 23:30:44 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN091124.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Sep 30 14:14:38 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.387	152	238162	20.000	ng/u1	0.00
20) Naphthalene-d8	10.140	136	942947	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.040	164	569546	20.000	ng/u1	0.00
64) Phenanthrene-d10	16.798	188	1088699	20.000	ng/u1	0.00
79) Chrysene-d12	21.028	240	1042608	20.000	ng/u1	0.00
88) Perylene-d12	23.739	264	1192187	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	2.999	96	49021	8.399	ng/uL	0.00
4) Pyridine-d5	3.387	84	309815	17.859	ng/u1	0.00
7) Phenol-d5	6.587	99	367621	17.392	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	6.740	67	231927	17.685	ng/u1	0.00
11) 2-Chlorophenol-d4	6.928	132	336919	20.250	ng/u1	0.00
15) 4-Methylphenol-d8	8.105	113	295850	17.011	ng/u1	0.00
21) Nitrobenzene-d5	8.528	128	149045	19.831	ng/u1	0.00
24) 2-Nitrophenol-d4	9.240	143	158053	20.450	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	9.775	165	302192	20.762	ng/u1	0.00
31) 4-Chloroaniline-d4	10.293	131	402213	18.253	ng/u1	0.00
46) Dimethylphthalate-d6	13.463	166	806467	18.947	ng/u1	0.00
49) Acenaphthylene-d8	13.722	160	961737	20.085	ng/u1	0.00
54) 4-Nitrophenol-d4	14.269	143	133752	15.312	ng/u1	0.00
60) Fluorene-d10	15.046	176	690197	19.133	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.175	200	116176	17.897	ng/u1	0.00
73) Anthracene-d10	16.892	188	1024909	20.024	ng/u1	0.00
81) Pyrene-d10	19.204	212	1162001	20.003	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.551	264	1268580	20.715	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.034	88	50282	8.167	ng/uL	89
5) Pyridine	3.405	79	317250	17.871	ng/u1	97
6) Benzaldehyde	6.546	77	228665	20.495	ng/u1	91
8) Phenol	6.611	94	377921	17.173	ng/u1	94
10) Bis(2-Chloroethyl)ether	6.834	93	319202	18.663	ng/u1	93
12) 2-Chlorophenol	6.958	128	335246	19.662	ng/u1	95
13) 2-Methylphenol	7.840	108	288260	17.525	ng/u1	100
14) 2,2'-oxybis(1-Chloropr...	7.911	45	430045	17.049	ng/u1	97
16) Acetophenone	8.205	105	456524	16.570	ng/u1	95
17) N-Nitroso-di-n-propyla...	8.193	70	213782	15.670	ng/u1#	96
18) 4-Methylphenol	8.164	108	312105	17.049	ng/u1	97
19) Hexachloroethane	8.440	117	139165	19.351	ng/u1#	83
22) Nitrobenzene	8.569	77	339666	18.056	ng/u1	92
23) Isophorone	9.099	82	604988	17.440	ng/u1	96
25) 2-Nitrophenol	9.269	139	175903	20.337	ng/u1	92
26) 2,4-Dimethylphenol	9.352	107	322065	18.099	ng/u1	91
27) Bis(2-Chloroethoxy)met...	9.581	93	402051	18.836	ng/u1	99
29) 2,4-Dichlorophenol	9.805	162	302768	20.544	ng/u1	94
30) Naphthalene	10.193	128	1022127	19.817	ng/u1	99
32) 4-Chloroaniline	10.310	127	392216	18.062	ng/u1	99
33) Hexachlorobutadiene	10.481	225	203016	21.472	ng/u1	99
34) Caprolactam	11.081	113	44450	7.575	ng/u1	87
35) 4-Chloro-3-methylphenol	11.452	107	286471	16.511	ng/u1	93
36) 2-Methylnaphthalene	11.816	142	680033	19.141	ng/u1	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN100124\
 Data File : BN034295.D
 Acq On : 01 Oct 2024 09:59
 Operator : RC/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1


Instrument :
 BNA_N
 ClientSampleId :
 SSTD020323


Quant Time: Oct 01 23:30:44 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN091124.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Sep 30 14:14:38 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.040	142	685673	19.066	ng/ul	98
39) 1,2,4,5-Tetrachloroben...	12.199	216	363848	22.129	ng/ul	97
40) Hexachlorocyclopentadiene	12.175	237	204068	20.171	ng/ul	100
41) 2,4,6-Trichlorophenol	12.452	196	222319	20.396	ng/ul	97
42) 2,4,5-Trichlorophenol	12.522	196	258322	21.575	ng/ul	94
43) 1,1'-Biphenyl	12.863	154	870079	20.263	ng/ul	97
44) 2-Chloronaphthalene	12.893	162	680471	20.343	ng/ul	96
45) 2-Nitroaniline	13.116	65	164256	15.196	ng/ul	88
47) Dimethylphthalate	13.516	163	819601	18.893	ng/ul	99
48) 2,6-Dinitrotoluene	13.628	165	157319	18.417	ng/ul	90
50) Acenaphthylene	13.751	152	1023457	19.533	ng/ul	100
51) 3-Nitroaniline	13.957	138	162217	17.184	ng/ul	95
52) Acenaphthene	14.104	153	725920	19.406	ng/ul	97
53) 2,4-Dinitrophenol	14.169	184	64491	12.836	ng/ul#	83
55) 4-Nitrophenol	14.287	109	99050	13.352	ng/ul	90
56) Dibenzofuran	14.446	168	995138	19.454	ng/ul	94
57) 2,4-Dinitrotoluene	14.422	165	231630	17.987	ng/ul	93
58) 2,3,4,6-Tetrachlorophenol	14.681	232	204554	20.588	ng/ul#	87
59) Diethylphthalate	14.898	149	796935	17.580	ng/ul	99
61) Fluorene	15.098	166	789724	18.872	ng/ul	96
62) 4-Chlorophenyl-phenyle...	15.104	204	398824	19.663	ng/ul	93
63) 4-Nitroaniline	15.128	138	159840	16.455	ng/ul	94
66) 4,6-Dinitro-2-methylph...	15.193	198	133431	18.511	ng/ul	91
67) N-Nitrosodiphenylamine	15.322	169	687502	21.074	ng/ul	98
68) 4-Bromophenyl-phenylether	15.998	248	246446	22.292	ng/ul	94
69) Hexachlorobenzene	16.104	284	289002	22.523	ng/ul#	91
70) Atrazine	16.293	200	231274	19.566	ng/ul	97
71) Pentachlorophenol	16.451	266	167703	19.623	ng/ul	98
72) Phenanthrene	16.840	178	1206690	20.101	ng/ul	99
74) Anthracene	16.928	178	1256492	20.498	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	12.816	216	365515	24.123	ng/uL	96
76) Pentachlorobenzene	14.363	250	352863	22.872	ng/uL	90
77) Carbazole	17.210	167	1099964	19.497	ng/ul	99
78) Di-n-butylphthalate	17.792	149	1303851	18.180	ng/ul	100
80) Fluoranthene	18.869	202	1362559	20.064	ng/ul	99
82) Pyrene	19.234	202	1471760	20.159	ng/ul	95
83) Butylbenzylphthalate	20.175	149	569007	17.379	ng/ul	96
84) 3,3'-Dichlorobenzidine	20.957	252	445471	19.022	ng/ul	99
85) Benzo(a)anthracene	21.010	228	1407811	18.938	ng/ul	98
86) Bis(2-ethylhexyl)phtha...	20.981	149	870145	18.157	ng/ul#	95
87) Chrysene	21.069	228	1395514	19.675	ng/ul	99
89) Di-n-octyl phthalate	22.016	149	1440835	18.163	ng/ul	100
90) Benzo(b)fluoranthene	22.892	252	1423241	19.580	ng/ul	99
91) Benzo(k)fluoranthene	22.951	252	1528891	21.322	ng/ul	96
93) Benzo(a)pyrene	23.610	252	1382546	20.135	ng/ul	98
94) Indeno(1,2,3-cd)pyrene	26.727	276	1801218	21.171	ng/ul	99
95) Dibenzo(a,h)anthracene	26.780	278	1438642	20.919	ng/ul	98
96) Benzo(g,h,i)perylene	27.657	276	1408270	21.323	ng/ul	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Oct 01 23:30:44 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN091124.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Mon Sep 30 14:14:38 2024
Response via : Initial Calibration









#2
 1,4-Dioxane
 Concen: 8.167 ng/uL
 RT: 3.034 min Scan# 21
 Delta R.T. 0.000 min
 Lab File: BN034295.D
 Acq: 01 Oct 2024 09:59




Instrument :
 BNA_N
 ClientSampleId :
 SST020323


Tgt Ion: 88 Resp: 50282
 Ion Ratio Lower Upper
 88 100
 43 27.0 24.7 37.1
 58 69.2 64.6 96.8



#5
 Pyridine
 Concen: 17.871 ng/uL
 RT: 3.405 min Scan# 88
 Delta R.T. 0.000 min
 Lab File: BN034295.D
 Acq: 01 Oct 2024 09:59

Tgt Ion: 79 Resp: 317250
 Ion Ratio Lower Upper
 79 100
 52 78.2 65.2 97.8
 51 35.8 29.7 44.5









#6
Benzaldehyde
Concen: 20.495 ng/ul
RT: 6.546 min Scan# 61
Delta R.T. 0.000 min
Lab File: BN034295.D
Acq: 01 Oct 2024 09:59




Instrument :
BNA_N
ClientSampleId :
SSTD020323


Tgt Ion: 77 Resp: 228665
Ion Ratio Lower Upper
77 100
105 103.0 76.4 114.6
106 97.5 69.8 104.6



#8
Phenol
Concen: 17.173 ng/ul
RT: 6.611 min Scan# 633
Delta R.T. 0.000 min
Lab File: BN034295.D
Acq: 01 Oct 2024 09:59

Tgt Ion: 94 Resp: 377921
Ion Ratio Lower Upper
94 100
65 26.8 24.3 36.5
66 36.2 31.8 47.6









#10
Bis(2-Chloroethyl)ether
Concen: 18.663 ng/ul
RT: 6.834 min Scan# 671
Delta R.T. 0.000 min
Lab File: BN034295.D
Acq: 01 Oct 2024 09:59




Instrument :
BNA_N
ClientSampleId :
SSTD020323


Tgt Ion: 93 Resp: 319202
Ion Ratio Lower Upper
93 100
63 71.9 63.6 95.4
95 32.9 25.7 38.5



#12
2-Chlorophenol
Concen: 19.662 ng/ul
RT: 6.958 min Scan# 692
Delta R.T. 0.000 min
Lab File: BN034295.D
Acq: 01 Oct 2024 09:59

Tgt Ion:128 Resp: 335246
Ion Ratio Lower Upper
128 100
64 43.2 39.1 58.7
130 31.7 25.8 38.6









#13
2-Methylphenol
Concen: 17.525 ng/ul
RT: 7.840 min Scan# 842
Delta R.T. 0.000 min
Lab File: BN034295.D
Acq: 01 Oct 2024 09:59




Instrument :
BNA_N
ClientSampleId :
SSTD020323


Tgt Ion:108 Resp: 288260
Ion Ratio Lower Upper
108 100
107 90.3 71.9 107.9



#14
2,2'-oxybis(1-Chloropropane)
Concen: 17.049 ng/ul
RT: 7.911 min Scan# 854
Delta R.T. 0.000 min
Lab File: BN034295.D
Acq: 01 Oct 2024 09:59

Tgt Ion: 45 Resp: 430045
Ion Ratio Lower Upper
45 100
77 15.0 11.0 16.6
79 11.1 7.8 11.8









#16
Acetophenone
Concen: 16.570 ng/ul
RT: 8.205 min Scan# 904
Delta R.T. 0.000 min
Lab File: BN034295.D
Acq: 01 Oct 2024 09:59




Instrument :
BNA_N
ClientSampleId :
SSTD020323


Tgt Ion:105 Resp: 456524
Ion Ratio Lower Upper
105 100
77 72.7 61.1 91.7
51 25.3 22.6 33.8



#17
N-Nitroso-di-n-propylamine
Concen: 15.670 ng/ul
RT: 8.193 min Scan# 902
Delta R.T. -0.006 min
Lab File: BN034295.D
Acq: 01 Oct 2024 09:59

Tgt Ion: 70 Resp: 213782
Ion Ratio Lower Upper
70 100
42 58.1 47.9 71.9
101 10.2 7.5 11.3
130 24.6 16.2 24.2#









#18
4-Methylphenol
Concen: 17.049 ng/ul
RT: 8.164 min Scan# 897
Delta R.T. 0.000 min
Lab File: BN034295.D
Acq: 01 Oct 2024 09:59




Instrument :
BNA_N
ClientSampleId :
SSTD020323


Tgt Ion:108 Resp: 312105
Ion Ratio Lower Upper
108 100
107 109.0 89.4 134.2



#19
Hexachloroethane
Concen: 19.351 ng/ul
RT: 8.440 min Scan# 944
Delta R.T. 0.000 min
Lab File: BN034295.D
Acq: 01 Oct 2024 09:59

Tgt Ion:117 Resp: 139165
Ion Ratio Lower Upper
117 100
201 107.3 73.9 110.9
199 68.2 44.0 66.0#









#22
Nitrobenzene
Concen: 18.056 ng/ul
RT: 8.569 min Scan# 966
Delta R.T. 0.000 min
Lab File: BN034295.D
Acq: 01 Oct 2024 09:59




Instrument :
BNA_N
ClientSampleId :
SSTD020323


Tgt Ion: 77 Resp: 339666
Ion Ratio Lower Upper
77 100
123 49.8 34.5 51.7
65 13.7 11.0 16.6



#23
Isophorone
Concen: 17.440 ng/ul
RT: 9.099 min Scan# 1056
Delta R.T. 0.000 min
Lab File: BN034295.D
Acq: 01 Oct 2024 09:59

Tgt Ion: 82 Resp: 604988
Ion Ratio Lower Upper
82 100
95 6.9 6.1 9.1
138 18.8 13.4 20.0









#25
2-Nitrophenol
Concen: 20.337 ng/ul
RT: 9.269 min Scan# 1085
Delta R.T. -0.006 min
Lab File: BN034295.D
Acq: 01 Oct 2024 09:59




Instrument :
BNA_N
ClientSampleId :
SSTD020323


Tgt Ion:139 Resp: 175903
Ion Ratio Lower Upper
139 100
65 46.8 42.0 63.0
109 28.4 26.5 39.7



#26
2,4-Dimethylphenol
Concen: 18.099 ng/ul
RT: 9.352 min Scan# 1099
Delta R.T. 0.000 min
Lab File: BN034295.D
Acq: 01 Oct 2024 09:59

Tgt Ion:107 Resp: 322065
Ion Ratio Lower Upper
107 100
121 53.4 38.4 57.6
122 87.1 62.9 94.3





#27

Bis(2-Chloroethoxy)methane

Concen: 18.836 ng/ul

RT: 9.581 min Scan# 1138

Delta R.T. -0.006 min

Lab File: BN034295.D

Acq: 01 Oct 2024 09:59

Instrument :

BNA_N

ClientSampleId :

SSTD020323





Tgt Ion: 93 Resp: 402051

Ion Ratio Lower Upper

93 100

95 31.7 24.9 37.3

123 11.4 8.8 13.2



#29

2,4-Dichlorophenol

Concen: 20.544 ng/ul

RT: 9.805 min Scan# 1176

Delta R.T. 0.000 min

Lab File: BN034295.D

Acq: 01 Oct 2024 09:59




Tgt Ion:162 Resp: 302768


Ion Ratio Lower Upper

162 100

164 62.6 53.0 79.6

98 33.0 30.6 46.0





#30

Naphthalene

Concen: 19.817 ng/ul

RT: 10.193 min Scan# 1242

Delta R.T. 0.000 min

Lab File: BN034295.D

Acq: 01 Oct 2024 09:59

Instrument :

BNA_N

ClientSampleId :

SSTD020323





Tgt Ion:128 Resp: 1022127

Ion Ratio Lower Upper

128 100

129 10.7 8.7 13.1

127 13.0 10.6 16.0



#32

4-Chloroaniline

Concen: 18.062 ng/ul

RT: 10.310 min Scan# 1262

Delta R.T. -0.006 min

Lab File: BN034295.D




Acq: 01 Oct 2024 09:59


Tgt Ion:127 Resp: 392216

Ion Ratio Lower Upper

127 100

129 32.3 26.2 39.2









#33
Hexachlorobutadiene
Concen: 21.472 ng/ul
RT: 10.481 min Scan# 1291
Delta R.T. -0.006 min
Lab File: BN034295.D
Acq: 01 Oct 2024 09:59




Instrument :
BNA_N
ClientSampleId :
SSTD020323


Tgt Ion:225 Resp: 203016
Ion Ratio Lower Upper
225 100
223 65.0 52.2 78.2
227 64.2 50.1 75.1



#34
Caprolactam
Concen: 7.575 ng/ul
RT: 11.081 min Scan# 1393
Delta R.T. -0.006 min
Lab File: BN034295.D
Acq: 01 Oct 2024 09:59

Tgt Ion:113 Resp: 44450
Ion Ratio Lower Upper
113 100
55 151.5 137.0 205.6
56 112.3 99.7 149.5









#35
 4-Chloro-3-methylphenol
 Concen: 16.511 ng/ul
 RT: 11.452 min Scan# 1456
 Delta R.T. 0.000 min
 Lab File: BN034295.D
 Acq: 01 Oct 2024 09:59




Instrument :
 BNA_N
 ClientSampleId :
 SST020323


Tgt Ion:107 Resp: 286471
 Ion Ratio Lower Upper
 107 100
 144 27.2 19.1 28.7
 142 83.2 61.4 92.0



#36
 2-Methylnaphthalene
 Concen: 19.141 ng/ul
 RT: 11.816 min Scan# 1518
 Delta R.T. 0.000 min
 Lab File: BN034295.D
 Acq: 01 Oct 2024 09:59

Tgt Ion:142 Resp: 680033
 Ion Ratio Lower Upper
 142 100
 141 91.5 72.2 108.2





#37

1-Methylnaphthalene

Concen: 19.066 ng/ul

RT: 12.040 min Scan# 1556

Delta R.T. 0.000 min

Lab File: BN034295.D

Acq: 01 Oct 2024 09:59

Instrument :

BNA_N

ClientSampleId :

SSTD020323





Tgt Ion:142 Resp: 685673

Ion Ratio Lower Upper

142 100

141 92.2 72.5 108.7

116 3.5 3.2 4.8



#39

1,2,4,5-Tetrachlorobenzene

Concen: 22.129 ng/ul

RT: 12.199 min Scan# 1583

Delta R.T. 0.000 min

Lab File: BN034295.D

Acq: 01 Oct 2024 09:59

Tgt Ion:216 Resp: 363848




Ion Ratio Lower Upper


216 100

214 77.3 63.0 94.6

179 20.2 18.2 27.2

108 16.2 15.0 22.4





#40

Hexachlorocyclopentadiene

Concen: 20.171 ng/ul

RT: 12.175 min Scan# 1579

Delta R.T. -0.006 min

Lab File: BN034295.D

Acq: 01 Oct 2024 09:59

Instrument :

BNA_N

ClientSampleId :

SSTD020323





Tgt Ion:237 Resp: 204068

Ion Ratio Lower Upper

237 100

235 63.1 50.7 76.1

272 12.8 10.1 15.1



#41

2,4,6-Trichlorophenol

Concen: 20.396 ng/ul

RT: 12.452 min Scan# 1626

Delta R.T. 0.000 min

Lab File: BN034295.D

Acq: 01 Oct 2024 09:59




Tgt Ion:196 Resp: 222319


Ion Ratio Lower Upper

196 100

198 97.6 75.0 112.4

200 31.4 24.7 37.1









#42
 2,4,5-Trichlorophenol
 Concen: 21.575 ng/ul
 RT: 12.522 min Scan# 1638
 Delta R.T. 0.000 min
 Lab File: BN034295.D
 Acq: 01 Oct 2024 09:59

Instrument :
 BNA_N
 ClientSampleId :
 SST020323

Tgt Ion:196 Resp: 258322




Ion	Ratio	Lower	Upper
196	100		
198	89.1	77.1	115.7
200	30.4	25.2	37.8




#43
 1,1'-Biphenyl
 Concen: 20.263 ng/ul
 RT: 12.863 min Scan# 1696
 Delta R.T. 0.000 min
 Lab File: BN034295.D
 Acq: 01 Oct 2024 09:59

Tgt Ion:154 Resp: 870079

Ion	Ratio	Lower	Upper
154	100		
153	39.7	33.3	49.9
76	11.8	10.2	15.4









#44
 2-Chloronaphthalene
 Concen: 20.343 ng/ul
 RT: 12.893 min Scan# 1701
 Delta R.T. -0.006 min
 Lab File: BN034295.D
 Acq: 01 Oct 2024 09:59

Instrument :
 BNA_N
 ClientSampleId :
 SSTD020323

Tgt Ion:162 Resp: 680471




Ion	Ratio	Lower	Upper
162	100		
164	33.2	25.1	37.7
127	38.4	32.6	48.8




#45
 2-Nitroaniline
 Concen: 15.196 ng/ul
 RT: 13.116 min Scan# 1739
 Delta R.T. 0.000 min
 Lab File: BN034295.D
 Acq: 01 Oct 2024 09:59

Tgt Ion: 65 Resp: 164256

Ion	Ratio	Lower	Upper
65	100		
92	72.5	53.2	79.8
138	116.2	80.9	121.3









#47
Dimethylphthalate
Concen: 18.893 ng/ul
RT: 13.516 min Scan# 1807
Delta R.T. 0.000 min
Lab File: BN034295.D
Acq: 01 Oct 2024 09:59




Instrument :
BNA_N
ClientSampleId :
SSTD020323


Tgt Ion:163 Resp: 819601
Ion Ratio Lower Upper
163 100
194 4.1 3.0 4.6
164 9.9 7.7 11.5



#48
2,6-Dinitrotoluene
Concen: 18.417 ng/ul
RT: 13.628 min Scan# 1826
Delta R.T. -0.006 min
Lab File: BN034295.D
Acq: 01 Oct 2024 09:59

Tgt Ion:165 Resp: 157319
Ion Ratio Lower Upper
165 100
89 52.7 48.6 72.8
121 20.1 19.0 28.6









#50
Acenaphthylene
Concen: 19.533 ng/ul
RT: 13.751 min Scan# 1847
Delta R.T. -0.006 min
Lab File: BN034295.D
Acq: 01 Oct 2024 09:59




Instrument :
BNA_N
ClientSampleId :
SSTD020323


Tgt Ion:152 Resp: 1023457
Ion Ratio Lower Upper
152 100
151 20.2 16.2 24.4
153 13.3 10.6 16.0



#51
3-Nitroaniline
Concen: 17.184 ng/ul
RT: 13.957 min Scan# 1882
Delta R.T. 0.000 min
Lab File: BN034295.D
Acq: 01 Oct 2024 09:59

Tgt Ion:138 Resp: 162217
Ion Ratio Lower Upper
138 100
108 10.2 9.0 13.4
92 112.7 94.6 141.8





#52

Acenaphthene

Concen: 19.406 ng/ul

RT: 14.104 min Scan# 1907

Delta R.T. 0.000 min

Lab File: BN034295.D

Acq: 01 Oct 2024 09:59

Instrument :

BNA_N

ClientSampleId :

SSTD020323





Tgt Ion:153 Resp: 725920

Ion Ratio Lower Upper

153 100

152 44.7 37.8 56.8

154 86.8 71.2 106.8



#53

2,4-Dinitrophenol

Concen: 12.836 ng/ul

RT: 14.169 min Scan# 1918

Delta R.T. 0.000 min

Lab File: BN034295.D

Acq: 01 Oct 2024 09:59




Tgt Ion:184 Resp: 64491


Ion Ratio Lower Upper

184 100

63 59.9 61.1 91.7#

154 61.3 58.5 87.7









#55
4-Nitrophenol
Concen: 13.352 ng/ul
RT: 14.287 min Scan# 1938
Delta R.T. 0.000 min
Lab File: BN034295.D
Acq: 01 Oct 2024 09:59




Instrument :
BNA_N
ClientSampleId :
SSTD020323


Tgt Ion:109 Resp: 99050
Ion Ratio Lower Upper
109 100
139 146.6 102.8 154.2
65 125.9 97.7 146.5



#56
Dibenzofuran
Concen: 19.454 ng/ul
RT: 14.446 min Scan# 1965
Delta R.T. 0.000 min
Lab File: BN034295.D
Acq: 01 Oct 2024 09:59

Tgt Ion:168 Resp: 995138
Ion Ratio Lower Upper
168 100
139 36.2 31.7 47.5









#57
2,4-Dinitrotoluene
Concen: 17.987 ng/ul
RT: 14.422 min Scan# 1961
Delta R.T. -0.006 min
Lab File: BN034295.D
Acq: 01 Oct 2024 09:59




Instrument :
BNA_N
ClientSampleId :
SSTD020323


Tgt Ion:165 Resp: 231630
Ion Ratio Lower Upper
165 100
63 42.2 37.8 56.8
182 3.1 2.3 3.5



#58
2,3,4,6-Tetrachlorophenol
Concen: 20.588 ng/ul
RT: 14.681 min Scan# 2005
Delta R.T. 0.000 min
Lab File: BN034295.D
Acq: 01 Oct 2024 09:59

Tgt Ion:232 Resp: 204554
Ion Ratio Lower Upper
232 100
131 40.0 41.0 61.6#
130 1.9 2.2 3.2#
166 28.0 27.1 40.7





#59

Diethylphthalate

Concen: 17.580 ng/ul

RT: 14.898 min Scan# 2042

Delta R.T. 0.000 min

Lab File: BN034295.D

Acq: 01 Oct 2024 09:59

Instrument :

BNA_N

ClientSampleId :

SSTD020323





Tgt Ion:149 Resp: 796935

Ion Ratio Lower Upper

149 100

177 22.1 16.9 25.3

150 12.4 9.9 14.9



#61

Fluorene

Concen: 18.872 ng/ul

RT: 15.098 min Scan# 2076

Delta R.T. 0.000 min

Lab File: BN034295.D

Acq: 01 Oct 2024 09:59




Tgt Ion:166 Resp: 789724


Ion Ratio Lower Upper

166 100

165 98.3 81.9 122.9

167 13.5 10.9 16.3









#62
4-Chlorophenyl-phenylether
Concen: 19.663 ng/ul
RT: 15.104 min Scan# 2077
Delta R.T. 0.000 min
Lab File: BN034295.D
Acq: 01 Oct 2024 09:59




Instrument :
BNA_N
ClientSampleId :
SSTD020323


Tgt Ion:204 Resp: 398824
Ion Ratio Lower Upper
204 100
206 30.8 26.2 39.4
141 56.1 50.4 75.6



#63
4-Nitroaniline
Concen: 16.455 ng/ul
RT: 15.128 min Scan# 2081
Delta R.T. -0.006 min
Lab File: BN034295.D
Acq: 01 Oct 2024 09:59

Tgt Ion:138 Resp: 159840
Ion Ratio Lower Upper
138 100
92 51.0 43.0 64.4
108 76.5 67.0 100.4









#66
 4,6-Dinitro-2-methylphenol
 Concen: 18.511 ng/ul
 RT: 15.193 min Scan# 2092
 Delta R.T. 0.000 min
 Lab File: BN034295.D
 Acq: 01 Oct 2024 09:59




Instrument :
 BNA_N
 ClientSampleId :
 SSTD020323


Tgt Ion:198 Resp: 133431
 Ion Ratio Lower Upper
 198 100
 182 4.7 3.2 4.8
 77 21.3 21.1 31.7



#67
 N-Nitrosodiphenylamine
 Concen: 21.074 ng/ul
 RT: 15.322 min Scan# 2114
 Delta R.T. 0.000 min
 Lab File: BN034295.D
 Acq: 01 Oct 2024 09:59

Tgt Ion:169 Resp: 687502
 Ion Ratio Lower Upper
 169 100
 168 64.0 52.8 79.2
 167 35.1 28.6 43.0









#68
4-Bromophenyl-phenylether
Concen: 22.292 ng/ul
RT: 15.998 min Scan# 211
Delta R.T. -0.006 min
Lab File: BN034295.D
Acq: 01 Oct 2024 09:59




Instrument :
BNA_N
ClientSampleId :
SSTD020323


Tgt Ion:248 Resp: 246446
Ion Ratio Lower Upper
248 100
250 103.4 79.8 119.6
141 76.1 67.0 100.4



#69
Hexachlorobenzene
Concen: 22.523 ng/ul
RT: 16.104 min Scan# 2247
Delta R.T. 0.000 min
Lab File: BN034295.D
Acq: 01 Oct 2024 09:59

Tgt Ion:284 Resp: 289002
Ion Ratio Lower Upper
284 100
142 30.8 31.3 46.9#
249 29.7 25.1 37.7









#70
 Atrazine
 Concen: 19.566 ng/ul
 RT: 16.293 min Scan# 21
 Delta R.T. 0.000 min
 Lab File: BN034295.D
 Acq: 01 Oct 2024 09:59




Instrument : BNA_N
 ClientSampleId : SST020323


Tgt Ion:	200	Resp:	231274
Ion Ratio	Lower	Upper	
200	100		
173	27.2	21.3	31.9
215	50.1	38.2	57.4



#71
 Pentachlorophenol
 Concen: 19.623 ng/ul
 RT: 16.451 min Scan# 2306
 Delta R.T. -0.006 min
 Lab File: BN034295.D
 Acq: 01 Oct 2024 09:59

Tgt Ion:	266	Resp:	167703
Ion Ratio	Lower	Upper	
266	100		
264	63.9	51.1	76.7
268	68.9	52.8	79.2









#72
 Phenanthrene
 Concen: 20.101 ng/ul
 RT: 16.840 min Scan# 2372
 Delta R.T. 0.000 min
 Lab File: BN034295.D
 Acq: 01 Oct 2024 09:59

Instrument :
 BNA_N
 ClientSampleId :
 SST020323

Tgt Ion:178 Resp: 1206690




Ion	Ratio	Lower	Upper
178	100		
179	15.4	11.9	17.9
176	19.0	15.2	22.8




#74
 Anthracene
 Concen: 20.498 ng/ul
 RT: 16.928 min Scan# 2387
 Delta R.T. -0.006 min
 Lab File: BN034295.D
 Acq: 01 Oct 2024 09:59

Tgt Ion:178 Resp: 1256492

Ion	Ratio	Lower	Upper
178	100		
179	15.2	12.2	18.2
176	18.2	15.0	22.6









#75
 1,2,3,4-Tetrachlorobenzene
 Concen: 24.123 ng/uL
 RT: 12.816 min Scan# 1688
 Delta R.T. -0.006 min
 Lab File: BN034295.D
 Acq: 01 Oct 2024 09:59




Instrument :
 BNA_N
 ClientSampleId :
 SST020323


Tgt Ion:	216	Resp:	365515
Ion Ratio	Lower	Upper	
216	100		
214	77.2	63.6	95.4
218	50.3	37.6	56.4



#76
 Pentachlorobenzene
 Concen: 22.872 ng/uL
 RT: 14.363 min Scan# 1951
 Delta R.T. 0.000 min
 Lab File: BN034295.D
 Acq: 01 Oct 2024 09:59

Tgt Ion:	250	Resp:	352863
Ion Ratio	Lower	Upper	
250	100		
248	66.6	49.3	73.9
252	71.1	49.1	73.7









#77
 Carbazole
 Concen: 19.497 ng/ul
 RT: 17.210 min Scan# 2435
 Delta R.T. 0.000 min
 Lab File: BN034295.D
 Acq: 01 Oct 2024 09:59




Instrument :
 BNA_N
 ClientSampleId :
 SSTD020323


Tgt Ion:167 Resp: 1099964
 Ion Ratio Lower Upper
 167 100
 166 21.7 17.4 26.0
 139 11.8 10.5 15.7



#78
 Di-n-butylphthalate
 Concen: 18.180 ng/ul
 RT: 17.792 min Scan# 2534
 Delta R.T. -0.006 min
 Lab File: BN034295.D
 Acq: 01 Oct 2024 09:59

Tgt Ion:149 Resp: 1303851
 Ion Ratio Lower Upper
 149 100
 150 9.4 7.4 11.2
 104 5.1 4.3 6.5





#80

Fluoranthene

Concen: 20.064 ng/ul

RT: 18.869 min Scan# 2717

Delta R.T. 0.000 min

Lab File: BN034295.D

Acq: 01 Oct 2024 09:59

Instrument :

BNA_N

ClientSampleId :

SSTD020323





Tgt Ion:202 Resp: 1362559

Ion Ratio Lower Upper

202 100

101 11.6 9.6 14.4

100 9.1 7.7 11.5



#82

Pyrene

Concen: 20.159 ng/ul

RT: 19.234 min Scan# 2779

Delta R.T. -0.006 min

Lab File: BN034295.D

Acq: 01 Oct 2024 09:59




Tgt Ion:202 Resp: 1471760


Ion Ratio Lower Upper

202 100

101 12.9 11.8 17.8

100 10.7 10.0 15.0





#83

Butylbenzylphthalate

Concen: 17.379 ng/ul

RT: 20.175 min Scan# 2939

Delta R.T. -0.006 min

Lab File: BN034295.D

Acq: 01 Oct 2024 09:59

Instrument :

BNA_N

ClientSampleId :

SSTD020323





Tgt Ion:149 Resp: 569007

Ion Ratio Lower Upper

149 100

91 67.0 56.3 84.5

206 20.7 15.9 23.9



#84

3,3'-Dichlorobenzidine

Concen: 19.022 ng/ul

RT: 20.957 min Scan# 3072

Delta R.T. 0.000 min

Lab File: BN034295.D

Acq: 01 Oct 2024 09:59




Tgt Ion:252 Resp: 445471


Ion Ratio Lower Upper

252 100

254 64.1 52.0 78.0

126 10.9 9.7 14.5





#85

Benzo(a)anthracene

Concen: 18.938 ng/ul

RT: 21.010 min Scan# 3081

Delta R.T. -0.006 min

Lab File: BN034295.D

Acq: 01 Oct 2024 09:59

Instrument :

BNA_N

ClientSampleId :

SSTD020323





Tgt Ion:228 Resp: 1407811

Ion Ratio Lower Upper

228 100

229 19.2 15.0 22.6

226 26.8 20.2 30.4



#86

Bis(2-ethylhexyl)phthalate

Concen: 18.157 ng/ul

RT: 20.981 min Scan# 3076

Delta R.T. 0.000 min

Lab File: BN034295.D

Acq: 01 Oct 2024 09:59




Tgt Ion:149 Resp: 870145


Ion Ratio Lower Upper

149 100

167 28.6 20.6 31.0

279 4.7 3.0 4.6#









#87
Chrysene
Concen: 19.675 ng/ul
RT: 21.069 min Scan# 3091
Delta R.T. -0.006 min
Lab File: BN034295.D
Acq: 01 Oct 2024 09:59




Instrument :
BNA_N
ClientSampleId :
SSTD020323


Tgt Ion:228 Resp: 1395514
Ion Ratio Lower Upper
228 100
226 28.9 22.5 33.7
229 19.1 15.8 23.6



#89
Di-n-octyl phthalate
Concen: 18.163 ng/ul
RT: 22.016 min Scan# 3252
Delta R.T. -0.006 min
Lab File: BN034295.D
Acq: 01 Oct 2024 09:59

Tgt Ion:149 Resp: 1440835





#90

Benzo(b)fluoranthene

Concen: 19.580 ng/ul

RT: 22.892 min Scan# 3401

Delta R.T. 0.000 min

Lab File: BN034295.D

Acq: 01 Oct 2024 09:59

Instrument :

BNA_N

ClientSampleId :

SSTD020323





Tgt Ion:252 Resp: 1423241

Ion Ratio Lower Upper

252 100

253 21.8 17.8 26.6

125 10.5 9.1 13.7



#91

Benzo(k)fluoranthene

Concen: 21.322 ng/ul

RT: 22.951 min Scan# 3411

Delta R.T. 0.006 min

Lab File: BN034295.D

Acq: 01 Oct 2024 09:59




Tgt Ion:252 Resp: 1528891


Ion Ratio Lower Upper

252 100

253 19.4 17.4 26.0

125 9.6 8.8 13.2





#93

Benzo(a)pyrene

Concen: 20.135 ng/ul

RT: 23.610 min Scan# 3523

Delta R.T. -0.006 min

Lab File: BN034295.D

Acq: 01 Oct 2024 09:59

Instrument :

BNA_N

ClientSampleId :

SSTD020323





Tgt Ion:252 Resp: 1382546

Ion Ratio Lower Upper

252 100

253 21.1 15.8 23.8

125 11.7 9.0 13.6



#94

Indeno(1,2,3-cd)pyrene

Concen: 21.171 ng/ul

RT: 26.727 min Scan# 4053

Delta R.T. -0.012 min

Lab File: BN034295.D

Acq: 01 Oct 2024 09:59




Tgt Ion:276 Resp: 1801218


Ion Ratio Lower Upper

276 100

138 21.5 17.7 26.5

277 23.0 18.4 27.6





#95

Dibenzo(a,h)anthracene

Concen: 20.919 ng/ul

RT: 26.780 min Scan# 4062

Delta R.T. -0.012 min

Lab File: BN034295.D

Acq: 01 Oct 2024 09:59

Instrument :

BNA_N

ClientSampleId :

SSTD020323





Tgt Ion:278 Resp: 1438642

Ion Ratio Lower Upper

278 100

139 16.3 14.2 21.2

279 23.6 18.1 27.1



#96

Benzo(g,h,i)perylene

Concen: 21.323 ng/ul

RT: 27.657 min Scan# 4211

Delta R.T. -0.006 min

Lab File: BN034295.D

Acq: 01 Oct 2024 09:59

Tgt Ion:276 Resp: 1408270

Ion Ratio Lower Upper

276 100

138 21.3 19.0 28.4

277 23.5 18.0 27.0

