

Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN100219\
 Data File : BN007982.D
 Acq On : 03 Oct 2019 01:04
 Operator : JU
 Sample : PB123477BL
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 PB123477BL

Quant Time: Oct 03 13:00:17 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN091819.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 18 18:24:40 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.69	152	5216	0.40	ng	-0.02
7) Naphthalene-d8	10.46	136	22069	0.40	ng	-0.01
13) Acenaphthene-d10	14.31	164	13594	0.40	ng	-0.02
19) Phenanthrene-d10	17.06	188	29120	0.40	ng	-0.01
27) Chrysene-d12	21.25	240	31968	0.40	ng	-0.02
34) Perylene-d12	23.46	264	34214	0.40	ng	-0.03

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
4) 2-Fluorophenol	5.31	112	5325	0.30	ng	0.00
5) Phenol-d6	6.87	99	7540	0.34	ng	0.00
8) Nitrobenzene-d5	8.82	82	6101	0.32	ng	-0.01
11) 2-Methylnaphthalene-d10	12.05	152	12563	0.34	ng	-0.01
14) 2,4,6-Tribromophenol	15.80	330	2286	0.31	ng	-0.02
15) 2-Fluorobiphenyl	12.93	172	17393	0.31	ng	-0.02
25) Fluoranthene-d10	19.09	212	35931	0.37	ng	-0.02
29) Terphenyl-d14	19.70	244	29839	0.38	ng	-0.02

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN100219\
Data File : BN007982.D
Acq On : 03 Oct 2019 01:04
Operator : JU
Sample : PB123477BL
Misc :
ALS Vial : 18 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
PB123477BL

Quant Time: Oct 03 13:00:17 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN091819.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Sep 18 18:24:40 2019
Response via : Initial Calibration

