

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN100224\
 Data File : BN034347.D
 Acq On : 02 Oct 2024 22:50
 Operator : RC/JU
 Sample : PB163513BL
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SBLK513

Quant Time: Oct 03 12:46:20 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN091124.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Sep 30 14:14:38 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.387	152	261184	20.000	ng/ul	0.00
20) Naphthalene-d8	10.140	136	1039657	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.039	164	623368	20.000	ng/ul	0.00
64) Phenanthrene-d10	16.798	188	1207724	20.000	ng/ul	0.00
79) Chrysene-d12	21.027	240	1167747	20.000	ng/ul	0.00
88) Perylene-d12	23.739	264	1415215	20.000	ng/ul	0.00

System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.005	96	33646	5.257	ng/uL	0.00
4) Pyridine-d5	3.387	84	422752	22.221	ng/ul	0.00
7) Phenol-d5	6.587	99	539651	23.281	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	6.740	67	325903	22.660	ng/ul	0.00
11) 2-Chlorophenol-d4	6.928	132	484743	26.567	ng/ul	0.00
15) 4-Methylphenol-d8	8.099	113	437291	22.927	ng/ul	0.00
21) Nitrobenzene-d5	8.528	128	228289	27.549	ng/ul	0.00
24) 2-Nitrophenol-d4	9.240	143	245759	28.841	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	9.775	165	421310	26.253	ng/ul	0.00
31) 4-Chloroaniline-d4	10.287	131	597135	24.578	ng/ul	0.00
46) Dimethylphthalate-d6	13.463	166	1222131	26.234	ng/ul	0.00
49) Acenaphthylene-d8	13.722	160	1446802	27.606	ng/ul	0.00
54) 4-Nitrophenol-d4	14.269	143	164676	17.224	ng/ul	0.00
60) Fluorene-d10	15.045	176	1054419	26.706	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.175	200	149962	20.825	ng/ul	0.00
73) Anthracene-d10	16.892	188	1618063	28.497	ng/ul	0.00
81) Pyrene-d10	19.204	212	1869216	28.729	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.557	264	2087360	28.713	ng/ul	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN100224\
Data File : BN034347.D
Acq On : 02 Oct 2024 22:50
Operator : RC/JU
Sample : PB163513BL
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SBLK513

Quant Time: Oct 03 12:46:20 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN091124.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Mon Sep 30 14:14:38 2024
Response via : Initial Calibration

