

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN100224\
 Data File : BN034348.D
 Acq On : 02 Oct 2024 23:29
 Operator : RC/JU
 Sample : P4019-06DL2 10X
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DDCA9DL2

Quant Time: Oct 03 00:10:51 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN091124.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Sep 30 14:14:38 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.381	152	233859	20.000	ng/ul	0.00
20) Naphthalene-d8	10.140	136	940305	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.040	164	582159	20.000	ng/ul	0.00
64) Phenanthrene-d10	16.798	188	1154004	20.000	ng/ul	0.00
79) Chrysene-d12	21.028	240	1113362	20.000	ng/ul	0.00
88) Perylene-d12	23.739	264	1340458	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	0.000	96	0d	0.000	ng/uL	
4) Pyridine-d5	0.000	84	0d	0.000	ng/ul	
7) Phenol-d5	6.587	99	26616	1.282	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	6.740	67	19805	1.538	ng/ul	0.00
11) 2-Chlorophenol-d4	6.922	132	24516	1.501	ng/ul	0.00
15) 4-Methylphenol-d8	8.105	113	20368	1.193	ng/ul	0.00
21) Nitrobenzene-d5	8.528	128	10535	1.406	ng/ul	0.00
24) 2-Nitrophenol-d4	9.240	143	9748	1.265	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	9.775	165	20633	1.422	ng/ul	0.00
31) 4-Chloroaniline-d4	10.293	131	24626	1.121	ng/ul	0.00
46) Dimethylphthalate-d6	13.463	166	80322	1.846	ng/ul	0.00
49) Acenaphthylene-d8	13.722	160	90370	1.846	ng/ul	0.00
54) 4-Nitrophenol-d4	0.000	143	0d	0.000	ng/ul	
60) Fluorene-d10	15.040	176	72639	1.970	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	0.000	200	0d	0.000	ng/ul	
73) Anthracene-d10	16.892	188	111164	2.049	ng/ul	0.00
81) Pyrene-d10	19.204	212	131164	2.114	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.551	264	131660	1.912	ng/ul	0.00
Target Compounds						
58) 2,3,4,6-Tetrachlorophenol	14.675	232	110456	10.876	ng/ul	96
71) Pentachlorophenol	16.457	266	979228	108.097	ng/ul	97


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN100224\
Data File : BN034348.D
Acq On : 02 Oct 2024 23:29
Operator : RC/JU
Sample : P4019-06DL2 10X
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DDCA9DL2


Quant Time: Oct 03 00:10:51 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN091124.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Mon Sep 30 14:14:38 2024
Response via : Initial Calibration








#58
2,3,4,6-Tetrachlorophenol
Concen: 10.876 ng/ul
RT: 14.675 min Scan# 20
Delta R.T. -0.006 min
Lab File: BN034348.D
Acq: 02 Oct 2024 23:29


Instrument :
BNA_N
ClientSampleId :
DDCA9DL2



Tgt Ion:232 Resp: 110456
Ion Ratio Lower Upper
232 100
131 47.4 41.0 61.6
130 2.6 2.2 3.2
166 32.1 27.1 40.7



#71
Pentachlorophenol
Concen: 108.097 ng/ul
RT: 16.457 min Scan# 2307
Delta R.T. 0.000 min
Lab File: BN034348.D
Acq: 02 Oct 2024 23:29



Tgt Ion:266 Resp: 979228
Ion Ratio Lower Upper
266 100
264 63.2 51.1 76.7
268 62.3 52.8 79.2

