

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN100319\
 Data File : BN007989.D
 Acq On : 03 Oct 2019 20:24
 Operator : JU
 Sample : K5077-08
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 PST-024-B278-092519

Manual Integrations
 APPROVED

mohammad
 10/7/2019 8:27:43 AM

Quant Time: Oct 04 08:38:04 2019

Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN091819.M

Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION

QLast Update : Thu Oct 03 18:39:49 2019

Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.69	152	4855	0.40	ng	-0.02
7) Naphthalene-d8	10.46	136	19978	0.40	ng	-0.01
13) Acenaphthene-d10	14.30	164	12618	0.40	ng	-0.02
19) Phenanthrene-d10	17.05	188	27814	0.40	ng	-0.02
27) Chrysene-d12	21.25	240	30783	0.40	ng	-0.02
34) Perylene-d12	23.46	264	33274	0.40	ng	-0.03

System Monitoring Compounds

4) 2-Fluorophenol	5.30	112	4790	0.29	ng	-0.02
5) Phenol-d6	6.86	99	6685	0.32	ng	0.00
8) Nitrobenzene-d5	8.82	82	4956	0.28	ng	-0.01
11) 2-Methylnaphthalene-d10	12.05	152	10477	0.31	ng	-0.02
14) 2,4,6-Tribromophenol	15.81	330	2070	0.30	ng	-0.01
15) 2-Fluorobiphenyl	12.93	172	14349	0.28	ng	-0.01
25) Fluoranthene-d10	19.09	212	29256	0.31	ng	-0.02
29) Terphenyl-d14	19.69	244	23323	0.31	ng	-0.02

Target Compounds

						Qvalue
9) Naphthalene	10.49	128	1802	0.032	ng	# 80
12) 2-Methylnaphthalene	12.12	142	1160	0.031	ng	# 87
16) Acenaphthylene	14.02	152	16166	0.242	ng	99
17) Acenaphthene	14.37	154	1373	0.032	ng	94
18) Fluorene	15.36	166	3999	0.072	ng	# 93
23) Phenanthrene	17.09	178	104155	1.202	ng	100
24) Anthracene	17.19	178	16190	0.207	ng	# 96
26) Fluoranthene	19.12	202	201643	1.872	ng	99
28) Pyrene	19.48	202	181467	1.728	ng	97
30) Benzo(a)anthracene	21.24	228	100612	0.888	ng	93
31) Chrysene	21.28	228	101400	0.918	ng	97
32) Bis(2-ethylhexyl)phthalate	21.17	149	17715	0.304	ng	97
33) Indeno(1,2,3-cd)pyrene	25.67	276	67618	0.489	ng	96
35) Benzo(b)fluoranthene	22.81	252	141750	1.227	ng	# 96
36) Benzo(k)fluoranthene	22.84	252	51922m	0.455	ng	
37) Benzo(a)pyrene	23.37	252	89610	0.825	ng	98
38) Dibenzo(a,h)anthracene	25.67	278	17706	0.159	ng	# 81
39) Benzo(a,h,i)perylene	26.34	276	68401	0.622	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN100319\
Data File : BN007989.D
Acq On : 03 Oct 2019 20:24
Operator : JU
Sample : K5077-08
Misc :
ALS Vial : 6 Sample Multiplier: 1


Instrument :
BNA_N
ClientSampleId :
PST-024-B278-092519

Quant Time: Oct 04 08:38:04 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN091819.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Oct 03 18:39:49 2019
Response via : Initial Calibration

Manual Integrations
APPROVED

mohammad
10/7/2019 8:27:43 AM





#1


1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.69 min Scan# 603
Delta R.T. -0.02 min
Lab File: BN007989.D
Acq: 03 Oct 2019 20:24

Instrument :
BNA_N
ClientSampleId :
PST-024-B278-092519

Tot Ion	Ratio	Lower	Upper
152	100		
150	149.1	125.6	188.4
115	63.2	53.3	79.9

Manual Integrations
APPROVED

mohammad
10/7/2019 8:27:43 AM






Abundance

Ion 152.00 (151.70 to 152.70): E

Ion 150.00 (149.70 to 150.70): E


Ion 115.00 (114.70 to 115.70): E



#4

2-Fluorophenol
Concen: 0.287 ng
RT: 5.30 min Scan# 307
Delta R.T. -0.02 min
Lab File: BN007989.D
Acq: 03 Oct 2019 20:24

Tgt Ion	Ratio	Lower	Upper
112	100		
64	57.7	46.6	70.0
63	30.0	23.1	34.7






Abundance

Ion 112.00 (111.70 to 112.70): E

Ion 64.00 (63.70 to 64.70): BNC

Ion 63.00 (62.70 to 63.70): BNC





#5

Phenol-d6

Concen: 0.320 ng

RT: 6.86 min Scan# 501

Delta R.T. -0.01 min

Lab File: BN007989.D

Acq: 03 Oct 2019 20:24

Instrument :

BNA_N


ClientSampleId :

PST-024-B278-092519

Tot Ion:	99	Resp:	6685
Ion	Ratio	Lower	Upper
99	100		
42	15.0	12.8	19.2
71	35.9	27.6	41.4

Manual Integrations
APPROVED

mohammad
10/7/2019 8:27:43 AM






Abundance Ion 99.00 (98.70 to 99.70): BN007706.D (-493) (-)

Ion 42.00 (41.70 to 42.70): BN007706.D (-493) (-)

Ion 71.00 (70.70 to 71.70): BN007706.D (-493) (-)

Time-->



#7

Naphthalene-d8

Concen: 0.400 ng


RT: 10.46 min Scan# 850

Delta R.T. -0.01 min

Lab File: BN007989.D

Acq: 03 Oct 2019 20:24

Tgt Ion:	136	Resp:	19978
Ion	Ratio	Lower	Upper
136	100		
137	11.4	9.2	13.8
54	5.1	5.8	8.8#
68	6.1	6.5	9.7#





Abundance Ion 136.00 (135.70 to 136.70): BN007706.D (-846) (-)


Ion 137.00 (136.70 to 137.70): BN007706.D (-846) (-)

Ion 54.00 (53.70 to 54.70): BN007706.D (-846) (-)

Ion 68.00 (67.70 to 68.70): BN007706.D (-846) (-)

Time-->





#8

Nitrobenzene-d5

Concen: 0.284 ng

RT: 8.82 min Scan# 721

Delta R.T. -0.01 min

Lab File: BN007989.D

Acq: 03 Oct 2019 20:24

Instrument :

BNA_N


ClientSampleId :

PST-024-B278-092519

Tot Ion:	82	Resp:	4956
Ion	Ratio	Lower	Upper
82	100		
128	37.3	24.5	36.7#
54	41.4	35.2	52.8

Manual Integrations
APPROVED

mohammad
10/7/2019 8:27:43 AM






Abundance

Ion 82.00 (81.70 to 82.70): BN0

Ion 128.00 (127.70 to 128.70): E

Ion 54.00 (53.70 to 54.70): BN0



#9

Naphthalene

Concen: 0.032 ng


RT: 10.49 min Scan# 853

Delta R.T. -0.03 min

Lab File: BN007989.D

Acq: 03 Oct 2019 20:24

Tgt Ion:	128	Resp:	1802
Ion	Ratio	Lower	Upper
128	100		
129	21.2	9.1	13.7#
127	19.5	10.6	15.8#






Abundance

Ion 128.00 (127.70 to 128.70): E

Ion 129.00 (128.70 to 129.70): E

Ion 127.00 (126.70 to 127.70): E









#11
2-Methylnaphthalene-d10
Concen: 0.311 ng
RT: 12.05 min Scan# 1054
Delta R.T. -0.02 min
Lab File: BN007989.D
Acq: 03 Oct 2019 20:24

Instrument :
BNA_N
ClientSampleId :
PST-024-B278-092519

Tot Ion	Ratio	Lower	Upper
152	100		
151	19.1	15.4	23.2




Manual Integrations
APPROVED


mohammad
10/7/2019 8:27:43 AM



#12
2-Methylnaphthalene
Concen: 0.031 ng
RT: 12.12 min Scan# 1070
Delta R.T. -0.02 min
Lab File: BN007989.D
Acq: 03 Oct 2019 20:24

Tgt Ion	Ratio	Lower	Upper
142	100		
141	93.8	70.6	106.0
115	52.7	28.3	42.5#









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.30 min Scan# 1319
Delta R.T. -0.02 min
Lab File: BN007989.D
Acq: 03 Oct 2019 20:24

Instrument :
BNA_N
ClientSampleId :
PST-024-B278-092519

Tot Ion	Ratio	Lower	Upper
164	100		
162	105.9	81.7	122.5
160	47.7	37.0	55.4




Manual Integrations
APPROVED


mohammad
10/7/2019 8:27:43 AM



#14
2,4,6-Tribromophenol
Concen: 0.301 ng
RT: 15.81 min Scan# 1454
Delta R.T. -0.01 min
Lab File: BN007989.D
Acq: 03 Oct 2019 20:24

Tgt Ion	Ratio	Lower	Upper
330	100		
332	0.0	0.0	0.0
141	35.8	28.7	43.1






#15
2-Fluorobiphenyl
Concen: 0.276 ng
RT: 12.93 min Scan# 1196
Delta R.T. -0.01 min
Lab File: BN007989.D
Acq: 03 Oct 2019 20:24

Instrument :
BNA_N
ClientSampleId :
PST-024-B278-092519

Tot Ion	Ratio	Lower	Upper
172	100		
171	34.7	28.8	43.2
170	22.2	19.5	29.3

Manual Integrations
APPROVED

mohammad
10/7/2019 8:27:43 AM






Abundance

Ion 172.00 (171.70 to 172.70): E

Ion 171.00 (170.70 to 171.70): E

Ion 170.00 (169.70 to 170.70): E



#16
Acenaphthylene
Concen: 0.242 ng
RT: 14.02 min Scan# 1294
Delta R.T. -0.02 min
Lab File: BN007989.D
Acq: 03 Oct 2019 20:24




Tgt Ion	Ratio	Lower	Upper
152	100		
151	19.7	15.5	23.3
153	13.1	10.4	15.6


Abundance

Ion 152.00 (151.70 to 152.70): E

Ion 151.00 (150.70 to 151.70): E

Ion 153.00 (152.70 to 153.70): E









#17
Acenaphthene
Concen: 0.032 ng
RT: 14.37 min Scan# 1325
Delta R.T. -0.02 min
Lab File: BN007989.D
Acq: 03 Oct 2019 20:24

Instrument :
BNA_N
ClientSampleId :
PST-024-B278-092519

Tot Ion	Ratio	Lower	Upper
154	100		
153	116.2	87.0	130.4
152	56.5	42.7	64.1




Manual Integrations
APPROVED


mohammad
10/7/2019 8:27:43 AM



#18
Fluorene
Concen: 0.072 ng
RT: 15.36 min Scan# 1414
Delta R.T. -0.01 min
Lab File: BN007989.D
Acq: 03 Oct 2019 20:24

Tgt Ion	Ratio	Lower	Upper
166	100		
165	104.0	78.0	117.0
167	16.9	10.9	16.3#









#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 17.05 min Scan# 1570
Delta R.T. -0.02 min
Lab File: BN007989.D
Acq: 03 Oct 2019 20:24

Instrument :
BNA_N
ClientSampleId :
PST-024-B278-092519

Tot Ion	Ratio	Resp	Lower	Upper
188	100	27814		
94	0.0	0.0	0.0	0.0
80	12.0	7.4	11.2	




Manual Integrations
APPROVED


mohammad
10/7/2019 8:27:43 AM



#23
Phenanthrene
Concen: 1.202 ng
RT: 17.09 min Scan# 1574
Delta R.T. -0.02 min
Lab File: BN007989.D
Acq: 03 Oct 2019 20:24

Tgt Ion	Ratio	Resp	Lower	Upper
178	100	104155		
176	18.6	14.9	22.3	
179	15.6	12.5	18.7	





#24

Anthracene

Concen: 0.207 ng

RT: 17.19 min Scan# 1583

Delta R.T. -0.01 min

Lab File: BN007989.D

Acq: 03 Oct 2019 20:24

Instrument :

BNA_N





ClientSampleId :

PST-024-B278-092519

Tot Ion:178	Resp:	16190
Ion Ratio	Lower	Upper
178	100	
176	18.2	14.6 21.8
179	19.1	12.2 18.4

Manual Integrations
APPROVED

mohammad
10/7/2019 8:27:43 AM



#25

Fluoranthene-d10

Concen: 0.312 ng




RT: 19.09 min Scan# 1878


Delta R.T. -0.02 min

Lab File: BN007989.D

Acq: 03 Oct 2019 20:24

Tgt Ion:212	Resp:	29256
Ion Ratio	Lower	Upper
212	100	
106	11.5	9.4 14.0
104	6.8	5.3 7.9









#26
Fluoranthene
 Concen: 1.872 ng
 RT: 19.12 min Scan# 1884
 Delta R.T. -0.02 min
 Lab File: BN007989.D
 Acq: 03 Oct 2019 20:24

Instrument :
 BNA_N
 ClientSampleId :
 PST-024-B278-092519

Tot Ion	Ratio	Lower	Upper
202	100		
101	10.2	8.4	12.6
203	17.6	13.8	20.8




Manual Integrations
 APPROVED


mohammad
 10/7/2019 8:27:43 AM



#27
Chrysene-d12
 Concen: 0.400 ng
 RT: 21.25 min Scan# 2180
 Delta R.T. -0.02 min
 Lab File: BN007989.D
 Acq: 03 Oct 2019 20:24

Tgt Ion	Ratio	Lower	Upper
240	100		
120	10.2	5.4	8.2#
236	27.6	21.0	31.4





#28

Pvrene

Concen: 1.728 ng

RT: 19.48 min Scan# 1957

Delta R.T. -0.02 min

Lab File: BN007989.D

Acq: 03 Oct 2019 20:24

Instrument :

BNA_N





ClientSampleId :

PST-024-B278-092519

Tot Ion:	202	Resp:	181467
Ion Ratio		Lower	Upper
202	100		
200	20.5	16.4	24.6
203	20.4	14.1	21.1

Manual Integrations
APPROVED

mohammad
10/7/2019 8:27:43 AM



#29

Terphenyl-d14

Concen: 0.308 ng




RT: 19.69 min Scan# 2000


Delta R.T. -0.02 min

Lab File: BN007989.D

Acq: 03 Oct 2019 20:24

Tgt Ion:	244	Resp:	23323
Ion Ratio		Lower	Upper
244	100		
212	7.8	5.9	8.9
122	10.1	7.4	11.2









#30
Benzo(a)anthracene
Concen: 0.888 ng
RT: 21.24 min Scan# 2179
Delta R.T. -0.01 min
Lab File: BN007989.D
Acq: 03 Oct 2019 20:24

Instrument :
BNA_N
ClientSampleId :
PST-024-B278-092519

Tot Ion	Ratio	Lower	Upper
228	100		
226	31.3	22.1	33.1
229	22.3	15.8	23.6




Manual Integrations
APPROVED


mohammad
10/7/2019 8:27:43 AM



#31
Chrysene
Concen: 0.918 ng
RT: 21.28 min Scan# 2183
Delta R.T. -0.02 min
Lab File: BN007989.D
Acq: 03 Oct 2019 20:24

Tgt Ion	Ratio	Lower	Upper
228	100		
226	30.4	23.9	35.9
229	21.9	15.8	23.6





#32

Bis(2-ethylhexyl)phthalate

Concen: 0.304 ng

RT: 21.17 min Scan# 2173

Delta R.T. -0.03 min

Lab File: BN007989.D

Acq: 03 Oct 2019 20:24

Instrument :

BNA_N


ClientSampleId :

PST-024-B278-092519




Tot Ion:	149	Resp:	17715
Ion Ratio	Lower	Upper	
149	100		
167	27.4	23.7	35.5
279	5.2	4.2	6.2

Manual Integrations
APPROVED

mohammad
10/7/2019 8:27:43 AM



Abundance Ion 149.00 (148.70 to 149.70): E
Ion 167.00 (166.70 to 167.70): E
Ion 279.00 (278.70 to 279.70): E



#33

Indeno(1,2,3-cd)pyrene

Concen: 0.489 ng

RT: 25.67 min Scan# 3241




Delta R.T. -0.04 min


Lab File: BN007989.D

Acq: 03 Oct 2019 20:24

Tgt Ion:	276	Resp:	67618
Ion Ratio	Lower	Upper	
276	100		
138	17.3	17.1	25.7
277	24.6	20.0	30.0

Abundance Ion 276.00 (275.70 to 276.70): E
Ion 138.00 (137.70 to 138.70): E
Ion 277.00 (276.70 to 277.70): E









#34
Pervlene-d12
Concen: 0.400 ng
RT: 23.46 min Scan# 2598
Delta R.T. -0.03 min
Lab File: BN007989.D
Acq: 03 Oct 2019 20:24

Instrument :
BNA_N
ClientSampleId :
PST-024-B278-092519

Tot Ion	Ratio	Lower	Upper
264	100		
260	25.4	19.3	28.9
265	40.2	26.9	40.3




Manual Integrations
APPROVED


mohammad
10/7/2019 8:27:43 AM



#35
Benzo(b)fluoranthene
Concen: 1.227 ng
RT: 22.81 min Scan# 2406
Delta R.T. -0.02 min
Lab File: BN007989.D
Acq: 03 Oct 2019 20:24

Tgt Ion	Ratio	Lower	Upper
252	100		
253	24.2	18.3	27.5
125	12.6	8.2	12.4





#36

Benzo(k)fluoranthene

Concen: 0.455 ng/mL

RT: 22.84 min Scan# 2417

Delta R.T. -0.03 min

Lab File: BN007989.D

Acq: 03 Oct 2019 20:24

Instrument :

BNA_N


ClientSampleId :

PST-024-B278-092519

Tot Ion:	252	Resp:	51922
Ion Ratio	Lower	Upper	
252	100		
253	25.4	18.5	27.7
125	18.5	8.4	12.6

Manual Integrations
APPROVED

mohammad
10/7/2019 8:27:43 AM






Abundance

Ion 252.00 (251.70 to 252.70): E

Ion 253.00 (252.70 to 253.70): E

Ion 125.00 (124.70 to 125.70): E



#37

Benzo(a)pyrene

Concen: 0.825 ng/mL

RT: 23.37 min Scan# 2569

Delta R.T. -0.03 min

Lab File: BN007989.D

Acq: 03 Oct 2019 20:24




Tgt Ion:	252	Resp:	89610
Ion Ratio	Lower	Upper	
252	100		
253	23.8	18.6	27.8
125	13.3	9.4	14.0


Abundance

Ion 252.00 (251.70 to 252.70): E

Ion 253.00 (252.70 to 253.70): E

Ion 125.00 (124.70 to 125.70): E









#38
Dibenzo(a,h)anthracene
Concen: 0.159 ng
RT: 25.67 min Scan# 3243
Delta R.T. -0.05 min
Lab File: BN007989.D
Acq: 03 Oct 2019 20:24

Instrument :
BNA_N
ClientSampleId :
PST-024-B278-092519

Tot Ion	Ratio	Lower	Upper
278	100		
139	24.9	11.8	17.6#
279	32.4	19.4	29.2#

Manual Integrations
APPROVED

mohammad
10/7/2019 8:27:43 AM



#39
Benzo(g,h,i)perylene
Concen: 0.622 ng
RT: 26.34 min Scan# 3438
Delta R.T. -0.04 min
Lab File: BN007989.D
Acq: 03 Oct 2019 20:24

Tgt Ion	Ratio	Lower	Upper
276	100		
277	25.6	19.8	29.8
138	19.7	15.1	22.7

