




Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN100418\
Data File : BN002919.D
Acq On : 04 Oct 2018 02:04
Operator : MJ/SJ
Sample : J5115-01
Misc :
ALS Vial : 22 Sample Multiplier: 1

Quant Time: Oct 04 04:30:23 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-BN091418MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Oct 04 04:26:52 2018
Response via : Initial Calibration



TIC: BN002919.D

(88) Benzo(k)fluoranthene

22.810min (+0.000) 1.52ng/ul m

response 47355

Ion	Exp%	Act%
252.00	100	100
253.00	22.30	25.51
125.00	10.40	28.44#
0.00	0.00	0.00