

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN100421\
 Data File : BN016730.D
 Acq On : 05 Oct 2021 04:38
 Operator : CG/JU
 Sample : M3833-04DL 2X
 Misc :
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 MW201D-20210919DL

Manual Integrations
 APPROVED

mohammad
 10/5/2021 12:08:37 PM

Quant Time: Oct 05 05:34:32 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN100421.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Oct 04 17:52:33 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.643	152	28362	0.40	ng	0.00
7) Naphthalene-d8	10.407	136	130413	0.40	ng	# 0.00
13) Acenaphthene-d10	14.269	164	68753	0.40	ng	0.00
19) Phenanthrene-d10	17.030	188	134173	0.40	ng	0.00
29) Chrysene-d12	21.238	240	131904	0.40	ng	# 0.00
35) Perylene-d12	23.488	264	133387	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.264	112	4487	0.06	ng	0.00
5) Phenol-d6	6.822	99	2939	0.04	ng	0.00
8) Nitrobenzene-d5	8.785	82	10079	0.11	ng	0.00
11) 2-Methylnaphthalene-d10	12.003	152	26874	0.14	ng	0.00
14) 2,4,6-Tribromophenol	15.779	330	4011	0.22	ng	0.00
15) 2-Fluorobiphenyl	12.899	172	37379	0.14	ng	0.00
27) Fluoranthene-d10	19.068	212	65110	0.18	ng	0.00
31) Terphenyl-d14	19.679	244	54938	0.19	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.253	88	36223m	2.74	ng	Qvalue
34) Bis(2-ethylhexyl)phtha...	21.164	149	16102	0.06	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN100421\
Data File : BN016730.D
Acq On : 05 Oct 2021 04:38
Operator : CG/JU
Sample : M3833-04DL 2X
Misc :
ALS Vial : 23 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
MW201D-20210919DL

Manual Integrations
APPROVED

mohammad
10/5/2021 12:08:37 PM

Quant Time: Oct 05 05:34:32 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN100421.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Oct 04 17:52:33 2021
Response via : Initial Calibration

