

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN100421\
 Data File : BN016749.D
 Acq On : 05 Oct 2021 18:12
 Operator : CG/JU
 Sample : M4040-07
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SW-RR-50

Quant Time: Oct 06 00:58:44 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN100421.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Oct 04 17:52:33 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.642	152	24220	0.400	ng	0.00
7) Naphthalene-d8	10.407	136	110334	0.400	ng	0.00
13) Acenaphthene-d10	14.269	164	59693	0.400	ng	0.00
19) Phenanthrene-d10	17.017	188	114589	0.400	ng	#-0.01
29) Chrysene-d12	21.238	240	114349	0.400	ng	0.00
35) Perylene-d12	23.488	264	120939	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.273	112	7376	0.119	ng	0.02
5) Phenol-d6	6.831	99	4925	0.072	ng	0.00
8) Nitrobenzene-d5	8.784	82	20187	0.261	ng	0.00
11) 2-Methylnaphthalene-d10	12.003	152	50212	0.306	ng	0.00
14) 2,4,6-Tribromophenol	15.766	330	9828	0.403	ng	-0.01
15) 2-Fluorobiphenyl	12.898	172	63960	0.268	ng	0.00
27) Fluoranthene-d10	19.068	212	121427	0.387	ng	0.00
31) Terphenyl-d14	19.678	244	93778	0.367	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.253	88	261	0.023	ng	# 63
9) Naphthalene	10.457	128	143890	0.477	ng	100
24) Pentachlorophenol	16.677	266	176	0.343	ng	# 85
34) Bis(2-ethylhexyl)phtha...	21.174	149	42499	0.198	ng	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN100421\
Data File : BN016749.D
Acq On : 05 Oct 2021 18:12
Operator : CG/JU
Sample : M4040-07
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SW-RR-50

Quant Time: Oct 06 00:58:44 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN100421.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Oct 04 17:52:33 2021
Response via : Initial Calibration

