

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN100421\
 Data File : BN016754.D
 Acq On : 05 Oct 2021 21:13
 Operator : CG/JU
 Sample : M3915-10
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 RE104D1-20210921

Manual Integrations
 APPROVED

mohammad
 10/7/2021 8:31:49 AM

Quant Time: Oct 06 00:59:22 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN100421.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Oct 04 17:52:33 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.642	152	26548	0.400	ng	0.00
7) Naphthalene-d8	10.406	136	120648	0.400	ng	0.00
13) Acenaphthene-d10	14.268	164	64206	0.400	ng	0.00
19) Phenanthrene-d10	17.029	188	123385	0.400	ng	0.00
29) Chrysene-d12	21.238	240	125212	0.400	ng	# 0.00
35) Perylene-d12	23.488	264	132639	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.272	112	7488	0.111	ng	0.02
5) Phenol-d6	6.831	99	4673	0.062	ng	0.00
8) Nitrobenzene-d5	8.784	82	17255	0.204	ng	0.00
11) 2-Methylnaphthalene-d10	12.003	152	45419	0.253	ng	0.00
14) 2,4,6-Tribromophenol	15.778	330	6617	0.296	ng	0.00
15) 2-Fluorobiphenyl	12.898	172	58045	0.226	ng	0.00
27) Fluoranthene-d10	19.068	212	112503	0.333	ng	0.00
31) Terphenyl-d14	19.678	244	96185	0.343	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.253	88	41561m	3.354	ng	
34) Bis(2-ethylhexyl)phtha...	21.163	149	38743	0.165	ng	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN100421\
Data File : BN016754.D
Acq On : 05 Oct 2021 21:13
Operator : CG/JU
Sample : M3915-10
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
RE104D1-20210921

Manual Integrations
APPROVED

mohammad
10/7/2021 8:31:49 AM

Quant Time: Oct 06 00:59:22 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN100421.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Oct 04 17:52:33 2021
Response via : Initial Calibration

