

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN100421\
 Data File : BN016740.D
 Acq On : 05 Oct 2021 10:36
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Oct 05 11:05:46 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN100421.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Oct 04 17:52:33 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	92	0.00
2	1,4-Dioxane	0.187	0.172	8.0	92	0.00
3	n-Nitrosodimethylamine	0.301	0.285	5.3	94	0.02
4 S	2-Fluorophenol	1.021	0.989	3.1	94	0.00
5 S	Phenol-d6	1.132	1.090	3.7	95	0.00
6	bis(2-Chloroethyl)ether	1.087	1.073	1.3	93	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	92	0.00
8 S	Nitrobenzene-d5	0.280	0.256	8.6	98	0.00
9	Naphthalene	1.094	1.021	6.7	91	0.00
10	Hexachlorobutadiene	0.205	0.196	4.4	92	0.00
11 SURR	2-Methylnaphthalene-d10	0.596	0.553	7.2	90	0.00
12	2-Methylnaphthalene	0.695	0.664	4.5	91	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	89	0.00
14 S	2,4,6-Tribromophenol	0.177	0.172	2.8	109	-0.01
15 S	2-Fluorobiphenyl	1.600	1.528	4.5	91	-0.01
16	Acenaphthylene	1.766	1.632	7.6	91	0.00
17	Acenaphthene	1.167	1.095	6.2	89	0.00
18	Fluorene	1.407	1.329	5.5	90	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	91	0.00
20	4,6-Dinitro-2-methylphenol	0.028	0.017	39.3#	376#	-0.01
21	4-Bromophenyl-phenylether	0.244	0.226	7.4	93	0.00
22	Hexachlorobenzene	0.276	0.265	4.0	93	0.00
23	Atrazine	0.193	0.165	14.5	95	0.00
24	Pentachlorophenol	0.076	0.083	-9.2	234#	-0.01
25	Phenanthrene	1.176	1.118	4.9	92	0.00
26	Anthracene	1.050	0.958	8.8	93	-0.01
27 SURR	Fluoranthene-d10	1.094	1.034	5.5	93	0.00
28	Fluoranthene	1.305	1.285	1.5	95	0.00
29 I	Chrysene-d12	1.000	1.000	0.0	102	0.00
30	Pyrene	1.315	1.265	3.8	96	0.00
31 S	Terphenyl-d14	0.895	0.860	3.9	96	0.00
32	Benzo(a)anthracene	1.241	1.234	0.6	102	0.00
33	Chrysene	1.259	1.294	-2.8	102	0.00
34	Bis(2-ethylhexyl)phthalate	0.752	0.592	21.3	92	0.00
35 I	Perylene-d12	1.000	1.000	0.0	114	0.00
36	Indeno(1,2,3-cd)pyrene	1.270	1.463	-15.2	139	-0.01
37	Benzo(b)fluoranthene	1.293	1.255	2.9	107	0.00
38	Benzo(k)fluoranthene	1.243	1.259	-1.3	114	0.00
39 C	Benzo(a)pyrene	1.157	1.155	0.2	113	0.00
40	Dibenzo(a,h)anthracene	1.008	1.177	-16.8	143	0.00
41	Benzo(g,h,i)perylene	1.127	1.260	-11.8	132	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

LabSampleId :

SSTDCCC0.4