

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN100421\
 Data File : BN016771.D
 Acq On : 06 Oct 2021 09:18
 Operator : CG/JU
 Sample : M3892-16MS 10X
 Misc :
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 S-832-J-SO-1-1.5-092221MS

Manual Integrations
 APPROVED

mohammad
 10/7/2021 8:32:11 AM

Quant Time: Oct 06 10:33:07 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN100421.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Oct 04 17:52:33 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.642	152	26236	0.40	ng	0.00
7) Naphthalene-d8	10.407	136	121396	0.40	ng	# 0.00
13) Acenaphthene-d10	14.269	164	70228	0.40	ng	0.00
19) Phenanthrene-d10	17.030	188	138845	0.40	ng	0.00
29) Chrysene-d12	21.270	240	95084	0.40	ng	# 0.03
35) Perylene-d12	23.512	264	141900	0.40	ng	# 0.02
System Monitoring Compounds						
4) 2-Fluorophenol	5.273	112	1875	0.03	ng	0.02
5) Phenol-d6	6.822	99	1941	0.03	ng	0.00
8) Nitrobenzene-d5	8.785	82	1954m	0.02	ng	0.00
11) 2-Methylnaphthalene-d10	12.003	152	6457m	0.04	ng	0.00
14) 2,4,6-Tribromophenol	15.766	330	1060	0.14	ng	-0.01
15) 2-Fluorobiphenyl	12.886	172	6119	0.02	ng	-0.01
27) Fluoranthene-d10	19.078	212	9524m	0.03	ng	0.01
31) Terphenyl-d14	19.718	244	8478m	0.04	ng	0.04
Target Compounds						Qvalue
3) n-Nitrosodimethylamine	3.576	42	730	0.04	ng	# 89
6) bis(2-Chloroethyl)ether	7.080	93	2632	0.04	ng	97
9) Naphthalene	10.457	128	46537	0.14	ng	99
10) Hexachlorobutadiene	10.749	225	2021	0.03	ng	# 99
12) 2-Methylnaphthalene	12.074	142	80032	0.38	ng	100
16) Acenaphthylene	13.990	152	35838	0.12	ng	# 91
17) Acenaphthene	14.332	154	805434	3.93	ng	99
18) Fluorene	15.322	166	984890	3.99	ng	99
20) 4,6-Dinitro-2-methylph...	15.474	198	101	0.27	ng	# 1
21) 4-Bromophenyl-phenylether	16.227	248	2889	0.03	ng	# 85
22) Hexachlorobenzene	16.336	284	3163	0.03	ng	# 89
23) Atrazine	16.506	200	2418	0.04	ng	# 65
24) Pentachlorophenol	16.689	266	2724	0.40	ng	96
25) Phenanthrene	17.078	178	11569748	28.35	ng	97
26) Anthracene	17.164	178	5244474	14.39	ng	96
28) Fluoranthene	19.118	202	23434704m	51.74	ng	
30) Pyrene	19.490	202	21915916	70.10	ng	94
32) Benzo(a)anthracene	21.249	228	18571285	62.97	ng	# 93
33) Chrysene	21.302	228	15310465	51.18	ng	# 91
34) Bis(2-ethylhexyl)phtha...	21.174	149	62511	0.35	ng	# 78
36) Indeno(1,2,3-cd)pyrene	25.805	276	4262788	9.46	ng	97
37) Benzo(b)fluoranthene	22.862	252	18179437	39.62	ng	96
38) Benzo(k)fluoranthene	22.896	252	6830512m	15.49	ng	
39) Benzo(a)pyrene	23.437	252	11769043	28.67	ng	96
40) Dibenzo(a,h)anthracene	25.816	278	1439983	4.03	ng	96
41) Benzo(g,h,i)perylene	26.507	276	3200571	8.00	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN100421\
Data File : BN016771.D
Acq On : 06 Oct 2021 09:18
Operator : CG/JU
Sample : M3892-16MS 10X
Misc :
ALS Vial : 29 Sample Multiplier: 1


Instrument :
BNA_N
ClientSampleId :
S-832-J-SO-1-1.5-092221MS

Manual Integrations
APPROVED

mohammad
10/7/2021 8:32:11 AM

Quant Time: Oct 06 10:33:07 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN100421.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Oct 04 17:52:33 2021
Response via : Initial Calibration





#1
1,4-Dichlorobenzene-d4
Concen: 0.40 ng
RT: 7.642 min Scan# 495
Delta R.T. 0.000 min
Lab File: BN016771.D
Acq: 06 Oct 2021 09:18





Tgt Ion: 152 Resp: 26236

Ion	Ratio	Lower	Upper
152	100		
150	153.1	123.3	184.9
115	58.2	47.0	70.6

Instrument :
BNA_N
ClientSampleId :
S-832-J-SO-1-1.5-092221MS

Manual Integrations
APPROVED




mohammad
10/7/2021 8:32:11 AM




#3
n-Nitrosodimethylamine
Concen: 0.04 ng
RT: 3.576 min Scan# 54
Delta R.T. 0.000 min
Lab File: BN016771.D
Acq: 06 Oct 2021 09:18

Tgt Ion: 42 Resp: 730

Ion	Ratio	Lower	Upper
42	100		
74	141.9	124.0	186.0
44	0.0	5.6	8.4









#4
2-Fluorophenol
Concen: 0.03 ng
RT: 5.273 min Scan# 238
Delta R.T. 0.019 min
Lab File: BN016771.D
Acq: 06 Oct 2021 09:18

Tgt Ion: 112 Resp: 1875
Ion Ratio Lower Upper
112 100
64 56.9 42.9 64.3
63 27.6 21.5 32.3

Instrument :
BNA_N
ClientSampleId :
S-832-J-SO-1-1.5-092221MS




Manual Integrations
APPROVED


mohammad
10/7/2021 8:32:11 AM



#5
Phenol-d6
Concen: 0.03 ng
RT: 6.822 min Scan# 406
Delta R.T. 0.000 min
Lab File: BN016771.D
Acq: 06 Oct 2021 09:18

Tgt Ion: 99 Resp: 1941
Ion Ratio Lower Upper
99 100
42 19.5 13.4 20.2
71 38.0 24.0 36.0#









#6
bis(2-Chloroethyl)ether
Concen: 0.04 ng
RT: 7.080 min Scan# 434
Delta R.T. 0.000 min
Lab File: BN016771.D
Acq: 06 Oct 2021 09:18

Tgt Ion:	93	Resp:	2632
Ion	Ratio	Lower	Upper
93	100		
63	70.6	59.0	88.6
95	32.4	26.4	39.6

Instrument :
BNA_N
ClientSampleId :
S-832-J-SO-1-1.5-092221MS




Manual Integrations
APPROVED


mohammad
10/7/2021 8:32:11 AM



#7
Naphthalene-d8
Concen: 0.40 ng
RT: 10.407 min Scan# 733
Delta R.T. 0.000 min
Lab File: BN016771.D
Acq: 06 Oct 2021 09:18


Tgt Ion:	136	Resp:	121396
Ion	Ratio	Lower	Upper
136	100		
137	11.2	8.9	13.3
54	6.0	6.5	9.7#
68	5.2	5.3	7.9#





#8
Nitrobenzene-d5
Concen: 0.02 ng m
RT: 8.785 min Scan# 605
Delta R.T. 0.000 min
Lab File: BN016771.D
Acq: 06 Oct 2021 09:18




Instrument :
BNA_N
ClientSampleId :
S-832-J-SO-1-1.5-092221MS



Tgt Ion: 82 Resp: 1954
Ion Ratio Lower Upper
82 100
128 50.6 33.8 50.6#
54 52.4 38.5 57.7




Manual Integrations
APPROVED


mohammad
10/7/2021 8:32:11 AM



#9
Naphthalene
Concen: 0.14 ng
RT: 10.457 min Scan# 737
Delta R.T. 0.000 min
Lab File: BN016771.D
Acq: 06 Oct 2021 09:18


Tgt Ion: 128 Resp: 46537
Ion Ratio Lower Upper
128 100
129 11.4 8.7 13.1
127 13.2 10.3 15.5





#10
Hexachlorobutadiene
Concen: 0.03 ng
RT: 10.749 min Scan# 760
Delta R.T. 0.000 min
Lab File: BN016771.D
Acq: 06 Oct 2021 09:18




Instrument :
BNA_N
ClientSampleId :
S-832-J-SO-1-1.5-092221MS



Tgt Ion:225 Resp: 2021
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 64.8 51.1 76.7




Manual Integrations
APPROVED


mohammad
10/7/2021 8:32:11 AM



#11
2-Methylnaphthalene-d10
Concen: 0.04 ng m
RT: 12.003 min Scan# 931
Delta R.T. -0.004 min
Lab File: BN016771.D
Acq: 06 Oct 2021 09:18

Tgt Ion:152 Resp: 6457
Ion Ratio Lower Upper
152 100
151 16.8 16.2 24.4









#12
2-Methylnaphthalene
Concen: 0.38 ng
RT: 12.074 min Scan# 947
Delta R.T. -0.008 min
Lab File: BN016771.D
Acq: 06 Oct 2021 09:18

Tgt Ion:142 Resp: 80032
Ion Ratio Lower Upper
142 100
141 88.2 70.6 106.0
115 33.3 26.1 39.1

Instrument :
BNA_N
ClientSampleId :
S-832-J-SO-1-1.5-092221MS




Manual Integrations
APPROVED


mohammad
10/7/2021 8:32:11 AM



#13
Acenaphthene-d10
Concen: 0.40 ng
RT: 14.269 min Scan# 1183
Delta R.T. 0.000 min
Lab File: BN016771.D
Acq: 06 Oct 2021 09:18

Tgt Ion:164 Resp: 70228
Ion Ratio Lower Upper
164 100
162 101.6 82.6 124.0
160 45.4 38.2 57.2









#14
2,4,6-Tribromophenol
Concen: 0.14 ng
RT: 15.766 min Scan# 1301
Delta R.T. -0.012 min
Lab File: BN016771.D
Acq: 06 Oct 2021 09:18

Tgt Ion: 330 Resp: 1060
Ion Ratio Lower Upper
330 100
332 0.0 0.0 0.0
141 79.5 24.2 36.2#

Instrument :
BNA_N
ClientSampleId :
S-832-J-SO-1-1.5-092221MS




Manual Integrations
APPROVED


mohammad
10/7/2021 8:32:11 AM



#15
2-Fluorobiphenyl
Concen: 0.02 ng
RT: 12.886 min Scan# 1074
Delta R.T. -0.012 min
Lab File: BN016771.D
Acq: 06 Oct 2021 09:18

Tgt Ion: 172 Resp: 6119
Ion Ratio Lower Upper
172 100
171 35.9 26.9 40.3
170 25.0 17.4 26.0









#16
Acenaphthylene
Concen: 0.12 ng
RT: 13.990 min Scan# 1161
Delta R.T. 0.001 min
Lab File: BN016771.D
Acq: 06 Oct 2021 09:18

Tgt Ion:152 Resp: 35838
Ion Ratio Lower Upper
152 100
151 20.9 15.5 23.3
153 19.6 10.4 15.6#

Instrument :
BNA_N
ClientSampleId :
S-832-J-SO-1-1.5-092221MS




Manual Integrations
APPROVED


mohammad
10/7/2021 8:32:11 AM



#17
Acenaphthene
Concen: 3.93 ng
RT: 14.332 min Scan# 1188
Delta R.T. 0.000 min
Lab File: BN016771.D
Acq: 06 Oct 2021 09:18


Tgt Ion:154 Resp: 805434
Ion Ratio Lower Upper
154 100
153 111.7 90.1 135.1
152 53.1 43.6 65.4





#18
Fluorene
Concen: 3.99 ng
RT: 15.322 min Scan# 1266
Delta R.T. 0.001 min
Lab File: BN016771.D
Acq: 06 Oct 2021 09:18




Instrument :
BNA_N
ClientSampleId :
S-832-J-SO-1-1.5-092221MS




Tgt Ion:166 Resp: 984890
Ion Ratio Lower Upper
166 100
165 96.6 78.1 117.1
167 14.6 10.7 16.1

Manual Integrations
APPROVED



mohammad
10/7/2021 8:32:11 AM




#19
Phenanthrene-d10
Concen: 0.40 ng
RT: 17.030 min Scan# 1404
Delta R.T. 0.000 min
Lab File: BN016771.D
Acq: 06 Oct 2021 09:18




Tgt Ion:188 Resp: 138845
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 8.0 5.8 8.8









#20
4,6-Dinitro-2-methylphenol
Concen: 0.27 ng
RT: 15.474 min Scan# 1278
Delta R.T. 0.051 min
Lab File: BN016771.D
Acq: 06 Oct 2021 09:18





Tgt Ion:198 Resp: 101
Ion Ratio Lower Upper
198 100
182 1049.5 26.2 39.2#
77 0.0 0.0 0.0



#21
4-Bromophenyl-phenylether
Concen: 0.03 ng
RT: 16.227 min Scan# 1338
Delta R.T. 0.001 min
Lab File: BN016771.D
Acq: 06 Oct 2021 09:18




Tgt Ion:248 Resp: 2889
Ion Ratio Lower Upper
248 100
250 101.5 76.2 114.4
141 78.4 45.6 68.4#



Instrument :
BNA_N
ClientSampleId :
S-832-J-SO-1-1.5-092221MS


Manual Integrations
APPROVED

mohammad
10/7/2021 8:32:11 AM



#22
Hexachlorobenzene
Concen: 0.03 ng
RT: 16.336 min Scan# 1347
Delta R.T. 0.000 min
Lab File: BN016771.D
Acq: 06 Oct 2021 09:18




Instrument :
BNA_N
ClientSampleId :
S-832-J-SO-1-1.5-092221MS




Tgt Ion:284 Resp: 3163
Ion Ratio Lower Upper
284 100
142 42.1 25.8 38.6#
249 30.6 23.3 34.9

Manual Integrations
APPROVED



mohammad
10/7/2021 8:32:11 AM




#23
Atrazine
Concen: 0.04 ng
RT: 16.506 min Scan# 1361
Delta R.T. 0.000 min
Lab File: BN016771.D
Acq: 06 Oct 2021 09:18




Tgt Ion:200 Resp: 2418
Ion Ratio Lower Upper
200 100
173 66.8 20.1 30.1#
215 55.3 38.8 58.2









#24
Pentachlorophenol
Concen: 0.40 ng
RT: 16.689 min Scan# 1376
Delta R.T. 0.000 min
Lab File: BN016771.D
Acq: 06 Oct 2021 09:18





Tgt Ion:266 Resp: 2724
Ion Ratio Lower Upper
266 100
264 65.5 49.3 73.9
268 62.0 51.8 77.8



#25
Phenanthrene
Concen: 28.35 ng
RT: 17.078 min Scan# 1408
Delta R.T. 0.013 min
Lab File: BN016771.D
Acq: 06 Oct 2021 09:18




Tgt Ion:178 Resp:11569748
Ion Ratio Lower Upper
178 100
176 19.7 15.0 22.6
179 17.3 12.4 18.6



Instrument :
BNA_N
ClientSampleId :
S-832-J-SO-1-1.5-092221MS


Manual Integrations
APPROVED

mohammad
10/7/2021 8:32:11 AM



#26
 Anthracene
 Concen: 14.39 ng
 RT: 17.164 min Scan# 1415
 Delta R.T. 0.000 min
 Lab File: BN016771.D
 Acq: 06 Oct 2021 09:18

Instrument :
 BNA_N
 ClientSampleId :
 S-832-J-SO-1-1.5-092221MS






Tgt Ion:178 Resp: 5244474


Ion	Ratio	Lower	Upper
178	100		
176	18.5	14.5	21.7
179	18.3	12.2	18.4

Manual Integrations
 APPROVED

mohammad
 10/7/2021 8:32:11 AM






#27
 Fluoranthene-d10
 Concen: 0.03 ng m
 RT: 19.078 min Scan# 1699
 Delta R.T. 0.010 min
 Lab File: BN016771.D
 Acq: 06 Oct 2021 09:18




Tgt Ion:212 Resp: 9524

Ion	Ratio	Lower	Upper
212	100		
106	14.5	10.6	16.0
104	0.0	5.8	8.8









#28
Fluoranthene
Concen: 51.74 ng m
RT: 19.118 min Scan# 1707
Delta R.T. 0.020 min
Lab File: BN016771.D
Acq: 06 Oct 2021 09:18





Tgt Ion:202 Resp:23434704
Ion Ratio Lower Upper
202 100
101 0.0 8.9 13.3#
203 0.0 13.8 20.8#



#29
Chrysene-d12
Concen: 0.40 ng
RT: 21.270 min Scan# 2005
Delta R.T. 0.032 min
Lab File: BN016771.D
Acq: 06 Oct 2021 09:18




Tgt Ion:240 Resp: 95084
Ion Ratio Lower Upper
240 100
120 20.7 6.6 10.0#
236 37.9 20.7 31.1#



Instrument :
BNA_N
ClientSampleId :
S-832-J-SO-1-1.5-092221MS

Manual Integrations
APPROVED

mohammad
10/7/2021 8:32:11 AM







#30
Pyrene
Concen: 70.10 ng
RT: 19.490 min Scan# 1782
Delta R.T. 0.025 min
Lab File: BN016771.D
Acq: 06 Oct 2021 09:18

Tgt Ion:202 Resp:21915916
Ion Ratio Lower Upper
202 100
200 24.0 16.5 24.7
203 19.3 13.9 20.9

Instrument :
BNA_N
ClientSampleId :
S-832-J-SO-1-1.5-092221MS




Manual Integrations
APPROVED


mohammad
10/7/2021 8:32:11 AM



#31
Terphenyl-d14
Concen: 0.04 ng m
RT: 19.718 min Scan# 1828
Delta R.T. 0.035 min
Lab File: BN016771.D
Acq: 06 Oct 2021 09:18


Tgt Ion:244 Resp: 8478
Ion Ratio Lower Upper
244 100
212 43.4 5.8 8.6#
122 140.7 8.9 13.3#





#32
Benzo(a)anthracene
Concen: 62.97 ng
RT: 21.249 min Scan# 2003
Delta R.T. 0.032 min
Lab File: BN016771.D
Acq: 06 Oct 2021 09:18




Instrument :
BNA_N
ClientSampleId :
S-832-J-SO-1-1.5-092221MS




Tgt Ion:228 Resp:18571285
Ion Ratio Lower Upper
228 100
226 29.6 22.3 33.5
229 24.8 15.4 23.0#

Manual Integrations
APPROVED



mohammad
10/7/2021 8:32:11 AM




#33
Chrysene
Concen: 51.18 ng
RT: 21.302 min Scan# 2008
Delta R.T. 0.032 min
Lab File: BN016771.D
Acq: 06 Oct 2021 09:18




Tgt Ion:228 Resp:15310465
Ion Ratio Lower Upper
228 100
226 32.9 24.2 36.4
229 26.9 15.4 23.0#





#34
Bis(2-ethylhexyl)phthalate
Concen: 0.35 ng
RT: 21.174 min Scan# 1996
Delta R.T. 0.000 min
Lab File: BN016771.D
Acq: 06 Oct 2021 09:18




Instrument :
BNA_N
ClientSampleId :
S-832-J-SO-1-1.5-092221MS




Tgt Ion:149 Resp: 62511
Ion Ratio Lower Upper
149 100
167 40.4 22.2 33.4#
279 8.3 3.4 5.2#

Manual Integrations
APPROVED



mohammad
10/7/2021 8:32:11 AM




#35
Perylene-d12
Concen: 0.40 ng
RT: 23.512 min Scan# 2435
Delta R.T. 0.018 min
Lab File: BN016771.D
Acq: 06 Oct 2021 09:18



Tgt Ion:264 Resp: 141900
Ion Ratio Lower Upper
264 100
260 27.6 19.8 29.6
265 103.1 20.3 30.5#









#36
 Indeno(1,2,3-cd)pyrene
 Concen: 9.46 ng
 RT: 25.805 min Scan# 3105
 Delta R.T. 0.024 min
 Lab File: BN016771.D
 Acq: 06 Oct 2021 09:18

Tgt Ion:276 Resp: 4262788
 Ion Ratio Lower Upper
 276 100
 138 21.3 19.2 28.8
 277 25.3 20.0 30.0

Instrument :
 BNA_N
 ClientSampleId :
 S-832-J-SO-1-1.5-092221MS




Manual Integrations
 APPROVED


mohammad
 10/7/2021 8:32:11 AM



#37
 Benzo(b)fluoranthene
 Concen: 39.62 ng
 RT: 22.862 min Scan# 2245
 Delta R.T. 0.048 min
 Lab File: BN016771.D
 Acq: 06 Oct 2021 09:18


Tgt Ion:252 Resp:18179437
 Ion Ratio Lower Upper
 252 100
 253 25.0 17.8 26.8
 125 10.8 8.3 12.5





#38
Benzo(k)fluoranthene
Concen: 15.49 ng m
RT: 22.896 min Scan# 2255
Delta R.T. 0.035 min
Lab File: BN016771.D
Acq: 06 Oct 2021 09:18




Instrument :
BNA_N
ClientSampleId :
S-832-J-SO-1-1.5-092221MS




Tgt Ion:252 Resp: 6830512
Ion Ratio Lower Upper
252 100
253 25.6 17.8 26.8
125 10.3 8.5 12.7

Manual Integrations
APPROVED



mohammad
10/7/2021 8:32:11 AM




#39
Benzo(a)pyrene
Concen: 28.67 ng
RT: 23.437 min Scan# 2413
Delta R.T. 0.045 min
Lab File: BN016771.D
Acq: 06 Oct 2021 09:18




Tgt Ion:252 Resp:11769043
Ion Ratio Lower Upper
252 100
253 24.9 17.9 26.9
125 11.0 9.4 14.2









#40
Dibenzo(a,h)anthracene
Concen: 4.03 ng
RT: 25.816 min Scan# 3108
Delta R.T. 0.018 min
Lab File: BN016771.D
Acq: 06 Oct 2021 09:18





Tgt Ion:278 Resp: 1439983
Ion Ratio Lower Upper
278 100
139 17.4 13.1 19.7
279 26.6 19.3 28.9



#41
Benzo(g,h,i)perylene
Concen: 8.00 ng
RT: 26.507 min Scan# 3310
Delta R.T. 0.031 min
Lab File: BN016771.D
Acq: 06 Oct 2021 09:18



Tgt Ion:276 Resp: 3200571
Ion Ratio Lower Upper
276 100
277 24.7 19.0 28.4
138 21.0 17.0 25.6



Instrument :
BNA_N
ClientSampleId :
S-832-J-SO-1-1.5-092221MS

Manual Integrations
APPROVED

mohammad
10/7/2021 8:32:11 AM