



Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN100421\
 Data File : BN016724.D
 Acq On : 05 Oct 2021 01:01
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN100421.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Mon Oct 04 17:52:33 2021






AutoFind: Scans 2448, 2449, 2450; Background Corrected with Scan 2441

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	30.5	60947	PASS
68	69	0.00	2	1.7	1126	PASS
69	198	0.00	100	32.7	65197	PASS
70	69	0.00	2	0.6	409	PASS
127	198	10	80	46.0	91760	PASS
197	198	0.00	2	0.6	1179	PASS
198	198	100	100	100.0	199509	PASS
199	198	5	9	7.0	13950	PASS
275	198	10	60	29.8	59501	PASS
365	198	1	100	4.1	8136	PASS
441	198	0.01	100	20.4	40645	PASS
442	442	50	100	100.0	258816	PASS
443	442	15	24	19.5	50437	PASS

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN100421\
Data File : BN016724.D
Acq On : 05 Oct 2021 01:01
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP

Quant Time: Oct 05 04:48:14 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Sep 30 18:14:09 2021
Response via : Initial Calibration



TIC: BN016724.D\data.ms

(70) Pentachlorophenol (C)

16.675min (-0.000) 27652.09 ng




response 262795

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	62.20	61.76
264.00	61.60	58.98
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN100421\
Data File : BN016724.D
Acq On : 05 Oct 2021 01:01
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP

Quant Time: Oct 05 04:48:14 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Sep 30 18:14:09 2021
Response via : Initial Calibration



TIC: BN016724.D\data.ms

(77) Benzidine

19.292min (-0.000) 0.00 ng m

response 1403909

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	15.50	13.53
183.00	13.20	11.54
0.00	0.00	0.00

DDT Breakdown

Date	Instrument Name	DFTPP Data File
10/4/2021	BNA_N	<u>BN016724.D</u>
Compound Name	Response	Retention Time
DDT	740169	20.533
DDD	5321	20.098
DDE	0	19.58
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
5321	745490	0.71