

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN100518\
 Data File : BN002943.D
 Acq On : 04 Oct 2018 21:22
 Operator : MJ/SJ
 Sample : J5183-08
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 JLC47

Quant Time: Oct 05 04:03:36 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-BN091418MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Oct 05 02:27:06 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.63	152	129427	20.00	ng/ul	0.00
18) Naphthalene-d8	10.39	136	614319	20.00	ng/ul	0.00
35) Acenaphthene-d10	14.26	164	385729	20.00	ng/ul	0.00
61) Phenanthrene-d10	17.02	188	840632	20.00	ng/ul	0.00
77) Chrysene-d12	21.22	240	672830	20.00	ng/ul	0.00
85) Perylene-d12	23.43	264	574909	20.00	ng/ul	0.00

System Monitoring Compounds

3) 1,4-Dioxane-d8	3.20	96	6143	2.19	ng/uL	0.00
5) Phenol-d5	6.82	99	199855	18.18	ng/ul	0.01
7) Bis-(2-Chloroethyl)ether-d	6.97	67	138407	20.28	ng/ul	0.00
9) 2-Chlorophenol-d4	7.17	132	182102	20.13	ng/ul	0.00
13) 4-Methylphenol-d8	8.34	113	164851	19.37	ng/ul	0.00
19) Nitrobenzene-d5	8.78	128	94566	19.87	ng/ul	0.00
22) 2-Nitrophenol-d4	9.49	143	106528	19.78	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	10.03	165	189774	19.52	ng/ul	0.00
29) 4-Chloroaniline-d4	10.55	131	179232	15.94	ng/ul	0.00
43) Dimethylphthalate-d6	13.68	166	574664	19.03	ng/ul	0.00
46) Acenaphthylene-d8	13.95	160	842295	21.36	ng/ul	0.00
51) 4-Nitrophenol-d4	14.52	143	47485	9.51	ng/ul	0.02
57) Fluorene-d10	15.26	176	597385	22.58	ng/ul	0.00
62) 4,6-Dinitro-2-methylphenol	15.42	200	32157	6.12	ng/ul	0.01
70) Anthracene-d10	17.12	188	891814	22.22	ng/ul	0.00
78) Pyrene-d10	19.42	212	913689	25.73	ng/ul	0.00
89) Benzo(a)pyrene-d12	23.29	264	671382	22.19	ng/ul	0.00

Target Compounds

					Ovalue	
44) Dimethylphthalate	13.73	163	300881	10.210	ng/ul	100
69) Phenanthrene	17.06	178	140008	3.023	ng/ul	98
76) Fluoranthene	19.09	202	202680	4.102	ng/ul	98
79) Pyrene	19.45	202	198762	4.468	ng/ul	99
80) Butylbenzylphthalate	20.36	149	20178	1.090	ng/ul	94
82) Benzo(a)anthracene	21.21	228	81423	1.945	ng/ul	97
83) Bis(2-ethylhexyl)phthalate	21.14	149	134709	5.013	ng/ul	98
84) Chrysene	21.26	228	183154	4.668	ng/ul	99
87) Benzo(b)fluoranthene	22.77	252	168506	4.799	ng/ul#	87
88) Benzo(k)fluoranthene	22.82	252	52595	1.606	ng/ul#	77
90) Benzo(a)pyrene	23.33	252	69014	2.079	ng/ul#	84
91) Indeno(1,2,3-cd)pyrene	25.63	276	85630	2.179	ng/ul	98
93) Benzo(g,h,i)perylene	26.31	276	99451	3.027	ng/ul	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN100518\
Data File : BN002943.D
Acq On : 04 Oct 2018 21:22
Operator : MJ/SJ
Sample : J5183-08
Misc :
ALS Vial : 21 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
JLC47

Quant Time: Oct 05 04:03:36 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-BN091418MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Oct 05 02:27:06 2018
Response via : Initial Calibration

