

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN100518\
 Data File : BN002924.D
 Acq On : 04 Oct 2018 10:14
 Operator : MJ/SJ
 Sample : PB113447BL
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SBLK47

Quant Time: Oct 05 02:01:05 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-BN091418MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Oct 05 02:00:32 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.62	152	112359	20.00	ng/ul	0.00
18) Naphthalene-d8	10.39	136	517231	20.00	ng/ul	0.00
35) Acenaphthene-d10	14.26	164	312537	20.00	ng/ul	0.00
61) Phenanthrene-d10	17.01	188	745042	20.00	ng/ul	0.00
77) Chrysene-d12	21.22	240	815454	20.00	ng/ul	0.00
85) Perylene-d12	23.42	264	702189	20.00	ng/ul	0.00

System Monitoring Compounds

3) 1,4-Dioxane-d8	3.20	96	11271	4.63	ng/uL	0.00
5) Phenol-d5	6.82	99	231997	24.31	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	6.97	67	148227	25.01	ng/ul	0.00
9) 2-Chlorophenol-d4	7.16	132	195451	24.89	ng/ul	0.00
13) 4-Methylphenol-d8	8.33	113	190289	25.75	ng/ul	0.00
19) Nitrobenzene-d5	8.78	128	98352	24.55	ng/ul	0.00
22) 2-Nitrophenol-d4	9.49	143	114754	25.31	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	10.03	165	191181	23.36	ng/ul	0.00
29) 4-Chloroaniline-d4	10.54	131	284018	29.99	ng/ul	0.00
43) Dimethylphthalate-d6	13.68	166	667026	27.27	ng/ul	0.00
46) Acenaphthylene-d8	13.95	160	837620	26.22	ng/ul	0.00
51) 4-Nitrophenol-d4	14.50	143	90657	22.41	ng/ul	0.00
57) Fluorene-d10	15.26	176	597123	27.86	ng/ul	0.00
62) 4,6-Dinitro-2-methylphenol	15.40	200	94019	20.17	ng/ul	0.00
70) Anthracene-d10	17.11	188	952030	26.76	ng/ul	0.00
78) Pyrene-d10	19.42	212	1076838	25.02	ng/ul	0.00
89) Benzo(a)pyrene-d12	23.29	264	993605	26.88	ng/ul	0.00

Target Compounds Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN100518\
Data File : BN002924.D
Acq On : 04 Oct 2018 10:14
Operator : MJ/SJ
Sample : PB113447BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SBLK47

Quant Time: Oct 05 02:01:05 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-BN091418MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Oct 05 02:00:32 2018
Response via : Initial Calibration

