

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN100520\
 Data File : BN012060.D
 Acq On : 05 Oct 2020 17:33
 Operator : CG/JU
 Sample : L4185-12
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 C0AR0

Quant Time: Oct 05 18:23:22 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-BN100220.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Oct 05 17:32:05 2020
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.658	152	248237	20.00	ng/ul	0.00
18) Naphthalene-d8	10.434	136	951386	20.00	ng/ul	0.00
36) Acenaphthene-d10	14.293	164	578708	20.00	ng/ul	0.00
62) Phenanthrene-d10	17.039	188	1215839	20.00	ng/ul	0.00
78) Chrysene-d12	21.245	240	1206341	20.00	ng/ul	0.00
86) Perylene-d12	23.451	264	1356465	20.00	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.199	96	34476	6.04	ng/uL	0.00
5) Phenol-d5	6.828	99	126760	6.12	ng/ul	0.00
7) Bis-(2-Chloroethyl)eth...	6.999	67	340119	26.72	ng/ul	0.00
9) 2-Chlorophenol-d4	7.193	132	372608	22.24	ng/ul	0.00
13) 4-Methylphenol-d8	8.358	113	234900	14.40	ng/ul	0.00
19) Nitrobenzene-d5	8.805	128	214450	28.39	ng/ul	0.00
22) 2-Nitrophenol-d4	9.522	143	221318	27.09	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	10.052	165	433827	27.69	ng/ul	0.00
29) 4-Chloroaniline-d4	10.575	131	35933	1.95	ng/ul	0.00
44) Dimethylphthalate-d6	13.710	166	1583705	34.89	ng/ul	0.00
47) Acenaphthylene-d8	13.987	160	1750317	32.30	ng/ul	0.00
52) 4-Nitrophenol-d4	14.493	143	42141	5.04	ng/ul	0.00
58) Fluorene-d10	15.293	176	1307360	33.00	ng/ul	0.00
63) 4,6-Dinitro-2-methylph...	15.410	200	237788	30.64	ng/ul	0.00
71) Anthracene-d10	17.139	188	2112368	35.83	ng/ul	0.00
79) Pyrene-d10	19.445	212	2464038	38.84	ng/ul	0.00
90) Benzo(a)pyrene-d12	23.316	264	2818397	37.02	ng/ul	0.00
Target Compounds						
45) Dimethylphthalate	13.757	163	141323	3.02	ng/ul	Qvalue 98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN100520\
Data File : BN012060.D
Acq On : 05 Oct 2020 17:33
Operator : CG/JU
Sample : L4185-12
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
C0AR0


Quant Time: Oct 05 18:23:22 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-BN100220.M
Quant Title : SVOA CALIBRATION
QLast Update : Mon Oct 05 17:32:05 2020
Response via : Initial Calibration





#45
Dimethylphthalate
Concen: 3.02 ng/ul
RT: 13.757 min Scan# 1831
Delta R.T. -0.005 min
Lab File: BN012060.D
Acq: 05 Oct 2020 17:33

Instrument :
BNA_N
ClientSampleId :
C0AR0



Tgt Ion:163 Resp: 141323
Ion Ratio Lower Upper
163 100
194 4.4 3.4 5.0
164 9.0 7.8 11.8

