

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN100520\
 Data File : BN012061.D
 Acq On : 05 Oct 2020 18:09
 Operator : CG/JU
 Sample : L4185-14
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 C0AR6

Quant Time: Oct 05 18:45:06 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-BN100220.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Oct 05 17:32:05 2020
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.658	152	249616	20.00	ng/ul	0.00
18) Naphthalene-d8	10.428	136	980765	20.00	ng/ul	0.00
36) Acenaphthene-d10	14.293	164	580325	20.00	ng/ul	0.00
62) Phenanthrene-d10	17.045	188	1181947	20.00	ng/ul	0.00
78) Chrysene-d12	21.245	240	1154171	20.00	ng/ul	0.00
86) Perylene-d12	23.451	264	1294332	20.00	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.199	96	29914	5.21	ng/uL	0.00
5) Phenol-d5	6.828	99	107711	5.17	ng/ul	0.00
7) Bis-(2-Chloroethyl)eth...	6.999	67	331566	25.90	ng/ul	0.00
9) 2-Chlorophenol-d4	7.193	132	352997	20.95	ng/ul	0.00
13) 4-Methylphenol-d8	8.358	113	215173	13.12	ng/ul	0.00
19) Nitrobenzene-d5	8.805	128	202455	26.00	ng/ul	0.00
22) 2-Nitrophenol-d4	9.522	143	221934	26.35	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	10.052	165	411136	25.45	ng/ul	0.00
29) 4-Chloroaniline-d4	10.569	131	168264	8.84	ng/ul	0.00
44) Dimethylphthalate-d6	13.710	166	1466591	32.22	ng/ul	0.00
47) Acenaphthylene-d8	13.987	160	1682148	30.96	ng/ul	0.00
52) 4-Nitrophenol-d4	14.487	143	29289	3.49	ng/ul	-0.01
58) Fluorene-d10	15.292	176	1213567	30.55	ng/ul	0.00
63) 4,6-Dinitro-2-methylph...	15.410	200	198783	26.35	ng/ul	0.00
71) Anthracene-d10	17.139	188	1898420	33.12	ng/ul	0.00
79) Pyrene-d10	19.445	212	2135752	35.18	ng/ul	0.00
90) Benzo(a)pyrene-d12	23.310	264	2358790	32.47	ng/ul	0.00
Target Compounds						
45) Dimethylphthalate	13.757	163	83336	1.78	ng/ul#	97


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN100520\
Data File : BN012061.D
Acq On : 05 Oct 2020 18:09
Operator : CG/JU
Sample : L4185-14
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
C0AR6


Quant Time: Oct 05 18:45:06 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-BN100220.M
Quant Title : SVOA CALIBRATION
QLast Update : Mon Oct 05 17:32:05 2020
Response via : Initial Calibration





#45
Dimethylphthalate
Concen: 1.78 ng/ul
RT: 13.757 min Scan# 1831
Delta R.T. -0.005 min
Lab File: BN012061.D
Acq: 05 Oct 2020 18:09

Instrument :
BNA_N
ClientSampleId :
C0AR6



Tgt Ion:163 Resp: 83336
Ion Ratio Lower Upper
163 100
194 5.7 3.4 5.0#
164 8.8 7.8 11.8

