

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN100621\
 Data File : BN016787.D
 Acq On : 06 Oct 2021 19:53
 Operator : CG/JU
 Sample : M3915-19
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 RE109D1-20210922

Manual Integrations
 APPROVED

mohammad
 10/7/2021 8:51:32 AM

Quant Time: Oct 07 00:59:07 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN100421.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Oct 06 14:22:07 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.633	152	20992	0.400	ng	0.00
7) Naphthalene-d8	10.407	136	97881	0.400	ng	# 0.00
13) Acenaphthene-d10	14.269	164	55092	0.400	ng	0.00
19) Phenanthrene-d10	17.018	188	108787	0.400	ng	# 0.00
29) Chrysene-d12	21.228	240	115813	0.400	ng	# 0.00
35) Perylene-d12	23.481	264	123172	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.264	112	6620	0.124	ng	0.02
5) Phenol-d6	6.822	99	4032	0.068	ng	0.00
8) Nitrobenzene-d5	8.772	82	17538	0.256	ng	0.00
11) 2-Methylnaphthalene-d10	11.998	152	43788	0.300	ng	0.00
14) 2,4,6-Tribromophenol	15.766	330	7935	0.367	ng	0.00
15) 2-Fluorobiphenyl	12.886	172	56438	0.256	ng	0.00
27) Fluoranthene-d10	19.058	212	113415	0.381	ng	0.00
31) Terphenyl-d14	19.674	244	93683	0.362	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.253	88	27461m	2.803	ng	Qvalue
34) Bis(2-ethylhexyl)phtha...	21.164	149	26979	0.124	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN100621\
Data File : BN016787.D
Acq On : 06 Oct 2021 19:53
Operator : CG/JU
Sample : M3915-19
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
RE109D1-20210922

Manual Integrations
APPROVED

mohammad
10/7/2021 8:51:32 AM

Quant Time: Oct 07 00:59:07 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN100421.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Oct 06 14:22:07 2021
Response via : Initial Calibration

