

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN100621\
 Data File : BN016793.D
 Acq On : 06 Oct 2021 23:31
 Operator : CG/JU
 Sample : M4063-01
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 P287-DITCH

Quant Time: Oct 07 00:59:50 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN100421.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Oct 06 14:22:07 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.633	152	21407	0.400	ng	0.00
7) Naphthalene-d8	10.406	136	100926	0.400	ng	# 0.00
13) Acenaphthene-d10	14.268	164	59343	0.400	ng	0.00
19) Phenanthrene-d10	17.017	188	117266	0.400	ng	# 0.00
29) Chrysene-d12	21.227	240	119353	0.400	ng	# 0.00
35) Perylene-d12	23.477	264	125315	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.263	112	8031	0.147	ng	0.02
5) Phenol-d6	6.821	99	5580	0.092	ng	0.00
8) Nitrobenzene-d5	8.771	82	21612	0.305	ng	0.00
11) 2-Methylnaphthalene-d10	11.998	152	54132	0.360	ng	0.00
14) 2,4,6-Tribromophenol	15.766	330	13511	0.512	ng	0.00
15) 2-Fluorobiphenyl	12.885	172	74810	0.315	ng	0.00
27) Fluoranthene-d10	19.063	212	132522	0.413	ng	0.00
31) Terphenyl-d14	19.673	244	117221	0.439	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.253	88	207	0.021	ng	# 1
24) Pentachlorophenol	16.676	266	141	0.342	ng	93
34) Bis(2-ethylhexyl)phtha...	21.163	149	24986	0.111	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN100621\
Data File : BN016793.D
Acq On : 06 Oct 2021 23:31
Operator : CG/JU
Sample : M4063-01
Misc :
ALS Vial : 18 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
P287-DITCH

Quant Time: Oct 07 00:59:50 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN100421.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Oct 06 14:22:07 2021
Response via : Initial Calibration

