

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN100621\
 Data File : BN016795.D
 Acq On : 07 Oct 2021 00:43
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Oct 07 01:21:57 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN100421.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Oct 06 14:22:07 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	85	0.00
2	1,4-Dioxane	0.400	0.436	-9.0	99	0.00
3	n-Nitrosodimethylamine	0.400	0.429	-7.2	98	0.00
4 S	2-Fluorophenol	0.400	0.433	-8.2	96	0.00
5 S	Phenol-d6	0.400	0.442	-10.5	100	0.00
6	bis(2-Chloroethyl)ether	0.400	0.414	-3.5	90	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	85	0.00
8 S	Nitrobenzene-d5	0.400	0.412	-3.0	102	0.00
9	Naphthalene	0.400	0.394	1.5	88	0.00
10	Hexachlorobutadiene	0.400	0.411	-2.7	91	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.393	1.8	87	0.00
12	2-Methylnaphthalene	0.400	0.409	-2.2	90	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	86	0.00
14 S	2,4,6-Tribromophenol	0.400	0.499	-24.7	134	0.00
15 S	2-Fluorobiphenyl	0.400	0.400	0.0	92	0.00
16	Acenaphthylene	0.400	0.403	-0.8	96	0.00
17	Acenaphthene	0.400	0.398	0.5	91	0.00
18	Fluorene	0.400	0.404	-1.0	93	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	88	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.806	-101.5#	1050	0.00
21	4-Bromophenyl-phenylether	0.400	0.404	-1.0	98	0.00
22	Hexachlorobenzene	0.400	0.411	-2.7	97	0.00
23	Atrazine	0.400	0.405	-1.3	109	0.00
24	Pentachlorophenol	0.400	0.671	-67.8#	303	0.00
25	Phenanthrene	0.400	0.403	-0.8	94	0.00
26	Anthracene	0.400	0.397	0.8	98	0.00
27 SURR	Fluoranthene-d10	0.400	0.405	-1.3	97	0.00
28	Fluoranthene	0.400	0.425	-6.2	99	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	102	0.00
30	Pyrene	0.400	0.404	-1.0	101	0.00
31 S	Terphenyl-d14	0.400	0.414	-3.5	104	0.00
32	Benzo(a)anthracene	0.400	0.428	-7.0	110	-0.01
33	Chrysene	0.400	0.425	-6.2	106	-0.01
34	Bis(2-ethylhexyl)phthalate	0.400	0.426	-6.5	125	0.00
35 I	Perylene-d12	0.400	0.400	0.0	113	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.493	-23.2	147	0.00
37	Benzo(b)fluoranthene	0.400	0.417	-4.2	114	0.00
38	Benzo(k)fluoranthene	0.400	0.435	-8.7	121	0.00
39 C	Benzo(a)pyrene	0.400	0.432	-8.0	121	0.00
40	Dibenzo(a,h)anthracene	0.400	0.508	-27.0#	154	0.00
41	Benzo(g,h,i)perylene	0.400	0.465	-16.3	136	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

LabSampleId :

SSTDCCC0.4