

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN100621\
 Data File : BN016797.D
 Acq On : 07 Oct 2021 02:31
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Oct 07 04:11:11 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN100421.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Oct 06 14:22:07 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	76	0.00
2	1,4-Dioxane	0.400	0.371	7.3	76	0.00
3	n-Nitrosodimethylamine	0.400	0.412	-3.0	84	0.00
4 S	2-Fluorophenol	0.400	0.404	-1.0	81	0.00
5 S	Phenol-d6	0.400	0.407	-1.7	83	0.00
6	bis(2-Chloroethyl)ether	0.400	0.414	-3.5	81	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	77	0.00
8 S	Nitrobenzene-d5	0.400	0.365	8.8	82	0.00
9	Naphthalene	0.400	0.390	2.5	79	0.00
10	Hexachlorobutadiene	0.400	0.405	-1.3	81	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.398	0.5	80	0.00
12	2-Methylnaphthalene	0.400	0.407	-1.7	81	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	78	0.00
14 S	2,4,6-Tribromophenol	0.400	0.458	-14.5	109	0.00
15 S	2-Fluorobiphenyl	0.400	0.393	1.8	82	0.00
16	Acenaphthylene	0.400	0.385	3.8	83	0.00
17	Acenaphthene	0.400	0.395	1.3	82	0.00
18	Fluorene	0.400	0.403	-0.8	84	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	80	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.685	-71.3#	716	0.00
21	4-Bromophenyl-phenylether	0.400	0.397	0.8	88	0.00
22	Hexachlorobenzene	0.400	0.417	-4.2	89	0.00
23	Atrazine	0.400	0.359	10.3	88	0.00
24	Pentachlorophenol	0.400	0.640	-60.0#	250	0.00
25	Phenanthrene	0.400	0.408	-2.0	87	0.00
26	Anthracene	0.400	0.386	3.5	87	0.00
27 SURR	Fluoranthene-d10	0.400	0.400	0.0	87	0.00
28	Fluoranthene	0.400	0.420	-5.0	89	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	92	0.00
30	Pyrene	0.400	0.400	0.0	90	0.00
31 S	Terphenyl-d14	0.400	0.407	-1.7	92	0.00
32	Benzo(a)anthracene	0.400	0.413	-3.2	95	-0.01
33	Chrysene	0.400	0.429	-7.2	96	-0.01
34	Bis(2-ethylhexyl)phthalate	0.400	0.296	26.0#	78	0.00
35 I	Perylene-d12	0.400	0.400	0.0	102	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.488	-22.0	132	0.00
37	Benzo(b)fluoranthene	0.400	0.411	-2.7	102	0.00
38	Benzo(k)fluoranthene	0.400	0.431	-7.7	109	0.00
39 C	Benzo(a)pyrene	0.400	0.417	-4.2	106	0.00
40	Dibenzo(a,h)anthracene	0.400	0.503	-25.7#	138	0.00
41	Benzo(g,h,i)perylene	0.400	0.454	-13.5	120	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

LabSampleId :

SSTDCCC0.4