

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN100621\
 Data File : BN016797.D
 Acq On : 07 Oct 2021 02:31
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4

Quant Time: Oct 07 04:11:11 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN100421.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Oct 06 14:22:07 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.633	152	19172	0.40	ng	0.00
7) Naphthalene-d8	10.406	136	86123	0.40	ng	# 0.00
13) Acenaphthene-d10	14.268	164	46378	0.40	ng	0.00
19) Phenanthrene-d10	17.017	188	89205	0.40	ng	0.00
29) Chrysene-d12	21.227	240	93961	0.40	ng	0.00
35) Perylene-d12	23.474	264	100959	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.245	112	19782	0.40	ng	0.00
5) Phenol-d6	6.812	99	22112	0.41	ng	0.00
8) Nitrobenzene-d5	8.771	82	22029	0.36	ng	0.00
11) 2-Methylnaphthalene-d10	11.998	152	51059	0.40	ng	0.00
14) 2,4,6-Tribromophenol	15.765	330	9122	0.46	ng	0.00
15) 2-Fluorobiphenyl	12.885	172	72900	0.39	ng	0.00
27) Fluoranthene-d10	19.058	212	97703	0.40	ng	0.00
31) Terphenyl-d14	19.673	244	85643	0.41	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.244	88	3317	0.37	ng	# 73
3) n-Nitrosodimethylamine	3.576	42	5951	0.41	ng	# 100
6) bis(2-Chloroethyl)ether	7.070	93	21551	0.41	ng	100
9) Naphthalene	10.444	128	91991	0.39	ng	100
10) Hexachlorobutadiene	10.736	225	17847	0.41	ng	# 100
12) 2-Methylnaphthalene	12.069	142	60906	0.41	ng	100
16) Acenaphthylene	13.977	152	78910	0.39	ng	100
17) Acenaphthene	14.332	154	53484	0.40	ng	100
18) Fluorene	15.321	166	65713	0.40	ng	100
20) 4,6-Dinitro-2-methylph...	15.410	198	3343	0.68	ng	# 66
21) 4-Bromophenyl-phenylether	16.214	248	21596	0.40	ng	# 89
22) Hexachlorobenzene	16.323	284	25667	0.42	ng	98
23) Atrazine	16.494	200	15449	0.36	ng	97
24) Pentachlorophenol	16.676	266	9031	0.64	ng	99
25) Phenanthrene	17.054	178	106998	0.41	ng	100
26) Anthracene	17.151	178	90338	0.39	ng	100
28) Fluoranthene	19.087	202	122359	0.42	ng	100
30) Pyrene	19.455	202	123566	0.40	ng	100
32) Benzo(a)anthracene	21.206	228	120463	0.41	ng	99
33) Chrysene	21.259	228	126932	0.43	ng	99
34) Bis(2-ethylhexyl)phtha...	21.163	149	52282	0.30	ng	100
36) Indeno(1,2,3-cd)pyrene	25.750	276	156384	0.49	ng	99
37) Benzo(b)fluoranthene	22.796	252	134328	0.41	ng	99
38) Benzo(k)fluoranthene	22.841	252	135107	0.43	ng	99
39) Benzo(a)pyrene	23.375	252	121944	0.42	ng	99
40) Dibenzo(a,h)anthracene	25.771	278	127896	0.50	ng	99
41) Benzo(g,h,i)perylene	26.445	276	129055	0.45	ng	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN100621\
Data File : BN016797.D
Acq On : 07 Oct 2021 02:31
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Quant Time: Oct 07 04:11:11 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN100421.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Oct 06 14:22:07 2021
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 0.40 ng
RT: 7.633 min Scan# 494
Delta R.T. -0.000 min
Lab File: BN016797.D
Acq: 07 Oct 2021 02:31




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:152 Resp: 19172
Ion Ratio Lower Upper
152 100
150 156.9 123.3 184.9
115 59.0 47.0 70.6



#2
1,4-Dioxane
Concen: 0.37 ng
RT: 3.244 min Scan# 18
Delta R.T. -0.000 min
Lab File: BN016797.D
Acq: 07 Oct 2021 02:31

Tgt Ion: 88 Resp: 3317
Ion Ratio Lower Upper
88 100
43 30.1 61.7 92.5#
58 77.3 62.2 93.2









#3
n-Nitrosodimethylamine
Concen: 0.41 ng
RT: 3.576 min Scan# 54
Delta R.T. -0.000 min
Lab File: BN016797.D
Acq: 07 Oct 2021 02:31




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion: 42 Resp: 5951
Ion Ratio Lower Upper
42 100
74 154.6 124.0 186.0
44 5.6 5.6 8.4#



#4
2-Fluorophenol
Concen: 0.40 ng
RT: 5.245 min Scan# 235
Delta R.T. -0.000 min
Lab File: BN016797.D
Acq: 07 Oct 2021 02:31

Tgt Ion: 112 Resp: 19782
Ion Ratio Lower Upper
112 100
64 54.5 42.9 64.3
63 27.3 21.5 32.3









#5
Phenol-d6
Concen: 0.41 ng
RT: 6.812 min Scan# 405
Delta R.T. -0.000 min
Lab File: BN016797.D
Acq: 07 Oct 2021 02:31




Instrument :
BNA_N
ClientSampled :
SSTDCCC0.4


Tgt Ion: 99 Resp: 22112
Ion Ratio Lower Upper
99 100
42 16.8 13.4 20.2
71 30.0 24.0 36.0



#6
bis(2-Chloroethyl)ether
Concen: 0.41 ng
RT: 7.070 min Scan# 433
Delta R.T. -0.000 min
Lab File: BN016797.D
Acq: 07 Oct 2021 02:31

Tgt Ion: 93 Resp: 21551
Ion Ratio Lower Upper
93 100
63 73.4 59.0 88.6
95 32.8 26.4 39.6









#7
Naphthalene-d8
Concen: 0.40 ng
RT: 10.406 min Scan# 733
Delta R.T. -0.000 min
Lab File: BN016797.D
Acq: 07 Oct 2021 02:31




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:	136	Resp:	86123
Ion	Ratio	Lower	Upper
136	100		
137	11.0	8.9	13.3
54	4.3	6.5	9.7#
68	3.8	5.3	7.9#



#8
Nitrobenzene-d5
Concen: 0.36 ng
RT: 8.771 min Scan# 604
Delta R.T. -0.000 min
Lab File: BN016797.D
Acq: 07 Oct 2021 02:31

Tgt Ion:	82	Resp:	22029
Ion	Ratio	Lower	Upper
82	100		
128	38.7	33.8	50.6
54	50.4	38.5	57.7









#9
Naphthalene
Concen: 0.39 ng
RT: 10.444 min Scan# 736
Delta R.T. -0.000 min
Lab File: BN016797.D
Acq: 07 Oct 2021 02:31




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:	128	Resp:	91991
Ion	Ratio	Lower	Upper
128	100		
129	11.0	8.7	13.1
127	13.2	10.3	15.5



#10
Hexachlorobutadiene
Concen: 0.41 ng
RT: 10.736 min Scan# 759
Delta R.T. -0.000 min
Lab File: BN016797.D
Acq: 07 Oct 2021 02:31

Tgt Ion:	225	Resp:	17847
Ion	Ratio	Lower	Upper
225	100		
223	0.0	0.0	0.0
227	63.8	51.1	76.7









#11
2-Methylnaphthalene-d10
Concen: 0.40 ng
RT: 11.998 min Scan# 930
Delta R.T. -0.000 min
Lab File: BN016797.D
Acq: 07 Oct 2021 02:31




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:152 Resp: 51059
Ion Ratio Lower Upper
152 100
151 20.2 16.2 24.4



#12
2-Methylnaphthalene
Concen: 0.41 ng
RT: 12.069 min Scan# 946
Delta R.T. -0.000 min
Lab File: BN016797.D
Acq: 07 Oct 2021 02:31

Tgt Ion:142 Resp: 60906
Ion Ratio Lower Upper
142 100
141 88.5 70.6 106.0
115 33.2 26.1 39.1









#13
Acenaphthene-d10
Concen: 0.40 ng
RT: 14.268 min Scan# 1183
Delta R.T. -0.000 min
Lab File: BN016797.D
Acq: 07 Oct 2021 02:31




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:164 Resp: 46378
Ion Ratio Lower Upper
164 100
162 98.4 82.6 124.0
160 43.1 38.2 57.2



#14
2,4,6-Tribromophenol
Concen: 0.46 ng
RT: 15.765 min Scan# 1301
Delta R.T. -0.000 min
Lab File: BN016797.D
Acq: 07 Oct 2021 02:31

Tgt Ion:330 Resp: 9122
Ion Ratio Lower Upper
330 100
332 0.0 0.0 0.0
141 28.3 24.2 36.2









#15
2-Fluorobiphenyl
Concen: 0.39 ng
RT: 12.885 min Scan# 1074
Delta R.T. -0.000 min
Lab File: BN016797.D
Acq: 07 Oct 2021 02:31




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:172 Resp: 72900
Ion Ratio Lower Upper
172 100
171 34.9 26.9 40.3
170 23.0 17.4 26.0



#16
Acenaphthylene
Concen: 0.39 ng
RT: 13.977 min Scan# 1160
Delta R.T. -0.000 min
Lab File: BN016797.D
Acq: 07 Oct 2021 02:31

Tgt Ion:152 Resp: 78910
Ion Ratio Lower Upper
152 100
151 19.4 15.5 23.3
153 13.0 10.4 15.6









#17
Acenaphthene
Concen: 0.40 ng
RT: 14.332 min Scan# 1188
Delta R.T. -0.000 min
Lab File: BN016797.D
Acq: 07 Oct 2021 02:31




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:154 Resp: 53484
Ion Ratio Lower Upper
154 100
153 113.0 90.1 135.1
152 54.9 43.6 65.4



#18
Fluorene
Concen: 0.40 ng
RT: 15.321 min Scan# 1266
Delta R.T. -0.000 min
Lab File: BN016797.D
Acq: 07 Oct 2021 02:31

Tgt Ion:166 Resp: 65713
Ion Ratio Lower Upper
166 100
165 97.3 78.1 117.1
167 13.6 10.7 16.1









#19
Phenanthrene-d10
Concen: 0.40 ng
RT: 17.017 min Scan# 1403
Delta R.T. -0.000 min
Lab File: BN016797.D
Acq: 07 Oct 2021 02:31




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:188 Resp: 89205
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 8.8 5.8 8.8



#20
4,6-Dinitro-2-methylphenol
Concen: 0.68 ng
RT: 15.410 min Scan# 1273
Delta R.T. -0.000 min
Lab File: BN016797.D
Acq: 07 Oct 2021 02:31

Tgt Ion:198 Resp: 3343
Ion Ratio Lower Upper
198 100
182 13.3 26.2 39.2#
77 0.0 0.0 0.0









#21
4-Bromophenyl-phenylether
Concen: 0.40 ng
RT: 16.214 min Scan# 1337
Delta R.T. -0.000 min
Lab File: BN016797.D
Acq: 07 Oct 2021 02:31




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:248 Resp: 21596
Ion Ratio Lower Upper
248 100
250 92.8 76.2 114.4
141 74.6 45.6 68.4#



#22
Hexachlorobenzene
Concen: 0.42 ng
RT: 16.323 min Scan# 1346
Delta R.T. -0.000 min
Lab File: BN016797.D
Acq: 07 Oct 2021 02:31

Tgt Ion:284 Resp: 25667
Ion Ratio Lower Upper
284 100
142 31.0 25.8 38.6
249 28.6 23.3 34.9









#23
Atrazine
Concen: 0.36 ng
RT: 16.494 min Scan# 1360
Delta R.T. -0.000 min
Lab File: BN016797.D
Acq: 07 Oct 2021 02:31




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:200 Resp: 15449
Ion Ratio Lower Upper
200 100
173 28.6 20.1 30.1
215 47.5 38.8 58.2



#24
Pentachlorophenol
Concen: 0.64 ng
RT: 16.676 min Scan# 1375
Delta R.T. -0.000 min
Lab File: BN016797.D
Acq: 07 Oct 2021 02:31

Tgt Ion:266 Resp: 9031
Ion Ratio Lower Upper
266 100
264 62.8 49.3 73.9
268 63.9 51.8 77.8









#25
Phenanthrene
Concen: 0.41 ng
RT: 17.054 min Scan# 1406
Delta R.T. -0.000 min
Lab File: BN016797.D
Acq: 07 Oct 2021 02:31




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:178 Resp: 106998
Ion Ratio Lower Upper
178 100
176 18.8 15.0 22.6
179 15.5 12.4 18.6



#26
Anthracene
Concen: 0.39 ng
RT: 17.151 min Scan# 1414
Delta R.T. -0.000 min
Lab File: BN016797.D
Acq: 07 Oct 2021 02:31

Tgt Ion:178 Resp: 90338
Ion Ratio Lower Upper
178 100
176 18.2 14.5 21.7
179 15.3 12.2 18.4









#27
Fluoranthene-d10
Concen: 0.40 ng
RT: 19.058 min Scan# 1695
Delta R.T. -0.005 min
Lab File: BN016797.D
Acq: 07 Oct 2021 02:31




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:212 Resp: 97703
Ion Ratio Lower Upper
212 100
106 13.2 10.6 16.0
104 7.2 5.8 8.8



#28
Fluoranthene
Concen: 0.42 ng
RT: 19.087 min Scan# 1701
Delta R.T. -0.005 min
Lab File: BN016797.D
Acq: 07 Oct 2021 02:31

Tgt Ion:202 Resp: 122359
Ion Ratio Lower Upper
202 100
101 10.9 8.9 13.3
203 17.3 13.8 20.8









#29
Chrysene-d12
Concen: 0.40 ng
RT: 21.227 min Scan# 2001
Delta R.T. -0.000 min
Lab File: BN016797.D
Acq: 07 Oct 2021 02:31




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:240 Resp: 93961
Ion Ratio Lower Upper
240 100
120 7.3 6.6 10.0
236 25.6 20.7 31.1



#30
Pyrene
Concen: 0.40 ng
RT: 19.455 min Scan# 1775
Delta R.T. -0.000 min
Lab File: BN016797.D
Acq: 07 Oct 2021 02:31

Tgt Ion:202 Resp: 123566
Ion Ratio Lower Upper
202 100
200 20.6 16.5 24.7
203 17.4 13.9 20.9









#31
Terphenyl-d14
Concen: 0.41 ng
RT: 19.673 min Scan# 1819
Delta R.T. -0.000 min
Lab File: BN016797.D
Acq: 07 Oct 2021 02:31




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:244 Resp: 85643
Ion Ratio Lower Upper
244 100
212 6.9 5.8 8.6
122 11.0 8.9 13.3



#32
Benzo(a)anthracene
Concen: 0.41 ng
RT: 21.206 min Scan# 1999
Delta R.T. -0.011 min
Lab File: BN016797.D
Acq: 07 Oct 2021 02:31

Tgt Ion:228 Resp: 120463
Ion Ratio Lower Upper
228 100
226 27.4 22.3 33.5
229 19.3 15.4 23.0









#33
Chrysene
Concen: 0.43 ng
RT: 21.259 min Scan# 2004
Delta R.T. -0.011 min
Lab File: BN016797.D
Acq: 07 Oct 2021 02:31

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Tgt Ion:228 Resp: 126932




Ion	Ratio	Lower	Upper
228	100		
226	29.8	24.2	36.4
229	19.2	15.4	23.0




#34
Bis(2-ethylhexyl)phthalate
Concen: 0.30 ng
RT: 21.163 min Scan# 1995
Delta R.T. -0.000 min
Lab File: BN016797.D
Acq: 07 Oct 2021 02:31

Tgt Ion:149 Resp: 52282


Ion	Ratio	Lower	Upper
149	100		
167	27.9	22.2	33.4
279	4.4	3.4	5.2








#35
Perylene-d12
Concen: 0.40 ng
RT: 23.474 min Scan# 2424
Delta R.T. -0.007 min
Lab File: BN016797.D
Acq: 07 Oct 2021 02:31

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4







Tgt Ion:264 Resp: 100959
Ion Ratio Lower Upper
264 100
260 24.3 19.8 29.6
265 23.4 20.3 30.5



#36
Indeno(1,2,3-cd)pyrene
Concen: 0.49 ng
RT: 25.750 min Scan# 3089
Delta R.T. -0.004 min
Lab File: BN016797.D
Acq: 07 Oct 2021 02:31


Tgt Ion:276 Resp: 156384
Ion Ratio Lower Upper
276 100
138 24.8 19.2 28.8
277 25.1 20.0 30.0







#37
Benzo(b)fluoranthene
Concen: 0.41 ng
RT: 22.796 min Scan# 2226
Delta R.T. -0.007 min
Lab File: BN016797.D
Acq: 07 Oct 2021 02:31


Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4





Tgt Ion:252 Resp: 134328
Ion Ratio Lower Upper
252 100
253 22.0 17.8 26.8
125 9.9 8.3 12.5




#38
Benzo(k)fluoranthene
Concen: 0.43 ng
RT: 22.841 min Scan# 2239
Delta R.T. -0.007 min
Lab File: BN016797.D
Acq: 07 Oct 2021 02:31




Tgt Ion:252 Resp: 135107
Ion Ratio Lower Upper
252 100
253 22.3 17.8 26.8
125 11.8 8.5 12.7








#39
Benzo(a)pyrene
Concen: 0.42 ng
RT: 23.375 min Scan# 2395
Delta R.T. -0.007 min
Lab File: BN016797.D
Acq: 07 Oct 2021 02:31

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4







Tgt Ion:252 Resp: 121944
Ion Ratio Lower Upper
252 100
253 22.4 17.9 26.9
125 10.9 9.4 14.2



#40
Dibenzo(a,h)anthracene
Concen: 0.50 ng
RT: 25.771 min Scan# 3095
Delta R.T. -0.007 min
Lab File: BN016797.D
Acq: 07 Oct 2021 02:31

Tgt Ion:278 Resp: 127896
Ion Ratio Lower Upper
278 100
139 16.8 13.1 19.7
279 23.9 19.3 28.9





#41
 Benzo(g,h,i)perylene
 Concen: 0.45 ng
 RT: 26.445 min Scan# 3292
 Delta R.T. -0.007 min
 Lab File: BN016797.D
 Acq: 07 Oct 2021 02:31

Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4

Tgt Ion:	276	Resp:	129055
Ion	Ratio	Lower	Upper
276	100		
277	23.7	19.0	28.4
138	20.6	17.0	25.6

