

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN100625\
 Data File : BN038013.D
 Acq On : 06 Oct 2025 17:22
 Operator : RC/JU
 Sample : Q3194-08
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 TT188S1-20250922

Quant Time: Oct 06 17:47:13 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN092325.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 24 11:00:01 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.804	152	8436	0.400	ng	-0.02
7) Naphthalene-d8	10.605	136	23277	0.400	ng	-0.02
13) Acenaphthene-d10	14.462	164	11715	0.400	ng	-0.01
19) Phenanthrene-d10	17.211	188	22848	0.400	ng	-0.01
29) Chrysene-d12	21.393	240	13580	0.400	ng	0.00
35) Perylene-d12	23.712	264	10518	0.400	ng	#-0.01
System Monitoring Compounds						
4) 2-Fluorophenol	5.377	112	2117	0.110	ng	0.00
5) Phenol-d6	6.966	99	1879	0.074	ng	0.00
8) Nitrobenzene-d5	8.950	82	5960	0.336	ng	-0.02
11) 2-Methylnaphthalene-d10	12.202	152	8482	0.262	ng	-0.02
14) 2,4,6-Tribromophenol	15.957	330	1206	0.271	ng	-0.01
15) 2-Fluorobiphenyl	13.083	172	17361	0.362	ng	-0.01
27) Fluoranthene-d10	19.239	212	19490	0.339	ng	-0.01
31) Terphenyl-d14	19.838	244	16058	0.594	ng	-0.01
Target Compounds						
2) 1,4-Dioxane	3.290	88	1054	0.123	ng	# 56
34) Bis(2-ethylhexyl)phtha...	21.313	149	1109	0.059	ng	# 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN100625\
Data File : BN038013.D
Acq On : 06 Oct 2025 17:22
Operator : RC/JU
Sample : Q3194-08
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
TT188S1-20250922

Quant Time: Oct 06 17:47:13 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN092325.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Sep 24 11:00:01 2025
Response via : Initial Calibration

