




Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN100722\
 Data File : BN022029.D
 Acq On : 07 Oct 2022 08:40
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Quant Time: Oct 10 04:25:22 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Oct 10 04:25:17 2022
 Response via : Initial Calibration



TIC: BN022029.D\data.ms

(70) Pentachlorophenol (C)




17.051min (0.000) 20631.17 ng

response 326407

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	62.20	61.64
264.00	61.60	60.52
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN100722\
Data File : BN022029.D
Acq On : 07 Oct 2022 08:40
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Oct 10 04:25:22 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Oct 10 04:25:17 2022
Response via : Initial Calibration



TIC: BN022029.D\data.ms

(77) Benzidine

19.663min (0.000) 0.00 ng

response 1758012

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	15.50	13.02
183.00	13.20	11.51
0.00	0.00	0.00