

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN100722\
 Data File : BN022030.D
 Acq On : 07 Oct 2022 09:16
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Oct 10 15:47:55 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN100322.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Oct 04 01:32:21 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	131	0.00
2	1,4-Dioxane	0.400	0.353	11.8	121	0.00
3	n-Nitrosodimethylamine	0.400	0.362	9.5	130	0.00
4 S	2-Fluorophenol	0.400	0.370	7.5	132	-0.01
5 S	Phenol-d6	0.400	0.371	7.3	135	-0.01
6	bis(2-Chloroethyl)ether	0.400	0.380	5.0	136	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	139	-0.01
8 S	Nitrobenzene-d5	0.400	0.362	9.5	138	-0.01
9	Naphthalene	0.400	0.369	7.8	138	-0.01
10	Hexachlorobutadiene	0.400	0.373	6.8	138	-0.01
11 SURR	2-Methylnaphthalene-d10	0.400	0.351	12.3	132	-0.01
12	2-Methylnaphthalene	0.400	0.386	3.5	160	-0.02
13 I	Acenaphthene-d10	0.400	0.400	0.0	141	-0.01
14 S	2,4,6-Tribromophenol	0.400	0.358	10.5	148	0.00
15 S	2-Fluorobiphenyl	0.400	0.381	4.8	140	-0.01
16	Acenaphthylene	0.400	0.347	13.3	138	-0.01
17	Acenaphthene	0.400	0.373	6.8	141	-0.01
18	Fluorene	0.400	0.401	-0.3	156	-0.01
19 I	Phenanthrene-d10	0.400	0.400	0.0	139	-0.01
20	4,6-Dinitro-2-methylphenol	0.400	0.437	-9.2	155	-0.01
21	4-Bromophenyl-phenylether	0.400	0.368	8.0	139	-0.01
22	Hexachlorobenzene	0.400	0.366	8.5	137	0.00
23	Atrazine	0.400	0.339	15.3	137	-0.01
24	Pentachlorophenol	0.400	0.352	12.0	153	-0.01
25	Phenanthrene	0.400	0.370	7.5	139	-0.01
26	Anthracene	0.400	0.352	12.0	139	0.00
27 SURR	Fluoranthene-d10	0.400	0.360	10.0	139	0.00
28	Fluoranthene	0.400	0.359	10.3	139	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	133	0.00
30	Pyrene	0.400	0.385	3.8	136	0.00
31 S	Terphenyl-d14	0.400	0.384	4.0	135	0.00
32	Benzo(a)anthracene	0.400	0.354	11.5	130	0.00
33	Chrysene	0.400	0.377	5.8	131	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.351	12.3	133	0.00
35 I	Perylene-d12	0.400	0.400	0.0	125	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.355	11.3	124	0.00
37	Benzo(b)fluoranthene	0.400	0.380	5.0	132	0.00
38	Benzo(k)fluoranthene	0.400	0.371	7.3	125	0.00
39 C	Benzo(a)pyrene	0.400	0.365	8.8	127	0.00
40	Dibenzo(a,h)anthracene	0.400	0.350	12.5	121	-0.01
41	Benzo(g,h,i)perylene	0.400	0.339	15.3	114	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4