

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN100722\
 Data File : BN022031.D
 Acq On : 07 Oct 2022 09:53
 Operator : CG/JU
 Sample : PB147964BL
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 PB147964BL

Manual Integrations
 APPROVED

Reviewed By :Christian Giraldo 10/10/2022
 Supervised By :Jagrut Upadhyay 10/10/2022

Quant Time: Oct 10 15:48:34 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN100322.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Oct 04 01:32:21 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.978	152	6418	0.400	ng	0.00
7) Naphthalene-d8	10.806	136	19542	0.400	ng	-0.01
13) Acenaphthene-d10	14.643	164	9847	0.400	ng	-0.01
19) Phenanthrene-d10	17.400	188	19401	0.400	ng	#-0.01
29) Chrysene-d12	21.600	240	13239	0.400	ng	# 0.00
35) Perylene-d12	24.081	264	11045	0.400	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.508	112	5976	0.402	ng	0.00
5) Phenol-d6	7.141	99	7339	0.380	ng	0.00
8) Nitrobenzene-d5	9.161	82	5273	0.336	ng	0.00
11) 2-Methylnaphthalene-d10	12.402	152	12791	0.356	ng	0.00
14) 2,4,6-Tribromophenol	16.146	330	945	0.252	ng	0.00
15) 2-Fluorobiphenyl	13.274	172	15002	0.349	ng	-0.01
27) Fluoranthene-d10	19.437	212	17572	0.341	ng	0.00
31) Terphenyl-d14	20.033	244	10020	0.376	ng	0.00
Target Compounds						
20) 4,6-Dinitro-2-methylph...	15.811	198	4	0.193	ng	# 1
34) Bis(2-ethylhexyl)phtha...	21.501	149	528	0.022	ng	# 91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN100722\
Data File : BN022031.D
Acq On : 07 Oct 2022 09:53
Operator : CG/JU
Sample : PB147964BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
PB147964BL

Quant Time: Oct 10 15:48:34 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN100322.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Oct 04 01:32:21 2022
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By : Christian Giraldo 10/10/2022
Supervised By : Jagrut Upadhyay 10/10/2022

