

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN100724\
 Data File : BN034443.D
 Acq On : 07 Oct 2024 18:15
 Operator : RC/JU
 Sample : PB163651BL
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SBLK651

Quant Time: Oct 08 00:38:09 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN091124.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Sep 30 14:14:38 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.375	152	230488	20.000	ng/ul	-0.01
20) Naphthalene-d8	10.134	136	904640	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.028	164	541902	20.000	ng/ul	-0.01
64) Phenanthrene-d10	16.792	188	1052923	20.000	ng/ul	0.00
79) Chrysene-d12	21.028	240	863028	20.000	ng/ul	0.00
88) Perylene-d12	23.733	264	990858	20.000	ng/ul	0.00

System Monitoring Compounds						
3) 1,4-Dioxane-d8	2.999	96	28089	4.973	ng/uL	0.00
4) Pyridine-d5	3.381	84	347646	20.707	ng/ul	0.00
7) Phenol-d5	6.581	99	445157	21.762	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	6.734	67	271609	21.400	ng/ul	0.00
11) 2-Chlorophenol-d4	6.917	132	401042	24.907	ng/ul	-0.01
15) 4-Methylphenol-d8	8.093	113	359116	21.336	ng/ul	-0.01
21) Nitrobenzene-d5	8.522	128	186209	25.825	ng/ul	0.00
24) 2-Nitrophenol-d4	9.234	143	202468	27.307	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	9.769	165	349059	24.997	ng/ul	0.00
31) 4-Chloroaniline-d4	10.281	131	492595	23.301	ng/ul	-0.01
46) Dimethylphthalate-d6	13.457	166	1003687	24.784	ng/ul	-0.01
49) Acenaphthylene-d8	13.716	160	1173107	25.749	ng/ul	-0.01
54) 4-Nitrophenol-d4	14.269	143	128528	15.464	ng/ul	0.00
60) Fluorene-d10	15.034	176	852279	24.832	ng/ul	-0.01
65) 4,6-Dinitro-2-methylph...	15.169	200	118584	18.889	ng/ul	0.00
73) Anthracene-d10	16.887	188	1255709	25.366	ng/ul	-0.01
81) Pyrene-d10	19.198	212	1300480	27.045	ng/ul	-0.01
92) Benzo(a)pyrene-d12	23.551	264	1347519	26.475	ng/ul	0.00

Target Compounds	Qvalue
-----	-----

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN100724\
Data File : BN034443.D
Acq On : 07 Oct 2024 18:15
Operator : RC/JU
Sample : PB163651BL
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SBLK651

Quant Time: Oct 08 00:38:09 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN091124.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Mon Sep 30 14:14:38 2024
Response via : Initial Calibration

