

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN100821\
 Data File : BN016843.D
 Acq On : 08 Oct 2021 14:18
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Oct 08 15:47:53 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN100721.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Oct 08 15:46:07 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	98	0.00
2	1,4-Dioxane	0.400	0.337	15.8	86	0.00
3	n-Nitrosodimethylamine	0.400	0.383	4.3	98	0.00
4 S	2-Fluorophenol	0.400	0.385	3.8	97	0.00
5 S	Phenol-d6	0.400	0.383	4.3	97	0.00
6	bis(2-Chloroethyl)ether	0.400	0.390	2.5	95	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	98	0.00
8 S	Nitrobenzene-d5	0.400	0.356	11.0	99	0.00
9	Naphthalene	0.400	0.377	5.8	95	0.00
10	Hexachlorobutadiene	0.400	0.384	4.0	96	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.377	5.8	94	0.00
12	2-Methylnaphthalene	0.400	0.383	4.3	95	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	99	0.00
14 S	2,4,6-Tribromophenol	0.400	0.350	12.5	98	0.00
15 S	2-Fluorobiphenyl	0.400	0.378	5.5	95	0.00
16	Acenaphthylene	0.400	0.368	8.0	96	0.00
17	Acenaphthene	0.400	0.381	4.8	96	0.00
18	Fluorene	0.400	0.383	4.3	95	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	99	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.281	29.8#	45	0.00
21	4-Bromophenyl-phenylether	0.400	0.370	7.5	95	0.00
22	Hexachlorobenzene	0.400	0.380	5.0	95	0.00
23	Atrazine	0.400	0.325	18.8	93	0.00
24	Pentachlorophenol	0.400	0.372	7.0	107	0.00
25	Phenanthrene	0.400	0.382	4.5	96	0.00
26	Anthracene	0.400	0.364	9.0	95	0.00
27 SURR	Fluoranthene-d10	0.400	0.374	6.5	96	0.00
28	Fluoranthene	0.400	0.388	3.0	97	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	102	0.00
30	Pyrene	0.400	0.402	-0.5	98	0.00
31 S	Terphenyl-d14	0.400	0.394	1.5	98	0.00
32	Benzo(a)anthracene	0.400	0.390	2.5	100	0.00
33	Chrysene	0.400	0.412	-3.0	98	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.339	15.3	87	0.00
35 I	Perylene-d12	0.400	0.400	0.0	102	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.393	1.8	94	0.00
37	Benzo(b)fluoranthene	0.400	0.400	0.0	95	0.00
38	Benzo(k)fluoranthene	0.400	0.417	-4.2	100	0.00
39 C	Benzo(a)pyrene	0.400	0.402	-0.5	97	0.00
40	Dibenzo(a,h)anthracene	0.400	0.393	1.8	95	0.00
41	Benzo(g,h,i)perylene	0.400	0.389	2.8	94	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

LabSampleId :

SSTDCCC0.4