

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN100821\
 Data File : BN016860.D
 Acq On : 09 Oct 2021 00:48
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4EC

Quant Time: Oct 09 05:42:53 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN100721.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Oct 08 15:46:07 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.615	152	21973	0.400	ng	0.00
7) Naphthalene-d8	10.381	136	97745	0.400	ng	# 0.00
13) Acenaphthene-d10	14.243	164	51416	0.400	ng	0.00
19) Phenanthrene-d10	16.993	188	99482	0.400	ng	#-0.01
29) Chrysene-d12	21.206	240	106543	0.400	ng	0.00
35) Perylene-d12	23.447	264	114595	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.226	112	21628	0.394	ng	0.00
5) Phenol-d6	6.794	99	24692	0.396	ng	0.00
8) Nitrobenzene-d5	8.759	82	25351	0.370	ng	0.00
11) 2-Methylnaphthalene-d10	11.976	152	54899	0.377	ng	0.00
14) 2,4,6-Tribromophenol	15.740	330	10046	0.378	ng	-0.01
15) 2-Fluorobiphenyl	12.873	172	77784	0.375	ng	0.00
27) Fluoranthene-d10	19.043	212	105220	0.381	ng	0.00
31) Terphenyl-d14	19.654	244	92002	0.396	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.226	88	3885	0.395	ng	# 65
3) n-Nitrosodimethylamine	3.548	42	6901	0.400	ng	99
6) bis(2-Chloroethyl)ether	7.052	93	23880	0.393	ng	100
9) Naphthalene	10.432	128	99877	0.375	ng	100
10) Hexachlorobutadiene	10.723	225	19517	0.384	ng	# 100
12) 2-Methylnaphthalene	12.052	142	65304	0.383	ng	100
16) Acenaphthylene	13.964	152	84622	0.373	ng	100
17) Acenaphthene	14.307	154	56571	0.375	ng	99
18) Fluorene	15.296	166	69639	0.380	ng	99
20) 4,6-Dinitro-2-methylph...	15.385	198	2079	0.284	ng	# 69
21) 4-Bromophenyl-phenylether	16.202	248	23077	0.372	ng	# 90
22) Hexachlorobenzene	16.312	284	26049	0.368	ng	99
23) Atrazine	16.482	200	17734	0.360	ng	98
24) Pentachlorophenol	16.652	266	10432	0.415	ng	99
25) Phenanthrene	17.042	178	113170	0.383	ng	100
26) Anthracene	17.127	178	98270	0.374	ng	100
28) Fluoranthene	19.073	202	129859	0.391	ng	100
30) Pyrene	19.435	202	133038	0.406	ng	100
32) Benzo(a)anthracene	21.195	228	133446	0.405	ng	97
33) Chrysene	21.238	228	136945	0.408	ng	99
34) Bis(2-ethylhexyl)phtha...	21.142	149	72601	0.434	ng	100
36) Indeno(1,2,3-cd)pyrene	25.709	276	170739	0.393	ng	98
37) Benzo(b)fluoranthene	22.772	252	146492	0.404	ng	99
38) Benzo(k)fluoranthene	22.820	252	146528	0.415	ng	99
39) Benzo(a)pyrene	23.347	252	135702	0.409	ng	99
40) Dibenzo(a,h)anthracene	25.733	278	140427	0.395	ng	99
41) Benzo(g,h,i)perylene	26.397	276	143102	0.379	ng	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN100821\
Data File : BN016860.D
Acq On : 09 Oct 2021 00:48
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 19 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC

Quant Time: Oct 09 05:42:53 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN100721.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 08 15:46:07 2021
Response via : Initial Calibration









#1
 1,4-Dichlorobenzene-d4
 Concen: 0.400 ng
 RT: 7.615 min Scan# 492
 Delta R.T. 0.001 min
 Lab File: BN016860.D
 Acq: 09 Oct 2021 00:48




Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4EC


Tgt Ion:152 Resp: 21973
 Ion Ratio Lower Upper
 152 100
 150 155.2 123.3 184.9
 115 58.8 47.0 70.6



#2
 1,4-Dioxane
 Concen: 0.395 ng
 RT: 3.226 min Scan# 16
 Delta R.T. 0.009 min
 Lab File: BN016860.D
 Acq: 09 Oct 2021 00:48

Tgt Ion: 88 Resp: 3885
 Ion Ratio Lower Upper
 88 100
 43 25.7 61.7 92.5#
 58 67.9 62.2 93.2









#3
n-Nitrosodimethylamine
Concen: 0.400 ng
RT: 3.548 min Scan# 51
Delta R.T. -0.018 min
Lab File: BN016860.D
Acq: 09 Oct 2021 00:48




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion: 42 Resp: 6901
Ion Ratio Lower Upper
42 100
74 153.2 124.0 186.0
44 6.9 5.6 8.4



#4
2-Fluorophenol
Concen: 0.394 ng
RT: 5.226 min Scan# 233
Delta R.T. -0.009 min
Lab File: BN016860.D
Acq: 09 Oct 2021 00:48

Tgt Ion: 112 Resp: 21628
Ion Ratio Lower Upper
112 100
64 55.1 42.9 64.3
63 27.9 21.5 32.3









#5
Phenol-d6
Concen: 0.396 ng
RT: 6.794 min Scan# 403
Delta R.T. 0.000 min
Lab File: BN016860.D
Acq: 09 Oct 2021 00:48




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:	99	Resp:	24692
Ion	Ratio	Lower	Upper
99	100		
42	16.8	13.4	20.2
71	30.2	24.0	36.0



#6
bis(2-Chloroethyl)ether
Concen: 0.393 ng
RT: 7.052 min Scan# 431
Delta R.T. 0.000 min
Lab File: BN016860.D
Acq: 09 Oct 2021 00:48

Tgt Ion:	93	Resp:	23880
Ion	Ratio	Lower	Upper
93	100		
63	74.2	59.0	88.6
95	32.9	26.4	39.6









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.381 min Scan# 731
Delta R.T. 0.000 min
Lab File: BN016860.D
Acq: 09 Oct 2021 00:48




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:	136	Resp:	97745
Ion	Ratio	Lower	Upper
136	100		
137	11.0	8.9	13.3
54	5.8	6.5	9.7#
68	5.0	5.3	7.9#



#8
Nitrobenzene-d5
Concen: 0.370 ng
RT: 8.759 min Scan# 603
Delta R.T. 0.000 min
Lab File: BN016860.D
Acq: 09 Oct 2021 00:48

Tgt Ion:	82	Resp:	25351
Ion	Ratio	Lower	Upper
82	100		
128	46.6	33.8	50.6
54	44.6	38.5	57.7









#9
Naphthalene
Concen: 0.375 ng
RT: 10.432 min Scan# 735
Delta R.T. 0.000 min
Lab File: BN016860.D
Acq: 09 Oct 2021 00:48




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:128 Resp: 99877
Ion Ratio Lower Upper
128 100
129 11.0 8.7 13.1
127 12.8 10.3 15.5



#10
Hexachlorobutadiene
Concen: 0.384 ng
RT: 10.723 min Scan# 758
Delta R.T. 0.000 min
Lab File: BN016860.D
Acq: 09 Oct 2021 00:48

Tgt Ion:225 Resp: 19517
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 64.1 51.1 76.7









#11
2-Methylnaphthalene-d10
Concen: 0.377 ng
RT: 11.976 min Scan# 925
Delta R.T. -0.004 min
Lab File: BN016860.D
Acq: 09 Oct 2021 00:48




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:152 Resp: 54899
Ion Ratio Lower Upper
152 100
151 20.4 16.2 24.4



#12
2-Methylnaphthalene
Concen: 0.383 ng
RT: 12.052 min Scan# 942
Delta R.T. 0.000 min
Lab File: BN016860.D
Acq: 09 Oct 2021 00:48


Tgt Ion:142 Resp: 65304
Ion Ratio Lower Upper
142 100
141 87.8 70.6 106.0
115 32.4 26.1 39.1








#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.243 min Scan# 1181
Delta R.T. 0.000 min
Lab File: BN016860.D
Acq: 09 Oct 2021 00:48


Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC





Tgt Ion:164 Resp: 51416
Ion Ratio Lower Upper
164 100
162 101.8 82.6 124.0
160 45.7 38.2 57.2




#14
2,4,6-Tribromophenol
Concen: 0.378 ng
RT: 15.740 min Scan# 1299
Delta R.T. -0.012 min
Lab File: BN016860.D
Acq: 09 Oct 2021 00:48



Tgt Ion:330 Resp: 10046
Ion Ratio Lower Upper
330 100
332 0.0 0.0 0.0
141 29.8 24.2 36.2









#15
2-Fluorobiphenyl
Concen: 0.375 ng
RT: 12.873 min Scan# 1073
Delta R.T. 0.000 min
Lab File: BN016860.D
Acq: 09 Oct 2021 00:48




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:172 Resp: 77784
Ion Ratio Lower Upper
172 100
171 33.2 26.9 40.3
170 21.3 17.4 26.0



#16
Acenaphthylene
Concen: 0.373 ng
RT: 13.964 min Scan# 1159
Delta R.T. 0.000 min
Lab File: BN016860.D
Acq: 09 Oct 2021 00:48

Tgt Ion:152 Resp: 84622
Ion Ratio Lower Upper
152 100
151 19.5 15.5 23.3
153 13.0 10.4 15.6









#17
Acenaphthene
Concen: 0.375 ng
RT: 14.307 min Scan# 1186
Delta R.T. 0.000 min
Lab File: BN016860.D
Acq: 09 Oct 2021 00:48




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:154 Resp: 56571
Ion Ratio Lower Upper
154 100
153 113.7 90.1 135.1
152 55.2 43.6 65.4



#18
Fluorene
Concen: 0.380 ng
RT: 15.296 min Scan# 1264
Delta R.T. 0.000 min
Lab File: BN016860.D
Acq: 09 Oct 2021 00:48

Tgt Ion:166 Resp: 69639
Ion Ratio Lower Upper
166 100
165 98.2 78.1 117.1
167 13.4 10.7 16.1









#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 16.993 min Scan# 1401
Delta R.T. -0.012 min
Lab File: BN016860.D
Acq: 09 Oct 2021 00:48




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:188 Resp: 99482
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 13.0 5.8 8.8#



#20
4,6-Dinitro-2-methylphenol
Concen: 0.284 ng
RT: 15.385 min Scan# 1271
Delta R.T. 0.000 min
Lab File: BN016860.D
Acq: 09 Oct 2021 00:48

Tgt Ion:198 Resp: 2079
Ion Ratio Lower Upper
198 100
182 15.1 26.2 39.2#
77 0.0 0.0 0.0









#21
4-Bromophenyl-phenylether
Concen: 0.372 ng
RT: 16.202 min Scan# 1336
Delta R.T. 0.000 min
Lab File: BN016860.D
Acq: 09 Oct 2021 00:48




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:248 Resp: 23077
Ion Ratio Lower Upper
248 100
250 99.6 76.2 114.4
141 42.8 45.6 68.4#



#22
Hexachlorobenzene
Concen: 0.368 ng
RT: 16.312 min Scan# 1345
Delta R.T. 0.000 min
Lab File: BN016860.D
Acq: 09 Oct 2021 00:48

Tgt Ion:284 Resp: 26049
Ion Ratio Lower Upper
284 100
142 31.8 25.8 38.6
249 29.5 23.3 34.9









#23
Atrazine
Concen: 0.360 ng
RT: 16.482 min Scan# 1359
Delta R.T. 0.000 min
Lab File: BN016860.D
Acq: 09 Oct 2021 00:48




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:200 Resp: 17734
Ion Ratio Lower Upper
200 100
173 23.2 20.1 30.1
215 49.7 38.8 58.2



#24
Pentachlorophenol
Concen: 0.415 ng
RT: 16.652 min Scan# 1373
Delta R.T. 0.000 min
Lab File: BN016860.D
Acq: 09 Oct 2021 00:48

Tgt Ion:266 Resp: 10432
Ion Ratio Lower Upper
266 100
264 62.7 49.3 73.9
268 63.9 51.8 77.8









#25
Phenanthrene
Concen: 0.383 ng
RT: 17.042 min Scan# 1405
Delta R.T. 0.000 min
Lab File: BN016860.D
Acq: 09 Oct 2021 00:48




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:178 Resp: 113170
Ion Ratio Lower Upper
178 100
176 18.8 15.0 22.6
179 15.5 12.4 18.6



#26
Anthracene
Concen: 0.374 ng
RT: 17.127 min Scan# 1412
Delta R.T. 0.000 min
Lab File: BN016860.D
Acq: 09 Oct 2021 00:48

Tgt Ion:178 Resp: 98270
Ion Ratio Lower Upper
178 100
176 18.2 14.5 21.7
179 15.3 12.2 18.4









#27
Fluoranthene-d10
Concen: 0.381 ng
RT: 19.043 min Scan# 1692
Delta R.T. 0.000 min
Lab File: BN016860.D
Acq: 09 Oct 2021 00:48




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:212 Resp: 105220
Ion Ratio Lower Upper
212 100
106 13.4 10.6 16.0
104 7.3 5.8 8.8



#28
Fluoranthene
Concen: 0.391 ng
RT: 19.073 min Scan# 1698
Delta R.T. 0.000 min
Lab File: BN016860.D
Acq: 09 Oct 2021 00:48

Tgt Ion:202 Resp: 129859
Ion Ratio Lower Upper
202 100
101 11.1 8.9 13.3
203 17.4 13.8 20.8









#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.206 min Scan# 1999
Delta R.T. 0.000 min
Lab File: BN016860.D
Acq: 09 Oct 2021 00:48




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:240 Resp: 106543
Ion Ratio Lower Upper
240 100
120 9.8 6.6 10.0
236 26.2 20.7 31.1



#30
Pyrene
Concen: 0.406 ng
RT: 19.435 min Scan# 1771
Delta R.T. 0.000 min
Lab File: BN016860.D
Acq: 09 Oct 2021 00:48

Tgt Ion:202 Resp: 133038
Ion Ratio Lower Upper
202 100
200 20.6 16.5 24.7
203 17.4 13.9 20.9









#31
Terphenyl-d14
Concen: 0.396 ng
RT: 19.654 min Scan# 1815
Delta R.T. -0.005 min
Lab File: BN016860.D
Acq: 09 Oct 2021 00:48




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:244	Resp:	92002
Ion Ratio	Lower	Upper
244	100	
212	7.1	5.8 8.6
122	11.8	8.9 13.3



#32
Benzo(a)anthracene
Concen: 0.405 ng
RT: 21.195 min Scan# 1998
Delta R.T. 0.000 min
Lab File: BN016860.D
Acq: 09 Oct 2021 00:48

Tgt Ion:228	Resp:	133446
Ion Ratio	Lower	Upper
228	100	
226	25.7	22.3 33.5
229	20.1	15.4 23.0









#33
Chrysene
Concen: 0.408 ng
RT: 21.238 min Scan# 2002
Delta R.T. -0.010 min
Lab File: BN016860.D
Acq: 09 Oct 2021 00:48




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:228	Resp: 136945
Ion Ratio	Lower Upper
228	100
226	30.7 24.2 36.4
229	19.1 15.4 23.0



#34
Bis(2-ethylhexyl)phthalate
Concen: 0.434 ng
RT: 21.142 min Scan# 1993
Delta R.T. -0.010 min
Lab File: BN016860.D
Acq: 09 Oct 2021 00:48


Tgt Ion:149	Resp: 72601
Ion Ratio	Lower Upper
149	100
167	28.0 22.2 33.4
279	4.4 3.4 5.2








#35
 Perylene-d12
 Concen: 0.400 ng
 RT: 23.447 min Scan# 2416
 Delta R.T. -0.007 min
 Lab File: BN016860.D
 Acq: 09 Oct 2021 00:48

Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4EC







Tgt Ion:264 Resp: 114595
 Ion Ratio Lower Upper
 264 100
 260 24.5 19.8 29.6
 265 22.9 20.3 30.5



#36
 Indeno(1,2,3-cd)pyrene
 Concen: 0.393 ng
 RT: 25.709 min Scan# 3077
 Delta R.T. -0.007 min
 Lab File: BN016860.D
 Acq: 09 Oct 2021 00:48

Tgt Ion:276 Resp: 170739
 Ion Ratio Lower Upper
 276 100
 138 25.3 19.2 28.8
 277 25.2 20.0 30.0









#37
Benzo(b)fluoranthene
Concen: 0.404 ng
RT: 22.772 min Scan# 2219
Delta R.T. -0.007 min
Lab File: BN016860.D
Acq: 09 Oct 2021 00:48




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:252 Resp: 146492
Ion Ratio Lower Upper
252 100
253 22.1 17.8 26.8
125 9.9 8.3 12.5



#38
Benzo(k)fluoranthene
Concen: 0.415 ng
RT: 22.820 min Scan# 2233
Delta R.T. -0.003 min
Lab File: BN016860.D
Acq: 09 Oct 2021 00:48


Tgt Ion:252 Resp: 146528
Ion Ratio Lower Upper
252 100
253 22.6 17.8 26.8
125 11.4 8.5 12.7








#39
Benzo(a)pyrene
Concen: 0.409 ng
RT: 23.347 min Scan# 2387
Delta R.T. -0.007 min
Lab File: BN016860.D
Acq: 09 Oct 2021 00:48


Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC





Tgt Ion:252 Resp: 135702
Ion Ratio Lower Upper
252 100
253 22.3 17.9 26.9
125 11.0 9.4 14.2




#40
Dibenzo(a,h)anthracene
Concen: 0.395 ng
RT: 25.733 min Scan# 3084
Delta R.T. -0.007 min
Lab File: BN016860.D
Acq: 09 Oct 2021 00:48




Tgt Ion:278 Resp: 140427
Ion Ratio Lower Upper
278 100
139 16.9 13.1 19.7
279 23.8 19.3 28.9





#41
Benzo(g,h,i)perylene
Concen: 0.379 ng
RT: 26.397 min Scan# 3278
Delta R.T. -0.010 min
Lab File: BN016860.D
Acq: 09 Oct 2021 00:48

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC



Tgt Ion:276 Resp: 143102
Ion Ratio Lower Upper
276 100
277 23.9 19.0 28.4
138 20.9 17.0 25.6

