

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN101024\
 Data File : BN034533.D
 Acq On : 10 Oct 2024 22:19
 Operator : RC/JU
 Sample : PB163828BSD
 Misc :
 ALS Vial : 109 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 PB163828BSD

Quant Time: Oct 11 00:23:42 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN091924.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Oct 11 00:21:08 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.373	152	5729	0.400	ng	0.00
7) Naphthalene-d8	10.127	136	16345	0.400	ng	0.00
13) Acenaphthene-d10	14.015	164	6026	0.400	ng	0.00
19) Phenanthrene-d10	16.778	188	14532	0.400	ng	# 0.00
29) Chrysene-d12	21.006	240	4039	0.400	ng	0.00
35) Perylene-d12	23.117	264	352	0.400	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.026	112	5262	0.324	ng	0.01
5) Phenol-d6	6.586	99	4508	0.238	ng	0.01
8) Nitrobenzene-d5	8.515	82	4018	0.321	ng	0.00
11) 2-Methylnaphthalene-d10	11.727	152	8534	0.354	ng	0.00
14) 2,4,6-Tribromophenol	15.537	330	1273	0.466	ng	0.00
15) 2-Fluorobiphenyl	12.636	172	11806	0.482	ng	0.00
27) Fluoranthene-d10	18.827	212	9246	0.252	ng	0.00
31) Terphenyl-d14	19.450	244	8541	1.059	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.032	88	7220	1.008	ng	# 91
3) n-Nitrosodimethylamine	3.328	42	2101	0.258	ng	# 85
6) bis(2-Chloroethyl)ether	6.824	93	5468	0.327	ng	92
9) Naphthalene	10.169	128	16001	0.357	ng	100
10) Hexachlorobutadiene	10.458	225	3409	0.403	ng	# 98
12) 2-Methylnaphthalene	11.803	142	10280	0.352	ng	98
16) Acenaphthylene	13.737	152	7213	0.280	ng	99
17) Acenaphthene	14.079	154	6905	0.373	ng	99
18) Fluorene	15.084	166	10804	0.445	ng	100
20) 4,6-Dinitro-2-methylph...	16.319	198	5	0.140	ng	# 1
21) 4-Bromophenyl-phenylether	15.984	248	3224	0.395	ng	# 76
22) Hexachlorobenzene	16.095	284	4594	0.502	ng	# 89
24) Pentachlorophenol	16.443	266	2996	0.989	ng	94
25) Phenanthrene	16.828	178	15913	0.381	ng	100
26) Anthracene	16.915	178	10019	0.294	ng	99
28) Fluoranthene	18.860	202	11975	0.248	ng	99
30) Pyrene	19.222	202	5001	0.293	ng	100
32) Benzo(a)anthracene	21.042	228	8058	0.631	ng	95
33) Chrysene	21.042	228	9261	0.608	ng	98
36) Indeno(1,2,3-cd)pyrene	25.175	276	3927	2.604	ng	98
37) Benzo(b)fluoranthene	22.494	252	5737	4.161	ng	# 80
38) Benzo(k)fluoranthene	22.538	252	4098	2.831	ng	# 71
39) Benzo(a)pyrene	23.032	252	1937	1.723	ng	# 31
40) Dibenzo(a,h)anthracene	25.192	278	5432	4.636	ng	# 87
41) Benzo(g,h,i)perylene	25.798	276	3994	3.035	ng	# 91


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN101024\
Data File : BN034533.D
Acq On : 10 Oct 2024 22:19
Operator : RC/JU
Sample : PB163828BSD
Misc :
ALS Vial : 109 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
PB163828BSD

Quant Time: Oct 11 00:23:42 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN091924.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 11 00:21:08 2024
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.373 min Scan# 61
Delta R.T. -0.000 min
Lab File: BN034533.D
Acq: 10 Oct 2024 22:19




Instrument :
BNA_N
ClientSampleId :
PB163828BSD


Tgt Ion:152 Resp: 5729
Ion Ratio Lower Upper
152 100
150 155.7 124.6 187.0
115 62.8 50.0 75.0



#2
1,4-Dioxane
Concen: 1.008 ng
RT: 3.032 min Scan# 21
Delta R.T. 0.007 min
Lab File: BN034533.D
Acq: 10 Oct 2024 22:19

Tgt Ion: 88 Resp: 7220
Ion Ratio Lower Upper
88 100
43 39.7 25.8 38.8#
58 67.9 58.8 88.2





#3

n-Nitrosodimethylamine

Concen: 0.258 ng

RT: 3.328 min Scan# 61

Delta R.T. 0.000 min

Lab File: BN034533.D

Acq: 10 Oct 2024 22:19

Instrument :

BNA_N

ClientSampleId :

PB163828BSD





Tgt Ion: 42 Resp: 2101

Ion Ratio Lower Upper

42 100

74 105.4 94.6 142.0

44 0.0 12.4 18.6#



#4

2-Fluorophenol

Concen: 0.324 ng

RT: 5.026 min Scan# 297

Delta R.T. 0.014 min

Lab File: BN034533.D

Acq: 10 Oct 2024 22:19




Tgt Ion: 112 Resp: 5262


Ion Ratio Lower Upper

112 100

64 55.0 48.6 72.8

63 29.2 25.6 38.4









#5
Phenol-d6
Concen: 0.238 ng
RT: 6.586 min Scan# 513
Delta R.T. 0.014 min
Lab File: BN034533.D
Acq: 10 Oct 2024 22:19




Instrument :
BNA_N
ClientSampleId :
PB163828BSD


Tgt Ion: 99 Resp: 4508
Ion Ratio Lower Upper
99 100
42 15.6 17.8 26.8#
71 25.6 26.2 39.2#



#6
bis(2-Chloroethyl)ether
Concen: 0.327 ng
RT: 6.824 min Scan# 546
Delta R.T. 0.007 min
Lab File: BN034533.D
Acq: 10 Oct 2024 22:19

Tgt Ion: 93 Resp: 5468
Ion Ratio Lower Upper
93 100
63 74.3 67.3 100.9
95 32.2 26.8 40.2









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.127 min Scan# 915
Delta R.T. 0.000 min
Lab File: BN034533.D
Acq: 10 Oct 2024 22:19




Instrument :
BNA_N
ClientSampleId :
PB163828BSD


Tgt Ion:	136	Resp:	16345
Ion Ratio	Lower	Upper	
136	100		
137	12.1	9.0	13.6
54	6.9	6.8	10.2
68	5.5	5.0	7.6



#8
Nitrobenzene-d5
Concen: 0.321 ng
RT: 8.515 min Scan# 771
Delta R.T. 0.000 min
Lab File: BN034533.D
Acq: 10 Oct 2024 22:19

Tgt Ion:	82	Resp:	4018
Ion Ratio	Lower	Upper	
82	100		
128	42.8	31.4	47.2
54	52.1	47.4	71.0





#9

Naphthalene

Concen: 0.357 ng

RT: 10.169 min Scan# 919

Delta R.T. 0.000 min

Lab File: BN034533.D

Acq: 10 Oct 2024 22:19





Instrument :

BNA_N

ClientSampleId :

PB163828BSD

Tgt Ion:	128	Resp:	16001
Ion	Ratio	Lower	Upper
128	100		
129	11.4	9.2	13.8
127	13.5	10.7	16.1



#10

Hexachlorobutadiene

Concen: 0.403 ng




RT: 10.458 min Scan# 953


Delta R.T. -0.011 min

Lab File: BN034533.D

Acq: 10 Oct 2024 22:19

Tgt Ion:	225	Resp:	3409
Ion	Ratio	Lower	Upper
225	100		
223	0.0	0.0	0.0
227	64.8	50.5	75.7









#11
2-Methylnaphthalene-d10
Concen: 0.354 ng
RT: 11.727 min Scan# 1110
Delta R.T. 0.000 min
Lab File: BN034533.D
Acq: 10 Oct 2024 22:19




Instrument :
BNA_N
ClientSampleId :
PB163828BSD


Tgt Ion:152 Resp: 8534
Ion Ratio Lower Upper
152 100
151 22.8 16.8 25.2



#12
2-Methylnaphthalene
Concen: 0.352 ng
RT: 11.803 min Scan# 1130
Delta R.T. 0.000 min
Lab File: BN034533.D
Acq: 10 Oct 2024 22:19

Tgt Ion:142 Resp: 10280
Ion Ratio Lower Upper
142 100
141 88.1 71.6 107.4
115 36.5 30.0 45.0





#13

Acenaphthene-d10

Concen: 0.400 ng

RT: 14.015 min Scan# 14

Delta R.T. 0.000 min

Lab File: BN034533.D

Acq: 10 Oct 2024 22:19

Instrument :

BNA_N

ClientSampleId :

PB163828BSD





Tgt Ion:164 Resp: 6026

Ion Ratio Lower Upper

164 100

162 108.4 84.2 126.2

160 52.3 41.7 62.5



#14

2,4,6-Tribromophenol

Concen: 0.466 ng

RT: 15.537 min Scan# 1616

Delta R.T. 0.000 min

Lab File: BN034533.D

Acq: 10 Oct 2024 22:19




Tgt Ion:330 Resp: 1273


Ion Ratio Lower Upper

330 100

332 124.9 77.4 116.0#

141 35.6 35.9 53.9#









#15
2-Fluorobiphenyl
Concen: 0.482 ng
RT: 12.636 min Scan# 11
Delta R.T. 0.000 min
Lab File: BN034533.D
Acq: 10 Oct 2024 22:19




Instrument :
BNA_N
ClientSampleId :
PB163828BSD


Tgt Ion:172 Resp: 11806
Ion Ratio Lower Upper
172 100
171 34.9 27.3 40.9
170 22.9 18.1 27.1



#16
Acenaphthylene
Concen: 0.280 ng
RT: 13.737 min Scan# 1448
Delta R.T. 0.000 min
Lab File: BN034533.D
Acq: 10 Oct 2024 22:19

Tgt Ion:152 Resp: 7213
Ion Ratio Lower Upper
152 100
151 20.0 15.6 23.4
153 13.2 10.3 15.5









#17
Acenaphthene
Concen: 0.373 ng
RT: 14.079 min Scan# 1472
Delta R.T. 0.000 min
Lab File: BN034533.D
Acq: 10 Oct 2024 22:19




Instrument :
BNA_N
ClientSampleId :
PB163828BSD


Tgt Ion:154 Resp: 6905
Ion Ratio Lower Upper
154 100
153 115.0 91.6 137.4
152 57.2 47.4 71.2



#18
Fluorene
Concen: 0.445 ng
RT: 15.084 min Scan# 1574
Delta R.T. 0.000 min
Lab File: BN034533.D
Acq: 10 Oct 2024 22:19

Tgt Ion:166 Resp: 10804
Ion Ratio Lower Upper
166 100
165 99.0 79.1 118.7
167 13.4 10.6 15.8





#19

Phenanthrene-d10

Concen: 0.400 ng

RT: 16.778 min Scan# 1

Delta R.T. 0.000 min

Lab File: BN034533.D

Acq: 10 Oct 2024 22:19

Instrument :

BNA_N

ClientSampleId :

PB163828BSD





Tgt Ion:188 Resp: 14532

Ion Ratio Lower Upper

188 100

94 0.0 0.0 0.0

80 12.1 7.4 11.0#



#20

4,6-Dinitro-2-methylphenol

Concen: 0.140 ng

RT: 16.319 min Scan# 1679

Delta R.T. 0.037 min

Lab File: BN034533.D

Acq: 10 Oct 2024 22:19




Tgt Ion:198 Resp: 5


Ion Ratio Lower Upper

198 100

51 305.9 106.4 159.6#

105 188.2 38.5 57.7#









#21
4-Bromophenyl-phenylether
Concen: 0.395 ng
RT: 15.984 min Scan# 1652
Delta R.T. 0.000 min
Lab File: BN034533.D
Acq: 10 Oct 2024 22:19




Instrument :
BNA_N
ClientSampleId :
PB163828BSD


Tgt Ion:248 Resp: 3224
Ion Ratio Lower Upper
248 100
250 94.7 80.5 120.7
141 89.2 37.1 55.7#



#22
Hexachlorobenzene
Concen: 0.502 ng
RT: 16.095 min Scan# 1661
Delta R.T. 0.000 min
Lab File: BN034533.D
Acq: 10 Oct 2024 22:19

Tgt Ion:284 Resp: 4594
Ion Ratio Lower Upper
284 100
142 33.3 34.5 51.7#
249 29.6 25.8 38.6





#24

Pentachlorophenol

Concen: 0.989 ng

RT: 16.443 min Scan# 1

Delta R.T. 0.000 min

Lab File: BN034533.D

Acq: 10 Oct 2024 22:19

Instrument :

BNA_N

ClientSampleId :

PB163828BSD





Tgt Ion:266 Resp: 2996

Ion Ratio Lower Upper

266 100

264 63.3 50.2 75.2

268 72.8 51.0 76.4



#25

Phenanthrene

Concen: 0.381 ng

RT: 16.828 min Scan# 1720

Delta R.T. 0.000 min

Lab File: BN034533.D

Acq: 10 Oct 2024 22:19




Tgt Ion:178 Resp: 15913


Ion Ratio Lower Upper

178 100

176 18.9 15.3 22.9

179 15.3 12.1 18.1





#26

Anthracene

Concen: 0.294 ng

RT: 16.915 min Scan# 11

Delta R.T. 0.000 min

Lab File: BN034533.D

Acq: 10 Oct 2024 22:19

Instrument :

BNA_N

ClientSampleId :

PB163828BSD





Tgt Ion:178 Resp: 10019

Ion Ratio Lower Upper

178 100

176 18.4 15.0 22.6

179 15.1 12.2 18.4



#27

Fluoranthene-d10

Concen: 0.252 ng

RT: 18.827 min Scan# 1923

Delta R.T. 0.000 min

Lab File: BN034533.D

Acq: 10 Oct 2024 22:19




Tgt Ion:212 Resp: 9246


Ion Ratio Lower Upper

212 100

106 15.0 13.4 20.2

104 8.6 7.8 11.6









#28
Fluoranthene
Concen: 0.248 ng
RT: 18.860 min Scan# 1930
Delta R.T. 0.005 min
Lab File: BN034533.D
Acq: 10 Oct 2024 22:19




Instrument :
BNA_N
ClientSampleId :
PB163828BSD


Tgt Ion:	202	Resp:	11975
Ion Ratio	Lower	Upper	
202	100		
101	12.0	10.1	15.1
203	17.0	13.6	20.4



#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.006 min Scan# 2392
Delta R.T. 0.000 min
Lab File: BN034533.D
Acq: 10 Oct 2024 22:19

Tgt Ion:	240	Resp:	4039
Ion Ratio	Lower	Upper	
240	100		
120	19.1	13.5	20.3
236	33.2	23.4	35.0









#30
Pyrene
Concen: 0.293 ng
RT: 19.222 min Scan# 2008
Delta R.T. 0.000 min
Lab File: BN034533.D
Acq: 10 Oct 2024 22:19




Instrument :
BNA_N
ClientSampleId :
PB163828BSD


Tgt Ion:	202	Resp:	5001
Ion	Ratio	Lower	Upper
202	100		
200	20.9	16.6	25.0
203	17.7	14.3	21.5



#31
Terphenyl-d14
Concen: 1.059 ng
RT: 19.450 min Scan# 2057
Delta R.T. 0.000 min
Lab File: BN034533.D
Acq: 10 Oct 2024 22:19

Tgt Ion:	244	Resp:	8541
Ion	Ratio	Lower	Upper
244	100		
212	9.3	7.8	11.6
122	15.3	14.8	22.2





#32

Benzo(a)anthracene

Concen: 0.631 ng

RT: 21.042 min Scan# 21

Delta R.T. 0.036 min

Lab File: BN034533.D

Acq: 10 Oct 2024 22:19

Instrument :

BNA_N

ClientSampleId :

PB163828BSD





Tgt Ion:228 Resp: 8058

Ion Ratio Lower Upper

228 100

226 31.4 22.3 33.5

229 20.4 15.7 23.5



#33

Chrysene

Concen: 0.608 ng

RT: 21.042 min Scan# 2396

Delta R.T. 0.000 min

Lab File: BN034533.D

Acq: 10 Oct 2024 22:19




Tgt Ion:228 Resp: 9261


Ion Ratio Lower Upper

228 100

226 31.4 24.6 37.0

229 20.4 15.5 23.3





#35

Perylene-d12

Concen: 0.400 ng

RT: 23.117 min Scan# 21

Delta R.T. 0.003 min

Lab File: BN034533.D

Acq: 10 Oct 2024 22:19

Instrument :

BNA_N

ClientSampleId :

PB163828BSD





Tgt Ion:264 Resp: 352

Ion Ratio Lower Upper

264 100

260 68.9 21.7 32.5#

265 1261.3 52.1 78.1#



#36

Indeno(1,2,3-cd)pyrene

Concen: 2.604 ng

RT: 25.175 min Scan# 3494

Delta R.T. 0.006 min

Lab File: BN034533.D

Acq: 10 Oct 2024 22:19




Tgt Ion:276 Resp: 3927


Ion Ratio Lower Upper

276 100

138 30.7 25.9 38.9

277 25.4 19.7 29.5









#37
Benzo(b)fluoranthene
Concen: 4.161 ng
RT: 22.494 min Scan# 21
Delta R.T. 0.000 min
Lab File: BN034533.D
Acq: 10 Oct 2024 22:19




Instrument :
BNA_N
ClientSampleId :
PB163828BSD


Tgt Ion:252 Resp: 5737
Ion Ratio Lower Upper
252 100
253 35.1 19.6 29.4#
125 25.5 13.8 20.8#



#38
Benzo(k)fluoranthene
Concen: 2.831 ng
RT: 22.538 min Scan# 2592
Delta R.T. 0.003 min
Lab File: BN034533.D
Acq: 10 Oct 2024 22:19

Tgt Ion:252 Resp: 4098
Ion Ratio Lower Upper
252 100
253 39.9 20.1 30.1#
125 34.5 16.8 25.2#





#39

Benzo(a)pyrene

Concen: 1.723 ng

RT: 23.032 min Scan# 21

Delta R.T. 0.009 min

Lab File: BN034533.D

Acq: 10 Oct 2024 22:19

Instrument :

BNA_N

ClientSampleId :

PB163828BSD





Tgt Ion:252 Resp: 1937

Ion Ratio Lower Upper

252 100

253 61.8 21.8 32.8#

125 56.3 17.5 26.3#



#40

Dibenzo(a,h)anthracene

Concen: 4.636 ng

RT: 25.192 min Scan# 3500

Delta R.T. 0.006 min

Lab File: BN034533.D

Acq: 10 Oct 2024 22:19




Tgt Ion:278 Resp: 5432


Ion Ratio Lower Upper

278 100

139 24.0 17.8 26.8


279 37.2 21.0 31.6#





#41
Benzo(g,h,i)perylene
Concen: 3.035 ng
RT: 25.798 min Scan# 31
Delta R.T. 0.003 min
Lab File: BN034533.D
Acq: 10 Oct 2024 22:19

Instrument :
BNA_N
ClientSampleId :
PB163828BSD



Tgt Ion:276 Resp: 3994
Ion Ratio Lower Upper
276 100
277 29.8 19.3 28.9#
138 31.6 22.6 34.0

