

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN101024\
 Data File : BN034543.D
 Acq On : 11 Oct 2024 05:38
 Operator : RC/JU
 Sample : P4282-02
 Misc :
 ALS Vial : 119 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 RW7-SP201-20241002

Manual Integrations
 APPROVED

Reviewed By :Jagrut Upadhyay 10/11/2024
 Supervised By :mohammad ahmed 10/14/2024

Quant Time: Oct 11 06:09:25 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN091924.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Oct 11 00:21:08 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.373	152	4054	0.400	ng	0.00
7) Naphthalene-d8	10.127	136	11158	0.400	ng	0.00
13) Acenaphthene-d10	14.019	164	6232	0.400	ng	0.00
19) Phenanthrene-d10	16.784	188	11826	0.400	ng	0.00
29) Chrysene-d12	21.008	240	6910	0.400	ng	0.00
35) Perylene-d12	23.119	264	5236	0.400	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.018	112	1161	0.101	ng	0.00
5) Phenol-d6	6.593	99	766m	0.057	ng	0.02
8) Nitrobenzene-d5	8.525	82	2278	0.267	ng	0.01
11) 2-Methylnaphthalene-d10	11.731	152	5663	0.344	ng	0.00
14) 2,4,6-Tribromophenol	15.542	330	389	0.138	ng	0.00
15) 2-Fluorobiphenyl	12.640	172	8388	0.331	ng	0.00
27) Fluoranthene-d10	18.829	212	12878	0.432	ng	0.00
31) Terphenyl-d14	19.447	244	10174	0.737	ng	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN101024\
Data File : BN034543.D
Acq On : 11 Oct 2024 05:38
Operator : RC/JU
Sample : P4282-02
Misc :
ALS Vial : 119 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
RW7-SP201-20241002

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 10/11/2024
Supervised By :mohammad ahmed 10/14/2024

Quant Time: Oct 11 06:09:25 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN091924.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 11 00:21:08 2024
Response via : Initial Calibration

