

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN101024\
 Data File : BN034536.D
 Acq On : 11 Oct 2024 01:26
 Operator : RC/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 112 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Oct 11 01:49:31 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN091924.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Oct 11 00:21:08 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	108	0.00
2	1,4-Dioxane	0.400	0.350	12.5	94	0.00
3	n-Nitrosodimethylamine	0.400	0.302	24.5	84	0.00
4 S	2-Fluorophenol	0.400	0.360	10.0	103	0.00
5 S	Phenol-d6	0.400	0.342	14.5	102	0.00
6	bis(2-Chloroethyl)ether	0.400	0.303	24.3	86	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	106	0.00
8 S	Nitrobenzene-d5	0.400	0.338	15.5	97	0.00
9	Naphthalene	0.400	0.372	7.0	102	0.00
10	Hexachlorobutadiene	0.400	0.448	-12.0	120	-0.01
11 SURR	2-Methylnaphthalene-d10	0.400	0.378	5.5	105	0.00
12	2-Methylnaphthalene	0.400	0.359	10.3	100	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	103	0.00
14 S	2,4,6-Tribromophenol	0.400	0.406	-1.5	123	0.00
15 S	2-Fluorobiphenyl	0.400	0.337	15.8	90	0.00
16	Acenaphthylene	0.400	0.353	11.8	102	0.00
17	Acenaphthene	0.400	0.362	9.5	98	0.00
18	Fluorene	0.400	0.362	9.5	98	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	94	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.169	57.8#	12	0.00
21	4-Bromophenyl-phenylether	0.400	0.377	5.8	94	0.00
22	Hexachlorobenzene	0.400	0.465	-16.3	112	0.00
23	Atrazine	0.400	0.350	12.5	90	0.00
24	Pentachlorophenol	0.400	0.413	-3.2	116	0.00
25	Phenanthrene	0.400	0.374	6.5	91	0.00
26	Anthracene	0.400	0.359	10.3	94	0.00
27 SURR	Fluoranthene-d10	0.400	0.380	5.0	95	0.00
28	Fluoranthene	0.400	0.368	8.0	92	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	49	0.00
30	Pyrene	0.400	0.763	-90.8#	94	0.00
31 S	Terphenyl-d14	0.400	0.800	-100.0#	99	0.00
32	Benzo(a)anthracene	0.400	0.081	79.8#	11	0.00
33	Chrysene	0.400	0.706	-76.5#	86	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.000	100.0#	0	-21.01#
35 I	Perylene-d12	0.400	0.400	0.0	89	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.382	4.5	83	0.00
37	Benzo(b)fluoranthene	0.400	0.355	11.3	85	0.00
38	Benzo(k)fluoranthene	0.400	0.387	3.3	87	0.00
39 C	Benzo(a)pyrene	0.400	0.372	7.0	88	0.00
40	Dibenzo(a,h)anthracene	0.400	0.381	4.8	84	0.00
41	Benzo(g,h,i)perylene	0.400	0.403	-0.8	89	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4