

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN101024\
 Data File : BN034550.D
 Acq On : 11 Oct 2024 10:27
 Operator : RC/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 126 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Oct 11 10:59:56 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN091924.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Oct 11 00:21:08 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	76	0.00
2	1,4-Dioxane	0.500	0.460	8.0	69	0.00
3	n-Nitrosodimethylamine	0.569	0.432	24.1	59	0.00
4 S	2-Fluorophenol	1.135	0.985	13.2	70	0.00
5 S	Phenol-d6	1.320	1.216	7.9	78	0.00
6	bis(2-Chloroethyl)ether	1.168	0.869	25.6#	60	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	74	0.00
8 S	Nitrobenzene-d5	0.306	0.230	24.8	61	0.00
9	Naphthalene	1.098	0.988	10.0	69	0.00
10	Hexachlorobutadiene	0.207	0.232	-12.1	84	-0.01
11 SURR	2-Methylnaphthalene-d10	0.591	0.516	12.7	68	0.00
12	2-Methylnaphthalene	0.714	0.626	12.3	68	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	74	0.00
14 S	2,4,6-Tribromophenol	0.181	0.144	20.4	69	0.00
15 S	2-Fluorobiphenyl	1.627	1.308	19.6	62	0.00
16	Acenaphthylene	1.712	1.437	16.1	69	0.00
17	Acenaphthene	1.228	1.097	10.7	69	0.00
18	Fluorene	1.612	1.386	14.0	67	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	66	0.00
20	4,6-Dinitro-2-methylphenol	0.057	0.005	91.2#	7#	0.00
21	4-Bromophenyl-phenylether	0.225	0.207	8.0	64	0.00
22	Hexachlorobenzene	0.252	0.301	-19.4	80	-0.01
23	Atrazine	0.186	0.146	21.5	57	0.00
24	Pentachlorophenol	0.083	0.068	18.1	64	0.01
25	Phenanthrene	1.149	1.058	7.9	63	0.00
26	Anthracene	0.937	0.805	14.1	63	0.00
27 SURR	Fluoranthene-d10	1.009	0.906	10.2	63	0.00
28	Fluoranthene	1.331	1.208	9.2	64	0.00
29 I	Chrysene-d12	1.000	1.000	0.0	51	0.00
30	Pyrene	1.688	2.132	-26.3#	64	0.00
31 S	Terphenyl-d14	0.799	0.950	-18.9	61	0.00
32	Benzo(a)anthracene	1.265	0.333	73.7#	14#	0.00
33	Chrysene	1.510	1.915	-26.8#	63	0.00
34	Bis(2-ethylhexyl)phthalate	0.527	0.000	100.0#	0#	-21.01#
35 I	Perylene-d12	1.000	1.000	0.0	63	0.00
36	Indeno(1,2,3-cd)pyrene	1.714	1.583	7.6	56	0.00
37	Benzo(b)fluoranthene	1.567	1.349	13.9	58	0.00
38	Benzo(k)fluoranthene	1.645	1.592	3.2	61	0.00
39 C	Benzo(a)pyrene	1.277	1.163	8.9	61	0.00
40	Dibenzo(a,h)anthracene	1.331	1.212	8.9	56	0.00
41	Benzo(g,h,i)perylene	1.495	1.496	-0.1	62	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4