

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN101122\
 Data File : BN022052.D
 Acq On : 11 Oct 2022 15:01
 Operator : CG/JU
 Sample : SST01030
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SST010239

Quant Time: Oct 12 03:06:53 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN101122.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Oct 12 03:00:49 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.969	152	90445	20.000	ng/ul	0.00
20) Naphthalene-d8	10.798	136	419464	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.639	164	281917	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.392	188	614760	20.000	ng/ul	0.00
79) Chrysene-d12	21.592	240	626810	20.000	ng/ul	0.00
88) Perylene-d12	24.062	264	675719	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.263	96	9734	4.041	ng/uL	0.00
4) Pyridine-d5	3.710	84	64878	9.882	ng/ul	0.00
7) Phenol-d5	7.122	99	75928	9.256	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.293	67	54222	10.920	ng/ul	0.00
11) 2-Chlorophenol-d4	7.492	132	55369	9.277	ng/ul	0.00
15) 4-Methylphenol-d8	8.687	113	59146	9.189	ng/ul	0.00
21) Nitrobenzene-d5	9.145	128	27273	9.543	ng/ul	0.00
24) 2-Nitrophenol-d4	9.875	143	26948	10.557	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.416	165	54600	9.658	ng/ul	0.00
31) 4-Chloroaniline-d4	10.939	131	95115	9.907	ng/ul	0.00
46) Dimethylphthalate-d6	14.045	166	200973	10.430	ng/ul	0.00
49) Acenaphthylene-d8	14.333	160	210435	9.368	ng/ul	0.00
54) 4-Nitrophenol-d4	14.833	143	36103	9.416	ng/ul	0.00
60) Fluorene-d10	15.627	176	170120	10.146	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.751	200	25715	10.458	ng/ul	0.00
73) Anthracene-d10	17.492	188	274537	10.134	ng/ul	0.00
81) Pyrene-d10	19.792	212	309620	10.387	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.903	264	320569	10.009	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.299	88	11026	4.444	ng/uL	93
5) Pyridine	3.728	79	67978	10.290	ng/ul	99
6) Benzaldehyde	7.098	77	42609	11.955	ng/ul	98
8) Phenol	7.151	94	83905	9.898	ng/ul	99
10) Bis(2-Chloroethyl)ether	7.387	93	71615	10.596	ng/ul	97
12) 2-Chlorophenol	7.522	128	63150	9.863	ng/ul	99
13) 2-Methylphenol	8.416	108	60516	9.396	ng/ul	100
14) 2,2'-oxybis(1-Chloropr...	8.504	45	110242	12.931	ng/ul	99
16) Acetophenone	8.804	105	110704	10.875	ng/ul	98
17) N-Nitroso-di-n-propyla...	8.787	70	55215	10.893	ng/ul	98
18) 4-Methylphenol	8.745	108	68526	9.686	ng/ul	98
19) Hexachloroethane	9.057	117	26194	10.264	ng/ul	92
22) Nitrobenzene	9.187	77	80830	11.477	ng/ul	98
23) Isophorone	9.716	82	151893	10.474	ng/ul	99
25) 2-Nitrophenol	9.904	139	32592	10.900	ng/ul	98
26) 2,4-Dimethylphenol	9.963	107	76152	10.285	ng/ul	98
27) Bis(2-Chloroethoxy)met...	10.204	93	102851	11.325	ng/ul	100
29) 2,4-Dichlorophenol	10.439	162	59843	10.103	ng/ul	97
30) Naphthalene	10.845	128	232842	10.661	ng/ul	98
32) 4-Chloroaniline	10.963	127	103301	10.382	ng/ul	96
33) Hexachlorobutadiene	11.128	225	34191	10.020	ng/ul#	92
34) Caprolactam	11.751	113	18909	8.346	ng/ul	95
35) 4-Chloro-3-methylphenol	12.086	107	71162	10.660	ng/ul	97
36) 2-Methylnaphthalene	12.463	142	153528	10.391	ng/ul	96

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN101122\
 Data File : BN022052.D
 Acq On : 11 Oct 2022 15:01
 Operator : CG/JU
 Sample : SSTD01030
 Misc :
 ALS Vial : 3 Sample Multiplier: 1


Instrument :
 BNA_N
 ClientSampleId :
 SSTD010239


Quant Time: Oct 12 03:06:53 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN101122.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Oct 12 03:00:49 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.680	142	158453	10.671	ng/ul	98
39) 1,2,4,5-Tetrachloroben...	12.828	216	72196	9.987	ng/ul	98
40) Hexachlorocyclopentadiene	12.804	237	33433	7.768	ng/ul	89
41) 2,4,6-Trichlorophenol	13.075	196	45550	10.155	ng/ul	98
42) 2,4,5-Trichlorophenol	13.139	196	50333	10.227	ng/ul	98
43) 1,1'-Biphenyl	13.475	154	205132	9.957	ng/ul	98
44) 2-Chloronaphthalene	13.516	162	163275	10.352	ng/ul	99
45) 2-Nitroaniline	13.727	65	43599	11.021	ng/ul	97
47) Dimethylphthalate	14.092	163	217574	11.039	ng/ul	99
48) 2,6-Dinitrotoluene	14.222	165	32628	9.649	ng/ul	95
50) Acenaphthylene	14.363	152	251832	9.711	ng/ul	98
51) 3-Nitroaniline	14.551	138	37612	9.985	ng/ul	97
52) Acenaphthene	14.704	153	183051	10.924	ng/ul	97
53) 2,4-Dinitrophenol	14.757	184	18245	11.022	ng/ul	94
55) 4-Nitrophenol	14.851	109	31955	10.396	ng/ul	99
56) Dibenzofuran	15.033	168	254008	10.551	ng/ul	100
57) 2,4-Dinitrotoluene	15.004	165	52652	10.721	ng/ul	99
58) 2,3,4,6-Tetrachlorophenol	15.263	232	43424	11.495	ng/ul	99
59) Diethylphthalate	15.457	149	220883	11.113	ng/ul	97
61) Fluorene	15.686	166	211962	11.163	ng/ul	91
62) 4-Chlorophenyl-phenyle...	15.680	204	96212	10.992	ng/ul	97
63) 4-Nitroaniline	15.716	138	38021	10.057	ng/ul	96
66) 4,6-Dinitro-2-methylph...	15.763	198	30282	11.625	ng/ul#	95
67) N-Nitrosodiphenylamine	15.898	169	181759	10.549	ng/ul	97
68) 4-Bromophenyl-phenylether	16.580	248	54891	10.040	ng/ul	99
69) Hexachlorobenzene	16.692	284	68000	10.448	ng/ul	99
70) Atrazine	16.851	200	58535	9.601	ng/ul	96
71) Pentachlorophenol	17.039	266	34619	9.554	ng/ul#	87
72) Phenanthrene	17.433	178	344219	10.600	ng/ul	98
74) Anthracene	17.527	178	346494	10.716	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.439	216	71997	9.202	ng/uL	95
76) Pentachlorobenzene	14.957	250	81314	10.909	ng/uL	98
77) Carbazole	17.798	167	318444	10.422	ng/ul	99
78) Di-n-butylphthalate	18.363	149	361434	10.492	ng/ul	100
80) Fluoranthene	19.457	202	390669	10.865	ng/ul	100
82) Pyrene	19.821	202	428580	11.437	ng/ul	98
83) Butylbenzylphthalate	20.721	149	156480	11.032	ng/ul	98
84) 3,3'-Dichlorobenzidine	21.509	252	128996	10.793	ng/ul	92
85) Benzo(a)anthracene	21.574	228	409315	10.845	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.498	149	233664	10.760	ng/ul	98
87) Chrysene	21.627	228	418991	11.467	ng/ul	98
89) Di-n-octyl phthalate	22.445	149	403864	10.203	ng/ul	100
90) Benzo(b)fluoranthene	23.303	252	433021	10.533	ng/ul	94
91) Benzo(k)fluoranthene	23.356	252	427761	11.632	ng/ul	98
93) Benzo(a)pyrene	23.950	252	360086	9.155	ng/ul#	91
94) Indeno(1,2,3-cd)pyrene	26.650	276	481685	9.975	ng/ul	98
95) Dibenzo(a,h)anthracene	26.680	278	431462	10.588	ng/ul	98
96) Benzo(g,h,i)perylene	27.450	276	226523	5.376	ng/ul	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Oct 12 03:06:53 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN101122.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Oct 12 03:00:49 2022
Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 4.444 ng/uL
RT: 3.299 min Scan# 51
Delta R.T. -0.000 min
Lab File: BN022052.D
Acq: 11 Oct 2022 15:01




Instrument :
BNA_N
ClientSampleId :
SSTD010239


Tgt Ion: 88 Resp: 11026
Ion Ratio Lower Upper
88 100
43 35.1 25.1 37.7
58 77.9 67.4 101.2



#5
Pyridine
Concen: 10.290 ng/uL
RT: 3.728 min Scan# 126
Delta R.T. 0.006 min
Lab File: BN022052.D
Acq: 11 Oct 2022 15:01

Tgt Ion: 79 Resp: 67978
Ion Ratio Lower Upper
79 100
52 79.4 64.1 96.1
51 37.6 28.9 43.3









#6
Benzaldehyde
Concen: 11.955 ng/ul
RT: 7.098 min Scan# 699
Delta R.T. -0.000 min
Lab File: BN022052.D
Acq: 11 Oct 2022 15:01




Instrument :
BNA_N
ClientSampleId :
SSTD010239


Tgt Ion: 77 Resp: 42609
Ion Ratio Lower Upper
77 100
105 97.1 77.0 115.6
106 93.3 72.3 108.5



#8
Phenol
Concen: 9.898 ng/ul
RT: 7.151 min Scan# 708
Delta R.T. -0.000 min
Lab File: BN022052.D
Acq: 11 Oct 2022 15:01

Tgt Ion: 94 Resp: 83905
Ion Ratio Lower Upper
94 100
65 28.3 22.8 34.2
66 35.2 29.3 43.9









#10
Bis(2-Chloroethyl)ether
Concen: 10.596 ng/ul
RT: 7.387 min Scan# 748
Delta R.T. -0.000 min
Lab File: BN022052.D
Acq: 11 Oct 2022 15:01




Instrument :
BNA_N
ClientSampleId :
SSTD010239


Tgt Ion: 93 Resp: 71615
Ion Ratio Lower Upper
93 100
63 76.6 63.2 94.8
95 30.6 25.9 38.9



#12
2-Chlorophenol
Concen: 9.863 ng/ul
RT: 7.522 min Scan# 771
Delta R.T. -0.000 min
Lab File: BN022052.D
Acq: 11 Oct 2022 15:01

Tgt Ion:128 Resp: 63150
Ion Ratio Lower Upper
128 100
64 50.4 40.9 61.3
130 32.0 25.9 38.9





#13

2-Methylphenol

Concen: 9.396 ng/ul

RT: 8.416 min Scan# 91

Delta R.T. -0.000 min

Lab File: BN022052.D

Acq: 11 Oct 2022 15:01

Instrument :

BNA_N

ClientSampleId :





SSTD010239

Tgt Ion:108 Resp: 60516

Ion Ratio Lower Upper

108 100

107 89.6 71.8 107.6



#14

2,2'-oxybis(1-Chloropropane)

Concen: 12.931 ng/ul

RT: 8.504 min Scan# 938

Delta R.T. -0.006 min

Lab File: BN022052.D

Acq: 11 Oct 2022 15:01




Tgt Ion: 45 Resp: 110242


Ion Ratio Lower Upper

45 100

77 14.3 11.3 16.9

79 9.8 8.6 12.8









#16
Acetophenone
Concen: 10.875 ng/ul
RT: 8.804 min Scan# 91
Delta R.T. -0.000 min
Lab File: BN022052.D
Acq: 11 Oct 2022 15:01




Instrument :
BNA_N
ClientSampleId :
SSTD010239


Tgt Ion:105 Resp: 110704
Ion Ratio Lower Upper
105 100
77 77.6 63.4 95.0
51 26.9 21.9 32.9



#17
N-Nitroso-di-n-propylamine
Concen: 10.893 ng/ul
RT: 8.787 min Scan# 986
Delta R.T. -0.000 min
Lab File: BN022052.D
Acq: 11 Oct 2022 15:01

Tgt Ion: 70 Resp: 55215
Ion Ratio Lower Upper
70 100
42 60.0 46.2 69.4
101 8.9 7.1 10.7
130 19.7 15.2 22.8









#18
4-Methylphenol
Concen: 9.686 ng/ul
RT: 8.745 min Scan# 91
Delta R.T. -0.006 min
Lab File: BN022052.D
Acq: 11 Oct 2022 15:01




Instrument :
BNA_N
ClientSampleId :
SSTD010239


Tgt Ion:108 Resp: 68526
Ion Ratio Lower Upper
108 100
107 111.3 87.3 130.9



#19
Hexachloroethane
Concen: 10.264 ng/ul
RT: 9.057 min Scan# 1032
Delta R.T. -0.000 min
Lab File: BN022052.D
Acq: 11 Oct 2022 15:01

Tgt Ion:117 Resp: 26194
Ion Ratio Lower Upper
117 100
201 87.9 76.8 115.2
199 52.6 46.2 69.2





#22

Nitrobenzene

Concen: 11.477 ng/ul

RT: 9.187 min Scan# 1054

Delta R.T. -0.000 min

Lab File: BN022052.D

Acq: 11 Oct 2022 15:01

Instrument :

BNA_N

ClientSampleId :

SSTD010239





Tgt Ion: 77 Resp: 80830

Ion Ratio Lower Upper

77 100

123 40.8 33.8 50.8

65 12.2 10.5 15.7



#23

Isophorone

Concen: 10.474 ng/ul

RT: 9.716 min Scan# 1144

Delta R.T. -0.000 min

Lab File: BN022052.D

Acq: 11 Oct 2022 15:01




Tgt Ion: 82 Resp: 151893


Ion Ratio Lower Upper

82 100

95 6.3 5.1 7.7

138 16.5 12.6 18.8









#25
2-Nitrophenol
Concen: 10.900 ng/ul
RT: 9.904 min Scan# 1176
Delta R.T. -0.000 min
Lab File: BN022052.D
Acq: 11 Oct 2022 15:01




Instrument :
BNA_N
ClientSampleId :
SSTD010239


Tgt Ion:139 Resp: 32592
Ion Ratio Lower Upper
139 100
65 55.4 44.9 67.3
109 30.9 23.5 35.3



#26
2,4-Dimethylphenol
Concen: 10.285 ng/ul
RT: 9.963 min Scan# 1186
Delta R.T. -0.000 min
Lab File: BN022052.D
Acq: 11 Oct 2022 15:01

Tgt Ion:107 Resp: 76152
Ion Ratio Lower Upper
107 100
121 53.1 41.0 61.4
122 88.0 71.8 107.6









#27
 Bis(2-Chloroethoxy)methane
 Concen: 11.325 ng/ul
 RT: 10.204 min Scan# 1227
 Delta R.T. -0.000 min
 Lab File: BN022052.D
 Acq: 11 Oct 2022 15:01




Instrument :
 BNA_N
 ClientSampleId :
 SSTD010239


Tgt Ion: 93 Resp: 102851
 Ion Ratio Lower Upper
 93 100
 95 31.4 25.0 37.6
 123 10.4 8.6 13.0



#29
 2,4-Dichlorophenol
 Concen: 10.103 ng/ul
 RT: 10.439 min Scan# 1267
 Delta R.T. -0.000 min
 Lab File: BN022052.D
 Acq: 11 Oct 2022 15:01

Tgt Ion: 162 Resp: 59843
 Ion Ratio Lower Upper
 162 100
 164 64.7 53.9 80.9
 98 39.0 32.5 48.7









#30
Naphthalene
Concen: 10.661 ng/ul
RT: 10.845 min Scan# 1336
Delta R.T. -0.000 min
Lab File: BN022052.D
Acq: 11 Oct 2022 15:01




Instrument :
BNA_N
ClientSampleId :
SSTD010239


Tgt Ion:128 Resp: 232842
Ion Ratio Lower Upper
128 100
129 10.7 9.1 13.7
127 12.8 11.0 16.4



#32
4-Chloroaniline
Concen: 10.382 ng/ul
RT: 10.963 min Scan# 1356
Delta R.T. -0.000 min
Lab File: BN022052.D
Acq: 11 Oct 2022 15:01

Tgt Ion:127 Resp: 103301
Ion Ratio Lower Upper
127 100
129 32.5 24.2 36.4









#33
Hexachlorobutadiene
Concen: 10.020 ng/ul
RT: 11.128 min Scan# 11
Delta R.T. -0.000 min
Lab File: BN022052.D
Acq: 11 Oct 2022 15:01




Instrument :
BNA_N
ClientSampleId :
SSTD010239


Tgt Ion:225 Resp: 34191
Ion Ratio Lower Upper
225 100
223 66.8 53.0 79.6
227 70.8 46.9 70.3#



#34
Caprolactam
Concen: 8.346 ng/ul
RT: 11.751 min Scan# 1490
Delta R.T. 0.006 min
Lab File: BN022052.D
Acq: 11 Oct 2022 15:01

Tgt Ion:113 Resp: 18909
Ion Ratio Lower Upper
113 100
55 167.2 143.4 215.0
56 132.0 104.6 157.0









#35
4-Chloro-3-methylphenol
Concen: 10.660 ng/ul
RT: 12.086 min Scan# 11
Delta R.T. -0.000 min
Lab File: BN022052.D
Acq: 11 Oct 2022 15:01




Instrument :
BNA_N
ClientSampleId :
SSTD010239


Tgt Ion:107 Resp: 71162
Ion Ratio Lower Upper
107 100
144 24.1 20.7 31.1
142 76.7 59.8 89.8



#36
2-Methylnaphthalene
Concen: 10.391 ng/ul
RT: 12.463 min Scan# 1611
Delta R.T. -0.000 min
Lab File: BN022052.D
Acq: 11 Oct 2022 15:01

Tgt Ion:142 Resp: 153528
Ion Ratio Lower Upper
142 100
141 89.4 68.7 103.1









#37
1-Methylnaphthalene
Concen: 10.671 ng/ul
RT: 12.680 min Scan# 1648
Delta R.T. -0.000 min
Lab File: BN022052.D
Acq: 11 Oct 2022 15:01




Instrument :
BNA_N
ClientSampleId :
SSTD010239


Tgt Ion:142 Resp: 158453
Ion Ratio Lower Upper
142 100
141 90.1 73.3 109.9
116 4.0 3.0 4.6



#39
1,2,4,5-Tetrachlorobenzene
Concen: 9.987 ng/ul
RT: 12.828 min Scan# 1673
Delta R.T. -0.000 min
Lab File: BN022052.D
Acq: 11 Oct 2022 15:01

Tgt Ion:216 Resp: 72196
Ion Ratio Lower Upper
216 100
214 73.8 60.4 90.6
179 22.3 18.4 27.6
108 21.2 17.8 26.6









#40
Hexachlorocyclopentadiene
Concen: 7.768 ng/ul
RT: 12.804 min Scan# 1669
Delta R.T. -0.000 min
Lab File: BN022052.D
Acq: 11 Oct 2022 15:01




Instrument :
BNA_N
ClientSampleId :
SSTD010239


Tgt Ion:237 Resp: 33433
Ion Ratio Lower Upper
237 100
235 58.8 56.0 84.0
272 13.3 10.6 15.8



#41
2,4,6-Trichlorophenol
Concen: 10.155 ng/ul
RT: 13.075 min Scan# 1715
Delta R.T. 0.006 min
Lab File: BN022052.D
Acq: 11 Oct 2022 15:01

Tgt Ion:196 Resp: 45550
Ion Ratio Lower Upper
196 100
198 92.9 74.3 111.5
200 33.0 22.3 33.5





#42

2,4,5-Trichlorophenol

Concen: 10.227 ng/ul

RT: 13.139 min Scan# 1726

Delta R.T. -0.000 min

Lab File: BN022052.D

Acq: 11 Oct 2022 15:01

Instrument :

BNA_N

ClientSampleId :

SSTD010239





Tgt Ion:196 Resp: 50333

Ion Ratio Lower Upper

196 100

198 98.2 76.8 115.2

200 31.0 23.9 35.9



#43

1,1'-Biphenyl

Concen: 9.957 ng/ul

RT: 13.475 min Scan# 1783

Delta R.T. -0.000 min

Lab File: BN022052.D

Acq: 11 Oct 2022 15:01




Tgt Ion:154 Resp: 205132


Ion Ratio Lower Upper

154 100

153 40.5 31.0 46.6

76 14.9 11.6 17.4





#44

2-Chloronaphthalene

Concen: 10.352 ng/ul

RT: 13.516 min Scan# 11

Delta R.T. -0.000 min

Lab File: BN022052.D

Acq: 11 Oct 2022 15:01

Instrument :

BNA_N

ClientSampleId :

SSTD010239





Tgt Ion:162 Resp: 163275

Ion Ratio Lower Upper

162 100

164 31.2 25.9 38.9

127 40.5 32.2 48.4



#45

2-Nitroaniline

Concen: 11.021 ng/ul

RT: 13.727 min Scan# 1826

Delta R.T. -0.000 min

Lab File: BN022052.D

Acq: 11 Oct 2022 15:01




Tgt Ion: 65 Resp: 43599


Ion Ratio Lower Upper

65 100

92 69.3 54.6 82.0

138 96.1 80.4 120.6









#47
Dimethylphthalate
Concen: 11.039 ng/ul
RT: 14.092 min Scan# 1888
Delta R.T. -0.006 min
Lab File: BN022052.D
Acq: 11 Oct 2022 15:01




Instrument :
BNA_N
ClientSampleId :
SSTD010239


Tgt Ion:163 Resp: 217574
Ion Ratio Lower Upper
163 100
194 4.1 3.1 4.7
164 10.3 8.0 12.0



#48
2,6-Dinitrotoluene
Concen: 9.649 ng/ul
RT: 14.222 min Scan# 1910
Delta R.T. -0.000 min
Lab File: BN022052.D
Acq: 11 Oct 2022 15:01

Tgt Ion:165 Resp: 32628
Ion Ratio Lower Upper
165 100
89 69.6 52.2 78.4
121 25.7 19.5 29.3









#50
Acenaphthylene
Concen: 9.711 ng/ul
RT: 14.363 min Scan# 1934
Delta R.T. -0.000 min
Lab File: BN022052.D
Acq: 11 Oct 2022 15:01




Instrument :
BNA_N
ClientSampleId :
SSTD010239


Tgt Ion:152 Resp: 251832
Ion Ratio Lower Upper
152 100
151 18.4 15.7 23.5
153 11.9 10.0 15.0



#51
3-Nitroaniline
Concen: 9.985 ng/ul
RT: 14.551 min Scan# 1966
Delta R.T. -0.000 min
Lab File: BN022052.D
Acq: 11 Oct 2022 15:01

Tgt Ion:138 Resp: 37612
Ion Ratio Lower Upper
138 100
108 12.9 9.1 13.7
92 125.8 97.7 146.5









#52
 Acenaphthene
 Concen: 10.924 ng/ul
 RT: 14.704 min Scan# 1992
 Delta R.T. -0.000 min
 Lab File: BN022052.D
 Acq: 11 Oct 2022 15:01

Instrument :
 BNA_N
 ClientSampleId :
 SST010239

Tgt Ion:153 Resp: 183051




Ion	Ratio	Lower	Upper
153	100		
152	49.0	37.8	56.8
154	86.6	71.8	107.6




#53
 2,4-Dinitrophenol
 Concen: 11.022 ng/ul
 RT: 14.757 min Scan# 2001
 Delta R.T. -0.000 min
 Lab File: BN022052.D
 Acq: 11 Oct 2022 15:01

Tgt Ion:184 Resp: 18245

Ion	Ratio	Lower	Upper
184	100		
63	77.2	58.9	88.3
154	59.2	51.8	77.6









#55
4-Nitrophenol
Concen: 10.396 ng/ul
RT: 14.851 min Scan# 2017
Delta R.T. -0.000 min
Lab File: BN022052.D
Acq: 11 Oct 2022 15:01




Instrument :
BNA_N
ClientSampleId :
SSTD010239


Tgt Ion:109 Resp: 31955
Ion Ratio Lower Upper
109 100
139 138.9 111.0 166.4
65 144.9 118.5 177.7



#56
Dibenzofuran
Concen: 10.551 ng/ul
RT: 15.033 min Scan# 2048
Delta R.T. -0.006 min
Lab File: BN022052.D
Acq: 11 Oct 2022 15:01

Tgt Ion:168 Resp: 254008
Ion Ratio Lower Upper
168 100
139 39.0 31.3 46.9









#57
2,4-Dinitrotoluene
Concen: 10.721 ng/ul
RT: 15.004 min Scan# 2043
Delta R.T. -0.000 min
Lab File: BN022052.D
Acq: 11 Oct 2022 15:01




Instrument :
BNA_N
ClientSampleId :
SSTD010239


Tgt Ion:165 Resp: 52652
Ion Ratio Lower Upper
165 100
63 50.2 40.6 60.8
182 2.4 1.9 2.9



#58
2,3,4,6-Tetrachlorophenol
Concen: 11.495 ng/ul
RT: 15.263 min Scan# 2087
Delta R.T. -0.000 min
Lab File: BN022052.D
Acq: 11 Oct 2022 15:01

Tgt Ion:232 Resp: 43424
Ion Ratio Lower Upper
232 100
131 56.0 44.0 66.0
130 2.6 2.3 3.5
166 34.5 27.8 41.8









#59
Diethylphthalate
Concen: 11.113 ng/ul
RT: 15.457 min Scan# 2120
Delta R.T. -0.000 min
Lab File: BN022052.D
Acq: 11 Oct 2022 15:01




Instrument :
BNA_N
ClientSampleId :
SSTD010239


Tgt Ion:149 Resp: 220883
Ion Ratio Lower Upper
149 100
177 21.4 15.9 23.9
150 12.9 9.8 14.8



#61
Fluorene
Concen: 11.163 ng/ul
RT: 15.686 min Scan# 2159
Delta R.T. -0.000 min
Lab File: BN022052.D
Acq: 11 Oct 2022 15:01

Tgt Ion:166 Resp: 211962
Ion Ratio Lower Upper
166 100
165 102.7 74.1 111.1
167 13.3 10.5 15.7









#62
4-Chlorophenyl-phenylether
Concen: 10.992 ng/ul
RT: 15.680 min Scan# 2158
Delta R.T. -0.000 min
Lab File: BN022052.D
Acq: 11 Oct 2022 15:01




Instrument :
BNA_N
ClientSampleId :
SSTD010239


Tgt Ion:204 Resp: 96212
Ion Ratio Lower Upper
204 100
206 34.2 26.1 39.1
141 66.4 50.9 76.3



#63
4-Nitroaniline
Concen: 10.057 ng/ul
RT: 15.716 min Scan# 2164
Delta R.T. -0.000 min
Lab File: BN022052.D
Acq: 11 Oct 2022 15:01

Tgt Ion:138 Resp: 38021
Ion Ratio Lower Upper
138 100
92 52.0 42.8 64.2
108 79.4 60.1 90.1









#66
4,6-Dinitro-2-methylphenol
Concen: 11.625 ng/ul
RT: 15.763 min Scan# 2172
Delta R.T. -0.006 min
Lab File: BN022052.D
Acq: 11 Oct 2022 15:01




Instrument :
BNA_N
ClientSampleId :
SSTD010239


Tgt Ion:198 Resp: 30282
Ion Ratio Lower Upper
198 100
182 8.8 3.9 5.9#
77 32.6 24.5 36.7



#67
N-Nitrosodiphenylamine
Concen: 10.549 ng/ul
RT: 15.898 min Scan# 2195
Delta R.T. -0.000 min
Lab File: BN022052.D
Acq: 11 Oct 2022 15:01

Tgt Ion:169 Resp: 181759
Ion Ratio Lower Upper
169 100
168 67.2 51.4 77.0
167 35.7 27.8 41.8









#68
4-Bromophenyl-phenylether
Concen: 10.040 ng/ul
RT: 16.580 min Scan# 2311
Delta R.T. -0.000 min
Lab File: BN022052.D
Acq: 11 Oct 2022 15:01




Instrument :
BNA_N
ClientSampleId :
SSTD010239


Tgt Ion:248 Resp: 54891
Ion Ratio Lower Upper
248 100
250 95.4 77.6 116.4
141 83.1 66.6 99.8



#69
Hexachlorobenzene
Concen: 10.448 ng/ul
RT: 16.692 min Scan# 2330
Delta R.T. -0.000 min
Lab File: BN022052.D
Acq: 11 Oct 2022 15:01

Tgt Ion:284 Resp: 68000
Ion Ratio Lower Upper
284 100
142 40.6 32.6 48.8
249 28.2 23.4 35.2









#70
Atrazine
Concen: 9.601 ng/ul
RT: 16.851 min Scan# 21
Delta R.T. -0.000 min
Lab File: BN022052.D
Acq: 11 Oct 2022 15:01




Instrument :
BNA_N
ClientSampleId :
SSTD010239


Tgt Ion:200 Resp: 58535
Ion Ratio Lower Upper
200 100
173 27.4 22.6 34.0
215 44.0 38.4 57.6



#71
Pentachlorophenol
Concen: 9.554 ng/ul
RT: 17.039 min Scan# 2389
Delta R.T. -0.000 min
Lab File: BN022052.D
Acq: 11 Oct 2022 15:01

Tgt Ion:266 Resp: 34619
Ion Ratio Lower Upper
266 100
264 72.8 53.8 80.6
268 70.9 45.1 67.7#









#72
 Phenanthrene
 Concen: 10.600 ng/ul
 RT: 17.433 min Scan# 2456
 Delta R.T. -0.000 min
 Lab File: BN022052.D
 Acq: 11 Oct 2022 15:01

Instrument :
 BNA_N
 ClientSampleId :
 SSTD010239

Tgt Ion:178 Resp: 344219




Ion	Ratio	Lower	Upper
178	100		
179	15.9	12.3	18.5
176	18.9	15.9	23.9




#74
 Anthracene
 Concen: 10.716 ng/ul
 RT: 17.527 min Scan# 2472
 Delta R.T. -0.000 min
 Lab File: BN022052.D
 Acq: 11 Oct 2022 15:01

Tgt Ion:178 Resp: 346494

Ion	Ratio	Lower	Upper
178	100		
179	14.9	12.2	18.2
176	18.8	15.2	22.8









#75
1,2,3,4-Tetrachlorobenzene
Concen: 9.202 ng/uL
RT: 13.439 min Scan# 1777
Delta R.T. -0.000 min
Lab File: BN022052.D
Acq: 11 Oct 2022 15:01




Instrument :
BNA_N
ClientSampleId :
SSTD010239


Tgt Ion:216 Resp: 71997
Ion Ratio Lower Upper
216 100
214 81.8 60.3 90.5
218 46.1 37.2 55.8



#76
Pentachlorobenzene
Concen: 10.909 ng/uL
RT: 14.957 min Scan# 2035
Delta R.T. -0.000 min
Lab File: BN022052.D
Acq: 11 Oct 2022 15:01

Tgt Ion:250 Resp: 81314
Ion Ratio Lower Upper
250 100
248 63.2 52.3 78.5
252 63.5 50.0 75.0









#77
 Carbazole
 Concen: 10.422 ng/ul
 RT: 17.798 min Scan# 2518
 Delta R.T. -0.000 min
 Lab File: BN022052.D
 Acq: 11 Oct 2022 15:01




Instrument :
 BNA_N
 ClientSampleId :
 SST010239


Tgt Ion:167 Resp: 318444
 Ion Ratio Lower Upper
 167 100
 166 21.0 17.2 25.8
 139 13.1 11.0 16.4



#78
 Di-n-butylphthalate
 Concen: 10.492 ng/ul
 RT: 18.363 min Scan# 2614
 Delta R.T. -0.000 min
 Lab File: BN022052.D
 Acq: 11 Oct 2022 15:01

Tgt Ion:149 Resp: 361434
 Ion Ratio Lower Upper
 149 100
 150 8.4 6.7 10.1
 104 5.7 4.3 6.5









#80
 Fluoranthene
 Concen: 10.865 ng/ul
 RT: 19.457 min Scan# 2800
 Delta R.T. -0.000 min
 Lab File: BN022052.D
 Acq: 11 Oct 2022 15:01




Instrument :
 BNA_N
 ClientSampleId :
 SST010239


Tgt Ion:202 Resp: 390669
 Ion Ratio Lower Upper
 202 100
 101 14.0 11.4 17.0
 100 10.8 8.6 13.0



#82
 Pyrene
 Concen: 11.437 ng/ul
 RT: 19.821 min Scan# 2862
 Delta R.T. -0.000 min
 Lab File: BN022052.D
 Acq: 11 Oct 2022 15:01

Tgt Ion:202 Resp: 428580
 Ion Ratio Lower Upper
 202 100
 101 16.2 13.8 20.6
 100 13.3 11.2 16.8





#83

Butylbenzylphthalate

Concen: 11.032 ng/ul

RT: 20.721 min Scan# 3015

Delta R.T. -0.000 min

Lab File: BN022052.D

Acq: 11 Oct 2022 15:01

Instrument :

BNA_N

ClientSampleId :

SSTD010239





Tgt Ion:149 Resp: 156480

Ion Ratio Lower Upper

149 100

91 74.5 60.9 91.3

206 19.1 14.9 22.3



#84

3,3'-Dichlorobenzidine

Concen: 10.793 ng/ul

RT: 21.509 min Scan# 3149

Delta R.T. -0.000 min

Lab File: BN022052.D

Acq: 11 Oct 2022 15:01




Tgt Ion:252 Resp: 128996


Ion Ratio Lower Upper

252 100

254 68.3 48.6 72.8

126 14.1 11.1 16.7









#85
Benzo(a)anthracene
Concen: 10.845 ng/ul
RT: 21.574 min Scan# 3160
Delta R.T. -0.000 min
Lab File: BN022052.D
Acq: 11 Oct 2022 15:01




Instrument :
BNA_N
ClientSampleId :
SSTD010239


Tgt Ion:228 Resp: 409315
Ion Ratio Lower Upper
228 100
229 21.3 16.7 25.1
226 28.3 22.4 33.6



#86
Bis(2-ethylhexyl)phthalate
Concen: 10.760 ng/ul
RT: 21.498 min Scan# 3147
Delta R.T. -0.000 min
Lab File: BN022052.D
Acq: 11 Oct 2022 15:01

Tgt Ion:149 Resp: 233664
Ion Ratio Lower Upper
149 100
167 26.1 21.8 32.6
279 3.1 3.0 4.4









#87
Chrysene
Concen: 11.467 ng/ul
RT: 21.627 min Scan# 3170
Delta R.T. -0.006 min
Lab File: BN022052.D
Acq: 11 Oct 2022 15:01




Instrument :
BNA_N
ClientSampleId :
SSTD010239


Tgt Ion:228 Resp: 418991
Ion Ratio Lower Upper
228 100
226 30.6 22.9 34.3
229 19.6 15.5 23.3



#89
Di-n-octyl phthalate
Concen: 10.203 ng/ul
RT: 22.445 min Scan# 3308
Delta R.T. -0.000 min
Lab File: BN022052.D
Acq: 11 Oct 2022 15:01

Tgt Ion:149 Resp: 403864









#90
Benzo(b)fluoranthene
Concen: 10.533 ng/ul
RT: 23.303 min Scan# 3454
Delta R.T. -0.000 min
Lab File: BN022052.D
Acq: 11 Oct 2022 15:01




Instrument :
BNA_N
ClientSampleId :
SSTD010239


Tgt Ion:252 Resp: 433021
Ion Ratio Lower Upper
252 100
253 23.4 15.7 23.5
125 13.9 10.0 15.0



#91
Benzo(k)fluoranthene
Concen: 11.632 ng/ul
RT: 23.356 min Scan# 3463
Delta R.T. -0.000 min
Lab File: BN022052.D
Acq: 11 Oct 2022 15:01

Tgt Ion:252 Resp: 427761
Ion Ratio Lower Upper
252 100
253 22.2 16.8 25.2
125 13.1 10.4 15.6





#93

Benzo(a)pyrene

Concen: 9.155 ng/ul

RT: 23.950 min Scan# 3

Delta R.T. -0.006 min

Lab File: BN022052.D

Acq: 11 Oct 2022 15:01

Instrument :

BNA_N

ClientSampleId :

SSTD010239





Tgt Ion:252 Resp: 360086

Ion Ratio Lower Upper

252 100

253 26.0 16.8 25.2#

125 16.6 11.2 16.8



#94

Indeno(1,2,3-cd)pyrene

Concen: 9.975 ng/ul

RT: 26.650 min Scan# 4023

Delta R.T. -0.018 min

Lab File: BN022052.D

Acq: 11 Oct 2022 15:01




Tgt Ion:276 Resp: 481685


Ion Ratio Lower Upper

276 100

138 30.5 25.4 38.0

277 24.3 18.8 28.2





#95

Dibenzo(a,h)anthracene

Concen: 10.588 ng/ul

RT: 26.680 min Scan# 41159

Delta R.T. 0.006 min

Lab File: BN022052.D

Acq: 11 Oct 2022 15:01

Instrument :

BNA_N

ClientSampleId :

SSTD010239





Tgt Ion:278 Resp: 431462

Ion Ratio Lower Upper

278 100

139 20.2 17.4 26.2

279 23.8 18.6 28.0



#96

Benzo(g,h,i)perylene

Concen: 5.376 ng/ul

RT: 27.450 min Scan# 4159

Delta R.T. -0.018 min

Lab File: BN022052.D

Acq: 11 Oct 2022 15:01

Tgt Ion:276 Resp: 226523

Ion Ratio Lower Upper

276 100

138 28.0 19.4 29.2

277 22.8 16.5 24.7

