

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN101122\
 Data File : BN022054.D
 Acq On : 11 Oct 2022 16:18
 Operator : CG/JU
 Sample : SST04032
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SST040241

Quant Time: Oct 12 03:07:59 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN101122.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Oct 12 03:00:49 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.969	152	97758	20.000	ng/ul	0.00
20) Naphthalene-d8	10.799	136	482767	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.640	164	317692	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.392	188	668381	20.000	ng/ul	0.00
79) Chrysene-d12	21.598	240	652142	20.000	ng/ul	0.00
88) Perylene-d12	24.069	264	697293	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.264	96	40986	15.742	ng/uL	0.00
4) Pyridine-d5	3.699	84	327157	46.103	ng/ul	0.00
7) Phenol-d5	7.122	99	391447	44.150	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.293	67	253901	47.309	ng/ul	0.00
11) 2-Chlorophenol-d4	7.493	132	285107	44.198	ng/ul	0.00
15) 4-Methylphenol-d8	8.687	113	317578	45.650	ng/ul	0.00
21) Nitrobenzene-d5	9.146	128	153969	46.811	ng/ul	0.00
24) 2-Nitrophenol-d4	9.875	143	162544	55.327	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.416	165	291281	44.769	ng/ul	0.00
31) 4-Chloroaniline-d4	10.940	131	464550	42.042	ng/ul	0.00
46) Dimethylphthalate-d6	14.051	166	949723	43.738	ng/ul	0.00
49) Acenaphthylene-d8	14.334	160	1074347	42.440	ng/ul	0.00
54) 4-Nitrophenol-d4	14.845	143	205598	47.581	ng/ul	0.00
60) Fluorene-d10	15.634	176	778238	41.189	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.757	200	159797	59.773	ng/ul	0.00
73) Anthracene-d10	17.492	188	1234026	41.896	ng/ul	0.00
81) Pyrene-d10	19.792	212	1399208	45.116	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.910	264	1482078	44.843	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.299	88	44290	16.517	ng/uL	89
5) Pyridine	3.723	79	323281	45.276	ng/ul	97
6) Benzaldehyde	7.099	77	215888	56.041	ng/ul	99
8) Phenol	7.152	94	419341	45.769	ng/ul	98
10) Bis(2-Chloroethyl)ether	7.387	93	335323	45.900	ng/ul	98
12) 2-Chlorophenol	7.522	128	311898	45.069	ng/ul	98
13) 2-Methylphenol	8.416	108	313585	45.046	ng/ul	99
14) 2,2'-oxybis(1-Chloropr...	8.511	45	503996	54.695	ng/ul	100
16) Acetophenone	8.805	105	522994	47.532	ng/ul	98
17) N-Nitroso-di-n-propyla...	8.793	70	284275	51.887	ng/ul	98
18) 4-Methylphenol	8.752	108	357680	46.774	ng/ul	99
19) Hexachloroethane	9.058	117	120949	43.848	ng/ul	94
22) Nitrobenzene	9.193	77	404092	49.855	ng/ul	100
23) Isophorone	9.722	82	804282	48.188	ng/ul	98
25) 2-Nitrophenol	9.905	139	188341	54.728	ng/ul	99
26) 2,4-Dimethylphenol	9.969	107	379943	44.585	ng/ul	95
27) Bis(2-Chloroethoxy)met...	10.210	93	480517	45.973	ng/ul	97
29) 2,4-Dichlorophenol	10.440	162	309105	45.343	ng/ul	91
30) Naphthalene	10.852	128	1085324	43.176	ng/ul	98
32) 4-Chloroaniline	10.963	127	490943	42.873	ng/ul	97
33) Hexachlorobutadiene	11.128	225	161046	41.007	ng/ul	97
34) Caprolactam	11.752	113	61477	23.576	ng/ul	96
35) 4-Chloro-3-methylphenol	12.093	107	375698	48.900	ng/ul	95
36) 2-Methylnaphthalene	12.463	142	744954	43.809	ng/ul	98


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN101122\
 Data File : BN022054.D
 Acq On : 11 Oct 2022 16:18
 Operator : CG/JU
 Sample : SST04032
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SST040241

Quant Time: Oct 12 03:07:59 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN101122.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Oct 12 03:00:49 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.681	142	754516	44.150	ng/ul	99
39) 1,2,4,5-Tetrachloroben...	12.828	216	343243	42.134	ng/ul	93
40) Hexachlorocyclopentadiene	12.804	237	179225	36.953	ng/ul#	92
41) 2,4,6-Trichlorophenol	13.075	196	247331	48.931	ng/ul	95
42) 2,4,5-Trichlorophenol	13.140	196	267162	48.172	ng/ul	96
43) 1,1'-Biphenyl	13.475	154	963405	41.496	ng/ul	98
44) 2-Chloronaphthalene	13.516	162	746384	41.993	ng/ul	99
45) 2-Nitroaniline	13.728	65	258737	58.039	ng/ul	98
47) Dimethylphthalate	14.098	163	987403	44.456	ng/ul	100
48) 2,6-Dinitrotoluene	14.222	165	200479	52.612	ng/ul	99
50) Acenaphthylene	14.363	152	1192124	40.793	ng/ul	100
51) 3-Nitroaniline	14.551	138	228830	53.905	ng/ul	96
52) Acenaphthene	14.704	153	847614	44.888	ng/ul	98
53) 2,4-Dinitrophenol	14.757	184	125890	67.486	ng/ul	97
55) 4-Nitrophenol	14.857	109	165354	47.739	ng/ul	96
56) Dibenzofuran	15.040	168	1134839	41.830	ng/ul	99
57) 2,4-Dinitrotoluene	15.010	165	298510	53.940	ng/ul	96
58) 2,3,4,6-Tetrachlorophenol	15.263	232	223269	52.449	ng/ul	99
59) Diethylphthalate	15.463	149	1020647	45.567	ng/ul	97
61) Fluorene	15.693	166	935704	43.731	ng/ul	98
62) 4-Chlorophenyl-phenyle...	15.681	204	423613	42.947	ng/ul	98
63) 4-Nitroaniline	15.716	138	227733	53.457	ng/ul	99
66) 4,6-Dinitro-2-methylph...	15.769	198	173230	61.164	ng/ul	99
67) N-Nitrosodiphenylamine	15.898	169	819764	43.760	ng/ul	98
68) 4-Bromophenyl-phenylether	16.581	248	258976	43.567	ng/ul	98
69) Hexachlorobenzene	16.692	284	289624	40.930	ng/ul	95
70) Atrazine	16.857	200	285906	43.131	ng/ul	100
71) Pentachlorophenol	17.045	266	183282	46.524	ng/ul	95
72) Phenanthrene	17.439	178	1529127	43.313	ng/ul	98
74) Anthracene	17.528	178	1541470	43.849	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.440	216	343204	40.344	ng/uL	95
76) Pentachlorobenzene	14.957	250	374745	46.243	ng/uL	95
77) Carbazole	17.804	167	1464252	44.076	ng/ul	99
78) Di-n-butylphthalate	18.363	149	1863121	49.744	ng/ul	99
80) Fluoranthene	19.457	202	1777817	47.525	ng/ul	99
82) Pyrene	19.822	202	1826322	46.842	ng/ul	99
83) Butylbenzylphthalate	20.722	149	870516	58.990	ng/ul	99
84) 3,3'-Dichlorobenzidine	21.516	252	631673	50.801	ng/ul	98
85) Benzo(a)anthracene	21.580	228	1756388	44.731	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.498	149	1259605	55.748	ng/ul	100
87) Chrysene	21.633	228	1701406	44.754	ng/ul	95
89) Di-n-octyl phthalate	22.445	149	2324058	56.899	ng/ul	100
90) Benzo(b)fluoranthene	23.310	252	1924236	45.359	ng/ul	94
91) Benzo(k)fluoranthene	23.363	252	1856658	48.926	ng/ul	99
93) Benzo(a)pyrene	23.963	252	1752711	43.182	ng/ul	98
94) Indeno(1,2,3-cd)pyrene	26.680	276	2162300	43.394	ng/ul	98
95) Dibenzo(a,h)anthracene	26.698	278	2017956	47.987	ng/ul	97
96) Benzo(g,h,i)perylene	27.474	276	1933532	44.467	ng/ul	93





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
1,4-Dioxane
Concen: 16.517 ng/uL
RT: 3.299 min Scan# 51
Delta R.T. 0.000 min
Lab File: BN022054.D
Acq: 11 Oct 2022 16:18




Instrument :
BNA_N
ClientSampleId :
SSTD040241


Tgt Ion: 88 Resp: 44290
Ion Ratio Lower Upper
88 100
43 27.6 25.1 37.7
58 72.9 67.4 101.2



#5
Pyridine
Concen: 45.276 ng/uL
RT: 3.723 min Scan# 125
Delta R.T. 0.000 min
Lab File: BN022054.D
Acq: 11 Oct 2022 16:18

Tgt Ion: 79 Resp: 323281
Ion Ratio Lower Upper
79 100
52 76.1 64.1 96.1
51 35.4 28.9 43.3









#6
Benzaldehyde
Concen: 56.041 ng/ul
RT: 7.099 min Scan# 699
Delta R.T. 0.000 min
Lab File: BN022054.D
Acq: 11 Oct 2022 16:18




Instrument :
BNA_N
ClientSampleId :
SSTD040241


Tgt Ion: 77 Resp: 215888
Ion Ratio Lower Upper
77 100
105 95.8 77.0 115.6
106 89.8 72.3 108.5



#8
Phenol
Concen: 45.769 ng/ul
RT: 7.152 min Scan# 708
Delta R.T. 0.000 min
Lab File: BN022054.D
Acq: 11 Oct 2022 16:18

Tgt Ion: 94 Resp: 419341
Ion Ratio Lower Upper
94 100
65 29.2 22.8 34.2
66 35.0 29.3 43.9









#10
Bis(2-Chloroethyl)ether
Concen: 45.900 ng/ul
RT: 7.387 min Scan# 748
Delta R.T. 0.000 min
Lab File: BN022054.D
Acq: 11 Oct 2022 16:18




Instrument :
BNA_N
ClientSampleId :
SSTD040241


Tgt Ion: 93 Resp: 335323
Ion Ratio Lower Upper
93 100
63 76.8 63.2 94.8
95 32.5 25.9 38.9



#12
2-Chlorophenol
Concen: 45.069 ng/ul
RT: 7.522 min Scan# 771
Delta R.T. 0.000 min
Lab File: BN022054.D
Acq: 11 Oct 2022 16:18

Tgt Ion:128 Resp: 311898
Ion Ratio Lower Upper
128 100
64 49.6 40.9 61.3
130 32.8 25.9 38.9





#13

2-Methylphenol

Concen: 45.046 ng/ul

RT: 8.416 min Scan# 91

Delta R.T. 0.000 min

Lab File: BN022054.D

Acq: 11 Oct 2022 16:18

Instrument :

BNA_N

ClientSampleId :





SSTD040241

Tgt Ion:108 Resp: 313585

Ion Ratio Lower Upper

108 100

107 88.6 71.8 107.6



#14

2,2'-oxybis(1-Chloropropane)

Concen: 54.695 ng/ul

RT: 8.511 min Scan# 939

Delta R.T. 0.000 min

Lab File: BN022054.D

Acq: 11 Oct 2022 16:18




Tgt Ion: 45 Resp: 503996


Ion Ratio Lower Upper

45 100

77 14.1 11.3 16.9

79 11.0 8.6 12.8









#16
Acetophenone
Concen: 47.532 ng/ul
RT: 8.805 min Scan# 989
Delta R.T. 0.000 min
Lab File: BN022054.D
Acq: 11 Oct 2022 16:18




Instrument :
BNA_N
ClientSampleId :
SSTD040241


Tgt Ion:105 Resp: 522994
Ion Ratio Lower Upper
105 100
77 81.2 63.4 95.0
51 28.2 21.9 32.9



#17
N-Nitroso-di-n-propylamine
Concen: 51.887 ng/ul
RT: 8.793 min Scan# 987
Delta R.T. 0.006 min
Lab File: BN022054.D
Acq: 11 Oct 2022 16:18

Tgt Ion: 70 Resp: 284275
Ion Ratio Lower Upper
70 100
42 59.3 46.2 69.4
101 9.5 7.1 10.7
130 20.7 15.2 22.8









#18
4-Methylphenol
Concen: 46.774 ng/ul
RT: 8.752 min Scan# 91
Delta R.T. 0.000 min
Lab File: BN022054.D
Acq: 11 Oct 2022 16:18




Instrument :
BNA_N
ClientSampleId :
SSTD040241


Tgt Ion:108 Resp: 357680
Ion Ratio Lower Upper
108 100
107 110.0 87.3 130.9



#19
Hexachloroethane
Concen: 43.848 ng/ul
RT: 9.058 min Scan# 1032
Delta R.T. 0.000 min
Lab File: BN022054.D
Acq: 11 Oct 2022 16:18

Tgt Ion:117 Resp: 120949
Ion Ratio Lower Upper
117 100
201 89.8 76.8 115.2
199 53.9 46.2 69.2





#22

Nitrobenzene

Concen: 49.855 ng/ul

RT: 9.193 min Scan# 1105

Delta R.T. 0.006 min

Lab File: BN022054.D

Acq: 11 Oct 2022 16:18

Instrument :

BNA_N

ClientSampleId :

SSTD040241





Tgt Ion: 77 Resp: 404092

Ion Ratio Lower Upper

77 100

123 42.3 33.8 50.8

65 13.3 10.5 15.7



#23

Isophorone

Concen: 48.188 ng/ul

RT: 9.722 min Scan# 1145

Delta R.T. 0.006 min

Lab File: BN022054.D

Acq: 11 Oct 2022 16:18




Tgt Ion: 82 Resp: 804282


Ion Ratio Lower Upper

82 100

95 6.7 5.1 7.7

138 16.6 12.6 18.8









#25
2-Nitrophenol
Concen: 54.728 ng/ul
RT: 9.905 min Scan# 1176
Delta R.T. 0.000 min
Lab File: BN022054.D
Acq: 11 Oct 2022 16:18




Instrument :
BNA_N
ClientSampleId :
SSTD040241


Tgt Ion:139 Resp: 188341
Ion Ratio Lower Upper
139 100
65 55.1 44.9 67.3
109 29.9 23.5 35.3



#26
2,4-Dimethylphenol
Concen: 44.585 ng/ul
RT: 9.969 min Scan# 1187
Delta R.T. 0.006 min
Lab File: BN022054.D
Acq: 11 Oct 2022 16:18

Tgt Ion:107 Resp: 379943
Ion Ratio Lower Upper
107 100
121 49.9 41.0 61.4
122 83.7 71.8 107.6









#27
 Bis(2-Chloroethoxy)methane
 Concen: 45.973 ng/ul
 RT: 10.210 min Scan# 1228
 Delta R.T. 0.006 min
 Lab File: BN022054.D
 Acq: 11 Oct 2022 16:18




Instrument :
 BNA_N
 ClientSampleId :
 SST040241


Tgt Ion: 93 Resp: 480517
 Ion Ratio Lower Upper
 93 100
 95 33.0 25.0 37.6
 123 11.2 8.6 13.0



#29
 2,4-Dichlorophenol
 Concen: 45.343 ng/ul
 RT: 10.440 min Scan# 1267
 Delta R.T. 0.000 min
 Lab File: BN022054.D
 Acq: 11 Oct 2022 16:18

Tgt Ion: 162 Resp: 309105
 Ion Ratio Lower Upper
 162 100
 164 57.9 53.9 80.9
 98 38.2 32.5 48.7





#30

Naphthalene

Concen: 43.176 ng/ul

RT: 10.852 min Scan# 1337

Delta R.T. 0.006 min

Lab File: BN022054.D

Acq: 11 Oct 2022 16:18

Instrument :

BNA_N

ClientSampleId :

SSTD040241





Tgt Ion:128 Resp: 1085324

Ion Ratio Lower Upper

128 100

129 10.5 9.1 13.7

127 12.8 11.0 16.4



#32

4-Chloroaniline

Concen: 42.873 ng/ul

RT: 10.963 min Scan# 1356

Delta R.T. 0.000 min

Lab File: BN022054.D




Acq: 11 Oct 2022 16:18


Tgt Ion:127 Resp: 490943

Ion Ratio Lower Upper

127 100

129 32.0 24.2 36.4









#33
Hexachlorobutadiene
Concen: 41.007 ng/ul
RT: 11.128 min Scan# 1111
Delta R.T. 0.000 min
Lab File: BN022054.D
Acq: 11 Oct 2022 16:18




Instrument :
BNA_N
ClientSampleId :
SSTD040241


Tgt Ion:225 Resp: 161046
Ion Ratio Lower Upper
225 100
223 62.5 53.0 79.6
227 58.3 46.9 70.3



#34
Caprolactam
Concen: 23.576 ng/ul
RT: 11.752 min Scan# 1490
Delta R.T. 0.006 min
Lab File: BN022054.D
Acq: 11 Oct 2022 16:18

Tgt Ion:113 Resp: 61477
Ion Ratio Lower Upper
113 100
55 174.7 143.4 215.0
56 124.5 104.6 157.0









#35
4-Chloro-3-methylphenol
Concen: 48.900 ng/ul
RT: 12.093 min Scan# 1548
Delta R.T. 0.006 min
Lab File: BN022054.D
Acq: 11 Oct 2022 16:18




Instrument :
BNA_N
ClientSampleId :
SSTD040241


Tgt Ion:107 Resp: 375698
Ion Ratio Lower Upper
107 100
144 25.4 20.7 31.1
142 79.7 59.8 89.8



#36
2-Methylnaphthalene
Concen: 43.809 ng/ul
RT: 12.463 min Scan# 1611
Delta R.T. 0.000 min
Lab File: BN022054.D
Acq: 11 Oct 2022 16:18

Tgt Ion:142 Resp: 744954
Ion Ratio Lower Upper
142 100
141 87.6 68.7 103.1









#37
1-Methylnaphthalene
Concen: 44.150 ng/ul
RT: 12.681 min Scan# 1648
Delta R.T. 0.000 min
Lab File: BN022054.D
Acq: 11 Oct 2022 16:18




Instrument :
BNA_N
ClientSampleId :
SSTD040241


Tgt Ion:142 Resp: 754516
Ion Ratio Lower Upper
142 100
141 92.8 73.3 109.9
116 3.8 3.0 4.6



#39
1,2,4,5-Tetrachlorobenzene
Concen: 42.134 ng/ul
RT: 12.828 min Scan# 1673
Delta R.T. 0.000 min
Lab File: BN022054.D
Acq: 11 Oct 2022 16:18


Tgt Ion:216 Resp: 343243
Ion Ratio Lower Upper
216 100
214 83.9 60.4 90.6
179 22.0 18.4 27.6
108 21.7 17.8 26.6








#40
Hexachlorocyclopentadiene
Concen: 36.953 ng/ul
RT: 12.804 min Scan# 1669
Delta R.T. 0.000 min
Lab File: BN022054.D
Acq: 11 Oct 2022 16:18


Instrument :
BNA_N
ClientSampleId :
SSTD040241





Tgt Ion:237 Resp: 179225
Ion Ratio Lower Upper
237 100
235 63.1 56.0 84.0
272 10.5 10.6 15.8#




#41
2,4,6-Trichlorophenol
Concen: 48.931 ng/ul
RT: 13.075 min Scan# 1715
Delta R.T. 0.006 min
Lab File: BN022054.D
Acq: 11 Oct 2022 16:18



Tgt Ion:196 Resp: 247331
Ion Ratio Lower Upper
196 100
198 97.6 74.3 111.5
200 30.2 22.3 33.5





#42

2,4,5-Trichlorophenol

Concen: 48.172 ng/ul

RT: 13.140 min Scan# 1726

Delta R.T. 0.000 min

Lab File: BN022054.D

Acq: 11 Oct 2022 16:18

Instrument :

BNA_N

ClientSampleId :

SSTD040241





Tgt Ion:196 Resp: 267162

Ion Ratio Lower Upper

196 100

198 91.8 76.8 115.2

200 29.4 23.9 35.9



#43

1,1'-Biphenyl

Concen: 41.496 ng/ul

RT: 13.475 min Scan# 1783

Delta R.T. 0.000 min

Lab File: BN022054.D

Acq: 11 Oct 2022 16:18




Tgt Ion:154 Resp: 963405


Ion Ratio Lower Upper

154 100

153 40.3 31.0 46.6

76 15.2 11.6 17.4









#44
2-Chloronaphthalene
Concen: 41.993 ng/ul
RT: 13.516 min Scan# 1790
Delta R.T. 0.000 min
Lab File: BN022054.D
Acq: 11 Oct 2022 16:18




Instrument :
BNA_N
ClientSampleId :
SSTD040241


Tgt Ion:162 Resp: 746384
Ion Ratio Lower Upper
162 100
164 33.0 25.9 38.9
127 40.6 32.2 48.4



#45
2-Nitroaniline
Concen: 58.039 ng/ul
RT: 13.728 min Scan# 1826
Delta R.T. 0.000 min
Lab File: BN022054.D
Acq: 11 Oct 2022 16:18

Tgt Ion: 65 Resp: 258737
Ion Ratio Lower Upper
65 100
92 69.3 54.6 82.0
138 102.7 80.4 120.6









#47
Dimethylphthalate
Concen: 44.456 ng/ul
RT: 14.098 min Scan# 1889
Delta R.T. 0.000 min
Lab File: BN022054.D
Acq: 11 Oct 2022 16:18




Instrument :
BNA_N
ClientSampleId :
SSTD040241


Tgt Ion:163 Resp: 987403
Ion Ratio Lower Upper
163 100
194 3.8 3.1 4.7
164 10.1 8.0 12.0



#48
2,6-Dinitrotoluene
Concen: 52.612 ng/ul
RT: 14.222 min Scan# 1910
Delta R.T. 0.000 min
Lab File: BN022054.D
Acq: 11 Oct 2022 16:18

Tgt Ion:165 Resp: 200479
Ion Ratio Lower Upper
165 100
89 66.6 52.2 78.4
121 24.1 19.5 29.3





#50

Acenaphthylene

Concen: 40.793 ng/ul

RT: 14.363 min Scan# 1934

Delta R.T. 0.000 min

Lab File: BN022054.D

Acq: 11 Oct 2022 16:18

Instrument :

BNA_N

ClientSampleId :

SSTD040241





Tgt Ion:152 Resp: 1192124

Ion Ratio Lower Upper

152 100

151 19.5 15.7 23.5

153 12.5 10.0 15.0



#51

3-Nitroaniline

Concen: 53.905 ng/ul

RT: 14.551 min Scan# 1966

Delta R.T. 0.000 min

Lab File: BN022054.D

Acq: 11 Oct 2022 16:18




Tgt Ion:138 Resp: 228830


Ion Ratio Lower Upper

138 100

108 11.5 9.1 13.7

92 126.5 97.7 146.5









#52
Acenaphthene
Concen: 44.888 ng/ul
RT: 14.704 min Scan# 1992
Delta R.T. 0.000 min
Lab File: BN022054.D
Acq: 11 Oct 2022 16:18




Instrument :
BNA_N
ClientSampleId :
SSTD040241


Tgt Ion:153 Resp: 847614
Ion Ratio Lower Upper
153 100
152 46.1 37.8 56.8
154 88.0 71.8 107.6



#53
2,4-Dinitrophenol
Concen: 67.486 ng/ul
RT: 14.757 min Scan# 2001
Delta R.T. 0.000 min
Lab File: BN022054.D
Acq: 11 Oct 2022 16:18

Tgt Ion:184 Resp: 125890
Ion Ratio Lower Upper
184 100
63 73.7 58.9 88.3
154 60.5 51.8 77.6









#55
4-Nitrophenol
Concen: 47.739 ng/ul
RT: 14.857 min Scan# 2017
Delta R.T. 0.006 min
Lab File: BN022054.D
Acq: 11 Oct 2022 16:18




Instrument :
BNA_N
ClientSampleId :
SSTD040241


Tgt Ion:109 Resp: 165354
Ion Ratio Lower Upper
109 100
139 139.6 111.0 166.4
65 140.1 118.5 177.7



#56
Dibenzofuran
Concen: 41.830 ng/ul
RT: 15.040 min Scan# 2049
Delta R.T. 0.000 min
Lab File: BN022054.D
Acq: 11 Oct 2022 16:18

Tgt Ion:168 Resp: 1134839
Ion Ratio Lower Upper
168 100
139 38.6 31.3 46.9









#57
2,4-Dinitrotoluene
Concen: 53.940 ng/ul
RT: 15.010 min Scan# 2044
Delta R.T. 0.006 min
Lab File: BN022054.D
Acq: 11 Oct 2022 16:18




Instrument :
BNA_N
ClientSampleId :
SSTD040241


Tgt Ion:165 Resp: 298510
Ion Ratio Lower Upper
165 100
63 47.9 40.6 60.8
182 2.7 1.9 2.9



#58
2,3,4,6-Tetrachlorophenol
Concen: 52.449 ng/ul
RT: 15.263 min Scan# 2087
Delta R.T. 0.000 min
Lab File: BN022054.D
Acq: 11 Oct 2022 16:18

Tgt Ion:232 Resp: 223269
Ion Ratio Lower Upper
232 100
131 55.5 44.0 66.0
130 2.8 2.3 3.5
166 34.0 27.8 41.8





#59

Diethylphthalate

Concen: 45.567 ng/ul

RT: 15.463 min Scan# 21

Delta R.T. 0.006 min

Lab File: BN022054.D

Acq: 11 Oct 2022 16:18

Instrument :

BNA_N

ClientSampleId :

SSTD040241





Tgt Ion:149 Resp: 1020647

Ion Ratio Lower Upper

149 100

177 21.6 15.9 23.9

150 13.0 9.8 14.8



#61

Fluorene

Concen: 43.731 ng/ul

RT: 15.693 min Scan# 2160

Delta R.T. 0.006 min

Lab File: BN022054.D

Acq: 11 Oct 2022 16:18




Tgt Ion:166 Resp: 935704


Ion Ratio Lower Upper

166 100

165 94.8 74.1 111.1

167 13.2 10.5 15.7









#62
4-Chlorophenyl-phenylether
Concen: 42.947 ng/ul
RT: 15.681 min Scan# 2158
Delta R.T. 0.000 min
Lab File: BN022054.D
Acq: 11 Oct 2022 16:18




Instrument :
BNA_N
ClientSampleId :
SSTD040241


Tgt Ion:204 Resp: 423613
Ion Ratio Lower Upper
204 100
206 31.1 26.1 39.1
141 62.8 50.9 76.3



#63
4-Nitroaniline
Concen: 53.457 ng/ul
RT: 15.716 min Scan# 2164
Delta R.T. 0.000 min
Lab File: BN022054.D
Acq: 11 Oct 2022 16:18

Tgt Ion:138 Resp: 227733
Ion Ratio Lower Upper
138 100
92 53.4 42.8 64.2
108 77.1 60.1 90.1









#66
4,6-Dinitro-2-methylphenol
Concen: 61.164 ng/ul
RT: 15.769 min Scan# 2173
Delta R.T. 0.000 min
Lab File: BN022054.D
Acq: 11 Oct 2022 16:18




Instrument :
BNA_N
ClientSampleId :
SSTD040241


Tgt Ion:198 Resp: 173230
Ion Ratio Lower Upper
198 100
182 4.0 3.9 5.9
77 29.9 24.5 36.7



#67
N-Nitrosodiphenylamine
Concen: 43.760 ng/ul
RT: 15.898 min Scan# 2195
Delta R.T. 0.000 min
Lab File: BN022054.D
Acq: 11 Oct 2022 16:18

Tgt Ion:169 Resp: 819764
Ion Ratio Lower Upper
169 100
168 65.7 51.4 77.0
167 36.0 27.8 41.8









#68
4-Bromophenyl-phenylether
Concen: 43.567 ng/ul
RT: 16.581 min Scan# 2311
Delta R.T. 0.000 min
Lab File: BN022054.D
Acq: 11 Oct 2022 16:18




Instrument :
BNA_N
ClientSampleId :
SSTD040241


Tgt Ion:248 Resp: 258976
Ion Ratio Lower Upper
248 100
250 99.6 77.6 116.4
141 82.7 66.6 99.8



#69
Hexachlorobenzene
Concen: 40.930 ng/ul
RT: 16.692 min Scan# 2330
Delta R.T. 0.000 min
Lab File: BN022054.D
Acq: 11 Oct 2022 16:18

Tgt Ion:284 Resp: 289624
Ion Ratio Lower Upper
284 100
142 43.9 32.6 48.8
249 32.3 23.4 35.2









#70
 Atrazine
 Concen: 43.131 ng/ul
 RT: 16.857 min Scan# 21
 Delta R.T. 0.006 min
 Lab File: BN022054.D
 Acq: 11 Oct 2022 16:18




Instrument :
 BNA_N
 ClientSampleId :
 SSTD040241


Tgt Ion:200 Resp: 285906
 Ion Ratio Lower Upper
 200 100
 173 28.3 22.6 34.0
 215 48.2 38.4 57.6



#71
 Pentachlorophenol
 Concen: 46.524 ng/ul
 RT: 17.045 min Scan# 2390
 Delta R.T. 0.006 min
 Lab File: BN022054.D
 Acq: 11 Oct 2022 16:18

Tgt Ion:266 Resp: 183282
 Ion Ratio Lower Upper
 266 100
 264 62.8 53.8 80.6
 268 60.1 45.1 67.7









#72
Phenanthrene
Concen: 43.313 ng/ul
RT: 17.439 min Scan# 2456
Delta R.T. 0.006 min
Lab File: BN022054.D
Acq: 11 Oct 2022 16:18




Instrument :
BNA_N
ClientSampleId :
SSTD040241


Tgt Ion:178 Resp: 1529127
Ion Ratio Lower Upper
178 100
179 15.6 12.3 18.5
176 18.6 15.9 23.9



#74
Anthracene
Concen: 43.849 ng/ul
RT: 17.528 min Scan# 2472
Delta R.T. 0.000 min
Lab File: BN022054.D
Acq: 11 Oct 2022 16:18

Tgt Ion:178 Resp: 1541470
Ion Ratio Lower Upper
178 100
179 16.2 12.2 18.2
176 18.8 15.2 22.8









#75
1,2,3,4-Tetrachlorobenzene
Concen: 40.344 ng/uL
RT: 13.440 min Scan# 1777
Delta R.T. 0.000 min
Lab File: BN022054.D
Acq: 11 Oct 2022 16:18




Instrument :
BNA_N
ClientSampleId :
SSTD040241


Tgt Ion:216 Resp: 343204
Ion Ratio Lower Upper
216 100
214 81.0 60.3 90.5
218 48.8 37.2 55.8



#76
Pentachlorobenzene
Concen: 46.243 ng/uL
RT: 14.957 min Scan# 2035
Delta R.T. 0.000 min
Lab File: BN022054.D
Acq: 11 Oct 2022 16:18

Tgt Ion:250 Resp: 374745
Ion Ratio Lower Upper
250 100
248 59.6 52.3 78.5
252 64.1 50.0 75.0





#77

Carbazole

Concen: 44.076 ng/ul

RT: 17.804 min Scan# 2518

Delta R.T. 0.006 min

Lab File: BN022054.D

Acq: 11 Oct 2022 16:18

Instrument :

BNA_N

ClientSampleId :

SSTD040241





Tgt Ion:167 Resp: 1464252

Ion Ratio Lower Upper

167 100

166 22.0 17.2 25.8

139 14.4 11.0 16.4



#78

Di-n-butylphthalate

Concen: 49.744 ng/ul

RT: 18.363 min Scan# 2614

Delta R.T. 0.000 min

Lab File: BN022054.D

Acq: 11 Oct 2022 16:18




Tgt Ion:149 Resp: 1863121


Ion Ratio Lower Upper

149 100

150 9.1 6.7 10.1

104 5.3 4.3 6.5





#80

Fluoranthene

Concen: 47.525 ng/ul

RT: 19.457 min Scan# 2800

Delta R.T. 0.000 min

Lab File: BN022054.D

Acq: 11 Oct 2022 16:18

Instrument :

BNA_N

ClientSampleId :

SSTD040241





Tgt Ion:202 Resp: 1777817

Ion Ratio Lower Upper

202 100

101 14.6 11.4 17.0

100 10.8 8.6 13.0



#82

Pyrene

Concen: 46.842 ng/ul

RT: 19.822 min Scan# 2862

Delta R.T. 0.000 min

Lab File: BN022054.D

Acq: 11 Oct 2022 16:18




Tgt Ion:202 Resp: 1826322


Ion Ratio Lower Upper

202 100

101 16.7 13.8 20.6

100 14.1 11.2 16.8





#83

Butylbenzylphthalate

Concen: 58.990 ng/ul

RT: 20.722 min Scan# 3015

Delta R.T. 0.000 min

Lab File: BN022054.D

Acq: 11 Oct 2022 16:18

Instrument :

BNA_N

ClientSampleId :

SSTD040241





Tgt Ion:149 Resp: 870516

Ion Ratio Lower Upper

149 100

91 75.1 60.9 91.3

206 17.8 14.9 22.3



#84

3,3'-Dichlorobenzidine

Concen: 50.801 ng/ul

RT: 21.516 min Scan# 3150

Delta R.T. 0.006 min

Lab File: BN022054.D

Acq: 11 Oct 2022 16:18




Tgt Ion:252 Resp: 631673


Ion Ratio Lower Upper

252 100

254 59.9 48.6 72.8

126 12.2 11.1 16.7





#85

Benzo(a)anthracene

Concen: 44.731 ng/ul

RT: 21.580 min Scan# 3161

Delta R.T. 0.006 min

Lab File: BN022054.D

Acq: 11 Oct 2022 16:18

Instrument :

BNA_N

ClientSampleId :

SSTD040241





Tgt Ion:228 Resp: 1756388

Ion Ratio Lower Upper

228 100

229 19.9 16.7 25.1

226 27.7 22.4 33.6



#86

Bis(2-ethylhexyl)phthalate

Concen: 55.748 ng/ul

RT: 21.498 min Scan# 3147

Delta R.T. 0.000 min

Lab File: BN022054.D

Acq: 11 Oct 2022 16:18




Tgt Ion:149 Resp: 1259605


Ion Ratio Lower Upper

149 100

167 27.3 21.8 32.6

279 3.3 3.0 4.4





#87

Chrysene

Concen: 44.754 ng/ul

RT: 21.633 min Scan# 3170

Delta R.T. 0.000 min

Lab File: BN022054.D

Acq: 11 Oct 2022 16:18

Instrument :

BNA_N

ClientSampleId :

SSTD040241





Tgt Ion:228 Resp: 1701406

Ion Ratio Lower Upper

228 100

226 32.3 22.9 34.3

229 20.3 15.5 23.3



#89

Di-n-octyl phthalate

Concen: 56.899 ng/ul




RT: 22.445 min Scan# 3308


Delta R.T. 0.000 min

Lab File: BN022054.D

Acq: 11 Oct 2022 16:18

Tgt Ion:149 Resp: 2324058





#90

Benzo(b)fluoranthene

Concen: 45.359 ng/ul

RT: 23.310 min Scan# 3455

Delta R.T. 0.006 min

Lab File: BN022054.D

Acq: 11 Oct 2022 16:18

Instrument :

BNA_N

ClientSampleId :

SSTD040241





Tgt Ion:252 Resp: 1924236

Ion Ratio Lower Upper

252 100

253 23.4 15.7 23.5

125 13.0 10.0 15.0



#91

Benzo(k)fluoranthene

Concen: 48.926 ng/ul

RT: 23.363 min Scan# 3464

Delta R.T. 0.006 min

Lab File: BN022054.D

Acq: 11 Oct 2022 16:18




Tgt Ion:252 Resp: 1856658


Ion Ratio Lower Upper

252 100

253 20.6 16.8 25.2

125 12.9 10.4 15.6





#93

Benzo(a)pyrene

Concen: 43.182 ng/ul

RT: 23.963 min Scan# 3565

Delta R.T. 0.006 min

Lab File: BN022054.D

Acq: 11 Oct 2022 16:18

Instrument :

BNA_N

ClientSampleId :

SSTD040241





Tgt Ion:252 Resp: 1752711

Ion Ratio Lower Upper

252 100

253 22.0 16.8 25.2

125 13.7 11.2 16.8



#94

Indeno(1,2,3-cd)pyrene

Concen: 43.394 ng/ul

RT: 26.680 min Scan# 4028

Delta R.T. 0.012 min

Lab File: BN022054.D

Acq: 11 Oct 2022 16:18




Tgt Ion:276 Resp: 2162300


Ion Ratio Lower Upper

276 100

138 30.6 25.4 38.0

277 23.0 18.8 28.2









#95
Dibenzo(a,h)anthracene
Concen: 47.987 ng/ul
RT: 26.698 min Scan# 4027
Delta R.T. 0.024 min
Lab File: BN022054.D
Acq: 11 Oct 2022 16:18

Instrument :
BNA_N
ClientSampleId :
SSTD040241

Tgt Ion:278 Resp: 2017956
Ion Ratio Lower Upper
278 100
139 19.4 17.4 26.2
279 22.8 18.6 28.0



#96
Benzo(g,h,i)perylene
Concen: 44.467 ng/ul
RT: 27.474 min Scan# 4163
Delta R.T. 0.006 min
Lab File: BN022054.D
Acq: 11 Oct 2022 16:18

Tgt Ion:276 Resp: 1933532
Ion Ratio Lower Upper
276 100
138 27.8 19.4 29.2
277 23.6 16.5 24.7

