

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN101122\
 Data File : BN022056.D
 Acq On : 11 Oct 2022 17:35
 Operator : CG/JU
 Sample : SSTD16034
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTD160243

Quant Time: Oct 12 03:08:57 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN101122.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Oct 12 03:00:49 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.969	152	117187	20.000	ng/u1	0.00
20) Naphthalene-d8	10.799	136	564750	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.646	164	353820	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.398	188	726276	20.000	ng/u1	0.00
79) Chrysene-d12	21.604	240	605657	20.000	ng/u1	0.01
88) Perylene-d12	24.080	264	849435	20.000	ng/u1	0.02
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.258	96	182466	58.461	ng/uL	0.00
4) Pyridine-d5	3.699	84	1536672	180.644	ng/u1	0.00
7) Phenol-d5	7.134	99	1928572	181.454	ng/u1	0.01
9) Bis-(2-Chloroethyl)eth...	7.299	67	1094188	170.075	ng/u1	0.00
11) 2-Chlorophenol-d4	7.499	132	1381604	178.670	ng/u1	0.00
15) 4-Methylphenol-d8	8.705	113	1460389	175.118	ng/u1	0.02
21) Nitrobenzene-d5	9.158	128	731620	190.143	ng/u1	0.01
24) 2-Nitrophenol-d4	9.881	143	795010	231.326	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.422	165	1298811	170.643	ng/u1	0.00
31) 4-Chloroaniline-d4	10.952	131	1961086	151.714	ng/u1	0.01
46) Dimethylphthalate-d6	14.063	166	3635226	150.319	ng/u1	0.02
49) Acenaphthylene-d8	14.340	160	3887968	137.904	ng/u1	0.00
54) 4-Nitrophenol-d4	14.875	143	817767	169.930	ng/u1	0.04
60) Fluorene-d10	15.640	176	2699301	128.276	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.775	200	670133	230.685	ng/u1	0.02
73) Anthracene-d10	17.504	188	4032792	126.002	ng/u1	0.01
81) Pyrene-d10	19.804	212	4361049	151.410	ng/u1	0.01
92) Benzo(a)pyrene-d12	23.933	264	5570026	138.344	ng/u1	0.03
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.293	88	193256	60.121	ng/uL	88
5) Pyridine	3.717	79	1508609	176.253	ng/u1	97
6) Benzaldehyde	7.099	77	707433	153.191	ng/u1	99
8) Phenol	7.164	94	1976696	179.976	ng/u1	100
10) Bis(2-Chloroethyl)ether	7.393	93	1477096	168.669	ng/u1	98
12) 2-Chlorophenol	7.528	128	1453353	175.192	ng/u1	98
13) 2-Methylphenol	8.428	108	1460186	174.979	ng/u1	100
14) 2,2'-oxybis(1-Chloropr...	8.511	45	2117834	191.729	ng/u1	100
16) Acetophenone	8.816	105	2102411	159.398	ng/u1	99
17) N-Nitroso-di-n-propyla...	8.828	70	1130932	172.200	ng/u1	95
18) 4-Methylphenol	8.775	108	1588768	173.316	ng/u1	96
19) Hexachloroethane	9.058	117	544865	164.783	ng/u1	94
22) Nitrobenzene	9.205	77	1755127	185.104	ng/u1	96
23) Isophorone	9.746	82	3493938	178.948	ng/u1	96
25) 2-Nitrophenol	9.916	139	861484	213.992	ng/u1	97
26) 2,4-Dimethylphenol	9.975	107	1655350	166.050	ng/u1	96
27) Bis(2-Chloroethoxy)met...	10.216	93	2022484	165.410	ng/u1	99
29) 2,4-Dichlorophenol	10.452	162	1304293	163.552	ng/u1	96
30) Naphthalene	10.857	128	4327270	147.155	ng/u1	98
32) 4-Chloroaniline	10.975	127	1981603	147.927	ng/u1	97
33) Hexachlorobutadiene	11.134	225	677399	147.445	ng/u1	95
34) Caprolactam	11.810	113	292915	96.025	ng/u1	93
35) 4-Chloro-3-methylphenol	12.116	107	1587800	176.664	ng/u1	97
36) 2-Methylnaphthalene	12.469	142	2901047	145.836	ng/u1	97

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN101122\
 Data File : BN022056.D
 Acq On : 11 Oct 2022 17:35
 Operator : CG/JU
 Sample : SSTD16034
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTD160243


Quant Time: Oct 12 03:08:57 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN101122.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Oct 12 03:00:49 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.693	142	2861839	143.150	ng/ul	98
39) 1,2,4,5-Tetrachloroben...	12.840	216	1337856	147.458	ng/ul	98
40) Hexachlorocyclopentadiene	12.810	237	750268	138.897	ng/ul	93
41) 2,4,6-Trichlorophenol	13.081	196	1003469	178.252	ng/ul	99
42) 2,4,5-Trichlorophenol	13.157	196	1058531	171.374	ng/ul	98
43) 1,1'-Biphenyl	13.481	154	3343405	129.304	ng/ul	100
44) 2-Chloronaphthalene	13.522	162	2789124	140.898	ng/ul	97
45) 2-Nitroaniline	13.740	65	1122815	226.149	ng/ul	98
47) Dimethylphthalate	14.110	163	3563894	144.072	ng/ul	99
48) 2,6-Dinitrotoluene	14.240	165	866466	204.170	ng/ul	95
50) Acenaphthylene	14.369	152	4234037	130.091	ng/ul	98
51) 3-Nitroaniline	14.569	138	891375	188.539	ng/ul	99
52) Acenaphthene	14.710	153	2891333	137.485	ng/ul	96
53) 2,4-Dinitrophenol	14.769	184	608890	293.079	ng/ul	97
55) 4-Nitrophenol	14.893	109	652811	169.227	ng/ul	91
56) Dibenzofuran	15.051	168	3738843	123.740	ng/ul	98
57) 2,4-Dinitrotoluene	15.028	165	1181134	191.635	ng/ul	97
58) 2,3,4,6-Tetrachlorophenol	15.275	232	838063	176.769	ng/ul	96
59) Diethylphthalate	15.475	149	3741014	149.964	ng/ul	98
61) Fluorene	15.704	166	2659066	111.584	ng/ul	100
62) 4-Chlorophenyl-phenyle...	15.687	204	1212162	110.345	ng/ul#	87
63) 4-Nitroaniline	15.751	138	810870	170.905	ng/ul	94
66) 4,6-Dinitro-2-methylph...	15.793	198	676147	219.703	ng/ul#	97
67) N-Nitrosodiphenylamine	15.910	169	2686564	131.981	ng/ul	97
68) 4-Bromophenyl-phenylether	16.587	248	921533	142.669	ng/ul	94
69) Hexachlorobenzene	16.698	284	1034231	134.508	ng/ul	95
70) Atrazine	16.875	200	1015919	141.040	ng/ul	98
71) Pentachlorophenol	17.051	266	754793	176.322	ng/ul	95
72) Phenanthrene	17.445	178	4899329	127.711	ng/ul	99
74) Anthracene	17.539	178	4635856	121.360	ng/ul	98
75) 1,2,3,4-Tetrachloroben...	13.446	216	1348125	145.841	ng/ul	94
76) Pentachlorobenzene	14.963	250	1297142	147.307	ng/ul	99
77) Carbazole	17.816	167	4803008	133.053	ng/ul	98
78) Di-n-butylphthalate	18.369	149	6098267	149.839	ng/ul	97
80) Fluoranthene	19.463	202	5355013	154.137	ng/ul	97
82) Pyrene	19.833	202	5121921	141.451	ng/ul	96
83) Butylbenzylphthalate	20.728	149	2936084	214.231	ng/ul	95
84) 3,3'-Dichlorobenzidine	21.522	252	1811975	156.908	ng/ul	91
85) Benzo(a)anthracene	21.592	228	5332732	146.235	ng/ul	98
86) Bis(2-ethylhexyl)phtha...	21.504	149	3805056	181.332	ng/ul#	99
87) Chrysene	21.651	228	4667460	132.196	ng/ul	98
89) Di-n-octyl phthalate	22.451	149	7623678	153.218	ng/ul	100
90) Benzo(b)fluoranthene	23.333	252	7073182	136.869	ng/ul	96
91) Benzo(k)fluoranthene	23.386	252	5451032	117.915	ng/ul	98
93) Benzo(a)pyrene	23.992	252	5803235	117.368	ng/ul	98
94) Indeno(1,2,3-cd)pyrene	26.698	276	3700954	60.969	ng/ul	94
95) Dibenzo(a,h)anthracene	26.751	278	6520426	127.284	ng/ul#	94
96) Benzo(g,h,i)perylene	27.545	276	7990352	150.846	ng/ul	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Oct 12 03:08:57 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN101122.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Oct 12 03:00:49 2022
Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 60.121 ng/uL
RT: 3.293 min Scan# 51
Delta R.T. -0.006 min
Lab File: BN022056.D
Acq: 11 Oct 2022 17:35




Instrument :
BNA_N
ClientSampleId :
SSTD160243


Tgt Ion: 88 Resp: 193256
Ion Ratio Lower Upper
88 100
43 28.0 25.1 37.7
58 71.6 67.4 101.2



#5
Pyridine
Concen: 176.253 ng/uL
RT: 3.717 min Scan# 124
Delta R.T. -0.006 min
Lab File: BN022056.D
Acq: 11 Oct 2022 17:35

Tgt Ion: 79 Resp: 1508609
Ion Ratio Lower Upper
79 100
52 76.7 64.1 96.1
51 34.5 28.9 43.3









#6
Benzaldehyde
Concen: 153.191 ng/ul
RT: 7.099 min Scan# 699
Delta R.T. 0.000 min
Lab File: BN022056.D
Acq: 11 Oct 2022 17:35




Instrument :
BNA_N
ClientSampleId :
SSTD160243


Tgt Ion: 77 Resp: 707433
Ion Ratio Lower Upper
77 100
105 96.2 77.0 115.6
106 92.3 72.3 108.5



#8
Phenol
Concen: 179.976 ng/ul
RT: 7.164 min Scan# 710
Delta R.T. 0.012 min
Lab File: BN022056.D
Acq: 11 Oct 2022 17:35

Tgt Ion: 94 Resp: 1976696
Ion Ratio Lower Upper
94 100
65 28.9 22.8 34.2
66 36.8 29.3 43.9









#10
Bis(2-Chloroethyl)ether
Concen: 168.669 ng/ul
RT: 7.393 min Scan# 749
Delta R.T. 0.006 min
Lab File: BN022056.D
Acq: 11 Oct 2022 17:35




Instrument :
BNA_N
ClientSampleId :
SSTD160243


Tgt Ion: 93 Resp: 1477096
Ion Ratio Lower Upper
93 100
63 76.7 63.2 94.8
95 32.2 25.9 38.9



#12
2-Chlorophenol
Concen: 175.192 ng/ul
RT: 7.528 min Scan# 772
Delta R.T. 0.006 min
Lab File: BN022056.D
Acq: 11 Oct 2022 17:35

Tgt Ion:128 Resp: 1453353
Ion Ratio Lower Upper
128 100
64 49.2 40.9 61.3
130 31.7 25.9 38.9





#13

2-Methylphenol

Concen: 174.979 ng/ul

RT: 8.428 min Scan# 91

Delta R.T. 0.012 min

Lab File: BN022056.D

Acq: 11 Oct 2022 17:35

Instrument :

BNA_N

ClientSampleId :





SSTD160243

Tgt Ion:108 Resp: 1460186

Ion Ratio Lower Upper

108 100

107 89.9 71.8 107.6



#14

2,2'-oxybis(1-Chloropropane)

Concen: 191.729 ng/ul

RT: 8.511 min Scan# 939

Delta R.T. 0.000 min

Lab File: BN022056.D

Acq: 11 Oct 2022 17:35




Tgt Ion: 45 Resp: 2117834


Ion Ratio Lower Upper

45 100

77 14.2 11.3 16.9

79 11.0 8.6 12.8









#16
 Acetophenone
 Concen: 159.398 ng/ul
 RT: 8.816 min Scan# 991
 Delta R.T. 0.012 min
 Lab File: BN022056.D
 Acq: 11 Oct 2022 17:35




Instrument :
 BNA_N
 ClientSampleId :
 SSTD160243


Tgt Ion:105 Resp: 2102411
 Ion Ratio Lower Upper
 105 100
 77 80.3 63.4 95.0
 51 26.9 21.9 32.9



#17
 N-Nitroso-di-n-propylamine
 Concen: 172.200 ng/ul
 RT: 8.828 min Scan# 993
 Delta R.T. 0.041 min
 Lab File: BN022056.D
 Acq: 11 Oct 2022 17:35

Tgt Ion: 70 Resp: 1130932
 Ion Ratio Lower Upper
 70 100
 42 52.8 46.2 69.4
 101 9.7 7.1 10.7
 130 20.4 15.2 22.8









#18
4-Methylphenol
Concen: 173.316 ng/ul
RT: 8.775 min Scan# 980
Delta R.T. 0.024 min
Lab File: BN022056.D
Acq: 11 Oct 2022 17:35




Instrument :
BNA_N
ClientSampleId :
SSTD160243


Tgt Ion:108 Resp: 1588768
Ion Ratio Lower Upper
108 100
107 112.9 87.3 130.9



#19
Hexachloroethane
Concen: 164.783 ng/ul
RT: 9.058 min Scan# 1032
Delta R.T. 0.000 min
Lab File: BN022056.D
Acq: 11 Oct 2022 17:35

Tgt Ion:117 Resp: 544865
Ion Ratio Lower Upper
117 100
201 86.8 76.8 115.2
199 56.6 46.2 69.2





#22

Nitrobenzene

Concen: 185.104 ng/ul

RT: 9.205 min Scan# 1105

Delta R.T. 0.018 min

Lab File: BN022056.D

Acq: 11 Oct 2022 17:35

Instrument :

BNA_N

ClientSampleId :

SSTD160243





Tgt Ion: 77 Resp: 1755127

Ion Ratio Lower Upper

77 100

123 45.5 33.8 50.8

65 13.4 10.5 15.7



#23

Isophorone

Concen: 178.948 ng/ul

RT: 9.746 min Scan# 1149

Delta R.T. 0.030 min

Lab File: BN022056.D

Acq: 11 Oct 2022 17:35




Tgt Ion: 82 Resp: 3493938


Ion Ratio Lower Upper

82 100

95 6.9 5.1 7.7

138 17.5 12.6 18.8









#25
2-Nitrophenol
Concen: 213.992 ng/ul
RT: 9.916 min Scan# 1178
Delta R.T. 0.012 min
Lab File: BN022056.D
Acq: 11 Oct 2022 17:35




Instrument :
BNA_N
ClientSampleId :
SSTD160243


Tgt Ion:139 Resp: 861484
Ion Ratio Lower Upper
139 100
65 52.8 44.9 67.3
109 29.7 23.5 35.3



#26
2,4-Dimethylphenol
Concen: 166.050 ng/ul
RT: 9.975 min Scan# 1188
Delta R.T. 0.012 min
Lab File: BN022056.D
Acq: 11 Oct 2022 17:35

Tgt Ion:107 Resp: 1655350
Ion Ratio Lower Upper
107 100
121 50.1 41.0 61.4
122 84.9 71.8 107.6









#27
 Bis(2-Chloroethoxy)methane
 Concen: 165.410 ng/ul
 RT: 10.216 min Scan# 1229
 Delta R.T. 0.012 min
 Lab File: BN022056.D
 Acq: 11 Oct 2022 17:35




Instrument :
 BNA_N
 ClientSampleId :
 SSTD160243


Tgt Ion: 93 Resp: 2022484
 Ion Ratio Lower Upper
 93 100
 95 31.7 25.0 37.6
 123 10.9 8.6 13.0



#29
 2,4-Dichlorophenol
 Concen: 163.552 ng/ul
 RT: 10.452 min Scan# 1269
 Delta R.T. 0.012 min
 Lab File: BN022056.D
 Acq: 11 Oct 2022 17:35

Tgt Ion:162 Resp: 1304293
 Ion Ratio Lower Upper
 162 100
 164 62.6 53.9 80.9
 98 40.0 32.5 48.7









#30
Naphthalene
Concen: 147.155 ng/ul
RT: 10.857 min Scan# 1338
Delta R.T. 0.012 min
Lab File: BN022056.D
Acq: 11 Oct 2022 17:35




Instrument :
BNA_N
ClientSampleId :
SSTD160243


Tgt Ion:128 Resp: 4327270
Ion Ratio Lower Upper
128 100
129 11.2 9.1 13.7
127 12.6 11.0 16.4



#32
4-Chloroaniline
Concen: 147.927 ng/ul
RT: 10.975 min Scan# 1358
Delta R.T. 0.012 min
Lab File: BN022056.D
Acq: 11 Oct 2022 17:35

Tgt Ion:127 Resp: 1981603
Ion Ratio Lower Upper
127 100
129 31.9 24.2 36.4





#33

Hexachlorobutadiene

Concen: 147.445 ng/ul

RT: 11.134 min Scan# 11

Delta R.T. 0.006 min

Lab File: BN022056.D

Acq: 11 Oct 2022 17:35

Instrument :

BNA_N

ClientSampleId :

SSTD160243





Tgt Ion:225 Resp: 677399

Ion Ratio Lower Upper

225 100

223 66.8 53.0 79.6

227 66.2 46.9 70.3



#34

Caprolactam

Concen: 96.025 ng/ul

RT: 11.810 min Scan# 1500

Delta R.T. 0.065 min

Lab File: BN022056.D

Acq: 11 Oct 2022 17:35




Tgt Ion:113 Resp: 292915


Ion Ratio Lower Upper

113 100

55 169.9 143.4 215.0

56 123.3 104.6 157.0









#35
 4-Chloro-3-methylphenol
 Concen: 176.664 ng/ul
 RT: 12.116 min Scan# 11
 Delta R.T. 0.030 min
 Lab File: BN022056.D
 Acq: 11 Oct 2022 17:35




Instrument :
 BNA_N
 ClientSampleId :
 SSTD160243


Tgt Ion:107 Resp: 1587800
 Ion Ratio Lower Upper
 107 100
 144 25.5 20.7 31.1
 142 77.7 59.8 89.8



#36
 2-Methylnaphthalene
 Concen: 145.836 ng/ul
 RT: 12.469 min Scan# 1612
 Delta R.T. 0.006 min
 Lab File: BN022056.D
 Acq: 11 Oct 2022 17:35

Tgt Ion:142 Resp: 2901047
 Ion Ratio Lower Upper
 142 100
 141 89.0 68.7 103.1









#37
1-Methylnaphthalene
Concen: 143.150 ng/ul
RT: 12.693 min Scan# 1650
Delta R.T. 0.012 min
Lab File: BN022056.D
Acq: 11 Oct 2022 17:35




Instrument :
BNA_N
ClientSampleId :
SSTD160243


Tgt Ion:142 Resp: 2861839
Ion Ratio Lower Upper
142 100
141 90.0 73.3 109.9
116 4.3 3.0 4.6



#39
1,2,4,5-Tetrachlorobenzene
Concen: 147.458 ng/ul
RT: 12.840 min Scan# 1675
Delta R.T. 0.012 min
Lab File: BN022056.D
Acq: 11 Oct 2022 17:35

Tgt Ion:216 Resp: 1337856
Ion Ratio Lower Upper
216 100
214 77.4 60.4 90.6
179 21.9 18.4 27.6
108 21.1 17.8 26.6









#40
Hexachlorocyclopentadiene
Concen: 138.897 ng/ul
RT: 12.810 min Scan# 1669
Delta R.T. 0.006 min
Lab File: BN022056.D
Acq: 11 Oct 2022 17:35




Instrument :
BNA_N
ClientSampleId :
SSTD160243


Tgt Ion:237 Resp: 750268
Ion Ratio Lower Upper
237 100
235 63.4 56.0 84.0
272 13.1 10.6 15.8



#41
2,4,6-Trichlorophenol
Concen: 178.252 ng/ul
RT: 13.081 min Scan# 1716
Delta R.T. 0.012 min
Lab File: BN022056.D
Acq: 11 Oct 2022 17:35

Tgt Ion:196 Resp: 1003469
Ion Ratio Lower Upper
196 100
198 93.1 74.3 111.5
200 30.4 22.3 33.5





#42

2,4,5-Trichlorophenol

Concen: 171.374 ng/ul

RT: 13.157 min Scan# 1

Delta R.T. 0.018 min

Lab File: BN022056.D

Acq: 11 Oct 2022 17:35

Instrument :

BNA_N

ClientSampleId :

SSTD160243





Tgt Ion:196 Resp: 1058531

Ion Ratio Lower Upper

196 100

198 97.6 76.8 115.2

200 28.1 23.9 35.9



#43

1,1'-Biphenyl

Concen: 129.304 ng/ul

RT: 13.481 min Scan# 1784

Delta R.T. 0.006 min

Lab File: BN022056.D

Acq: 11 Oct 2022 17:35




Tgt Ion:154 Resp: 3343405


Ion Ratio Lower Upper

154 100

153 38.9 31.0 46.6

76 14.6 11.6 17.4









#44
2-Chloronaphthalene
Concen: 140.898 ng/ul
RT: 13.522 min Scan# 1791
Delta R.T. 0.006 min
Lab File: BN022056.D
Acq: 11 Oct 2022 17:35




Instrument :
BNA_N
ClientSampleId :
SSTD160243


Tgt Ion:162 Resp: 2789124
Ion Ratio Lower Upper
162 100
164 33.2 25.9 38.9
127 42.8 32.2 48.4



#45
2-Nitroaniline
Concen: 226.149 ng/ul
RT: 13.740 min Scan# 1828
Delta R.T. 0.012 min
Lab File: BN022056.D
Acq: 11 Oct 2022 17:35

Tgt Ion: 65 Resp: 1122815
Ion Ratio Lower Upper
65 100
92 69.4 54.6 82.0
138 102.2 80.4 120.6









#47
Dimethylphthalate
Concen: 144.072 ng/ul
RT: 14.110 min Scan# 1889
Delta R.T. 0.012 min
Lab File: BN022056.D
Acq: 11 Oct 2022 17:35




Instrument :
BNA_N
ClientSampleId :
SSTD160243


Tgt Ion:163 Resp: 3563894
Ion Ratio Lower Upper
163 100
194 4.0 3.1 4.7
164 10.7 8.0 12.0



#48
2,6-Dinitrotoluene
Concen: 204.170 ng/ul
RT: 14.240 min Scan# 1913
Delta R.T. 0.018 min
Lab File: BN022056.D
Acq: 11 Oct 2022 17:35

Tgt Ion:165 Resp: 866466
Ion Ratio Lower Upper
165 100
89 60.8 52.2 78.4
121 22.1 19.5 29.3









#50
Acenaphthylene
Concen: 130.091 ng/ul
RT: 14.369 min Scan# 1934
Delta R.T. 0.006 min
Lab File: BN022056.D
Acq: 11 Oct 2022 17:35




Instrument :
BNA_N
ClientSampleId :
SSTD160243


Tgt Ion:152 Resp: 4234037
Ion Ratio Lower Upper
152 100
151 20.6 15.7 23.5
153 13.2 10.0 15.0



#51
3-Nitroaniline
Concen: 188.539 ng/ul
RT: 14.569 min Scan# 1969
Delta R.T. 0.018 min
Lab File: BN022056.D
Acq: 11 Oct 2022 17:35

Tgt Ion:138 Resp: 891375
Ion Ratio Lower Upper
138 100
108 10.9 9.1 13.7
92 120.8 97.7 146.5









#52
Acenaphthene
Concen: 137.485 ng/ul
RT: 14.710 min Scan# 1992
Delta R.T. 0.006 min
Lab File: BN022056.D
Acq: 11 Oct 2022 17:35




Instrument :
BNA_N
ClientSampleId :
SSTD160243


Tgt Ion:153 Resp: 2891333
Ion Ratio Lower Upper
153 100
152 49.1 37.8 56.8
154 94.3 71.8 107.6



#53
2,4-Dinitrophenol
Concen: 293.079 ng/ul
RT: 14.769 min Scan# 2003
Delta R.T. 0.012 min
Lab File: BN022056.D
Acq: 11 Oct 2022 17:35

Tgt Ion:184 Resp: 608890
Ion Ratio Lower Upper
184 100
63 75.6 58.9 88.3
154 67.7 51.8 77.6









#55
4-Nitrophenol
Concen: 169.227 ng/ul
RT: 14.893 min Scan# 2017
Delta R.T. 0.041 min
Lab File: BN022056.D
Acq: 11 Oct 2022 17:35




Instrument :
BNA_N
ClientSampleId :
SSTD160243


Tgt Ion:109 Resp: 652811
Ion Ratio Lower Upper
109 100
139 135.9 111.0 166.4
65 129.8 118.5 177.7



#56
Dibenzofuran
Concen: 123.740 ng/ul
RT: 15.051 min Scan# 2051
Delta R.T. 0.012 min
Lab File: BN022056.D
Acq: 11 Oct 2022 17:35

Tgt Ion:168 Resp: 3738843
Ion Ratio Lower Upper
168 100
139 40.1 31.3 46.9





#57

2,4-Dinitrotoluene

Concen: 191.635 ng/ul

RT: 15.028 min Scan# 2047

Delta R.T. 0.024 min

Lab File: BN022056.D

Acq: 11 Oct 2022 17:35

Instrument :

BNA_N

ClientSampleId :

SSTD160243





Tgt Ion:165 Resp: 1181134

Ion Ratio Lower Upper

165 100

63 48.5 40.6 60.8

182 2.8 1.9 2.9



#58

2,3,4,6-Tetrachlorophenol

Concen: 176.769 ng/ul

RT: 15.275 min Scan# 2089

Delta R.T. 0.012 min

Lab File: BN022056.D

Acq: 11 Oct 2022 17:35

Tgt Ion:232 Resp: 838063




Ion Ratio Lower Upper


232 100

131 57.9 44.0 66.0

130 3.0 2.3 3.5

166 36.6 27.8 41.8









#59
Diethylphthalate
Concen: 149.964 ng/ul
RT: 15.475 min Scan# 2120
Delta R.T. 0.018 min
Lab File: BN022056.D
Acq: 11 Oct 2022 17:35




Instrument :
BNA_N
ClientSampleId :
SSTD160243


Tgt Ion:149 Resp: 3741014
Ion Ratio Lower Upper
149 100
177 20.8 15.9 23.9
150 12.6 9.8 14.8



#61
Fluorene
Concen: 111.584 ng/ul
RT: 15.704 min Scan# 2162
Delta R.T. 0.018 min
Lab File: BN022056.D
Acq: 11 Oct 2022 17:35

Tgt Ion:166 Resp: 2659066
Ion Ratio Lower Upper
166 100
165 92.5 74.1 111.1
167 12.6 10.5 15.7









#62
4-Chlorophenyl-phenylether
Concen: 110.345 ng/ul
RT: 15.687 min Scan# 2158
Delta R.T. 0.006 min
Lab File: BN022056.D
Acq: 11 Oct 2022 17:35




Instrument :
BNA_N
ClientSampleId :
SSTD160243


Tgt Ion:204 Resp: 1212162
Ion Ratio Lower Upper
204 100
206 33.8 26.1 39.1
141 78.3 50.9 76.3#



#63
4-Nitroaniline
Concen: 170.905 ng/ul
RT: 15.751 min Scan# 2170
Delta R.T. 0.035 min
Lab File: BN022056.D
Acq: 11 Oct 2022 17:35

Tgt Ion:138 Resp: 810870
Ion Ratio Lower Upper
138 100
92 52.9 42.8 64.2
108 82.9 60.1 90.1









#66
4,6-Dinitro-2-methylphenol
Concen: 219.703 ng/ul
RT: 15.793 min Scan# 2177
Delta R.T. 0.024 min
Lab File: BN022056.D
Acq: 11 Oct 2022 17:35




Instrument :
BNA_N
ClientSampleId :
SSTD160243


Tgt Ion:198 Resp: 676147
Ion Ratio Lower Upper
198 100
182 3.4 3.9 5.9#
77 29.0 24.5 36.7



#67
N-Nitrosodiphenylamine
Concen: 131.981 ng/ul
RT: 15.910 min Scan# 2197
Delta R.T. 0.012 min
Lab File: BN022056.D
Acq: 11 Oct 2022 17:35

Tgt Ion:169 Resp: 2686564
Ion Ratio Lower Upper
169 100
168 67.1 51.4 77.0
167 36.6 27.8 41.8









#68
4-Bromophenyl-phenylether
Concen: 142.669 ng/ul
RT: 16.587 min Scan# 211
Delta R.T. 0.006 min
Lab File: BN022056.D
Acq: 11 Oct 2022 17:35




Instrument :
BNA_N
ClientSampleId :
SSTD160243


Tgt Ion:248 Resp: 921533
Ion Ratio Lower Upper
248 100
250 99.7 77.6 116.4
141 92.9 66.6 99.8



#69
Hexachlorobenzene
Concen: 134.508 ng/ul
RT: 16.698 min Scan# 2331
Delta R.T. 0.006 min
Lab File: BN022056.D
Acq: 11 Oct 2022 17:35

Tgt Ion:284 Resp: 1034231
Ion Ratio Lower Upper
284 100
142 43.7 32.6 48.8
249 31.9 23.4 35.2





#70

Atrazine

Concen: 141.040 ng/ul

RT: 16.875 min Scan# 21

Delta R.T. 0.024 min

Lab File: BN022056.D

Acq: 11 Oct 2022 17:35

Instrument :

BNA_N

ClientSampleId :

SSTD160243





Tgt Ion:200 Resp: 1015919

Ion Ratio Lower Upper

200 100

173 26.8 22.6 34.0

215 46.8 38.4 57.6



#71

Pentachlorophenol

Concen: 176.322 ng/ul

RT: 17.051 min Scan# 2391

Delta R.T. 0.012 min

Lab File: BN022056.D

Acq: 11 Oct 2022 17:35




Tgt Ion:266 Resp: 754793


Ion Ratio Lower Upper

266 100

264 64.7 53.8 80.6

268 61.5 45.1 67.7









#72
 Phenanthrene
 Concen: 127.711 ng/uI
 RT: 17.445 min Scan# 2458
 Delta R.T. 0.012 min
 Lab File: BN022056.D
 Acq: 11 Oct 2022 17:35

Instrument :
 BNA_N
 ClientSampleId :
 SSTD160243

Tgt Ion:178 Resp: 4899329




Ion	Ratio	Lower	Upper
178	100		
179	15.9	12.3	18.5
176	19.8	15.9	23.9




#74
 Anthracene
 Concen: 121.360 ng/uI
 RT: 17.539 min Scan# 2474
 Delta R.T. 0.012 min
 Lab File: BN022056.D
 Acq: 11 Oct 2022 17:35

Tgt Ion:178 Resp: 4635856

Ion	Ratio	Lower	Upper
178	100		
179	16.1	12.2	18.2
176	19.5	15.2	22.8





#75

1,2,3,4-Tetrachlorobenzene

Concen: 145.841 ng/uL

RT: 13.446 min Scan# 1

Delta R.T. 0.006 min

Lab File: BN022056.D

Acq: 11 Oct 2022 17:35

Instrument :

BNA_N

ClientSampleId :

SSTD160243





Tgt Ion:216 Resp: 1348125

Ion Ratio Lower Upper

216 100

214 80.7 60.3 90.5

218 50.6 37.2 55.8



#76

Pentachlorobenzene

Concen: 147.307 ng/uL

RT: 14.963 min Scan# 2036

Delta R.T. 0.006 min

Lab File: BN022056.D

Acq: 11 Oct 2022 17:35




Tgt Ion:250 Resp: 1297142


Ion Ratio Lower Upper

250 100

248 65.0 52.3 78.5

252 64.3 50.0 75.0









#77
 Carbazole
 Concen: 133.053 ng/ul
 RT: 17.816 min Scan# 2521
 Delta R.T. 0.018 min
 Lab File: BN022056.D
 Acq: 11 Oct 2022 17:35




Instrument :
 BNA_N
 ClientSampleId :
 SSTD160243


Tgt Ion:167 Resp: 4803008
 Ion Ratio Lower Upper
 167 100
 166 22.2 17.2 25.8
 139 14.9 11.0 16.4



#78
 Di-n-butylphthalate
 Concen: 149.839 ng/ul
 RT: 18.369 min Scan# 2615
 Delta R.T. 0.006 min
 Lab File: BN022056.D
 Acq: 11 Oct 2022 17:35

Tgt Ion:149 Resp: 6098267
 Ion Ratio Lower Upper
 149 100
 150 9.3 6.7 10.1
 104 6.3 4.3 6.5









#80
Fluoranthene
Concen: 154.137 ng/ul
RT: 19.463 min Scan# 2800
Delta R.T. 0.006 min
Lab File: BN022056.D
Acq: 11 Oct 2022 17:35




Instrument :
BNA_N
ClientSampleId :
SSTD160243


Tgt Ion:202 Resp: 5355013
Ion Ratio Lower Upper
202 100
101 13.0 11.4 17.0
100 10.0 8.6 13.0



#82
Pyrene
Concen: 141.451 ng/ul
RT: 19.833 min Scan# 2864
Delta R.T. 0.012 min
Lab File: BN022056.D
Acq: 11 Oct 2022 17:35

Tgt Ion:202 Resp: 5121921
Ion Ratio Lower Upper
202 100
101 15.3 13.8 20.6
100 12.5 11.2 16.8





#83

Butylbenzylphthalate

Concen: 214.231 ng/ul

RT: 20.728 min Scan# 3015

Delta R.T. 0.006 min

Lab File: BN022056.D

Acq: 11 Oct 2022 17:35

Instrument :

BNA_N

ClientSampleId :

SSTD160243





Tgt Ion:149 Resp: 2936084

Ion Ratio Lower Upper

149 100

91 80.8 60.9 91.3

206 19.9 14.9 22.3



#84

3,3'-Dichlorobenzidine

Concen: 156.908 ng/ul

RT: 21.522 min Scan# 3151

Delta R.T. 0.012 min

Lab File: BN022056.D

Acq: 11 Oct 2022 17:35




Tgt Ion:252 Resp: 1811975


Ion Ratio Lower Upper

252 100

254 68.6 48.6 72.8

126 14.1 11.1 16.7





#85

Benzo(a)anthracene

Concen: 146.235 ng/ul

RT: 21.592 min Scan# 3163

Delta R.T. 0.018 min

Lab File: BN022056.D

Acq: 11 Oct 2022 17:35

Instrument :

BNA_N

ClientSampleId :

SSTD160243





Tgt Ion:228 Resp: 5332732

Ion Ratio Lower Upper

228 100

229 18.9 16.7 25.1

226 28.2 22.4 33.6



#86

Bis(2-ethylhexyl)phthalate

Concen: 181.332 ng/ul

RT: 21.504 min Scan# 3148

Delta R.T. 0.006 min

Lab File: BN022056.D

Acq: 11 Oct 2022 17:35




Tgt Ion:149 Resp: 3805056


Ion Ratio Lower Upper

149 100

167 27.7 21.8 32.6

279 5.2 3.0 4.4#





#87

Chrysene

Concen: 132.196 ng/ul

RT: 21.651 min Scan# 3173

Delta R.T. 0.018 min

Lab File: BN022056.D

Acq: 11 Oct 2022 17:35

Instrument :

BNA_N

ClientSampleId :

SSTD160243





Tgt Ion:228 Resp: 4667460

Ion Ratio Lower Upper

228 100

226 30.7 22.9 34.3

229 19.5 15.5 23.3



#89

Di-n-octyl phthalate

Concen: 153.218 ng/ul




RT: 22.451 min Scan# 3309


Delta R.T. 0.006 min

Lab File: BN022056.D

Acq: 11 Oct 2022 17:35

Tgt Ion:149 Resp: 7623678





#90

Benzo(b)fluoranthene

Concen: 136.869 ng/ul

RT: 23.333 min Scan# 3454

Delta R.T. 0.030 min

Lab File: BN022056.D

Acq: 11 Oct 2022 17:35

Instrument :

BNA_N

ClientSampleId :

SSTD160243





Tgt Ion:252 Resp: 7073182

Ion Ratio Lower Upper

252 100

253 21.6 15.7 23.5

125 11.6 10.0 15.0



#91

Benzo(k)fluoranthene

Concen: 117.915 ng/ul

RT: 23.386 min Scan# 3468

Delta R.T. 0.030 min

Lab File: BN022056.D

Acq: 11 Oct 2022 17:35




Tgt Ion:252 Resp: 5451032


Ion Ratio Lower Upper

252 100

253 20.9 16.8 25.2

125 11.0 10.4 15.6









#93
Benzo(a)pyrene
Concen: 117.368 ng/ul
RT: 23.992 min Scan# 31
Delta R.T. 0.035 min
Lab File: BN022056.D
Acq: 11 Oct 2022 17:35




Instrument :
BNA_N
ClientSampleId :
SSTD160243


Tgt Ion:252 Resp: 5803235
Ion Ratio Lower Upper
252 100
253 21.2 16.8 25.2
125 11.7 11.2 16.8



#94
Indeno(1,2,3-cd)pyrene
Concen: 60.969 ng/ul
RT: 26.698 min Scan# 4031
Delta R.T. 0.030 min
Lab File: BN022056.D
Acq: 11 Oct 2022 17:35

Tgt Ion:276 Resp: 3700954
Ion Ratio Lower Upper
276 100
138 27.5 25.4 38.0
277 26.0 18.8 28.2





#95

Dibenzo(a,h)anthracene

Concen: 127.284 ng/ul

RT: 26.751 min Scan# 4027

Delta R.T. 0.077 min

Lab File: BN022056.D

Acq: 11 Oct 2022 17:35

Instrument :

BNA_N

ClientSampleId :

SSTD160243





Tgt Ion:278 Resp: 6520426

Ion Ratio Lower Upper

278 100

139 16.8 17.4 26.2#

279 24.3 18.6 28.0



#96

Benzo(g,h,i)perylene

Concen: 150.846 ng/ul

RT: 27.545 min Scan# 4175

Delta R.T. 0.077 min

Lab File: BN022056.D

Acq: 11 Oct 2022 17:35

Tgt Ion:276 Resp: 7990352

Ion Ratio Lower Upper

276 100

138 25.1 19.4 29.2

277 22.8 16.5 24.7

