

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN101122\
 Data File : BN022057.D
 Acq On : 11 Oct 2022 18:53
 Operator : CG/JU
 Sample : SSTDICV020
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SICV244

Quant Time: Oct 12 04:43:10 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN101122.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Oct 12 04:33:09 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.969	152	82532	20.000	ng/u1	0.00
20) Naphthalene-d8	10.799	136	402702	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.640	164	284663	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.392	188	594750	20.000	ng/u1	0.00
79) Chrysene-d12	21.592	240	648610	20.000	ng/u1	0.00
88) Perylene-d12	24.063	264	687830	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.264	96	17521	8.035	ng/uL	0.00
4) Pyridine-d5	3.705	84	129067	19.578	ng/u1	0.00
7) Phenol-d5	7.122	99	157794	19.775	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.293	67	101775	19.923	ng/u1	0.00
11) 2-Chlorophenol-d4	7.493	132	115187	20.927	ng/u1	0.00
15) 4-Methylphenol-d8	8.687	113	127382	20.207	ng/u1	0.00
21) Nitrobenzene-d5	9.146	128	61711	21.589	ng/u1	0.00
24) 2-Nitrophenol-d4	9.875	143	63436	21.769	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.416	165	118884	21.572	ng/u1	0.00
31) 4-Chloroaniline-d4	10.940	131	197551	21.227	ng/u1	0.00
46) Dimethylphthalate-d6	14.051	166	417998	20.894	ng/u1	0.00
49) Acenaphthylene-d8	14.334	160	467909	21.498	ng/u1	0.00
54) 4-Nitrophenol-d4	14.840	143	87376	20.743	ng/u1	0.00
60) Fluorene-d10	15.634	176	350196	20.678	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.751	200	66809	20.507	ng/u1	0.00
73) Anthracene-d10	17.492	188	566678	22.061	ng/u1	0.00
81) Pyrene-d10	19.792	212	636834	20.194	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.904	264	714500	21.482	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.299	88	19357	8.193	ng/uL	88
5) Pyridine	3.723	79	133259	20.112	ng/u1	97
6) Benzaldehyde	7.099	77	99484	24.890	ng/u1	99
8) Phenol	7.146	94	169913	20.009	ng/u1	98
10) Bis(2-Chloroethyl)ether	7.387	93	138380	20.436	ng/u1	99
12) 2-Chlorophenol	7.522	128	126863	20.949	ng/u1	98
13) 2-Methylphenol	8.416	108	124196	19.746	ng/u1	99
14) 2,2'-oxybis(1-Chloropr...	8.505	45	198214	19.577	ng/u1	100
16) Acetophenone	8.805	105	219335	21.180	ng/u1	98
17) N-Nitroso-di-n-propyla...	8.787	70	115718	21.352	ng/u1	96
18) 4-Methylphenol	8.752	108	145158	20.562	ng/u1	99
19) Hexachloroethane	9.057	117	49084	20.218	ng/u1	94
22) Nitrobenzene	9.187	77	164608	20.959	ng/u1	100
23) Isophorone	9.716	82	328884	21.426	ng/u1	98
25) 2-Nitrophenol	9.905	139	75901	21.975	ng/u1	93
26) 2,4-Dimethylphenol	9.963	107	159046	21.347	ng/u1	99
27) Bis(2-Chloroethoxy)met...	10.204	93	201487	20.733	ng/u1	97
29) 2,4-Dichlorophenol	10.440	162	127209	21.641	ng/u1	94
30) Naphthalene	10.846	128	453666	20.756	ng/u1	99
32) 4-Chloroaniline	10.963	127	206781	21.107	ng/u1	94
33) Hexachlorobutadiene	11.128	225	68838	21.163	ng/u1	94
34) Caprolactam	11.746	113	48950	21.410	ng/u1	94
35) 4-Chloro-3-methylphenol	12.087	107	155485	21.742	ng/u1	98
36) 2-Methylnaphthalene	12.463	142	317306	21.323	ng/u1	97

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN101122\
 Data File : BN022057.D
 Acq On : 11 Oct 2022 18:53
 Operator : CG/JU
 Sample : SSTDICV020
 Misc :
 ALS Vial : 9 Sample Multiplier: 1


Instrument :
 BNA_N
 ClientSampleId :
 SICV244


Quant Time: Oct 12 04:43:10 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN101122.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Oct 12 04:33:09 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.681	142	323444	21.506	ng/ul	97
39) 1,2,4,5-Tetrachloroben...	12.828	216	148582	20.450	ng/ul	97
40) Hexachlorocyclopentadiene	12.804	237	76169	19.924	ng/ul#	84
41) 2,4,6-Trichlorophenol	13.069	196	100522	20.529	ng/ul	90
42) 2,4,5-Trichlorophenol	13.140	196	113863	21.133	ng/ul	96
43) 1,1'-Biphenyl	13.475	154	417338	20.107	ng/ul	98
44) 2-Chloronaphthalene	13.516	162	327811	20.259	ng/ul	100
45) 2-Nitroaniline	13.722	65	105321	21.130	ng/ul	97
47) Dimethylphthalate	14.098	163	437819	20.597	ng/ul	99
48) 2,6-Dinitrotoluene	14.222	165	81534	21.491	ng/ul	98
50) Acenaphthylene	14.363	152	524907	20.987	ng/ul	99
51) 3-Nitroaniline	14.551	138	95138	20.818	ng/ul	96
52) Acenaphthene	14.704	153	377436	20.740	ng/ul	95
53) 2,4-Dinitrophenol	14.757	184	47144	18.042	ng/ul	97
55) 4-Nitrophenol	14.851	109	70750	20.527	ng/ul	95
56) Dibenzofuran	15.040	168	502021	20.465	ng/ul	100
57) 2,4-Dinitrotoluene	15.004	165	127292	22.190	ng/ul	94
58) 2,3,4,6-Tetrachlorophenol	15.263	232	97843	21.379	ng/ul	98
59) Diethylphthalate	15.457	149	459162	20.834	ng/ul	97
61) Fluorene	15.687	166	422613	20.841	ng/ul	98
62) 4-Chlorophenyl-phenyle...	15.681	204	194754	21.154	ng/ul	99
63) 4-Nitroaniline	15.710	138	100884	22.187	ng/ul	97
66) 4,6-Dinitro-2-methylph...	15.763	198	75995	21.739	ng/ul	99
67) N-Nitrosodiphenylamine	15.898	169	365330	21.226	ng/ul	96
68) 4-Bromophenyl-phenylether	16.575	248	116236	21.741	ng/ul	97
69) Hexachlorobenzene	16.687	284	135587	21.390	ng/ul	97
70) Atrazine	16.851	200	126962	21.821	ng/ul	99
71) Pentachlorophenol	17.039	266	82708	21.497	ng/ul	89
72) Phenanthrene	17.434	178	691313	21.346	ng/ul	99
74) Anthracene	17.522	178	697798	21.637	ng/ul	97
75) 1,2,3,4-Tetrachloroben...	13.434	216	147254	20.442	ng/uL	92
76) Pentachlorobenzene	14.957	250	163298	20.949	ng/uL	97
77) Carbazole	17.798	167	657528	22.089	ng/ul	100
78) Di-n-butylphthalate	18.363	149	802831	22.025	ng/ul	98
80) Fluoranthene	19.457	202	773346	19.309	ng/ul	98
82) Pyrene	19.822	202	834885	19.754	ng/ul	98
83) Butylbenzylphthalate	20.722	149	365362	20.273	ng/ul	98
84) 3,3'-Dichlorobenzidine	21.510	252	297390	21.087	ng/ul	98
85) Benzo(a)anthracene	21.574	228	846098	20.765	ng/ul	96
86) Bis(2-ethylhexyl)phtha...	21.498	149	553908	21.167	ng/ul	99
87) Chrysene	21.633	228	801080	19.981	ng/ul	96
89) Di-n-octyl phthalate	22.445	149	1000617	20.860	ng/ul	100
90) Benzo(b)fluoranthene	23.351	252	864559	19.908	ng/ul	96
91) Benzo(k)fluoranthene	23.351	252	864559	20.184	ng/ul	99
93) Benzo(a)pyrene	23.957	252	797870	20.808	ng/ul	96
94) Indeno(1,2,3-cd)pyrene	26.651	276	456414	9.376	ng/ul	95
95) Dibenzo(a,h)anthracene	26.680	278	929855	21.121	ng/ul	97
96) Benzo(g,h,i)perylene	27.457	276	922383	21.021	ng/ul	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Oct 12 04:43:10 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN101122.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Oct 12 04:33:09 2022
Response via : Initial Calibration









#2
 1,4-Dioxane
 Concen: 8.193 ng/uL
 RT: 3.299 min Scan# 51
 Delta R.T. -0.000 min
 Lab File: BN022057.D
 Acq: 11 Oct 2022 18:53




Instrument :
 BNA_N
 ClientSampleId :
 SICV244


Tgt Ion: 88 Resp: 19357
 Ion Ratio Lower Upper
 88 100
 43 27.2 25.1 37.7
 58 72.5 67.4 101.2



#5
 Pyridine
 Concen: 20.112 ng/uL
 RT: 3.723 min Scan# 125
 Delta R.T. 0.000 min
 Lab File: BN022057.D
 Acq: 11 Oct 2022 18:53

Tgt Ion: 79 Resp: 133259
 Ion Ratio Lower Upper
 79 100
 52 77.5 64.1 96.1
 51 34.2 28.9 43.3









#6
Benzaldehyde
Concen: 24.890 ng/ul
RT: 7.099 min Scan# 699
Delta R.T. 0.000 min
Lab File: BN022057.D
Acq: 11 Oct 2022 18:53




Instrument :
BNA_N
ClientSampleId :
SICV244


Tgt Ion: 77 Resp: 99484
Ion Ratio Lower Upper
77 100
105 94.5 77.0 115.6
106 90.3 72.3 108.5



#8
Phenol
Concen: 20.009 ng/ul
RT: 7.146 min Scan# 707
Delta R.T. -0.006 min
Lab File: BN022057.D
Acq: 11 Oct 2022 18:53

Tgt Ion: 94 Resp: 169913
Ion Ratio Lower Upper
94 100
65 27.6 22.8 34.2
66 35.7 29.3 43.9









#10
Bis(2-Chloroethyl)ether
Concen: 20.436 ng/ul
RT: 7.387 min Scan# 748
Delta R.T. 0.000 min
Lab File: BN022057.D
Acq: 11 Oct 2022 18:53




Instrument :
BNA_N
ClientSampleId :
SICV244


Tgt Ion: 93 Resp: 138380
Ion Ratio Lower Upper
93 100
63 77.3 63.2 94.8
95 32.6 25.9 38.9



#12
2-Chlorophenol
Concen: 20.949 ng/ul
RT: 7.522 min Scan# 771
Delta R.T. 0.000 min
Lab File: BN022057.D
Acq: 11 Oct 2022 18:53

Tgt Ion:128 Resp: 126863
Ion Ratio Lower Upper
128 100
64 49.8 40.9 61.3
130 31.4 25.9 38.9





#13

2-Methylphenol

Concen: 19.746 ng/ul

RT: 8.416 min Scan# 91

Delta R.T. 0.000 min

Lab File: BN022057.D

Acq: 11 Oct 2022 18:53

Instrument :

BNA_N

ClientSampleId :





SICV244

Tgt Ion:108 Resp: 124196

Ion Ratio Lower Upper

108 100

107 91.0 71.8 107.6



#14

2,2'-oxybis(1-Chloropropane)

Concen: 19.577 ng/ul

RT: 8.505 min Scan# 938

Delta R.T. -0.006 min

Lab File: BN022057.D

Acq: 11 Oct 2022 18:53




Tgt Ion: 45 Resp: 198214


Ion Ratio Lower Upper

45 100

77 14.1 11.3 16.9

79 10.9 8.6 12.8









#16
Acetophenone
Concen: 21.180 ng/ul
RT: 8.805 min Scan# 989
Delta R.T. 0.000 min
Lab File: BN022057.D
Acq: 11 Oct 2022 18:53




Instrument :
BNA_N
ClientSampleId :
SICV244


Tgt Ion:105 Resp: 219335
Ion Ratio Lower Upper
105 100
77 80.7 63.4 95.0
51 27.9 21.9 32.9



#17
N-Nitroso-di-n-propylamine
Concen: 21.352 ng/ul
RT: 8.787 min Scan# 986
Delta R.T. 0.000 min
Lab File: BN022057.D
Acq: 11 Oct 2022 18:53

Tgt Ion: 70 Resp: 115718
Ion Ratio Lower Upper
70 100
42 54.4 46.2 69.4
101 9.6 7.1 10.7
130 19.9 15.2 22.8









#18
4-Methylphenol
Concen: 20.562 ng/ul
RT: 8.752 min Scan# 91
Delta R.T. 0.000 min
Lab File: BN022057.D
Acq: 11 Oct 2022 18:53




Instrument :
BNA_N
ClientSampleId :
SICV244


Tgt Ion:108 Resp: 145158
Ion Ratio Lower Upper
108 100
107 109.7 87.3 130.9



#19
Hexachloroethane
Concen: 20.218 ng/ul
RT: 9.057 min Scan# 1032
Delta R.T. 0.000 min
Lab File: BN022057.D
Acq: 11 Oct 2022 18:53

Tgt Ion:117 Resp: 49084
Ion Ratio Lower Upper
117 100
201 87.5 76.8 115.2
199 57.1 46.2 69.2





#22

Nitrobenzene

Concen: 20.959 ng/ul

RT: 9.187 min Scan# 1054

Delta R.T. 0.000 min

Lab File: BN022057.D

Acq: 11 Oct 2022 18:53

Instrument :

BNA_N

ClientSampleId :

SICV244





Tgt Ion: 77 Resp: 164608

Ion Ratio Lower Upper

77 100

123 42.3 33.8 50.8

65 13.1 10.5 15.7



#23

Isophorone

Concen: 21.426 ng/ul

RT: 9.716 min Scan# 1144

Delta R.T. 0.000 min

Lab File: BN022057.D

Acq: 11 Oct 2022 18:53




Tgt Ion: 82 Resp: 328884


Ion Ratio Lower Upper

82 100

95 6.5 5.1 7.7

138 17.1 12.6 18.8









#25
2-Nitrophenol
Concen: 21.975 ng/ul
RT: 9.905 min Scan# 1176
Delta R.T. 0.000 min
Lab File: BN022057.D
Acq: 11 Oct 2022 18:53




Instrument :
BNA_N
ClientSampleId :
SICV244


Tgt Ion:139 Resp: 75901
Ion Ratio Lower Upper
139 100
65 49.4 44.9 67.3
109 27.8 23.5 35.3



#26
2,4-Dimethylphenol
Concen: 21.347 ng/ul
RT: 9.963 min Scan# 1186
Delta R.T. 0.000 min
Lab File: BN022057.D
Acq: 11 Oct 2022 18:53

Tgt Ion:107 Resp: 159046
Ion Ratio Lower Upper
107 100
121 52.6 41.0 61.4
122 89.5 71.8 107.6









#27
 Bis(2-Chloroethoxy)methane
 Concen: 20.733 ng/ul
 RT: 10.204 min Scan# 1227
 Delta R.T. 0.000 min
 Lab File: BN022057.D
 Acq: 11 Oct 2022 18:53




Instrument :
 BNA_N
 ClientSampleId :
 SICV244


Tgt Ion: 93 Resp: 201487
 Ion Ratio Lower Upper
 93 100
 95 33.8 25.0 37.6
 123 10.8 8.6 13.0



#29
 2,4-Dichlorophenol
 Concen: 21.641 ng/ul
 RT: 10.440 min Scan# 1267
 Delta R.T. 0.000 min
 Lab File: BN022057.D
 Acq: 11 Oct 2022 18:53

Tgt Ion:162 Resp: 127209
 Ion Ratio Lower Upper
 162 100
 164 61.5 53.9 80.9
 98 38.3 32.5 48.7









#30
Naphthalene
Concen: 20.756 ng/ul
RT: 10.846 min Scan# 1336
Delta R.T. 0.000 min
Lab File: BN022057.D
Acq: 11 Oct 2022 18:53




Instrument :
BNA_N
ClientSampleId :
SICV244


Tgt Ion:128 Resp: 453666
Ion Ratio Lower Upper
128 100
129 11.3 9.1 13.7
127 13.1 11.0 16.4



#32
4-Chloroaniline
Concen: 21.107 ng/ul
RT: 10.963 min Scan# 1356
Delta R.T. 0.000 min
Lab File: BN022057.D
Acq: 11 Oct 2022 18:53

Tgt Ion:127 Resp: 206781
Ion Ratio Lower Upper
127 100
129 33.6 24.2 36.4





#33

Hexachlorobutadiene

Concen: 21.163 ng/ul

RT: 11.128 min Scan# 1384

Delta R.T. 0.000 min

Lab File: BN022057.D

Acq: 11 Oct 2022 18:53

Instrument :

BNA_N

ClientSampleId :

SICV244





Tgt Ion:225 Resp: 68838

Ion Ratio Lower Upper

225 100

223 62.2 53.0 79.6

227 64.2 46.9 70.3



#34

Caprolactam

Concen: 21.410 ng/ul

RT: 11.746 min Scan# 1489

Delta R.T. 0.000 min

Lab File: BN022057.D

Acq: 11 Oct 2022 18:53




Tgt Ion:113 Resp: 48950


Ion Ratio Lower Upper

113 100

55 169.8 143.4 215.0

56 124.6 104.6 157.0









#35
 4-Chloro-3-methylphenol
 Concen: 21.742 ng/ul
 RT: 12.087 min Scan# 1547
 Delta R.T. 0.000 min
 Lab File: BN022057.D
 Acq: 11 Oct 2022 18:53




Instrument :
 BNA_N
 ClientSampleId :
 SICV244


Tgt Ion:107 Resp: 155485
 Ion Ratio Lower Upper
 107 100
 144 25.2 20.7 31.1
 142 77.0 59.8 89.8



#36
 2-Methylnaphthalene
 Concen: 21.323 ng/ul
 RT: 12.463 min Scan# 1611
 Delta R.T. 0.000 min
 Lab File: BN022057.D
 Acq: 11 Oct 2022 18:53

Tgt Ion:142 Resp: 317306
 Ion Ratio Lower Upper
 142 100
 141 83.4 68.7 103.1





#37

1-Methylnaphthalene

Concen: 21.506 ng/ul

RT: 12.681 min Scan# 1648

Delta R.T. 0.000 min

Lab File: BN022057.D

Acq: 11 Oct 2022 18:53

Instrument :

BNA_N

ClientSampleId :

SICV244





Tgt Ion:142 Resp: 323444

Ion Ratio Lower Upper

142 100

141 88.4 73.3 109.9

116 3.8 3.0 4.6



#39

1,2,4,5-Tetrachlorobenzene

Concen: 20.450 ng/ul

RT: 12.828 min Scan# 1673

Delta R.T. 0.000 min

Lab File: BN022057.D

Acq: 11 Oct 2022 18:53

Tgt Ion:216 Resp: 148582




Ion Ratio Lower Upper


216 100

214 77.2 60.4 90.6

179 21.8 18.4 27.6

108 20.0 17.8 26.6









#40
Hexachlorocyclopentadiene
Concen: 19.924 ng/ul
RT: 12.804 min Scan# 1669
Delta R.T. 0.000 min
Lab File: BN022057.D
Acq: 11 Oct 2022 18:53




Instrument :
BNA_N
ClientSampleId :
SICV244


Tgt Ion:237 Resp: 76169
Ion Ratio Lower Upper
237 100
235 54.2 56.0 84.0#
272 12.7 10.6 15.8



#41
2,4,6-Trichlorophenol
Concen: 20.529 ng/ul
RT: 13.069 min Scan# 1714
Delta R.T. 0.000 min
Lab File: BN022057.D
Acq: 11 Oct 2022 18:53

Tgt Ion:196 Resp: 100522
Ion Ratio Lower Upper
196 100
198 104.1 74.3 111.5
200 29.8 22.3 33.5





#42

2,4,5-Trichlorophenol

Concen: 21.133 ng/ul

RT: 13.140 min Scan# 1

Delta R.T. 0.000 min

Lab File: BN022057.D

Acq: 11 Oct 2022 18:53

Instrument :

BNA_N

ClientSampleId :

SICV244





Tgt Ion:196 Resp: 113863

Ion Ratio Lower Upper

196 100

198 91.4 76.8 115.2

200 28.9 23.9 35.9



#43

1,1'-Biphenyl

Concen: 20.107 ng/ul

RT: 13.475 min Scan# 1783

Delta R.T. 0.000 min

Lab File: BN022057.D

Acq: 11 Oct 2022 18:53




Tgt Ion:154 Resp: 417338


Ion Ratio Lower Upper

154 100

153 40.5 31.0 46.6

76 14.6 11.6 17.4









#44
2-Chloronaphthalene
Concen: 20.259 ng/ul
RT: 13.516 min Scan# 1790
Delta R.T. 0.000 min
Lab File: BN022057.D
Acq: 11 Oct 2022 18:53




Instrument :
BNA_N
ClientSampleId :
SICV244


Tgt Ion:162 Resp: 327811
Ion Ratio Lower Upper
162 100
164 31.9 25.9 38.9
127 40.2 32.2 48.4



#45
2-Nitroaniline
Concen: 21.130 ng/ul
RT: 13.722 min Scan# 1825
Delta R.T. -0.006 min
Lab File: BN022057.D
Acq: 11 Oct 2022 18:53

Tgt Ion: 65 Resp: 105321
Ion Ratio Lower Upper
65 100
92 68.7 54.6 82.0
138 95.1 80.4 120.6









#47
Dimethylphthalate
Concen: 20.597 ng/ul
RT: 14.098 min Scan# 1889
Delta R.T. 0.000 min
Lab File: BN022057.D
Acq: 11 Oct 2022 18:53




Instrument :
BNA_N
ClientSampleId :
SICV244


Tgt Ion:163 Resp: 437819
Ion Ratio Lower Upper
163 100
194 4.0 3.1 4.7
164 9.4 8.0 12.0



#48
2,6-Dinitrotoluene
Concen: 21.491 ng/ul
RT: 14.222 min Scan# 1910
Delta R.T. 0.000 min
Lab File: BN022057.D
Acq: 11 Oct 2022 18:53

Tgt Ion:165 Resp: 81534
Ion Ratio Lower Upper
165 100
89 64.1 52.2 78.4
121 23.3 19.5 29.3









#50
Acenaphthylene
Concen: 20.987 ng/ul
RT: 14.363 min Scan# 1934
Delta R.T. 0.000 min
Lab File: BN022057.D
Acq: 11 Oct 2022 18:53




Instrument :
BNA_N
ClientSampleId :
SICV244


Tgt Ion:152 Resp: 524907
Ion Ratio Lower Upper
152 100
151 20.0 15.7 23.5
153 12.4 10.0 15.0



#51
3-Nitroaniline
Concen: 20.818 ng/ul
RT: 14.551 min Scan# 1966
Delta R.T. 0.000 min
Lab File: BN022057.D
Acq: 11 Oct 2022 18:53

Tgt Ion:138 Resp: 95138
Ion Ratio Lower Upper
138 100
108 12.2 9.1 13.7
92 126.9 97.7 146.5









#52
 Acenaphthene
 Concen: 20.740 ng/ul
 RT: 14.704 min Scan# 1992
 Delta R.T. 0.000 min
 Lab File: BN022057.D
 Acq: 11 Oct 2022 18:53




Instrument :
 BNA_N
 ClientSampleId :
 SICV244


Tgt Ion:	153	Resp:	377436
Ion Ratio	Lower	Upper	
153	100		
152	46.7	37.8	56.8
154	83.6	71.8	107.6



#53
 2,4-Dinitrophenol
 Concen: 18.042 ng/ul
 RT: 14.757 min Scan# 2001
 Delta R.T. 0.000 min
 Lab File: BN022057.D
 Acq: 11 Oct 2022 18:53

Tgt Ion:	184	Resp:	47144
Ion Ratio	Lower	Upper	
184	100		
63	69.9	58.9	88.3
154	63.0	51.8	77.6









#55
4-Nitrophenol
Concen: 20.527 ng/ul
RT: 14.851 min Scan# 2017
Delta R.T. 0.000 min
Lab File: BN022057.D
Acq: 11 Oct 2022 18:53




Instrument :
BNA_N
ClientSampleId :
SICV244


Tgt Ion:109 Resp: 70750
Ion Ratio Lower Upper
109 100
139 138.3 111.0 166.4
65 135.7 118.5 177.7



#56
Dibenzofuran
Concen: 20.465 ng/ul
RT: 15.040 min Scan# 2049
Delta R.T. 0.000 min
Lab File: BN022057.D
Acq: 11 Oct 2022 18:53

Tgt Ion:168 Resp: 502021
Ion Ratio Lower Upper
168 100
139 39.4 31.3 46.9









#57
2,4-Dinitrotoluene
Concen: 22.190 ng/ul
RT: 15.004 min Scan# 2043
Delta R.T. 0.000 min
Lab File: BN022057.D
Acq: 11 Oct 2022 18:53




Instrument :
BNA_N
ClientSampleId :
SICV244


Tgt Ion:165 Resp: 127292
Ion Ratio Lower Upper
165 100
63 46.3 40.6 60.8
182 2.7 1.9 2.9



#58
2,3,4,6-Tetrachlorophenol
Concen: 21.379 ng/ul
RT: 15.263 min Scan# 2087
Delta R.T. 0.000 min
Lab File: BN022057.D
Acq: 11 Oct 2022 18:53

Tgt Ion:232 Resp: 97843
Ion Ratio Lower Upper
232 100
131 52.8 44.0 66.0
130 2.8 2.3 3.5
166 35.0 27.8 41.8





#59

Diethylphthalate

Concen: 20.834 ng/ul

RT: 15.457 min Scan# 2120

Delta R.T. 0.000 min

Lab File: BN022057.D

Acq: 11 Oct 2022 18:53

Instrument :

BNA_N

ClientSampleId :

SICV244





Tgt Ion:149 Resp: 459162

Ion Ratio Lower Upper

149 100

177 21.7 15.9 23.9

150 12.8 9.8 14.8



#61

Fluorene

Concen: 20.841 ng/ul

RT: 15.687 min Scan# 2159

Delta R.T. 0.000 min

Lab File: BN022057.D

Acq: 11 Oct 2022 18:53




Tgt Ion:166 Resp: 422613


Ion Ratio Lower Upper

166 100

165 94.1 74.1 111.1

167 13.7 10.5 15.7









#62
4-Chlorophenyl-phenylether
Concen: 21.154 ng/ul
RT: 15.681 min Scan# 2158
Delta R.T. 0.000 min
Lab File: BN022057.D
Acq: 11 Oct 2022 18:53




Instrument :
BNA_N
ClientSampleId :
SICV244


Tgt Ion:204 Resp: 194754
Ion Ratio Lower Upper
204 100
206 33.0 26.1 39.1
141 65.0 50.9 76.3



#63
4-Nitroaniline
Concen: 22.187 ng/ul
RT: 15.710 min Scan# 2163
Delta R.T. -0.006 min
Lab File: BN022057.D
Acq: 11 Oct 2022 18:53

Tgt Ion:138 Resp: 100884
Ion Ratio Lower Upper
138 100
92 54.8 42.8 64.2
108 78.6 60.1 90.1









#66
4,6-Dinitro-2-methylphenol
Concen: 21.739 ng/ul
RT: 15.763 min Scan# 2172
Delta R.T. -0.006 min
Lab File: BN022057.D
Acq: 11 Oct 2022 18:53




Instrument :
BNA_N
ClientSampleId :
SICV244


Tgt Ion:198 Resp: 75995
Ion Ratio Lower Upper
198 100
182 5.3 3.9 5.9
77 30.9 24.5 36.7



#67
N-Nitrosodiphenylamine
Concen: 21.226 ng/ul
RT: 15.898 min Scan# 2195
Delta R.T. 0.000 min
Lab File: BN022057.D
Acq: 11 Oct 2022 18:53

Tgt Ion:169 Resp: 365330
Ion Ratio Lower Upper
169 100
168 67.5 51.4 77.0
167 37.5 27.8 41.8









#68
4-Bromophenyl-phenylether
Concen: 21.741 ng/ul
RT: 16.575 min Scan# 21
Delta R.T. -0.006 min
Lab File: BN022057.D
Acq: 11 Oct 2022 18:53




Instrument :
BNA_N
ClientSampleId :
SICV244


Tgt Ion:248 Resp: 116236
Ion Ratio Lower Upper
248 100
250 98.1 77.6 116.4
141 88.0 66.6 99.8



#69
Hexachlorobenzene
Concen: 21.390 ng/ul
RT: 16.687 min Scan# 2329
Delta R.T. -0.006 min
Lab File: BN022057.D
Acq: 11 Oct 2022 18:53

Tgt Ion:284 Resp: 135587
Ion Ratio Lower Upper
284 100
142 42.2 32.6 48.8
249 31.7 23.4 35.2





#70

Atrazine

Concen: 21.821 ng/ul

RT: 16.851 min Scan# 21

Delta R.T. 0.000 min

Lab File: BN022057.D

Acq: 11 Oct 2022 18:53

Instrument :

BNA_N

ClientSampleId :

SICV244





Tgt Ion:200 Resp: 126962

Ion Ratio Lower Upper

200 100

173 28.3 22.6 34.0

215 48.6 38.4 57.6



#71

Pentachlorophenol

Concen: 21.497 ng/ul

RT: 17.039 min Scan# 2389

Delta R.T. 0.000 min

Lab File: BN022057.D

Acq: 11 Oct 2022 18:53




Tgt Ion:266 Resp: 82708


Ion Ratio Lower Upper

266 100

264 57.4 53.8 80.6

268 63.0 45.1 67.7









#72
Phenanthrene
Concen: 21.346 ng/ul
RT: 17.434 min Scan# 2456
Delta R.T. 0.000 min
Lab File: BN022057.D
Acq: 11 Oct 2022 18:53




Instrument :
BNA_N
ClientSampleId :
SICV244


Tgt Ion:178 Resp: 691313
Ion Ratio Lower Upper
178 100
179 15.6 12.3 18.5
176 19.2 15.9 23.9



#74
Anthracene
Concen: 21.637 ng/ul
RT: 17.522 min Scan# 2471
Delta R.T. -0.006 min
Lab File: BN022057.D
Acq: 11 Oct 2022 18:53

Tgt Ion:178 Resp: 697798
Ion Ratio Lower Upper
178 100
179 14.3 12.2 18.2
176 17.7 15.2 22.8









#75
1,2,3,4-Tetrachlorobenzene
Concen: 20.442 ng/uL
RT: 13.434 min Scan# 1776
Delta R.T. -0.006 min
Lab File: BN022057.D
Acq: 11 Oct 2022 18:53




Instrument :
BNA_N
ClientSampleId :
SICV244


Tgt Ion:216 Resp: 147254
Ion Ratio Lower Upper
216 100
214 82.0 60.3 90.5
218 51.8 37.2 55.8



#76
Pentachlorobenzene
Concen: 20.949 ng/uL
RT: 14.957 min Scan# 2035
Delta R.T. 0.000 min
Lab File: BN022057.D
Acq: 11 Oct 2022 18:53

Tgt Ion:250 Resp: 163298
Ion Ratio Lower Upper
250 100
248 60.8 52.3 78.5
252 62.8 50.0 75.0





#77

Carbazole

Concen: 22.089 ng/ul

RT: 17.798 min Scan# 2518

Delta R.T. 0.000 min

Lab File: BN022057.D

Acq: 11 Oct 2022 18:53

Instrument :

BNA_N

ClientSampleId :

SICV244





Tgt Ion:167 Resp: 657528

Ion Ratio Lower Upper

167 100

166 21.6 17.2 25.8

139 13.9 11.0 16.4



#78

Di-n-butylphthalate

Concen: 22.025 ng/ul

RT: 18.363 min Scan# 2614

Delta R.T. 0.000 min

Lab File: BN022057.D

Acq: 11 Oct 2022 18:53




Tgt Ion:149 Resp: 802831


Ion Ratio Lower Upper

149 100

150 9.1 6.7 10.1

104 5.1 4.3 6.5









#80
Fluoranthene
Concen: 19.309 ng/u1
RT: 19.457 min Scan# 2800
Delta R.T. 0.000 min
Lab File: BN022057.D
Acq: 11 Oct 2022 18:53




Instrument :
BNA_N
ClientSampleId :
SICV244


Tgt Ion:202 Resp: 773346
Ion Ratio Lower Upper
202 100
101 15.4 11.4 17.0
100 11.3 8.6 13.0



#82
Pyrene
Concen: 19.754 ng/u1
RT: 19.822 min Scan# 2862
Delta R.T. 0.000 min
Lab File: BN022057.D
Acq: 11 Oct 2022 18:53

Tgt Ion:202 Resp: 834885
Ion Ratio Lower Upper
202 100
101 16.3 13.8 20.6
100 12.9 11.2 16.8









#83
Butylbenzylphthalate
Concen: 20.273 ng/ul
RT: 20.722 min Scan# 3015
Delta R.T. 0.000 min
Lab File: BN022057.D
Acq: 11 Oct 2022 18:53




Instrument :
BNA_N
ClientSampleId :
SICV244


Tgt Ion:149 Resp: 365362
Ion Ratio Lower Upper
149 100
91 74.7 60.9 91.3
206 19.9 14.9 22.3



#84
3,3'-Dichlorobenzidine
Concen: 21.087 ng/ul
RT: 21.510 min Scan# 3149
Delta R.T. 0.000 min
Lab File: BN022057.D
Acq: 11 Oct 2022 18:53

Tgt Ion:252 Resp: 297390
Ion Ratio Lower Upper
252 100
254 62.7 48.6 72.8
126 13.6 11.1 16.7





#85

Benzo(a)anthracene

Concen: 20.765 ng/ul

RT: 21.574 min Scan# 3160

Delta R.T. 0.000 min

Lab File: BN022057.D

Acq: 11 Oct 2022 18:53

Instrument :

BNA_N

ClientSampleId :

SICV244





Tgt Ion:228 Resp: 846098

Ion Ratio Lower Upper

228 100

229 19.0 16.7 25.1

226 26.3 22.4 33.6



#86

Bis(2-ethylhexyl)phthalate

Concen: 21.167 ng/ul

RT: 21.498 min Scan# 3147

Delta R.T. 0.000 min

Lab File: BN022057.D

Acq: 11 Oct 2022 18:53




Tgt Ion:149 Resp: 553908


Ion Ratio Lower Upper

149 100

167 26.8 21.8 32.6

279 3.4 3.0 4.4









#87
Chrysene
Concen: 19.981 ng/ul
RT: 21.633 min Scan# 3170
Delta R.T. 0.000 min
Lab File: BN022057.D
Acq: 11 Oct 2022 18:53




Instrument :
BNA_N
ClientSampleId :
SICV244


Tgt Ion:228 Resp: 801080
Ion Ratio Lower Upper
228 100
226 31.2 22.9 34.3
229 20.4 15.5 23.3



#89
Di-n-octyl phthalate
Concen: 20.860 ng/ul
RT: 22.445 min Scan# 3308
Delta R.T. 0.000 min
Lab File: BN022057.D
Acq: 11 Oct 2022 18:53

Tgt Ion:149 Resp: 1000617





#90

Benzo(b)fluoranthene

Concen: 19.908 ng/ul

RT: 23.351 min Scan# 3462

Delta R.T. 0.047 min

Lab File: BN022057.D

Acq: 11 Oct 2022 18:53

Instrument :

BNA_N

ClientSampleId :

SICV244





Tgt Ion:252 Resp: 864559

Ion Ratio Lower Upper

252 100

253 21.8 15.7 23.5

125 13.3 10.0 15.0



#91

Benzo(k)fluoranthene

Concen: 20.184 ng/ul

RT: 23.351 min Scan# 3462

Delta R.T. -0.006 min

Lab File: BN022057.D

Acq: 11 Oct 2022 18:53




Tgt Ion:252 Resp: 864559


Ion Ratio Lower Upper

252 100

253 21.8 16.8 25.2

125 13.3 10.4 15.6





#93

Benzo(a)pyrene

Concen: 20.808 ng/ul

RT: 23.957 min Scan# 3565

Delta R.T. 0.000 min

Lab File: BN022057.D

Acq: 11 Oct 2022 18:53

Instrument :

BNA_N

ClientSampleId :

SICV244





Tgt Ion:252 Resp: 797870

Ion Ratio Lower Upper

252 100

253 22.6 16.8 25.2

125 15.6 11.2 16.8



#94

Indeno(1,2,3-cd)pyrene

Concen: 9.376 ng/ul

RT: 26.651 min Scan# 4023

Delta R.T. -0.018 min

Lab File: BN022057.D

Acq: 11 Oct 2022 18:53




Tgt Ion:276 Resp: 456414


Ion Ratio Lower Upper

276 100

138 27.2 25.4 38.0

277 23.3 18.8 28.2





#95

Dibenzo(a,h)anthracene

Concen: 21.121 ng/ul

RT: 26.680 min Scan# 4027

Delta R.T. 0.006 min

Lab File: BN022057.D

Acq: 11 Oct 2022 18:53

Instrument :

BNA_N

ClientSampleId :

SICV244





Tgt Ion:278 Resp: 929855

Ion Ratio Lower Upper

278 100

139 20.0 17.4 26.2

279 24.7 18.6 28.0



#96

Benzo(g,h,i)perylene

Concen: 21.021 ng/ul

RT: 27.457 min Scan# 4160

Delta R.T. -0.012 min

Lab File: BN022057.D

Acq: 11 Oct 2022 18:53

Tgt Ion:276 Resp: 922383

Ion Ratio Lower Upper

276 100

138 26.7 19.4 29.2

277 20.7 16.5 24.7

