


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN101122\
 Data File : BN022050.D
 Acq On : 11 Oct 2022 13:45
 Operator : CG/JU
 Sample : DFTPP252
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DFTPP252

Integration File: LSCINT.P

Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN101122.M
 Title : SVOA CALIBRATION
 Last Update : Wed Oct 12 04:33:09 2022



AutoFind: Scans 2522, 2523, 2524; Background Corrected with Scan 2516

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	45.8	134884	PASS
68	69	0.00	2	1.6	2151	PASS
69	198	0.00	100	46.4	136915	PASS
70	69	0.00	2	0.5	728	PASS
127	198	10	80	54.8	161605	PASS
197	198	0.00	2	0.3	818	PASS
198	198	100	100	100.0	294827	PASS
199	198	5	9	6.7	19739	PASS
275	198	10	60	24.7	72693	PASS
365	198	1	100	3.1	9159	PASS
441	198	0.01	100	12.0	35243	PASS
442	198	50	100	71.4	210368	PASS
443	442	15	24	20.7	43448	PASS

m/z	Abundance
38.90	163.0
39.90	1020.0
41.00	247.0
43.00	294.0
44.00	1981.0
45.00	165.0
48.90	201.0
50.90	156.0
55.00	206.0
57.00	177.0
73.00	585.0
74.90	248.0
77.90	217.0
83.90	174.0
132.90	232.0
146.90	156.0
207.00	1341.0
208.00	285.0
281.10	445.0
282.10	166.0

Instrument :
BNA_N
ClientSampleId :
DFTPP252

m/z	Abundance
37.00	736.0
38.10	1811.0
39.00	10878.0
40.00	1121.0
41.00	546.0
43.00	233.0
44.00	1874.0
44.90	373.0
49.10	1158.0
50.00	38656.0
51.00	150784.0
52.00	7683.0
53.00	362.0
55.00	1061.0
56.00	4738.0
57.00	9360.0
57.90	457.0
58.90	177.0
61.00	1929.0
62.00	2301.0
63.00	5913.0
64.00	840.0
65.00	3121.0
66.00	176.0
67.00	314.0
68.00	2579.0
69.00	151424.0
70.00	797.0
71.00	285.0
73.00	1648.0
74.00	15972.0
75.00	24880.0
76.00	8893.0
77.00	182912.0
78.00	14075.0
79.00	11497.0
80.00	8644.0
81.00	12249.0
82.00	3191.0
83.00	2799.0
84.00	453.0
85.00	1869.0
86.00	3550.0
87.00	1627.0
88.00	663.0
89.00	394.0
90.90	2674.0
92.00	3310.0
93.00	20040.0
93.90	1194.0
94.90	379.0
96.00	950.0
96.90	376.0
98.00	15537.0
99.00	14375.0
99.90	1167.0
101.00	6484.0
101.90	392.0
103.00	1885.0
104.00	3876.0
105.00	3952.0
106.00	1256.0
107.00	50400.0
108.00	7858.0
109.00	1439.0
110.00	90208.0
111.00	14046.0
112.00	1855.0
112.90	577.0
114.90	231.0
116.00	2685.0
117.00	46680.0
118.00	3619.0
118.90	380.0
119.90	633.0
120.90	266.0
122.00	3056.0
123.00	5043.0
123.90	2287.0
125.00	2173.0
126.00	622.0
127.00	169152.0
128.00	13457.0
129.00	62928.0
130.00	5528.0
131.00	1035.0
132.00	531.0
133.00	485.0
134.00	2109.0
135.00	5231.0
136.00	2080.0
137.00	3695.0

Instrument :
BNA_N
ClientSampleId :
DFTPP252

138.00	723.0
138.90	302.0
139.90	914.0
141.00	8046.0
142.00	2476.0
143.00	1941.0
144.00	534.0
145.10	441.0
146.00	1588.0
147.00	4394.0
148.00	10808.0
149.00	2636.0
150.00	624.0
151.10	1153.0
152.90	2699.0
153.90	1949.0
155.00	5070.0
156.00	6469.0
157.00	1257.0
157.90	1265.0
159.00	1204.0
160.00	2325.0
161.00	3591.0
162.00	1151.0
163.00	341.0
163.80	313.0
165.00	3086.0
166.00	2530.0
167.00	18112.0
168.00	13833.0
169.00	1737.0
169.90	513.0
170.90	724.0
172.00	1441.0
173.00	1906.0
174.00	3479.0
175.00	6264.0
176.00	1813.0
177.00	2657.0
177.90	1131.0
179.00	11695.0
180.00	9026.0
181.00	3470.0
182.00	658.0
182.90	293.0
183.90	1069.0
185.00	5572.0
186.00	43144.0
187.00	12553.0
188.00	1786.0
189.00	2648.0
190.00	510.0
191.00	1370.0
192.00	3736.0
193.00	4506.0
194.00	804.0
195.00	467.0
196.00	8237.0
196.70	1551.0
198.00	294528.0
198.90	20408.0
200.00	1515.0
201.60	1210.0
203.00	2295.0
204.00	11148.0
205.00	18656.0
206.00	79984.0
207.10	10942.0
208.00	2774.0
209.00	1175.0
210.10	1322.0
211.00	2995.0
212.90	235.0
215.00	764.0
216.00	1785.0
217.00	20192.0
218.00	2784.0
219.00	209.0
221.10	11338.0
221.80	3193.0
223.00	4734.0
224.10	42352.0
225.00	11302.0
226.00	1642.0
227.00	16952.0
228.00	2607.0
229.00	3539.0
230.00	479.0
231.00	1409.0
232.00	302.0
232.90	281.0
234.00	1296.0
235.00	1393.0
236.00	799.0
237.00	2516.0
238.00	314.0

Instrument :
BNA_N
ClientSampleId :
DFTPP252

239.00	717.0
239.90	564.0
241.00	1021.0
242.00	1910.0
243.00	2428.0
244.00	31344.0
245.00	3984.0
246.00	6233.0
246.90	1460.0
247.90	255.0
248.90	915.0
250.00	190.0
250.90	333.0
252.00	300.0
253.00	665.0
254.00	1070.0
255.00	160384.0
256.00	24472.0
257.10	2088.0
258.00	9368.0
259.00	1499.0
260.00	282.0
261.00	302.0
264.00	217.0
265.00	3611.0
265.90	579.0
266.90	165.0
271.00	658.0
272.00	499.0
273.00	4951.0
274.00	12588.0
275.00	63128.0
276.00	9096.0
277.00	5890.0
277.90	1245.0
279.00	331.0
281.00	500.0
281.90	198.0
283.00	593.0
284.10	574.0
285.00	1049.0
286.00	196.0
289.10	274.0
290.00	232.0
291.90	312.0
293.00	1569.0
294.00	605.0
295.00	393.0
296.00	15303.0
297.00	2479.0
298.10	253.0
301.00	227.0
302.00	382.0
303.00	2245.0
304.00	570.0
308.10	220.0
309.00	198.0
310.00	244.0
312.90	254.0
314.00	1021.0
315.00	2106.0
316.00	1248.0
317.00	247.0
321.00	539.0
321.90	334.0
323.00	5596.0
324.00	936.0
326.90	1078.0
327.90	574.0
328.90	175.0
332.00	354.0
333.00	491.0
334.10	4359.0
335.00	799.0
336.00	157.0
341.00	790.0
342.00	202.0
345.90	1207.0
346.90	326.0
352.00	1406.0
353.00	1073.0
354.10	1837.0
355.00	501.0
365.00	8555.0
366.00	1270.0
370.10	180.0
370.90	489.0
372.00	2593.0
373.00	725.0
383.00	649.0
383.90	266.0
389.90	388.0
390.90	264.0
392.00	176.0
401.10	188.0
402.00	991.0

Instrument :
BNA_N
ClientSampleId :
DFTPP252

403.00	1930.0
404.00	759.0
405.10	198.0
421.00	1311.0
422.00	1279.0
423.00	9461.0
424.00	3423.0
425.00	561.0
441.10	25840.0
442.00	150016.0
443.00	30912.0
444.10	3242.0
445.00	172.0

Instrument :
BNA_N
ClientSampleId :
DFTPP252

m/z	Abundance
37.00	717.0
38.00	1848.0
39.10	11686.0
40.00	1311.0
41.10	397.0
43.10	265.0
44.00	1607.0
44.90	409.0
49.10	1173.0
50.00	40744.0
51.00	158016.0
52.00	7984.0
53.00	427.0
55.00	848.0
56.00	4247.0
57.00	10162.0
57.90	544.0
59.00	157.0
59.90	221.0
60.90	1994.0
62.00	2239.0
63.00	6222.0
64.00	827.0
65.00	3336.0
66.00	251.0
66.90	276.0
68.00	2378.0
69.00	160128.0
70.00	782.0
70.90	202.0
73.00	1999.0
74.00	17328.0
75.00	25976.0
76.00	9055.0
77.00	191360.0
78.00	13602.0
79.00	11421.0
80.00	9092.0
81.00	12827.0
82.00	3019.0
83.00	3271.0
84.00	533.0
85.00	2295.0
86.00	4030.0
87.00	1767.0
87.90	597.0
88.90	492.0
91.00	2922.0
92.00	3321.0
93.00	21320.0
94.00	1505.0
95.00	440.0
96.00	1098.0
97.10	488.0
98.00	15977.0
99.00	14625.0
100.00	1217.0
101.00	7201.0
101.80	392.0
103.00	2021.0
104.00	4509.0
105.00	4239.0
106.00	1278.0
107.00	54640.0
108.00	8897.0
109.00	1367.0
110.00	97544.0
111.00	15266.0
112.00	1820.0
113.00	735.0
114.90	215.0
116.00	3039.0
117.00	51192.0
118.00	4005.0
118.90	425.0
119.90	753.0
121.00	225.0
122.00	3552.0
123.00	5450.0
124.00	2857.0
125.00	2556.0
126.10	597.0
127.00	192896.0
128.00	14465.0
129.00	73536.0
130.00	6496.0
131.00	1226.0
132.00	624.0
132.90	549.0
134.00	1879.0
135.00	5820.0
136.00	2515.0

Instrument :
BNA_N
ClientSampleId :
DFTPP252

137.00	3826.0
138.00	745.0
138.90	426.0
140.00	893.0
141.00	9283.0
142.00	3180.0
143.00	1979.0
144.00	602.0
145.00	543.0
146.00	1902.0
147.00	4993.0
148.00	12991.0
148.90	3032.0
150.00	836.0
151.10	1233.0
151.80	840.0
153.00	2867.0
154.00	2228.0
155.00	5435.0
156.00	8563.0
157.00	1611.0
157.90	1719.0
159.00	1289.0
160.00	2764.0
161.00	4053.0
161.90	1134.0
162.90	380.0
163.90	460.0
164.90	3254.0
166.00	2890.0
167.00	20024.0
168.00	12842.0
169.00	2065.0
169.90	658.0
170.90	715.0
172.00	1611.0
173.00	2428.0
174.00	3987.0
175.00	7098.0
176.00	2117.0
177.00	3387.0
177.90	1180.0
179.00	14634.0
180.00	10302.0
181.00	4342.0
182.00	721.0
182.90	386.0
184.00	1144.0
185.00	6655.0
186.00	49736.0
187.00	13855.0
188.00	1727.0
189.00	3238.0
189.90	471.0
191.00	1609.0
192.00	4723.0
193.00	4932.0
194.00	993.0
194.90	455.0
196.00	9645.0
196.90	1349.0
198.00	347648.0
199.00	23304.0
200.00	2197.0
201.40	1547.0
203.00	2866.0
204.00	12837.0
205.00	22176.0
206.00	87312.0
207.00	13552.0
208.10	3666.0
209.00	1167.0
210.00	1435.0
211.00	3630.0
212.90	198.0
215.00	931.0
216.00	2330.0
217.00	23744.0
217.90	3121.0
219.00	358.0
220.00	395.0
221.10	14473.0
221.90	3208.0
223.00	5697.0
224.00	53592.0
225.00	13367.0
226.00	1374.0
227.00	22864.0
228.00	3229.0
229.00	4520.0
230.00	716.0
231.00	1907.0
232.00	377.0
232.90	453.0
234.00	1235.0
235.00	1577.0

Instrument :
BNA_N
ClientSampleId :
DFTPP252

235.90	1167.0
237.00	2503.0
238.00	414.0
239.00	926.0
240.00	702.0
241.00	1184.0
242.00	2663.0
243.00	2856.0
244.00	40584.0
245.00	5922.0
246.00	7717.0
247.00	1575.0
248.00	342.0
249.00	1204.0
250.00	299.0
251.00	308.0
252.00	337.0
253.00	1008.0
254.00	1191.0
255.00	208000.0
256.00	28024.0
257.00	2321.0
258.00	12012.0
259.00	1929.0
259.90	359.0
261.00	324.0
263.00	164.0
264.00	357.0
265.00	4874.0
266.00	722.0
266.90	185.0
269.90	268.0
270.90	582.0
272.00	574.0
273.00	6317.0
274.00	15960.0
275.00	88480.0
276.10	12213.0
277.00	7864.0
278.00	1717.0
278.90	308.0
281.00	428.0
282.00	272.0
283.00	993.0
283.90	734.0
285.10	1155.0
286.00	249.0
288.90	285.0
290.00	201.0
291.10	243.0
292.00	447.0
293.00	1951.0
294.00	694.0
295.00	444.0
296.00	26216.0
297.00	3169.0
297.90	278.0
300.90	401.0
302.00	395.0
303.00	3056.0
304.00	830.0
305.10	229.0
307.90	341.0
308.90	213.0
309.90	297.0
312.90	215.0
314.00	1306.0
315.00	2840.0
316.00	1737.0
317.00	350.0
321.10	783.0
321.90	411.0
323.00	8903.0
324.00	1528.0
325.00	189.0
327.00	1470.0
328.00	805.0
329.00	156.0
332.10	632.0
333.00	741.0
334.10	5267.0
335.10	1163.0
335.90	155.0
341.10	1089.0
341.90	226.0
346.00	1864.0
347.00	288.0
351.00	154.0
352.00	2550.0
353.00	1645.0
354.10	2143.0
355.00	631.0
358.90	193.0
364.90	10162.0
365.90	1732.0
366.90	160.0

Instrument :
BNA_N
ClientSampleId :
DFTPP252

369.90	222.0
370.90	529.0
372.00	4523.0
373.00	1158.0
376.90	161.0
383.00	1174.0
384.00	341.0
384.80	150.0
390.00	582.0
390.90	277.0
392.00	233.0
401.00	312.0
402.00	1925.0
403.00	2702.0
403.90	814.0
405.00	266.0
414.90	163.0
421.00	2098.0
422.00	1981.0
423.00	13059.0
424.00	4692.0
425.00	687.0
441.00	42328.0
442.00	284480.0
443.10	53528.0
444.00	3786.0
445.00	266.0

Instrument :
BNA_N
ClientSampleId :
DFTPP252

m/z	Abundance
36.00	152.0
37.00	519.0
38.00	1127.0
39.10	7409.0
40.00	1204.0
41.00	451.0
42.10	164.0
43.00	301.0
44.00	1959.0
45.00	320.0
49.00	848.0
50.00	25136.0
51.00	96320.0
52.00	5060.0
53.00	189.0
55.00	676.0
56.00	2769.0
57.00	6720.0
57.90	342.0
61.00	1236.0
62.00	1407.0
63.00	3960.0
64.00	436.0
65.00	2101.0
65.90	198.0
66.90	225.0
68.00	1497.0
69.00	99192.0
70.00	605.0
71.00	184.0
73.00	1408.0
74.00	10072.0
75.00	16068.0
76.00	5676.0
77.00	120200.0
78.00	8660.0
79.00	7279.0
80.00	5483.0
81.00	7907.0
82.00	2058.0
83.00	1891.0
83.90	355.0
85.00	1539.0
86.00	2642.0
87.00	1056.0
88.00	464.0
88.90	191.0
90.90	1669.0
91.90	2023.0
93.00	12983.0
93.90	858.0
94.90	284.0
96.00	712.0
96.90	366.0
98.00	10052.0
99.00	9148.0
100.00	811.0
101.00	4637.0
102.00	242.0
103.00	1312.0
104.00	2680.0
105.00	2748.0
106.00	936.0
107.00	34416.0
108.00	5794.0
109.00	1029.0
110.00	64112.0
111.00	8958.0
112.00	1153.0
113.00	428.0
114.90	222.0
116.00	1698.0
117.00	32176.0
118.00	2289.0
118.90	282.0
120.00	420.0
122.00	2279.0
123.00	3538.0
124.00	1638.0
125.00	1430.0
126.00	374.0
127.00	122768.0
128.00	8768.0
129.00	46032.0
130.00	4341.0
131.00	687.0
132.10	341.0
132.90	341.0
134.00	1219.0
135.00	3688.0
136.00	1405.0
137.00	2706.0

Instrument :
BNA_N
ClientSampleId :
DFTPP252

137.90	426.0
138.90	272.0
139.90	624.0
141.00	6123.0
142.00	1910.0
143.00	1272.0
144.00	349.0
145.00	292.0
146.00	1011.0
147.00	3690.0
148.00	8190.0
149.00	1923.0
150.00	597.0
151.10	830.0
151.80	545.0
152.90	1969.0
154.00	1537.0
155.00	3219.0
156.00	5058.0
157.10	883.0
157.90	1049.0
159.00	941.0
160.00	1815.0
161.00	2847.0
161.90	818.0
162.90	214.0
163.90	264.0
165.00	2292.0
165.90	2011.0
167.00	14317.0
168.00	9527.0
169.00	1448.0
170.00	362.0
171.00	507.0
171.90	1213.0
173.00	1551.0
174.00	2482.0
175.00	5050.0
176.00	1388.0
176.90	2313.0
177.90	786.0
179.00	9274.0
180.00	7122.0
181.00	2761.0
181.90	454.0
183.00	337.0
184.00	743.0
185.00	4026.0
186.00	33712.0
187.00	9738.0
188.00	1263.0
189.00	2001.0
190.00	339.0
191.00	1185.0
192.00	3102.0
193.00	3640.0
194.00	716.0
195.00	454.0
196.00	6964.0
196.90	1104.0
198.00	242304.0
199.00	15506.0
200.00	1392.0
201.40	882.0
202.90	1842.0
204.00	9090.0
205.00	15284.0
206.10	66384.0
207.10	9272.0
208.00	2272.0
209.00	918.0
210.10	1111.0
211.00	2615.0
213.00	256.0
215.00	756.0
216.00	1643.0
217.00	15626.0
218.00	2215.0
218.90	228.0
221.00	9653.0
221.90	2478.0
223.00	4007.0
224.00	38184.0
225.00	9582.0
226.00	1006.0
227.00	16066.0
228.00	2351.0
229.00	3416.0
230.00	520.0
231.00	1393.0
231.90	221.0
233.00	230.0
234.00	987.0
235.00	1015.0
236.00	700.0
237.00	1458.0

Instrument :
BNA_N
ClientSampleId :
DFTPP252

238.00	285.0
239.00	592.0
239.90	466.0
241.00	912.0
242.00	1883.0
243.00	2031.0
244.10	30968.0
245.00	3753.0
246.00	6457.0
247.00	1385.0
247.90	279.0
249.00	1079.0
250.00	218.0
251.00	276.0
252.00	355.0
253.00	609.0
254.10	965.0
255.00	154048.0
256.00	21624.0
257.00	1740.0
258.00	9137.0
259.00	1602.0
260.00	266.0
260.90	204.0
264.00	287.0
265.00	3461.0
266.00	398.0
270.00	221.0
271.00	329.0
272.00	473.0
273.00	4826.0
274.00	12528.0
275.00	66472.0
276.00	8235.0
277.00	6113.0
278.00	1152.0
279.00	260.0
281.10	547.0
282.00	251.0
283.00	659.0
284.00	443.0
285.00	1355.0
286.00	208.0
290.90	171.0
291.90	298.0
293.00	1400.0
294.00	469.0
295.00	407.0
296.00	21608.0
297.00	3092.0
297.90	191.0
301.00	238.0
302.00	395.0
303.00	2472.0
304.00	536.0
308.00	434.0
309.00	220.0
310.00	328.0
312.90	164.0
314.00	993.0
315.00	2743.0
316.00	1227.0
316.90	153.0
321.00	669.0
322.00	247.0
323.00	6866.0
324.10	1322.0
327.00	1316.0
328.00	651.0
328.90	160.0
332.00	262.0
333.00	702.0
334.00	4393.0
335.00	1186.0
341.00	743.0
341.90	235.0
346.00	1708.0
346.90	393.0
351.00	155.0
352.00	1805.0
353.00	1429.0
354.00	1745.0
355.00	366.0
365.00	8760.0
365.90	1252.0
370.00	221.0
371.00	538.0
372.00	3537.0
373.00	845.0
383.00	798.0
384.00	293.0
390.00	423.0
391.00	327.0
391.90	163.0
400.90	209.0
402.00	1558.0

Instrument :
BNA_N
ClientSampleId :
DFTPP252

403.00	2274.0
404.00	820.0
405.00	155.0
421.00	1914.0
422.00	1911.0
423.10	10039.0
424.00	3614.0
425.00	473.0
441.00	37560.0
442.00	196608.0
443.10	45904.0
444.10	3321.0
445.00	203.0

Instrument :
BNA_N
ClientSampleId :
DFTPP252