

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN101220\
 Data File : BN012144.D
 Acq On : 12 Oct 2020 10:34
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Oct 12 11:21:51 2020
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN092320.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Oct 12 11:15:56 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	59	0.00
2	1,4-Dioxane	0.400	0.327	18.3	54	0.00
3	n-Nitrosodimethylamine	0.400	0.382	4.5	66	0.00
4 S	2-Fluorophenol	0.400	0.355	11.3	58	0.00
5 S	Phenol-d6	0.400	0.353	11.8	59	0.00
6	bis(2-Chloroethyl)ether	0.400	0.335	16.3	56	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	61	0.00
8 S	Nitrobenzene-d5	0.400	0.365	8.8	63	0.00
9	Naphthalene	0.400	0.343	14.2	58	0.00
10	Hexachlorobutadiene	0.400	0.416	-4.0	68	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.359	10.3	61	0.00
12	2-Methylnaphthalene	0.400	0.346	13.5	59	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	68	0.00
14 S	2,4,6-Tribromophenol	0.400	0.451	-12.7	87	0.00
15 S	2-Fluorobiphenyl	0.400	0.347	13.3	64	0.00
16	Acenaphthylene	0.400	0.347	13.3	67	0.00
17	Acenaphthene	0.400	0.330	17.5	63	0.00
18	Fluorene	0.400	0.337	15.8	63	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	71	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.390	2.5	82	0.00
21	4-Bromophenyl-phenylether	0.400	0.361	9.8	71	0.00
22	Hexachlorobenzene	0.400	0.381	4.8	74	0.00
23	Atrazine	0.400	0.376	6.0	78	0.00
24	Pentachlorophenol	0.400	0.375	6.3	76	0.00
25	Phenanthrene	0.400	0.329	17.8	66	0.00
26	Anthracene	0.400	0.337	15.8	68	0.00
27 SURR	Fluoranthene-d10	0.400	0.360	10.0	72	0.00
28	Fluoranthene	0.400	0.348	13.0	70	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	79	0.00
30	Pyrene	0.400	0.326	18.5	70	0.00
31 S	Terphenyl-d14	0.400	0.344	14.0	74	0.00
32	Benzo(a)anthracene	0.400	0.333	16.8	73	0.00
33	Chrysene	0.400	0.343	14.2	76	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.359	10.3	81	0.00
35	Indeno(1,2,3-cd)pyrene	0.400	0.392	2.0	86	0.00
36 I	Perylene-d12	0.400	0.400	0.0	89	0.00
37	Benzo(b)fluoranthene	0.400	0.331	17.3	83	0.00
38	Benzo(k)fluoranthene	0.400	0.325	18.8	77	0.00
39 C	Benzo(a)pyrene	0.400	0.326	18.5	79	0.00
40	Dibenzo(a,h)anthracene	0.400	0.348	13.0	86	0.00
41	Benzo(g,h,i)perylene	0.400	0.344	14.0	83	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

LabSampleId :

SSTDCCC0.4