

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN101220\
 Data File : BN012151.D
 Acq On : 12 Oct 2020 17:45
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Oct 12 18:20:16 2020
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN092320.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Oct 12 11:15:56 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	52	0.00
2	1,4-Dioxane	0.400	0.334	16.5	48	0.00
3	n-Nitrosodimethylamine	0.400	0.413	-3.2	63	0.00
4 S	2-Fluorophenol	0.400	0.359	10.3	52	0.00
5 S	Phenol-d6	0.400	0.354	11.5	52	0.00
6	bis(2-Chloroethyl)ether	0.400	0.332	17.0	49	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	53	0.00
8 S	Nitrobenzene-d5	0.400	0.370	7.5	55	-0.01
9	Naphthalene	0.400	0.347	13.3	50	0.00
10	Hexachlorobutadiene	0.400	0.405	-1.3	57	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.361	9.8	53	0.00
12	2-Methylnaphthalene	0.400	0.354	11.5	52	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	57	0.00
14 S	2,4,6-Tribromophenol	0.400	0.415	-3.7	68	0.00
15 S	2-Fluorobiphenyl	0.400	0.351	12.3	54	0.00
16	Acenaphthylene	0.400	0.346	13.5	56	0.00
17	Acenaphthene	0.400	0.331	17.3	53	0.00
18	Fluorene	0.400	0.340	15.0	54	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	57	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.366	8.5	62	0.00
21	4-Bromophenyl-phenylether	0.400	0.368	8.0	59	-0.01
22	Hexachlorobenzene	0.400	0.388	3.0	61	-0.01
23	Atrazine	0.400	0.376	6.0	63	0.00
24	Pentachlorophenol	0.400	0.351	12.3	58	0.00
25	Phenanthrene	0.400	0.326	18.5	53	0.00
26	Anthracene	0.400	0.333	16.8	54	0.00
27 SURR	Fluoranthene-d10	0.400	0.356	11.0	57	0.00
28	Fluoranthene	0.400	0.344	14.0	56	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	64	0.00
30	Pyrene	0.400	0.318	20.5	55	0.00
31 S	Terphenyl-d14	0.400	0.337	15.8	59	0.00
32	Benzo(a)anthracene	0.400	0.319	20.3	57	0.01
33	Chrysene	0.400	0.348	13.0	62	0.01
34	Bis(2-ethylhexyl)phthalate	0.400	0.336	16.0	62	0.00
35	Indeno(1,2,3-cd)pyrene	0.400	0.399	0.3	71	0.00
36 I	Perylene-d12	0.400	0.400	0.0	75	0.00
37	Benzo(b)fluoranthene	0.400	0.314	21.5	66	0.00
38	Benzo(k)fluoranthene	0.400	0.312	22.0	62	0.00
39 C	Benzo(a)pyrene	0.400	0.325	18.8	66	0.00
40	Dibenzo(a,h)anthracene	0.400	0.342	14.5	71	0.00
41	Benzo(g,h,i)perylene	0.400	0.341	14.8	70	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

LabSampleId :

SSTDCCC0.4