

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN101221\
 Data File : BN016894.D
 Acq On : 11 Oct 2021 23:23
 Operator : CG/JU
 Sample : PB139657BL
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 PB139657BL

Quant Time: Oct 12 02:35:58 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN101221.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Oct 12 02:31:22 2021
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.698	152	23103	0.400	ng	0.00
7) Naphthalene-d8	10.457	136	103199	0.400	ng	0.00
13) Acenaphthene-d10	14.294	164	52863	0.400	ng	0.00
19) Phenanthrene-d10	17.042	188	97314	0.400	ng	0.00
29) Chrysene-d12	21.227	240	89703	0.400	ng	#-0.01
35) Perylene-d12	23.474	264	92144	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.337	112	21750	0.368	ng	0.04
5) Phenol-d6	6.886	99	20372	0.307	ng	0.02
8) Nitrobenzene-d5	8.835	82	23969	0.344	ng	0.00
11) 2-Methylnaphthalene-d10	12.043	152	60020	0.401	ng	0.00
14) 2,4,6-Tribromophenol	15.791	330	4111	0.160	ng	0.00
15) 2-Fluorobiphenyl	12.924	172	75602	0.374	ng	0.00
27) Fluoranthene-d10	19.073	212	102655	0.402	ng	0.00
31) Terphenyl-d14	19.678	244	85903	0.421	ng	0.00
Target Compounds						
3) n-Nitrosodimethylamine	3.567	42	659	0.035	ng	# 1
6) bis(2-Chloroethyl)ether	7.126	93	30	0.000	ng	# 48
9) Naphthalene	10.495	128	950	0.003	ng	# 76
10) Hexachlorobutadiene	10.787	225	8	0.000	ng	# 83
12) 2-Methylnaphthalene	12.114	142	194	0.001	ng	# 75
16) Acenaphthylene	14.015	152	41	0.000	ng	# 28
17) Acenaphthene	14.357	154	30	0.000	ng	# 81
18) Fluorene	15.347	166	50	0.000	ng	# 80
20) 4,6-Dinitro-2-methylph...	15.449	198	11	0.190	ng	# 1
21) 4-Bromophenyl-phenylether	16.251	248	4	0.000	ng	# 50
22) Hexachlorobenzene	16.348	284	6	0.000	ng	# 66
23) Atrazine	16.616	200	4	0.000	ng	# 1
24) Pentachlorophenol	16.701	266	79	0.161	ng	# 85
25) Phenanthrene	17.103	178	532	0.002	ng	# 1
26) Anthracene	17.164	178	42	0.000	ng	# 80
28) Fluoranthene	19.103	202	92	0.000	ng	# 64
30) Pyrene	19.465	202	99	0.000	ng	99
32) Benzo(a)anthracene	21.227	228	375	0.001	ng	# 76
33) Chrysene	21.270	228	180	0.001	ng	# 52
34) Bis(2-ethylhexyl)phtha...	21.164	149	730	0.005	ng	97
36) Indeno(1,2,3-cd)pyrene	25.754	276	260	0.001	ng	# 70
37) Benzo(b)fluoranthene	22.803	252	306	0.001	ng	# 1
39) Benzo(a)pyrene	23.375	252	82	0.000	ng	# 1
40) Dibenzo(a,h)anthracene	25.781	278	196	0.001	ng	# 1
41) Benzo(g,h,i)perylene	26.438	276	219	0.001	ng	# 1


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN101221\
 Data File : BN016894.D
 Acq On : 11 Oct 2021 23:23
 Operator : CG/JU
 Sample : PB139657BL
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 PB139657BL

Quant Time: Oct 12 02:35:58 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN101221.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Oct 12 02:31:22 2021
 Response via : Initial Calibration









#1
 1,4-Dichlorobenzene-d4
 Concen: 0.400 ng
 RT: 7.698 min Scan# 50
 Delta R.T. 0.010 min
 Lab File: BN016894.D
 Acq: 11 Oct 2021 23:23




Instrument :
 BNA_N
 ClientSampleId :
 PB139657BL


Tgt Ion:152 Resp: 23103
 Ion Ratio Lower Upper
 152 100
 150 151.0 124.2 186.4
 115 56.5 47.0 70.6



#3
 n-Nitrosodimethylamine
 Concen: 0.035 ng
 RT: 3.567 min Scan# 53
 Delta R.T. -0.046 min
 Lab File: BN016894.D
 Acq: 11 Oct 2021 23:23

Tgt Ion: 42 Resp: 659
 Ion Ratio Lower Upper
 42 100
 74 0.0 117.2 175.8#
 44 164.5 5.4 8.0#









#4
2-Fluorophenol
Concen: 0.368 ng
RT: 5.337 min Scan# 241
Delta R.T. 0.037 min
Lab File: BN016894.D
Acq: 11 Oct 2021 23:23




Instrument :
BNA_N
ClientSampleId :
PB139657BL


Tgt Ion:112 Resp: 21750
Ion Ratio Lower Upper
112 100
64 55.8 44.5 66.7
63 28.1 22.1 33.1



#5
Phenol-d6
Concen: 0.307 ng
RT: 6.886 min Scan# 413
Delta R.T. 0.018 min
Lab File: BN016894.D
Acq: 11 Oct 2021 23:23

Tgt Ion: 99 Resp: 20372
Ion Ratio Lower Upper
99 100
42 17.2 13.9 20.9
71 31.7 24.5 36.7









#6
bis(2-Chloroethyl)ether
Concen: 0.000 ng
RT: 7.126 min Scan# 439
Delta R.T. -0.000 min
Lab File: BN016894.D
Acq: 11 Oct 2021 23:23




Instrument :
BNA_N
ClientSampleId :
PB139657BL


Tgt Ion: 93 Resp: 30
Ion Ratio Lower Upper
93 100
63 33.3 59.8 89.6#
95 0.0 26.4 39.6#



#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.457 min Scan# 737
Delta R.T. -0.000 min
Lab File: BN016894.D
Acq: 11 Oct 2021 23:23

Tgt Ion: 136 Resp: 103199
Ion Ratio Lower Upper
136 100
137 11.0 8.8 13.2
54 4.7 3.8 5.6
68 4.1 3.4 5.0









#8
Nitrobenzene-d5
Concen: 0.344 ng
RT: 8.835 min Scan# 609
Delta R.T. -0.000 min
Lab File: BN016894.D
Acq: 11 Oct 2021 23:23




Instrument :
BNA_N
ClientSampleId :
PB139657BL


Tgt Ion: 82 Resp: 23969
Ion Ratio Lower Upper
82 100
128 47.6 37.8 56.6
54 44.1 35.3 52.9



#9
Naphthalene
Concen: 0.003 ng
RT: 10.495 min Scan# 740
Delta R.T. -0.000 min
Lab File: BN016894.D
Acq: 11 Oct 2021 23:23

Tgt Ion: 128 Resp: 950
Ion Ratio Lower Upper
128 100
129 22.3 8.8 13.2#
127 21.3 10.7 16.1#









#10
Hexachlorobutadiene
Concen: 0.000 ng
RT: 10.787 min Scan# 70
Delta R.T. -0.000 min
Lab File: BN016894.D
Acq: 11 Oct 2021 23:23




Instrument :
BNA_N
ClientSampleId :
PB139657BL


Tgt Ion:225 Resp: 8
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 50.0 50.9 76.3#



#11
2-Methylnaphthalene-d10
Concen: 0.401 ng
RT: 12.043 min Scan# 940
Delta R.T. -0.000 min
Lab File: BN016894.D
Acq: 11 Oct 2021 23:23

Tgt Ion:152 Resp: 60020
Ion Ratio Lower Upper
152 100
151 20.4 16.3 24.5









#12
2-Methylnaphthalene
Concen: 0.001 ng
RT: 12.114 min Scan# 91
Delta R.T. -0.000 min
Lab File: BN016894.D
Acq: 11 Oct 2021 23:23




Instrument :
BNA_N
ClientSampleId :
PB139657BL


Tgt Ion:142 Resp: 194
Ion Ratio Lower Upper
142 100
141 97.0 71.0 106.4
115 72.8 26.6 40.0#



#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.294 min Scan# 1185
Delta R.T. -0.000 min
Lab File: BN016894.D
Acq: 11 Oct 2021 23:23

Tgt Ion:164 Resp: 52863
Ion Ratio Lower Upper
164 100
162 102.2 81.8 122.8
160 45.7 37.0 55.4









#14
2,4,6-Tribromophenol
Concen: 0.160 ng
RT: 15.791 min Scan# 11
Delta R.T. -0.000 min
Lab File: BN016894.D
Acq: 11 Oct 2021 23:23




Instrument :
BNA_N
ClientSampleId :
PB139657BL


Tgt Ion:330 Resp: 4111
Ion Ratio Lower Upper
330 100
332 0.0 0.0 0.0
141 31.3 24.3 36.5



#15
2-Fluorobiphenyl
Concen: 0.374 ng
RT: 12.924 min Scan# 1077
Delta R.T. -0.000 min
Lab File: BN016894.D
Acq: 11 Oct 2021 23:23

Tgt Ion:172 Resp: 75602
Ion Ratio Lower Upper
172 100
171 34.2 27.7 41.5
170 22.3 18.3 27.5









#16
Acenaphthylene
Concen: 0.000 ng
RT: 14.015 min Scan# 1163
Delta R.T. -0.000 min
Lab File: BN016894.D
Acq: 11 Oct 2021 23:23




Instrument :
BNA_N
ClientSampleId :
PB139657BL


Tgt Ion:152 Resp: 41
Ion Ratio Lower Upper
152 100
151 0.0 15.5 23.3#
153 58.5 10.4 15.6#



#17
Acenaphthene
Concen: 0.000 ng
RT: 14.357 min Scan# 1190
Delta R.T. -0.000 min
Lab File: BN016894.D
Acq: 11 Oct 2021 23:23


Tgt Ion:154 Resp: 30
Ion Ratio Lower Upper
154 100
153 93.3 90.6 135.8
152 40.0 43.7 65.5#








#18
Fluorene
Concen: 0.000 ng
RT: 15.347 min Scan# 1268
Delta R.T. -0.000 min
Lab File: BN016894.D
Acq: 11 Oct 2021 23:23


Instrument :
BNA_N
ClientSampleId :
PB139657BL





Tgt Ion:166 Resp: 50
Ion Ratio Lower Upper
166 100
165 80.0 78.2 117.2
167 28.0 10.8 16.2#




#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 17.042 min Scan# 1405
Delta R.T. -0.000 min
Lab File: BN016894.D
Acq: 11 Oct 2021 23:23



Tgt Ion:188 Resp: 97314
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 6.7 5.4 8.2









#20
4,6-Dinitro-2-methylphenol
Concen: 0.190 ng
RT: 15.449 min Scan# 11
Delta R.T. 0.013 min
Lab File: BN016894.D
Acq: 11 Oct 2021 23:23




Instrument :
BNA_N
ClientSampleId :
PB139657BL


Tgt Ion:198 Resp: 11
Ion Ratio Lower Upper
198 100
182 98.0 11.1 16.7#
77 0.0 0.0 0.0



#21
4-Bromophenyl-phenylether
Concen: 0.000 ng
RT: 16.251 min Scan# 1340
Delta R.T. 0.012 min
Lab File: BN016894.D
Acq: 11 Oct 2021 23:23

Tgt Ion:248 Resp: 4
Ion Ratio Lower Upper
248 100
250 106.3 75.8 113.6
141 141.7 47.4 71.2#









#22
Hexachlorobenzene
Concen: 0.000 ng
RT: 16.348 min Scan# 11
Delta R.T. -0.000 min
Lab File: BN016894.D
Acq: 11 Oct 2021 23:23




Instrument :
BNA_N
ClientSampleId :
PB139657BL


Tgt Ion:284 Resp: 6
Ion Ratio Lower Upper
284 100
142 0.0 25.1 37.7#
249 33.3 23.0 34.4



#23
Atrazine
Concen: 0.000 ng
RT: 16.616 min Scan# 1370
Delta R.T. 0.097 min
Lab File: BN016894.D
Acq: 11 Oct 2021 23:23

Tgt Ion:200 Resp: 4
Ion Ratio Lower Upper
200 100
173 102.0 19.7 29.5#
215 100.0 39.5 59.3#









#24
Pentachlorophenol
Concen: 0.161 ng
RT: 16.701 min Scan# 1410
Delta R.T. 0.012 min
Lab File: BN016894.D
Acq: 11 Oct 2021 23:23




Instrument :
BNA_N
ClientSampleId :
PB139657BL


Tgt Ion:266 Resp: 79
Ion Ratio Lower Upper
266 100
264 49.4 50.6 76.0#
268 73.4 51.3 76.9



#25
Phenanthrene
Concen: 0.002 ng
RT: 17.103 min Scan# 1410
Delta R.T. 0.024 min
Lab File: BN016894.D
Acq: 11 Oct 2021 23:23

Tgt Ion:178 Resp: 532
Ion Ratio Lower Upper
178 100
176 0.0 15.0 22.6#
179 896.6 12.3 18.5#









#26
 Anthracene
 Concen: 0.000 ng
 RT: 17.164 min Scan# 1415
 Delta R.T. -0.000 min
 Lab File: BN016894.D
 Acq: 11 Oct 2021 23:23




Instrument :
 BNA_N
 ClientSampleId :
 PB139657BL


Tgt Ion:	178	Resp:	42
Ion Ratio	Lower	Upper	
178	100		
176	21.4	14.6	22.0
179	0.0	12.2	18.4



#27
 Fluoranthene-d10
 Concen: 0.402 ng
 RT: 19.073 min Scan# 1698
 Delta R.T. -0.000 min
 Lab File: BN016894.D
 Acq: 11 Oct 2021 23:23

Tgt Ion:	212	Resp:	102655
Ion Ratio	Lower	Upper	
212	100		
106	13.5	10.8	16.2
104	7.4	5.9	8.9









#28
Fluoranthene
Concen: 0.000 ng
RT: 19.103 min Scan# 11
Delta R.T. -0.000 min
Lab File: BN016894.D
Acq: 11 Oct 2021 23:23




Instrument :
BNA_N
ClientSampleId :
PB139657BL


Tgt Ion:202 Resp: 92
Ion Ratio Lower Upper
202 100
101 0.0 9.0 13.6#
203 0.0 13.8 20.8#



#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.227 min Scan# 2001
Delta R.T. -0.011 min
Lab File: BN016894.D
Acq: 11 Oct 2021 23:23

Tgt Ion:240 Resp: 89703
Ion Ratio Lower Upper
240 100
120 10.5 4.1 6.1#
236 26.3 20.1 30.1









#30
Pyrene
Concen: 0.000 ng
RT: 19.465 min Scan# 11
Delta R.T. -0.000 min
Lab File: BN016894.D
Acq: 11 Oct 2021 23:23




Instrument :
BNA_N
ClientSampleId :
PB139657BL


Tgt Ion:202 Resp: 99
Ion Ratio Lower Upper
202 100
200 20.2 16.4 24.6
203 18.2 14.0 21.0



#31
Terphenyl-d14
Concen: 0.421 ng
RT: 19.678 min Scan# 1820
Delta R.T. -0.005 min
Lab File: BN016894.D
Acq: 11 Oct 2021 23:23

Tgt Ion:244 Resp: 85903
Ion Ratio Lower Upper
244 100
212 7.1 5.5 8.3
122 12.2 8.6 13.0









#32
Benzo(a)anthracene
Concen: 0.001 ng
RT: 21.227 min Scan# 2001
Delta R.T. 0.011 min
Lab File: BN016894.D
Acq: 11 Oct 2021 23:23




Instrument :
BNA_N
ClientSampleId :
PB139657BL


Tgt Ion:228 Resp: 375
Ion Ratio Lower Upper
228 100
226 29.3 21.2 31.8
229 42.2 15.8 23.6#



#33
Chrysene
Concen: 0.001 ng
RT: 21.270 min Scan# 2005
Delta R.T. -0.000 min
Lab File: BN016894.D
Acq: 11 Oct 2021 23:23

Tgt Ion:228 Resp: 180
Ion Ratio Lower Upper
228 100
226 50.0 23.3 34.9#
229 47.2 15.8 23.6#









#34
 Bis(2-ethylhexyl)phthalate
 Concen: 0.005 ng
 RT: 21.164 min Scan# 1995
 Delta R.T. -0.000 min
 Lab File: BN016894.D
 Acq: 11 Oct 2021 23:23




Instrument :
 BNA_N
 ClientSampleId :
 PB139657BL


Tgt Ion:149	Resp:	730
Ion Ratio	Lower	Upper
149	100	
167	26.2	22.2 33.2
279	4.2	3.5 5.3



#35
 Perylene-d12
 Concen: 0.400 ng
 RT: 23.474 min Scan# 2424
 Delta R.T. -0.003 min
 Lab File: BN016894.D
 Acq: 11 Oct 2021 23:23

Tgt Ion:264	Resp:	92144
Ion Ratio	Lower	Upper
264	100	
260	24.2	19.4 29.2
265	23.9	19.2 28.8









#36
 Indeno(1,2,3-cd)pyrene
 Concen: 0.001 ng
 RT: 25.754 min Scan# 3089
 Delta R.T. 0.003 min
 Lab File: BN016894.D
 Acq: 11 Oct 2021 23:23




Instrument :
 BNA_N
 ClientSampleId :
 PB139657BL


Tgt Ion:276 Resp: 260
 Ion Ratio Lower Upper
 276 100
 138 0.0 20.1 30.1#
 277 19.6 20.1 30.1#



#37
 Benzo(b)fluoranthene
 Concen: 0.001 ng
 RT: 22.803 min Scan# 2228
 Delta R.T. -0.000 min
 Lab File: BN016894.D
 Acq: 11 Oct 2021 23:23

Tgt Ion:252 Resp: 306
 Ion Ratio Lower Upper
 252 100
 253 110.4 17.8 26.8#
 125 97.1 8.1 12.1#





#39

Benzo(a)pyrene

Concen: 0.000 ng

RT: 23.375 min Scan# 21

Delta R.T. -0.003 min

Lab File: BN016894.D


Acq: 11 Oct 2021 23:23

Instrument :

BNA_N

ClientSampleId :

PB139657BL






Tgt Ion:252 Resp: 82

Ion Ratio Lower Upper

252 100

253 152.6 17.8 26.8#

125 127.6 9.0 13.6#



#40

Dibenzo(a,h)anthracene

Concen: 0.001 ng

RT: 25.781 min Scan# 3098

Delta R.T. 0.010 min

Lab File: BN016894.D

Acq: 11 Oct 2021 23:23




Tgt Ion:278 Resp: 196


Ion Ratio Lower Upper

278 100

139 123.4 13.4 20.0#


279 144.1 19.2 28.8#





#41
 Benzo(g,h,i)perylene
 Concen: 0.001 ng
 RT: 26.438 min Scan# 31
 Delta R.T. -0.003 min
 Lab File: BN016894.D
 Acq: 11 Oct 2021 23:23

Instrument :
 BNA_N
 ClientSampleId :
 PB139657BL



Tgt Ion: 276 Resp: 219
 Ion Ratio Lower Upper
 276 100
 277 75.7 19.0 28.6#
 138 91.3 16.7 25.1#

