

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN101221\
 Data File : BN016893.D
 Acq On : 11 Oct 2021 22:10
 Operator : CG/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 ICVBN101221

Quant Time: Oct 12 02:35:49 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN101221.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Oct 12 02:31:22 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	91	0.00
2	1,4-Dioxane	0.400	0.389	2.8	95	0.00
3	n-Nitrosodimethylamine	0.400	0.374	6.5	91	0.00
4 S	2-Fluorophenol	0.400	0.366	8.5	87	0.00
5 S	Phenol-d6	0.400	0.365	8.8	88	0.00
6	bis(2-Chloroethyl)ether	0.400	0.386	3.5	91	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	91	0.00
8 S	Nitrobenzene-d5	0.400	0.367	8.3	96	-0.01
9	Naphthalene	0.400	0.375	6.3	90	0.00
10	Hexachlorobutadiene	0.400	0.387	3.3	93	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.379	5.3	92	0.00
12	2-Methylnaphthalene	0.400	0.386	3.5	92	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	93	0.00
14 S	2,4,6-Tribromophenol	0.400	0.326	18.5	90	0.00
15 S	2-Fluorobiphenyl	0.400	0.380	5.0	92	0.00
16	Acenaphthylene	0.400	0.368	8.0	92	0.00
17	Acenaphthene	0.400	0.375	6.3	92	0.00
18	Fluorene	0.400	0.372	7.0	91	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	93	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.359	10.3	87	0.00
21	4-Bromophenyl-phenylether	0.400	0.373	6.8	95	0.00
22	Hexachlorobenzene	0.400	0.384	4.0	93	0.00
23	Atrazine	0.400	0.341	14.8	94	0.00
24	Pentachlorophenol	0.400	0.362	9.5	82	0.00
25	Phenanthrene	0.400	0.381	4.8	94	0.00
26	Anthracene	0.400	0.371	7.3	94	0.00
27 SURR	Fluoranthene-d10	0.400	0.370	7.5	94	0.00
28	Fluoranthene	0.400	0.387	3.3	94	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	95	-0.01
30	Pyrene	0.400	0.373	6.8	94	0.00
31 S	Terphenyl-d14	0.400	0.369	7.8	95	0.00
32	Benzo(a)anthracene	0.400	0.372	7.0	94	0.00
33	Chrysene	0.400	0.386	3.5	100	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.292	27.0#	91	0.00
35 I	Perylene-d12	0.400	0.400	0.0	98	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.380	5.0	99	-0.01
37	Benzo(b)fluoranthene	0.400	0.381	4.8	99	0.00
38	Benzo(k)fluoranthene	0.400	0.380	5.0	96	0.00
39 C	Benzo(a)pyrene	0.400	0.377	5.8	97	0.00
40	Dibenzo(a,h)anthracene	0.400	0.380	5.0	99	0.00
41	Benzo(g,h,i)perylene	0.400	0.377	5.8	99	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

ClientSampleId :

ICVBN101221