

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN101222\
 Data File : BN022059.D
 Acq On : 12 Oct 2022 10:50
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTD020340

Quant Time: Oct 13 02:13:56 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN101222.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Oct 13 02:12:15 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.957	152	106431	20.000	ng/u1	0.00
20) Naphthalene-d8	10.793	136	514949	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.634	164	332396	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.386	188	726312	20.000	ng/u1	0.00
79) Chrysene-d12	21.592	240	723403	20.000	ng/u1	0.00
88) Perylene-d12	24.068	264	793522	20.000	ng/u1	0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.252	96	23392	8.318	ng/uL	0.00
4) Pyridine-d5	3.699	84	176020	20.704	ng/u1	0.00
7) Phenol-d5	7.116	99	207795	20.194	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.281	67	132686	20.141	ng/u1	0.00
11) 2-Chlorophenol-d4	7.481	132	150325	21.179	ng/u1	0.00
15) 4-Methylphenol-d8	8.681	113	167628	20.620	ng/u1	0.00
21) Nitrobenzene-d5	9.140	128	78504	21.478	ng/u1	0.00
24) 2-Nitrophenol-d4	9.869	143	83416	22.386	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.410	165	151431	21.489	ng/u1	0.00
31) 4-Chloroaniline-d4	10.934	131	248539	20.885	ng/u1	0.00
46) Dimethylphthalate-d6	14.045	166	489773	20.966	ng/u1	0.00
49) Acenaphthylene-d8	14.328	160	563940	22.189	ng/u1	0.00
54) 4-Nitrophenol-d4	14.834	143	106750	21.703	ng/u1	0.00
60) Fluorene-d10	15.628	176	408948	20.679	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.751	200	81924	20.591	ng/u1	0.00
73) Anthracene-d10	17.492	188	650408	20.734	ng/u1	0.00
81) Pyrene-d10	19.792	212	763089	21.695	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.904	264	821961	21.421	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.293	88	25350	8.321	ng/uL	93
5) Pyridine	3.716	79	176295	20.633	ng/u1	99
6) Benzaldehyde	7.087	77	123640	23.987	ng/u1	99
8) Phenol	7.140	94	222014	20.274	ng/u1	99
10) Bis(2-Chloroethyl)ether	7.381	93	179812	20.592	ng/u1	99
12) 2-Chlorophenol	7.516	128	167216	21.413	ng/u1	99
13) 2-Methylphenol	8.410	108	164626	20.297	ng/u1	98
14) 2,2'-oxybis(1-Chloropr...	8.499	45	254062	19.459	ng/u1	99
16) Acetophenone	8.799	105	276605	20.712	ng/u1	97
17) N-Nitroso-di-n-propyla...	8.781	70	145523	20.822	ng/u1	99
18) 4-Methylphenol	8.746	108	186613	20.498	ng/u1	98
19) Hexachloroethane	9.052	117	65221	20.832	ng/u1	99
22) Nitrobenzene	9.181	77	210238	20.934	ng/u1	99
23) Isophorone	9.710	82	413805	21.083	ng/u1	99
25) 2-Nitrophenol	9.899	139	97215	22.011	ng/u1	98
26) 2,4-Dimethylphenol	9.957	107	198223	20.806	ng/u1	97
27) Bis(2-Chloroethoxy)met...	10.198	93	253458	20.396	ng/u1	99
29) 2,4-Dichlorophenol	10.434	162	159988	21.285	ng/u1	98
30) Naphthalene	10.840	128	573018	20.502	ng/u1	99
32) 4-Chloroaniline	10.957	127	257734	20.574	ng/u1	99
33) Hexachlorobutadiene	11.122	225	85191	20.482	ng/u1	96
34) Caprolactam	11.763	113	60732	20.773	ng/u1	94
35) 4-Chloro-3-methylphenol	12.081	107	192394	21.039	ng/u1	97
36) 2-Methylnaphthalene	12.457	142	397668	20.898	ng/u1	97

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN101222\
 Data File : BN022059.D
 Acq On : 12 Oct 2022 10:50
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1


Instrument :
 BNA_N
 ClientSampleId :
 SSTD020340


Quant Time: Oct 13 02:13:56 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN101222.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Oct 13 02:12:15 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.675	142	399666	20.781	ng/ul	93
39) 1,2,4,5-Tetrachloroben...	12.822	216	180996	21.334	ng/ul	94
40) Hexachlorocyclopentadiene	12.798	237	100707	22.559	ng/ul	98
41) 2,4,6-Trichlorophenol	13.069	196	129400	22.632	ng/ul	100
42) 2,4,5-Trichlorophenol	13.134	196	136218	21.651	ng/ul	99
43) 1,1'-Biphenyl	13.469	154	503236	20.764	ng/ul	97
44) 2-Chloronaphthalene	13.510	162	394355	20.871	ng/ul	93
45) 2-Nitroaniline	13.722	65	131328	22.564	ng/ul	97
47) Dimethylphthalate	14.092	163	525765	21.182	ng/ul	100
48) 2,6-Dinitrotoluene	14.216	165	100742	22.741	ng/ul	99
50) Acenaphthylene	14.357	152	644120	22.055	ng/ul	99
51) 3-Nitroaniline	14.545	138	118867	22.275	ng/ul	100
52) Acenaphthene	14.698	153	434814	20.462	ng/ul	98
53) 2,4-Dinitrophenol	14.751	184	65290	21.398	ng/ul	99
55) 4-Nitrophenol	14.851	109	84266	20.937	ng/ul	92
56) Dibenzofuran	15.034	168	604284	21.096	ng/ul	98
57) 2,4-Dinitrotoluene	15.004	165	151269	22.583	ng/ul	96
58) 2,3,4,6-Tetrachlorophenol	15.263	232	116105	21.726	ng/ul	96
59) Diethylphthalate	15.457	149	546184	21.224	ng/ul	98
61) Fluorene	15.686	166	482865	20.393	ng/ul	93
62) 4-Chlorophenyl-phenyle...	15.681	204	217660	20.247	ng/ul	94
63) 4-Nitroaniline	15.710	138	120483	22.692	ng/ul	94
66) 4,6-Dinitro-2-methylph...	15.763	198	88663	20.768	ng/ul	98
67) N-Nitrosodiphenylamine	15.892	169	425277	20.233	ng/ul	97
68) 4-Bromophenyl-phenylether	16.575	248	135048	20.685	ng/ul	86
69) Hexachlorobenzene	16.686	284	157450	20.340	ng/ul	95
70) Atrazine	16.851	200	151462	21.316	ng/ul	96
71) Pentachlorophenol	17.033	266	98038	20.866	ng/ul	91
72) Phenanthrene	17.433	178	815981	20.631	ng/ul	99
74) Anthracene	17.528	178	814703	20.686	ng/ul	97
75) 1,2,3,4-Tetrachloroben...	13.434	216	184519	20.975	ng/uL	97
76) Pentachlorobenzene	14.951	250	198343	20.835	ng/uL	96
77) Carbazole	17.798	167	770947	21.208	ng/ul	99
78) Di-n-butylphthalate	18.363	149	989948	22.239	ng/ul	98
80) Fluoranthene	19.457	202	963181	21.562	ng/ul	99
82) Pyrene	19.822	202	965951	20.492	ng/ul	100
83) Butylbenzylphthalate	20.721	149	486275	24.193	ng/ul	95
84) 3,3'-Dichlorobenzidine	21.510	252	366530	23.303	ng/ul	95
85) Benzo(a)anthracene	21.580	228	868249	19.105	ng/ul	97
86) Bis(2-ethylhexyl)phtha...	21.498	149	907271	31.086	ng/ul	100
87) Chrysene	21.633	228	920673	20.590	ng/ul	98
89) Di-n-octyl phthalate	22.445	149	1203699	21.752	ng/ul	100
90) Benzo(b)fluoranthene	23.304	252	1061625	21.190	ng/ul	96
91) Benzo(k)fluoranthene	23.357	252	1054965	21.349	ng/ul	95
93) Benzo(a)pyrene	23.962	252	959470	21.690	ng/ul	96
94) Indeno(1,2,3-cd)pyrene	26.668	276	1187543	21.145	ng/ul	99
95) Dibenzo(a,h)anthracene	26.686	278	1088178	21.426	ng/ul#	95
96) Benzo(g,h,i)perylene	27.468	276	1064069	21.020	ng/ul	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Quant Time: Oct 13 02:13:56 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN101222.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Oct 13 02:12:15 2022
Response via : Initial Calibration








#2
1,4-Dioxane
Concen: 8.321 ng/uL
RT: 3.293 min Scan# 51
Delta R.T. -0.000 min
Lab File: BN022059.D
Acq: 12 Oct 2022 10:50

Instrument :
BNA_N
ClientSampleId :
SSTD020340







Tgt Ion: 88 Resp: 25350
Ion Ratio Lower Upper
88 100
43 29.8 22.6 34.0
58 67.4 59.9 89.9



#5
Pyridine
Concen: 20.633 ng/uL
RT: 3.716 min Scan# 124
Delta R.T. -0.000 min
Lab File: BN022059.D
Acq: 12 Oct 2022 10:50

Tgt Ion: 79 Resp: 176295
Ion Ratio Lower Upper
79 100
52 76.1 61.7 92.5
51 34.5 27.8 41.6









#6
Benzaldehyde
Concen: 23.987 ng/ul
RT: 7.087 min Scan# 698
Delta R.T. -0.006 min
Lab File: BN022059.D
Acq: 12 Oct 2022 10:50




Instrument :
BNA_N
ClientSampleId :
SSTD020340


Tgt Ion: 77 Resp: 123640
Ion Ratio Lower Upper
77 100
105 97.7 78.6 118.0
106 92.4 75.4 113.2



#8
Phenol
Concen: 20.274 ng/ul
RT: 7.140 min Scan# 706
Delta R.T. -0.000 min
Lab File: BN022059.D
Acq: 12 Oct 2022 10:50

Tgt Ion: 94 Resp: 222014
Ion Ratio Lower Upper
94 100
65 27.2 22.1 33.1
66 35.8 29.0 43.6









#10
Bis(2-Chloroethyl)ether
Concen: 20.592 ng/ul
RT: 7.381 min Scan# 747
Delta R.T. -0.000 min
Lab File: BN022059.D
Acq: 12 Oct 2022 10:50




Instrument :
BNA_N
ClientSampleId :
SSTD020340


Tgt Ion: 93 Resp: 179812
Ion Ratio Lower Upper
93 100
63 77.2 61.1 91.7
95 32.3 26.4 39.6



#12
2-Chlorophenol
Concen: 21.413 ng/ul
RT: 7.516 min Scan# 770
Delta R.T. -0.000 min
Lab File: BN022059.D
Acq: 12 Oct 2022 10:50

Tgt Ion:128 Resp: 167216
Ion Ratio Lower Upper
128 100
64 48.7 39.7 59.5
130 33.4 26.7 40.1





#13

2-Methylphenol

Concen: 20.297 ng/ul

RT: 8.410 min Scan# 91

Delta R.T. -0.000 min

Lab File: BN022059.D

Acq: 12 Oct 2022 10:50

Instrument :

BNA_N

ClientSampleId :





SSTD020340

Tgt Ion:108 Resp: 164626

Ion Ratio Lower Upper

108 100

107 92.5 72.4 108.6



#14

2,2'-oxybis(1-Chloropropane)

Concen: 19.459 ng/ul

RT: 8.499 min Scan# 937

Delta R.T. -0.006 min

Lab File: BN022059.D

Acq: 12 Oct 2022 10:50




Tgt Ion: 45 Resp: 254062


Ion Ratio Lower Upper

45 100

77 14.5 11.7 17.5

79 11.4 8.7 13.1









#16
Acetophenone
Concen: 20.712 ng/ul
RT: 8.799 min Scan# 988
Delta R.T. -0.000 min
Lab File: BN022059.D
Acq: 12 Oct 2022 10:50




Instrument :
BNA_N
ClientSampleId :
SSTD020340


Tgt Ion:105 Resp: 276605
Ion Ratio Lower Upper
105 100
77 77.6 64.9 97.3
51 26.3 21.8 32.6



#17
N-Nitroso-di-n-propylamine
Concen: 20.822 ng/ul
RT: 8.781 min Scan# 985
Delta R.T. -0.000 min
Lab File: BN022059.D
Acq: 12 Oct 2022 10:50

Tgt Ion: 70 Resp: 145523
Ion Ratio Lower Upper
70 100
42 56.9 45.6 68.4
101 9.0 7.4 11.2
130 20.0 15.4 23.0









#18
4-Methylphenol
Concen: 20.498 ng/ul
RT: 8.746 min Scan# 91
Delta R.T. -0.000 min
Lab File: BN022059.D
Acq: 12 Oct 2022 10:50




Instrument :
BNA_N
ClientSampleId :
SSTD020340


Tgt Ion:108 Resp: 186613
Ion Ratio Lower Upper
108 100
107 107.4 87.2 130.8



#19
Hexachloroethane
Concen: 20.832 ng/ul
RT: 9.052 min Scan# 1031
Delta R.T. -0.000 min
Lab File: BN022059.D
Acq: 12 Oct 2022 10:50

Tgt Ion:117 Resp: 65221
Ion Ratio Lower Upper
117 100
201 87.5 69.8 104.6
199 57.6 45.2 67.8









#22
Nitrobenzene
Concen: 20.934 ng/ul
RT: 9.181 min Scan# 1053
Delta R.T. -0.000 min
Lab File: BN022059.D
Acq: 12 Oct 2022 10:50




Instrument :
BNA_N
ClientSampleId :
SSTD020340


Tgt Ion: 77 Resp: 210238
Ion Ratio Lower Upper
77 100
123 43.9 34.9 52.3
65 13.0 10.2 15.2



#23
Isophorone
Concen: 21.083 ng/ul
RT: 9.710 min Scan# 1143
Delta R.T. -0.000 min
Lab File: BN022059.D
Acq: 12 Oct 2022 10:50

Tgt Ion: 82 Resp: 413805
Ion Ratio Lower Upper
82 100
95 6.6 5.4 8.0
138 17.6 13.8 20.8









#25
2-Nitrophenol
Concen: 22.011 ng/ul
RT: 9.899 min Scan# 1175
Delta R.T. -0.000 min
Lab File: BN022059.D
Acq: 12 Oct 2022 10:50




Instrument :
BNA_N
ClientSampleId :
SSTD020340


Tgt Ion:139 Resp: 97215
Ion Ratio Lower Upper
139 100
65 54.4 42.7 64.1
109 28.1 23.8 35.6



#26
2,4-Dimethylphenol
Concen: 20.806 ng/ul
RT: 9.957 min Scan# 1185
Delta R.T. -0.000 min
Lab File: BN022059.D
Acq: 12 Oct 2022 10:50

Tgt Ion:107 Resp: 198223
Ion Ratio Lower Upper
107 100
121 49.7 41.0 61.6
122 85.6 71.5 107.3





#27

Bis(2-Chloroethoxy)methane

Concen: 20.396 ng/ul

RT: 10.198 min Scan# 11

Delta R.T. -0.000 min

Lab File: BN022059.D

Acq: 12 Oct 2022 10:50

Instrument :

BNA_N

ClientSampleId :

SSTD020340





Tgt Ion: 93 Resp: 253458

Ion Ratio Lower Upper

93 100

95 32.4 25.2 37.8

123 10.9 8.6 12.8



#29

2,4-Dichlorophenol

Concen: 21.285 ng/ul

RT: 10.434 min Scan# 1266

Delta R.T. -0.000 min

Lab File: BN022059.D

Acq: 12 Oct 2022 10:50




Tgt Ion:162 Resp: 159988


Ion Ratio Lower Upper

162 100

164 63.9 52.9 79.3

98 39.0 30.5 45.7









#30
Naphthalene
Concen: 20.502 ng/ul
RT: 10.840 min Scan# 1335
Delta R.T. -0.000 min
Lab File: BN022059.D
Acq: 12 Oct 2022 10:50




Instrument :
BNA_N
ClientSampleId :
SSTD020340


Tgt Ion:128 Resp: 573018
Ion Ratio Lower Upper
128 100
129 10.8 8.5 12.7
127 13.1 10.8 16.2



#32
4-Chloroaniline
Concen: 20.574 ng/ul
RT: 10.957 min Scan# 1355
Delta R.T. -0.000 min
Lab File: BN022059.D
Acq: 12 Oct 2022 10:50

Tgt Ion:127 Resp: 257734
Ion Ratio Lower Upper
127 100
129 32.2 26.2 39.4





#33

Hexachlorobutadiene

Concen: 20.482 ng/ul

RT: 11.122 min Scan# 1492

Delta R.T. -0.000 min

Lab File: BN022059.D

Acq: 12 Oct 2022 10:50

Instrument :

BNA_N

ClientSampleId :

SSTD020340





Tgt Ion:225 Resp: 85191

Ion Ratio Lower Upper

225 100

223 65.6 47.7 71.5

227 64.7 51.4 77.0



#34

Caprolactam

Concen: 20.773 ng/ul

RT: 11.763 min Scan# 1492

Delta R.T. 0.023 min

Lab File: BN022059.D

Acq: 12 Oct 2022 10:50




Tgt Ion:113 Resp: 60732


Ion Ratio Lower Upper

113 100

55 167.0 138.7 208.1

56 118.8 101.4 152.0









#35
 4-Chloro-3-methylphenol
 Concen: 21.039 ng/ul
 RT: 12.081 min Scan# 1546
 Delta R.T. -0.000 min
 Lab File: BN022059.D
 Acq: 12 Oct 2022 10:50




Instrument :
 BNA_N
 ClientSampleId :
 SSTD020340


Tgt Ion:107 Resp: 192394
 Ion Ratio Lower Upper
 107 100
 144 23.8 20.3 30.5
 142 77.4 64.4 96.6



#36
 2-Methylnaphthalene
 Concen: 20.898 ng/ul
 RT: 12.457 min Scan# 1610
 Delta R.T. -0.000 min
 Lab File: BN022059.D
 Acq: 12 Oct 2022 10:50

Tgt Ion:142 Resp: 397668
 Ion Ratio Lower Upper
 142 100
 141 89.5 69.7 104.5





#37

1-Methylnaphthalene

Concen: 20.781 ng/ul

RT: 12.675 min Scan# 1647

Delta R.T. -0.000 min

Lab File: BN022059.D

Acq: 12 Oct 2022 10:50

Instrument :

BNA_N

ClientSampleId :

SSTD020340





Tgt Ion:142 Resp: 399666

Ion Ratio Lower Upper

142 100

141 89.0 76.7 115.1

116 3.7 3.1 4.7



#39

1,2,4,5-Tetrachlorobenzene

Concen: 21.334 ng/ul

RT: 12.822 min Scan# 1672

Delta R.T. -0.000 min

Lab File: BN022059.D

Acq: 12 Oct 2022 10:50

Tgt Ion:216 Resp: 180996




Ion Ratio Lower Upper


216 100

214 74.3 65.4 98.2

179 22.8 17.4 26.0


108 21.1 16.4 24.6








#40
Hexachlorocyclopentadiene
Concen: 22.559 ng/ul
RT: 12.798 min Scan# 1668
Delta R.T. -0.000 min
Lab File: BN022059.D
Acq: 12 Oct 2022 10:50


Instrument :
BNA_N
ClientSampleId :
SSTD020340





Tgt Ion:237 Resp: 100707
Ion Ratio Lower Upper
237 100
235 55.5 43.4 65.0
272 11.9 10.2 15.2




#41
2,4,6-Trichlorophenol
Concen: 22.632 ng/ul
RT: 13.069 min Scan# 1714
Delta R.T. 0.006 min
Lab File: BN022059.D
Acq: 12 Oct 2022 10:50



Tgt Ion:196 Resp: 129400
Ion Ratio Lower Upper
196 100
198 91.6 73.4 110.2
200 30.0 24.3 36.5





#42

2,4,5-Trichlorophenol

Concen: 21.651 ng/ul

RT: 13.134 min Scan# 1725

Delta R.T. -0.000 min

Lab File: BN022059.D

Acq: 12 Oct 2022 10:50

Instrument :

BNA_N

ClientSampleId :

SSTD020340





Tgt Ion:196 Resp: 136218

Ion Ratio Lower Upper

196 100

198 95.9 77.0 115.6

200 29.5 24.7 37.1



#43

1,1'-Biphenyl

Concen: 20.764 ng/ul

RT: 13.469 min Scan# 1782

Delta R.T. -0.000 min

Lab File: BN022059.D

Acq: 12 Oct 2022 10:50




Tgt Ion:154 Resp: 503236


Ion Ratio Lower Upper

154 100

153 41.3 30.9 46.3

76 14.7 11.6 17.4









#44
2-Chloronaphthalene
Concen: 20.871 ng/ul
RT: 13.510 min Scan# 1789
Delta R.T. -0.000 min
Lab File: BN022059.D
Acq: 12 Oct 2022 10:50




Instrument :
BNA_N
ClientSampleId :
SSTD020340


Tgt Ion:162 Resp: 394355
Ion Ratio Lower Upper
162 100
164 34.8 23.7 35.5
127 41.8 30.8 46.2



#45
2-Nitroaniline
Concen: 22.564 ng/ul
RT: 13.722 min Scan# 1825
Delta R.T. -0.000 min
Lab File: BN022059.D
Acq: 12 Oct 2022 10:50

Tgt Ion: 65 Resp: 131328
Ion Ratio Lower Upper
65 100
92 67.9 56.2 84.2
138 102.3 85.0 127.6









#47
Dimethylphthalate
Concen: 21.182 ng/ul
RT: 14.092 min Scan# 1888
Delta R.T. -0.000 min
Lab File: BN022059.D
Acq: 12 Oct 2022 10:50




Instrument :
BNA_N
ClientSampleId :
SSTD020340


Tgt Ion:163 Resp: 525765
Ion Ratio Lower Upper
163 100
194 3.8 3.1 4.7
164 10.0 8.2 12.2



#48
2,6-Dinitrotoluene
Concen: 22.741 ng/ul
RT: 14.216 min Scan# 1909
Delta R.T. -0.000 min
Lab File: BN022059.D
Acq: 12 Oct 2022 10:50

Tgt Ion:165 Resp: 100742
Ion Ratio Lower Upper
165 100
89 64.5 52.1 78.1
121 23.8 19.0 28.4





#50

Acenaphthylene

Concen: 22.055 ng/ul

RT: 14.357 min Scan# 1933

Delta R.T. -0.000 min

Lab File: BN022059.D

Acq: 12 Oct 2022 10:50

Instrument :

BNA_N

ClientSampleId :

SSTD020340





Tgt Ion:152 Resp: 644120

Ion Ratio Lower Upper

152 100

151 19.9 15.5 23.3

153 13.0 9.8 14.8



#51

3-Nitroaniline

Concen: 22.275 ng/ul

RT: 14.545 min Scan# 1965

Delta R.T. -0.000 min

Lab File: BN022059.D

Acq: 12 Oct 2022 10:50




Tgt Ion:138 Resp: 118867


Ion Ratio Lower Upper

138 100

108 10.6 9.0 13.6

92 120.6 96.4 144.6





#52

Acenaphthene

Concen: 20.462 ng/ul

RT: 14.698 min Scan# 1991

Delta R.T. -0.000 min

Lab File: BN022059.D

Acq: 12 Oct 2022 10:50

Instrument :

BNA_N

ClientSampleId :

SSTD020340





Tgt Ion:153 Resp: 434814

Ion Ratio Lower Upper

153 100

152 44.3 38.0 57.0

154 89.3 71.1 106.7



#53

2,4-Dinitrophenol

Concen: 21.398 ng/ul

RT: 14.751 min Scan# 2000

Delta R.T. -0.000 min

Lab File: BN022059.D

Acq: 12 Oct 2022 10:50




Tgt Ion:184 Resp: 65290


Ion Ratio Lower Upper

184 100

63 72.1 56.8 85.2

154 61.7 49.0 73.6









#55
4-Nitrophenol
Concen: 20.937 ng/ul
RT: 14.851 min Scan# 2017
Delta R.T. -0.000 min
Lab File: BN022059.D
Acq: 12 Oct 2022 10:50




Instrument :
BNA_N
ClientSampleId :
SSTD020340


Tgt Ion:109 Resp: 84266
Ion Ratio Lower Upper
109 100
139 145.2 108.3 162.5
65 144.7 107.9 161.9



#56
Dibenzofuran
Concen: 21.096 ng/ul
RT: 15.034 min Scan# 2048
Delta R.T. -0.000 min
Lab File: BN022059.D
Acq: 12 Oct 2022 10:50

Tgt Ion:168 Resp: 604284
Ion Ratio Lower Upper
168 100
139 37.3 30.6 45.8









#57
2,4-Dinitrotoluene
Concen: 22.583 ng/ul
RT: 15.004 min Scan# 2043
Delta R.T. -0.000 min
Lab File: BN022059.D
Acq: 12 Oct 2022 10:50




Instrument :
BNA_N
ClientSampleId :
SSTD020340


Tgt Ion:165 Resp: 151269
Ion Ratio Lower Upper
165 100
63 48.1 36.2 54.4
182 3.2 2.2 3.2



#58
2,3,4,6-Tetrachlorophenol
Concen: 21.726 ng/ul
RT: 15.263 min Scan# 2087
Delta R.T. 0.006 min
Lab File: BN022059.D
Acq: 12 Oct 2022 10:50

Tgt Ion:232 Resp: 116105
Ion Ratio Lower Upper
232 100
131 51.6 44.4 66.6
130 2.6 2.2 3.4
166 34.2 28.8 43.2





#59

Diethylphthalate

Concen: 21.224 ng/ul

RT: 15.457 min Scan# 2120

Delta R.T. -0.000 min

Lab File: BN022059.D

Acq: 12 Oct 2022 10:50

Instrument :

BNA_N

ClientSampleId :

SSTD020340





Tgt Ion:149 Resp: 546184

Ion Ratio Lower Upper

149 100

177 20.3 17.5 26.3

150 12.1 9.8 14.8



#61

Fluorene

Concen: 20.393 ng/ul

RT: 15.686 min Scan# 2159

Delta R.T. -0.000 min

Lab File: BN022059.D

Acq: 12 Oct 2022 10:50




Tgt Ion:166 Resp: 482865


Ion Ratio Lower Upper

166 100

165 98.3 72.8 109.2

167 13.5 10.6 16.0









#62
4-Chlorophenyl-phenylether
Concen: 20.247 ng/ul
RT: 15.681 min Scan# 2158
Delta R.T. -0.000 min
Lab File: BN022059.D
Acq: 12 Oct 2022 10:50




Instrument :
BNA_N
ClientSampleId :
SSTD020340


Tgt Ion:204 Resp: 217660
Ion Ratio Lower Upper
204 100
206 30.9 26.2 39.2
141 59.4 52.0 78.0



#63
4-Nitroaniline
Concen: 22.692 ng/ul
RT: 15.710 min Scan# 2163
Delta R.T. -0.000 min
Lab File: BN022059.D
Acq: 12 Oct 2022 10:50

Tgt Ion:138 Resp: 120483
Ion Ratio Lower Upper
138 100
92 55.1 41.8 62.8
108 81.8 60.4 90.6









#66
 4,6-Dinitro-2-methylphenol
 Concen: 20.768 ng/ul
 RT: 15.763 min Scan# 2172
 Delta R.T. -0.000 min
 Lab File: BN022059.D
 Acq: 12 Oct 2022 10:50




Instrument :
 BNA_N
 ClientSampleId :
 SSTD020340


Tgt Ion:198 Resp: 88663
 Ion Ratio Lower Upper
 198 100
 182 5.6 3.8 5.8
 77 31.3 24.4 36.6



#67
 N-Nitrosodiphenylamine
 Concen: 20.233 ng/ul
 RT: 15.892 min Scan# 2194
 Delta R.T. -0.000 min
 Lab File: BN022059.D
 Acq: 12 Oct 2022 10:50

Tgt Ion:169 Resp: 425277
 Ion Ratio Lower Upper
 169 100
 168 66.8 50.9 76.3
 167 36.6 28.2 42.4









#68
4-Bromophenyl-phenylether
Concen: 20.685 ng/ul
RT: 16.575 min Scan# 2310
Delta R.T. -0.000 min
Lab File: BN022059.D
Acq: 12 Oct 2022 10:50




Instrument :
BNA_N
ClientSampleId :
SSTD020340


Tgt Ion:248 Resp: 135048
Ion Ratio Lower Upper
248 100
250 90.1 86.3 129.5
141 82.1 74.2 111.4



#69
Hexachlorobenzene
Concen: 20.340 ng/ul
RT: 16.686 min Scan# 2329
Delta R.T. -0.000 min
Lab File: BN022059.D
Acq: 12 Oct 2022 10:50

Tgt Ion:284 Resp: 157450
Ion Ratio Lower Upper
284 100
142 42.0 35.7 53.5
249 29.6 27.0 40.6





#70

Atrazine

Concen: 21.316 ng/ul

RT: 16.851 min Scan# 21

Delta R.T. -0.000 min

Lab File: BN022059.D

Acq: 12 Oct 2022 10:50

Instrument :

BNA_N

ClientSampleId :

SSTD020340





Tgt Ion:200 Resp: 151462

Ion Ratio Lower Upper

200 100

173 30.4 22.8 34.2

215 50.2 38.1 57.1



#71

Pentachlorophenol

Concen: 20.866 ng/ul

RT: 17.033 min Scan# 2388

Delta R.T. -0.000 min

Lab File: BN022059.D

Acq: 12 Oct 2022 10:50




Tgt Ion:266 Resp: 98038


Ion Ratio Lower Upper

266 100

264 63.6 52.7 79.1

268 56.6 54.6 82.0









#72
Phenanthrene
Concen: 20.631 ng/u1
RT: 17.433 min Scan# 2456
Delta R.T. -0.000 min
Lab File: BN022059.D
Acq: 12 Oct 2022 10:50




Instrument :
BNA_N
ClientSampleId :
SSTD020340


Tgt Ion:178 Resp: 815981
Ion Ratio Lower Upper
178 100
179 15.3 12.3 18.5
176 20.2 16.5 24.7



#74
Anthracene
Concen: 20.686 ng/u1
RT: 17.528 min Scan# 2472
Delta R.T. 0.006 min
Lab File: BN022059.D
Acq: 12 Oct 2022 10:50

Tgt Ion:178 Resp: 814703
Ion Ratio Lower Upper
178 100
179 15.2 13.0 19.6
176 18.4 15.8 23.6









#75
1,2,3,4-Tetrachlorobenzene
Concen: 20.975 ng/uL
RT: 13.434 min Scan# 1776
Delta R.T. -0.000 min
Lab File: BN022059.D
Acq: 12 Oct 2022 10:50




Instrument :
BNA_N
ClientSampleId :
SSTD020340


Tgt Ion:216 Resp: 184519
Ion Ratio Lower Upper
216 100
214 75.7 61.2 91.8
218 45.0 39.4 59.0



#76
Pentachlorobenzene
Concen: 20.835 ng/uL
RT: 14.951 min Scan# 2034
Delta R.T. -0.000 min
Lab File: BN022059.D
Acq: 12 Oct 2022 10:50

Tgt Ion:250 Resp: 198343
Ion Ratio Lower Upper
250 100
248 61.6 46.8 70.2
252 64.9 53.9 80.9









#77
 Carbazole
 Concen: 21.208 ng/ul
 RT: 17.798 min Scan# 2518
 Delta R.T. -0.000 min
 Lab File: BN022059.D
 Acq: 12 Oct 2022 10:50




Instrument :
 BNA_N
 ClientSampleId :
 SSTD020340


Tgt Ion:167 Resp: 770947
 Ion Ratio Lower Upper
 167 100
 166 21.5 17.3 25.9
 139 13.9 10.7 16.1



#78
 Di-n-butylphthalate
 Concen: 22.239 ng/ul
 RT: 18.363 min Scan# 2614
 Delta R.T. 0.006 min
 Lab File: BN022059.D
 Acq: 12 Oct 2022 10:50

Tgt Ion:149 Resp: 989948
 Ion Ratio Lower Upper
 149 100
 150 9.7 7.1 10.7
 104 5.4 4.2 6.2





#80

Fluoranthene

Concen: 21.562 ng/ul

RT: 19.457 min Scan# 2800

Delta R.T. -0.000 min

Lab File: BN022059.D

Acq: 12 Oct 2022 10:50

Instrument :

BNA_N

ClientSampleId :

SSTD020340





Tgt Ion:202 Resp: 963181

Ion Ratio Lower Upper

202 100

101 13.6 10.3 15.5

100 10.0 7.8 11.8



#82

Pyrene

Concen: 20.492 ng/ul

RT: 19.822 min Scan# 2862

Delta R.T. 0.006 min

Lab File: BN022059.D

Acq: 12 Oct 2022 10:50




Tgt Ion:202 Resp: 965951


Ion Ratio Lower Upper

202 100

101 16.6 13.4 20.0

100 13.7 10.9 16.3









#83
Butylbenzylphthalate
Concen: 24.193 ng/ul
RT: 20.721 min Scan# 3014
Delta R.T. 0.006 min
Lab File: BN022059.D
Acq: 12 Oct 2022 10:50




Instrument :
BNA_N
ClientSampleId :
SSTD020340


Tgt Ion:149 Resp: 486275
Ion Ratio Lower Upper
149 100
91 72.6 61.7 92.5
206 19.0 14.7 22.1



#84
3,3'-Dichlorobenzidine
Concen: 23.303 ng/ul
RT: 21.510 min Scan# 3149
Delta R.T. -0.000 min
Lab File: BN022059.D
Acq: 12 Oct 2022 10:50

Tgt Ion:252 Resp: 366530
Ion Ratio Lower Upper
252 100
254 61.7 52.9 79.3
126 13.7 11.2 16.8









#85
Benzo(a)anthracene
Concen: 19.105 ng/ul
RT: 21.580 min Scan# 3161
Delta R.T. 0.006 min
Lab File: BN022059.D
Acq: 12 Oct 2022 10:50




Instrument :
BNA_N
ClientSampleId :
SSTD020340


Tgt Ion:228 Resp: 868249
Ion Ratio Lower Upper
228 100
229 18.8 13.5 20.3
226 25.3 21.0 31.6



#86
Bis(2-ethylhexyl)phthalate
Concen: 31.086 ng/ul
RT: 21.498 min Scan# 3147
Delta R.T. 0.006 min
Lab File: BN022059.D
Acq: 12 Oct 2022 10:50

Tgt Ion:149 Resp: 907271
Ion Ratio Lower Upper
149 100
167 26.6 21.3 31.9
279 3.6 2.9 4.3





#87

Chrysene

Concen: 20.590 ng/ul

RT: 21.633 min Scan# 3169

Delta R.T. 0.006 min

Lab File: BN022059.D

Acq: 12 Oct 2022 10:50

Instrument :

BNA_N

ClientSampleId :

SSTD020340





Tgt Ion:228 Resp: 920673

Ion Ratio Lower Upper

228 100

226 29.6 23.8 35.8

229 18.8 16.5 24.7



#89

Di-n-octyl phthalate

Concen: 21.752 ng/ul




RT: 22.445 min Scan# 3308


Delta R.T. 0.006 min

Lab File: BN022059.D

Acq: 12 Oct 2022 10:50

Tgt Ion:149 Resp: 1203699









#90
Benzo(b)fluoranthene
Concen: 21.190 ng/ul
RT: 23.304 min Scan# 3453
Delta R.T. 0.006 min
Lab File: BN022059.D
Acq: 12 Oct 2022 10:50




Instrument :
BNA_N
ClientSampleId :
SSTD020340


Tgt Ion:252 Resp: 1061625
Ion Ratio Lower Upper
252 100
253 19.7 17.6 26.4
125 12.8 10.7 16.1



#91
Benzo(k)fluoranthene
Concen: 21.349 ng/ul
RT: 23.357 min Scan# 3463
Delta R.T. 0.006 min
Lab File: BN022059.D
Acq: 12 Oct 2022 10:50

Tgt Ion:252 Resp: 1054965
Ion Ratio Lower Upper
252 100
253 18.7 17.7 26.5
125 12.2 10.6 15.8





#93

Benzo(a)pyrene

Concen: 21.690 ng/ul

RT: 23.962 min Scan# 3566

Delta R.T. 0.012 min

Lab File: BN022059.D

Acq: 12 Oct 2022 10:50

Instrument :

BNA_N

ClientSampleId :

SSTD020340





Tgt Ion:252 Resp: 959470

Ion Ratio Lower Upper

252 100

253 18.8 17.2 25.8

125 12.8 10.8 16.2



#94

Indeno(1,2,3-cd)pyrene

Concen: 21.145 ng/ul

RT: 26.668 min Scan# 4026

Delta R.T. 0.006 min

Lab File: BN022059.D

Acq: 12 Oct 2022 10:50




Tgt Ion:276 Resp: 1187543


Ion Ratio Lower Upper

276 100

138 30.0 24.6 36.8

277 23.1 18.4 27.6









#95
 Dibenzo(a,h)anthracene
 Concen: 21.426 ng/ul
 RT: 26.686 min Scan# 4027
 Delta R.T. 0.012 min
 Lab File: BN022059.D
 Acq: 12 Oct 2022 10:50

Instrument :
 BNA_N
 ClientSampleId :
 SSTD020340

Tgt Ion:278 Resp: 1088178
 Ion Ratio Lower Upper
 278 100
 139 20.0 15.7 23.5
 279 25.0 16.5 24.7#



#96
 Benzo(g,h,i)perylene
 Concen: 21.020 ng/ul
 RT: 27.468 min Scan# 4162
 Delta R.T. 0.012 min
 Lab File: BN022059.D
 Acq: 12 Oct 2022 10:50

Tgt Ion:276 Resp: 1064069
 Ion Ratio Lower Upper
 276 100
 138 26.4 21.2 31.8
 277 24.6 17.9 26.9

